

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3157

Analytical Method: SW8270E

Client: Always Available Construction LLC.

DataFile: BG064423.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q3133-02MS	Client Sample ID:	VNJ-238MS								
Pyridine	500	0	310	ug/L	62				10	109	
1,4-Dichlorobenzene	500	0	490	ug/L	98				55	125	
2-Methylphenol	500	0	470	ug/L	94				60	131	
3+4-Methylphenols	500	0	480	ug/L	96				54	136	
Hexachloroethane	500	0	450	ug/L	90				19	146	
Nitrobenzene	500	0	510	ug/L	102				62	112	
Hexachlorobutadiene	500	0	610	ug/L	122				52	125	
2,4,6-Trichlorophenol	500	0	580	ug/L	116	*			78	112	
2,4,5-Trichlorophenol	500	0	620	ug/L	124	*			71	111	
2,4-Dinitrotoluene	500	0	610	ug/L	122				74	137	
Hexachlorobenzene	500	0	620	ug/L	124	*			72	115	
Pentachlorophenol	1000	0	1300	ug/L	130				52	162	