

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3157

Analytical Method: SW8270E

Client: Always Available Construction LLC.

DataFile: BG064424.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q3133-02MSD		Client Sample ID: VNJ-238MSD									
Pyridine	500	0	330	ug/L	66		6		10	109	20
1,4-Dichlorobenzene	500	0	490	ug/L	98		0		55	125	20
2-Methylphenol	500	0	520	ug/L	104		10		60	131	20
3+4-Methylphenols	500	0	500	ug/L	100		4		54	136	20
Hexachloroethane	500	0	480	ug/L	96		6		19	146	20
Nitrobenzene	500	0	520	ug/L	104		2		62	112	20
Hexachlorobutadiene	500	0	610	ug/L	122		0		52	125	20
2,4,6-Trichlorophenol	500	0	620	ug/L	124	*	7		78	112	20
2,4,5-Trichlorophenol	500	0	620	ug/L	124	*	0		71	111	20
2,4-Dinitrotoluene	500	0	650	ug/L	130		6		74	137	20
Hexachlorobenzene	500	0	620	ug/L	124	*	0		72	115	20
Pentachlorophenol	1000	0	1400	ug/L	140		7		52	162	20