

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3157

Analytical Method: 8270E

Client: Always Available Construction LLC.

DataFile: BG064420.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB169783BS	Pyridine	50	30.1	ug/L	60				29	97	
	1,4-Dichlorobenzene	50	41.9	ug/L	84				76	103	
	2-Methylphenol	50	39.9	ug/L	80				69	109	
	3+4-Methylphenols	50	40.1	ug/L	80				67	106	
	Hexachloroethane	50	41.2	ug/L	82				76	118	
	Nitrobenzene	50	41.4	ug/L	83				58	106	
	Hexachlorobutadiene	50	51.4	ug/L	103		*		69	101	
	2,4,6-Trichlorophenol	50	44.6	ug/L	89				61	110	
	2,4,5-Trichlorophenol	50	44.8	ug/L	90				70	106	
	2,4-Dinitrotoluene	50	48.4	ug/L	97				60	115	
	Hexachlorobenzene	50	47.2	ug/L	94				73	106	
	Pentachlorophenol	100	100	ug/L	100				47	114	