

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX091625

Review By	John Carlone	Review On	9/17/2025 8:27:00 AM
Supervise By	Maresh Dadoda	Supervise On	9/18/2025 3:22:52 PM
SubDirectory	VX091625	HP Acquire Method	HP Processing Method 82X091625W.M

STD. NAME	STD REF.#
Tune/Reschk	VP135487
Initial Calibration Stds	VP135489,VP135490,VP135491,VP135492,VP135493,VP135494
CCC	VP135488
Internal Standard/PEM	
ICV/I.BLK	VP135495
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX047584.D	16 Sep 2025 08:56		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX047585.D	16 Sep 2025 09:23		JC/MD	Ok,M
3	VSTDIC005	VSTDIC005	VX047586.D	16 Sep 2025 10:16	QR-58	JC/MD	Ok,M
4	VSTDIC020	VSTDIC020	VX047587.D	16 Sep 2025 10:37		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX047588.D	16 Sep 2025 10:58		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX047589.D	16 Sep 2025 11:40		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX047590.D	16 Sep 2025 12:01		JC/MD	Ok,M
8	IBLK	IBLK	VX047591.D	16 Sep 2025 12:22		JC/MD	Ok
9	VSTDICV050	ICVVX091625	VX047592.D	16 Sep 2025 13:35		JC/MD	Ok,M
10	VX0916MBL01	VX0916MBL01	VX047593.D	16 Sep 2025 14:03		JC/MD	Ok
11	VX0916WBL01	VX0916WBL01	VX047594.D	16 Sep 2025 14:53		JC/MD	Ok
12	VX0916WBS01	VX0916WBS01	VX047595.D	16 Sep 2025 15:21		JC/MD	Ok,M
13	VX0916WBSD01	VX0916WBSD01	VX047596.D	16 Sep 2025 15:46		JC/MD	Ok,M
14	Q3015-01	WP0925-PT-VOA-WP	VX047597.D	16 Sep 2025 16:07		JC/MD	Ok,M
15	VIBLK	VIBLK	VX047598.D	16 Sep 2025 16:30		JC/MD	Ok
16	VSTDIC050	VSTDIC050EC	VX047599.D	16 Sep 2025 16:51		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX092225

Review By	John Carlone	Review On	9/23/2025 8:29:39 AM
Supervise By	Maresh Dadoda	Supervise On	9/24/2025 2:20:47 AM
SubDirectory	VX092225	HP Acquire Method	HP Processing Method 82X091625W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135556,VP135576
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135557,VP135558,VP135577,VP135578

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX047688.D	22 Sep 2025 09:07		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX047689.D	22 Sep 2025 09:38	pH#Lot#V12668	JC/MD	Ok,M
3	VX0922MBL01	VX0922MBL01	VX047690.D	22 Sep 2025 10:07		JC/MD	Ok
4	VX0922WBL01	VX0922WBL01	VX047691.D	22 Sep 2025 10:27		JC/MD	Ok
5	VX0922WBS01	VX0922WBS01	VX047692.D	22 Sep 2025 10:54		JC/MD	Ok,M
6	Q3097-08DL	RE125D2-20250911DL	VX047693.D	22 Sep 2025 11:21	vial B pH<2	JC/MD	Ok
7	VX0922WBSD01	VX0922WBSD01	VX047694.D	22 Sep 2025 11:42		JC/MD	Ok,M
8	Q3097-07DL	RE108D2-20250911DL	VX047695.D	22 Sep 2025 12:03	vial B pH<2	JC/MD	Ok
9	Q3097-14DL	RE105D2-20250911DL	VX047696.D	22 Sep 2025 12:23	vial B pH<2	JC/MD	Ok
10	Q3097-09	RE108D1-20250911	VX047697.D	22 Sep 2025 12:44	vial B pH<2	JC/MD	Ok
11	Q3097-10	RE120D3-20250911	VX047698.D	22 Sep 2025 13:05	vial B pH<2	JC/MD	Ok
12	Q3097-26	EB01-20250912	VX047699.D	22 Sep 2025 13:25	vial B pH<2 EB	JC/MD	Ok
13	PB169755TB	PB169755TB	VX047700.D	22 Sep 2025 13:46		JC/MD	Ok
14	Q3133-02	VNJ-238	VX047701.D	22 Sep 2025 14:07	vial A pH#5.0	JC/MD	Ok
15	Q3133-04	VNJ-222	VX047702.D	22 Sep 2025 14:28	vial A pH#5.0	JC/MD	Ok,M
16	Q3133-06	361	VX047703.D	22 Sep 2025 14:48	vial A pH#5.0	JC/MD	Ok,M
17	Q3133-08	VNJ-210	VX047704.D	22 Sep 2025 15:09	vial A pH#5.0	JC/MD	Ok,M
18	Q3153-04	OR-500-COMP-39	VX047705.D	22 Sep 2025 15:30	vial A pH#5.0	JC/MD	Ok,M

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX092225

Review By	John Carlone	Review On	9/23/2025 8:29:39 AM
Supervise By	Maresh Dadoda	Supervise On	9/24/2025 2:20:47 AM
SubDirectory	VX092225	HP Acquire Method	HP Processing Method 82X091625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135556,VP135576		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135557,VP135558,VP135577,VP135578		

19	Q3153-08	OR-500-COMP-40	VX047706.D	22 Sep 2025 15:51	vial A pH#5.0	JC/MD	Ok
20	Q3157-02	20-DEAD-1	VX047707.D	22 Sep 2025 16:11	vial A pH#5.0	JC/MD	Ok,M
21	Q3136-01	34922	VX047708.D	22 Sep 2025 16:32	vial A pH#5.0	JC/MD	Ok
22	Q3136-02	34923	VX047709.D	22 Sep 2025 16:53	vial A pH#5.0	JC/MD	Ok,M
23	Q3136-03	34924	VX047710.D	22 Sep 2025 17:14	vial A pH#5.0	JC/MD	Ok
24	Q3136-04	34925	VX047711.D	22 Sep 2025 17:35	vial A pH#5.0	JC/MD	Ok,M
25	Q3136-05	34926	VX047712.D	22 Sep 2025 17:56	vial A pH#5.0	JC/MD	Ok
26	IBLK	IBLK	VX047713.D	22 Sep 2025 18:16		JC/MD	Ok
27	Q3097-01	RE103D2-20250911	VX047714.D	22 Sep 2025 18:37	vial A pH<2 Surrogate Fail;Need 5X Dilution	JC/MD	Dilution
28	Q3097-02MS	RE103D2-20250911MS	VX047715.D	22 Sep 2025 18:58	vial A pH<2	JC/MD	Ok,M
29	Q3097-03MSD	RE103D2-20250911MS	VX047716.D	22 Sep 2025 19:19	vial A pH<2	JC/MD	Ok,M
30	VSTDCCC050	VSTDCCC050EC	VX047717.D	22 Sep 2025 19:40		JC/MD	Ok,M

M : Manual Integration