

Surrogate Summary

SDG No.: Q3157

Client: Always Available Construction LLC.

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3157-02	20-DEAD-1	1,2-Dichloroethane-d4	50	55.3	111		74	125
		Dibromofluoromethane	50	48.7	97		75	124
		Toluene-d8	50	48.7	97		86	113
		4-Bromofluorobenzene	50	54.9	110		77	121

Surrogate Summary

SDG No.: Q3157

Client: Always Available Construction LLC.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
VX0922WBL01	VX0922WBL01	1,2-Dichloroethane-d4	50	47.2	94		74	125
		Dibromofluoromethane	50	48.1	96		75	124
		Toluene-d8	50	49.2	98		86	113
		4-Bromofluorobenzene	50	49.5	99		77	121
VX0922WBS01	VX0922WBS01	1,2-Dichloroethane-d4	50	55.1	110		74	125
		Dibromofluoromethane	50	55.0	110		75	124
		Toluene-d8	50	54.9	110		86	113
		4-Bromofluorobenzene	50	56.8	114		77	121
VX0922WBSD0	VX0922WBSD01	1,2-Dichloroethane-d4	50	53.1	106		74	125
		Dibromofluoromethane	50	54.8	110		75	124
		Toluene-d8	50	54.3	109		86	113
		4-Bromofluorobenzene	50	54.8	110		77	121