

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3157</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Always Available Construction LLC.</u>	Datafile :	<u>VX047692.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0922WBS01	Vinyl chloride	20	23.8	ug/L	119		*	65	117	
	1,1-Dichloroethene	20	22.4	ug/L	112		*	74	110	
	2-Butanone	100	95.1	ug/L	95			65	122	
	Carbon Tetrachloride	20	20.8	ug/L	104			77	113	
	Chloroform	20	22.1	ug/L	111			79	113	
	Benzene	20	21.5	ug/L	108			82	109	
	1,2-Dichloroethane	20	20.4	ug/L	102			80	115	
	Trichloroethene	20	21.7	ug/L	109			77	113	
	Tetrachloroethene	20	21.3	ug/L	106			67	123	
	Chlorobenzene	20	21.1	ug/L	106			82	109	