

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3157</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Always Available Construction LLC.</u>	Datafile :	<u>VX047694.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0922WBSD01	Vinyl chloride	20	22.2	ug/L	111	7		65	117	19
	1,1-Dichloroethene	20	21.8	ug/L	109	3		74	110	20
	2-Butanone	100	98.9	ug/L	99	4		65	122	26
	Carbon Tetrachloride	20	20.5	ug/L	103	1		77	113	15
	Chloroform	20	21.5	ug/L	108	3		79	113	20
	Benzene	20	21.6	ug/L	108	0		82	109	15
	1,2-Dichloroethane	20	21.5	ug/L	108	6		80	115	20
	Trichloroethene	20	21.7	ug/L	109	0		77	113	15
	Tetrachloroethene	20	21.7	ug/L	109	3		67	123	15
	Chlorobenzene	20	21.8	ug/L	109	3		82	109	15