

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ALWA01
 Lab Code: ACE SDG NO.: Q3157
 Lab File ID: VX047689.D Date Analyzed: 09/22/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:38
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	186088	5.53	326997	6.74	282237	10.04
UPPER LIMIT	372176	6.032	653994	7.239	564474	10.537
LOWER LIMIT	93044	5.032	163499	6.239	141119	9.537
EPA SAMPLE NO.						
20-DEAD-1	155439	5.54	330407	6.75	332863	10.04
VX0922WBL01	207294	5.54	374581	6.75	340485	10.04
VX0922WBS01	172067	5.53	311516	6.75	283214	10.04
VX0922WBSD01	166680	5.54	297578	6.75	263948	10.04

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.