



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: ALWA01
Lab Code: ACE SDG No.: Q3157
Instrument ID: MSVOA_W Calibration Date/Time: 09/22/2025 08:47
Lab File ID: VW032223.D Init. Calib. Date(s): 08/26/2025 08/26/2025
Heated Purge: (Y/N) Y Init. Calib. Time(s): 09:39 12:34
GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Methylene Chloride	50.000	45.43	9.148003	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.