

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	Q3157	Analytical Method:	SW8260D
Client:	Always Available Construction LLC.	Datafile :	VW032226.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VW0922SBSD01	Dichlorodifluoromethane	20	23.9	ug/Kg	119	3		64	136	20
	Chloromethane	20	25.9	ug/Kg	130	9		52	151	20
	Vinyl chloride	20	28.3	ug/Kg	142	7		56	148	20
	Bromomethane	20	26.7	ug/Kg	134	10		58	141	20
	Chloroethane	20	27.6	ug/Kg	138	4	*	69	130	20
	Trichlorofluoromethane	20	27.5	ug/Kg	138	34	* *	69	134	20
	1,1,2-Trichlorotrifluoroethane	20	27.1	ug/Kg	136	15	*	81	123	20
	1,1-Dichloroethene	20	26.5	ug/Kg	133	15	*	79	121	20
	Acetone	100	150	ug/Kg	150	14		40	171	20
	Carbon disulfide	20	22.7	ug/Kg	114	7		59	130	20
	Methyl tert-butyl Ether	20	23.5	ug/Kg	117	6		77	129	20
	Methyl Acetate	20	23.3	ug/Kg	117	7		69	149	20
	Methylene Chloride	20	21.7	ug/Kg	109	5		72	131	20
	trans-1,2-Dichloroethene	20	23.3	ug/Kg	117	6		80	123	20
	1,1-Dichloroethane	20	23.5	ug/Kg	117	6		82	123	20
	Cyclohexane	20	22.6	ug/Kg	113	8		76	122	20
	2-Butanone	100	120	ug/Kg	120	0		69	131	20
	Carbon Tetrachloride	20	22.1	ug/Kg	111	4		76	129	20
	cis-1,2-Dichloroethene	20	22.7	ug/Kg	114	5		82	123	20
	Bromochloromethane	20	23.0	ug/Kg	115	4		80	127	20
	Chloroform	20	24.1	ug/Kg	121	10		82	125	20
	1,1,1-Trichloroethane	20	24.1	ug/Kg	121	11		80	126	20
	Methylcyclohexane	20	20.7	ug/Kg	104	5		77	123	20
	Benzene	20	22.3	ug/Kg	112	4		84	121	20
	1,2-Dichloroethane	20	22.3	ug/Kg	112	2		81	126	20
	Trichloroethene	20	21.6	ug/Kg	108	7		83	122	20
	1,2-Dichloropropane	20	22.4	ug/Kg	112	1		83	122	20
	Bromodichloromethane	20	22.1	ug/Kg	111	0		82	123	20
	4-Methyl-2-Pentanone	100	120	ug/Kg	120	0		70	135	20
	Toluene	20	21.7	ug/Kg	109	1		83	122	20
	t-1,3-Dichloropropene	20	21.5	ug/Kg	108	0		78	124	20
	cis-1,3-Dichloropropene	20	21.2	ug/Kg	106	6		81	122	20
	1,1,2-Trichloroethane	20	22.7	ug/Kg	114	4		82	125	20
	2-Hexanone	100	120	ug/Kg	120	0		66	138	20
	Dibromochloromethane	20	22.0	ug/Kg	110	1		79	125	20
	1,2-Dibromoethane	20	22.3	ug/Kg	112	1		80	125	20
	Tetrachloroethene	20	20.9	ug/Kg	104	7		83	125	20
	Chlorobenzene	20	20.9	ug/Kg	104	7		84	122	20
	Ethyl Benzene	20	20.9	ug/Kg	104	8		82	124	20
	m/p-Xylenes	40	41.9	ug/Kg	105	7		83	124	20
	o-Xylene	20	20.8	ug/Kg	104	6		83	123	20
	Styrene	20	20.8	ug/Kg	104	10		82	124	20
	Bromoform	20	20.8	ug/Kg	104	2		75	127	20
	Isopropylbenzene	20	21.0	ug/Kg	105	0		82	124	20
	1,1,2,2-Tetrachloroethane	20	22.1	ug/Kg	111	4		77	127	20
	1,3-Dichlorobenzene	20	21.0	ug/Kg	105	5		83	122	20
	1,4-Dichlorobenzene	20	21.1	ug/Kg	106	2		84	121	20
	1,2-Dichlorobenzene	20	22.1	ug/Kg	111	1		83	124	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3157</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Always Available Construction LLC.</u>	Datafile :	<u>VW032226.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VW0922SBSD01	1,2-Dibromo-3-Chloropropane	20	23.5	ug/Kg	117	8		66	134	20
	1,2,4-Trichlorobenzene	20	19.8	ug/Kg	99	2		78	127	20
	1,2,3-Trichlorobenzene	20	20.4	ug/Kg	102	5		70	137	20