

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ALWA01  
 Lab Code: ACE SDG NO.: Q3157  
 Lab File ID: VW032223.D Date Analyzed: 09/22/2025  
 Instrument ID: MSVOA\_W Time Analyzed: 08:47  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|----------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD    | 201565        | 7.97  | 338923        | 8.86  | 319016        | 11.63  |
| UPPER LIMIT    | 403130        | 8.465 | 677846        | 9.355 | 638032        | 12.129 |
| LOWER LIMIT    | 100783        | 7.465 | 169462        | 8.355 | 159508        | 11.129 |
| EPA SAMPLE NO. |               |       |               |       |               |        |
| 20-DEAD-1      | 145332        | 7.97  | 298948        | 8.86  | 293437        | 11.64  |
| VW0922SBL01    | 157714        | 7.97  | 321952        | 8.85  | 326905        | 11.63  |
| VW0922SBS01    | 187927        | 7.96  | 311645        | 8.85  | 278316        | 11.64  |
| VW0922SBSD01   | 177120        | 7.97  | 322196        | 8.86  | 306484        | 11.64  |

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.