

Initial and Continuing Calibration Verification

Client: Tanagro Corp.

SDG No.: Q3158

Project: Semi-Annual 2025

RunNo.: LB137275

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.095	0.099	96	85-115	09/23/2025
Sample ID: Cyanide	CCV1	mg/L	0.24	0.25	96	90-110	09/23/2025
Sample ID: Cyanide	CCV2	mg/L	0.26	0.25	104	90-110	09/23/2025