

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q3163
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113960.D **Time Analyzed:** 10:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.741	4.643	4.843	489.840	500.000	-2.0
Aroclor-1016-2	4.759	4.661	4.861	500.820	500.000	0.2
Aroclor-1016-3	4.816	4.718	4.918	482.330	500.000	-3.5
Aroclor-1016-4	4.934	4.837	5.037	480.820	500.000	-3.8
Aroclor-1016-5	5.191	5.093	5.293	514.930	500.000	3.0
Aroclor-1260-1	6.226	6.129	6.329	526.080	500.000	5.2
Aroclor-1260-2	6.416	6.319	6.519	508.020	500.000	1.6
Aroclor-1260-3	6.781	6.683	6.883	511.130	500.000	2.2
Aroclor-1260-4	7.039	6.941	7.141	590.510	500.000	18.1
Aroclor-1260-5	7.282	7.185	7.385	594.700	500.000	18.9
Decachlorobiphenyl	8.668	8.570	8.770	53.460	50.000	6.9
Tetrachloro-m-xylene	3.659	3.560	3.760	52.240	50.000	4.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q3163
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113960.D **Time Analyzed:** 10:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.719	4.623	4.823	473.430	500.000	-5.3
Aroclor-1016-2	4.737	4.641	4.841	474.680	500.000	-5.1
Aroclor-1016-3	4.912	4.816	5.016	472.750	500.000	-5.5
Aroclor-1016-4	4.955	4.858	5.058	476.570	500.000	-4.7
Aroclor-1016-5	5.166	5.070	5.270	487.430	500.000	-2.5
Aroclor-1260-1	6.192	6.096	6.296	468.900	500.000	-6.2
Aroclor-1260-2	6.381	6.284	6.484	454.800	500.000	-9.0
Aroclor-1260-3	6.532	6.435	6.635	442.380	500.000	-11.5
Aroclor-1260-4	7.000	6.903	7.103	459.590	500.000	-8.1
Aroclor-1260-5	7.243	7.146	7.346	459.980	500.000	-8.0
Decachlorobiphenyl	8.611	8.514	8.714	52.940	50.000	5.9
Tetrachloro-m-xylene	3.648	3.551	3.751	50.250	50.000	0.5