

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ATCE02  
**Lab Code:** ACE **SDG NO.:** Q3163  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

**Client Sample No.:** CCAL02 **Date Analyzed:** 09/23/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO113972.D **Time Analyzed:** 15:11

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.741 | 4.643     | 4.843 | 515.720            | 500.000           | 3.1  |
| Aroclor-1016-2       | 4.759 | 4.661     | 4.861 | 524.950            | 500.000           | 5.0  |
| Aroclor-1016-3       | 4.816 | 4.718     | 4.918 | 511.280            | 500.000           | 2.3  |
| Aroclor-1016-4       | 4.935 | 4.837     | 5.037 | 508.590            | 500.000           | 1.7  |
| Aroclor-1016-5       | 5.191 | 5.093     | 5.293 | 580.130            | 500.000           | 16.0 |
| Aroclor-1260-1       | 6.226 | 6.129     | 6.329 | 540.510            | 500.000           | 8.1  |
| Aroclor-1260-2       | 6.416 | 6.319     | 6.519 | 523.260            | 500.000           | 4.7  |
| Aroclor-1260-3       | 6.781 | 6.683     | 6.883 | 522.740            | 500.000           | 4.5  |
| Aroclor-1260-4       | 7.039 | 6.941     | 7.141 | 582.090            | 500.000           | 16.4 |
| Aroclor-1260-5       | 7.283 | 7.185     | 7.385 | 598.830            | 500.000           | 19.8 |
| Decachlorobiphenyl   | 8.667 | 8.570     | 8.770 | 51.740             | 50.000            | 3.5  |
| Tetrachloro-m-xylene | 3.659 | 3.560     | 3.760 | 54.710             | 50.000            | 9.4  |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ATCE02  
**Lab Code:** ACE **SDG NO.:** Q3163  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

**Client Sample No.:** CCAL02 **Date Analyzed:** 09/23/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO113972.D **Time Analyzed:** 15:11

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.721 | 4.623     | 4.823 | 501.440            | 500.000           | 0.3  |
| Aroclor-1016-2       | 4.739 | 4.641     | 4.841 | 497.040            | 500.000           | -0.6 |
| Aroclor-1016-3       | 4.913 | 4.816     | 5.016 | 507.060            | 500.000           | 1.4  |
| Aroclor-1016-4       | 4.956 | 4.858     | 5.058 | 505.430            | 500.000           | 1.1  |
| Aroclor-1016-5       | 5.167 | 5.070     | 5.270 | 525.210            | 500.000           | 5.0  |
| Aroclor-1260-1       | 6.193 | 6.096     | 6.296 | 482.140            | 500.000           | -3.6 |
| Aroclor-1260-2       | 6.381 | 6.284     | 6.484 | 471.100            | 500.000           | -5.8 |
| Aroclor-1260-3       | 6.532 | 6.435     | 6.635 | 451.170            | 500.000           | -9.8 |
| Aroclor-1260-4       | 7.000 | 6.903     | 7.103 | 463.830            | 500.000           | -7.2 |
| Aroclor-1260-5       | 7.243 | 7.146     | 7.346 | 458.420            | 500.000           | -8.3 |
| Decachlorobiphenyl   | 8.612 | 8.514     | 8.714 | 51.480             | 50.000            | 3.0  |
| Tetrachloro-m-xylene | 3.650 | 3.551     | 3.751 | 52.710             | 50.000            | 5.4  |