

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q3163
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113979.D **Time Analyzed:** 09:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.746	4.643	4.843	492.480	500.000	-1.5
Aroclor-1016-2	4.765	4.661	4.861	483.870	500.000	-3.2
Aroclor-1016-3	4.822	4.718	4.918	468.430	500.000	-6.3
Aroclor-1016-4	4.941	4.837	5.037	471.510	500.000	-5.7
Aroclor-1016-5	5.197	5.093	5.293	505.290	500.000	1.1
Aroclor-1260-1	6.234	6.129	6.329	502.060	500.000	0.4
Aroclor-1260-2	6.423	6.319	6.519	480.200	500.000	-4.0
Aroclor-1260-3	6.787	6.683	6.883	494.970	500.000	-1.0
Aroclor-1260-4	7.046	6.941	7.141	563.880	500.000	12.8
Aroclor-1260-5	7.290	7.185	7.385	587.770	500.000	17.6
Decachlorobiphenyl	8.675	8.570	8.770	51.380	50.000	2.8
Tetrachloro-m-xylene	3.663	3.560	3.760	50.810	50.000	1.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q3163
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113979.D **Time Analyzed:** 09:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.719	4.623	4.823	454.940	500.000	-9.0
Aroclor-1016-2	4.737	4.641	4.841	450.430	500.000	-9.9
Aroclor-1016-3	4.912	4.816	5.016	450.960	500.000	-9.8
Aroclor-1016-4	4.955	4.858	5.058	450.370	500.000	-9.9
Aroclor-1016-5	5.166	5.070	5.270	476.690	500.000	-4.7
Aroclor-1260-1	6.192	6.096	6.296	454.820	500.000	-9.0
Aroclor-1260-2	6.381	6.284	6.484	447.230	500.000	-10.6
Aroclor-1260-3	6.532	6.435	6.635	434.380	500.000	-13.1
Aroclor-1260-4	7.001	6.903	7.103	466.580	500.000	-6.7
Aroclor-1260-5	7.243	7.146	7.346	467.570	500.000	-6.5
Decachlorobiphenyl	8.612	8.514	8.714	48.740	50.000	-2.5
Tetrachloro-m-xylene	3.648	3.551	3.751	47.510	50.000	-5.0