

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q3163
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113994.D **Time Analyzed:** 16:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.740	4.643	4.843	504.670	500.000	0.9
Aroclor-1016-2	4.758	4.661	4.861	498.540	500.000	-0.3
Aroclor-1016-3	4.815	4.718	4.918	481.500	500.000	-3.7
Aroclor-1016-4	4.934	4.837	5.037	471.150	500.000	-5.8
Aroclor-1016-5	5.191	5.093	5.293	577.480	500.000	15.5
Aroclor-1260-1	6.225	6.129	6.329	497.220	500.000	-0.6
Aroclor-1260-2	6.415	6.319	6.519	457.420	500.000	-8.5
Aroclor-1260-3	6.779	6.683	6.883	453.040	500.000	-9.4
Aroclor-1260-4	7.039	6.941	7.141	515.780	500.000	3.2
Aroclor-1260-5	7.282	7.185	7.385	513.380	500.000	2.7
Decachlorobiphenyl	8.667	8.570	8.770	45.000	50.000	-10.0
Tetrachloro-m-xylene	3.658	3.560	3.760	53.770	50.000	7.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q3163
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113994.D **Time Analyzed:** 16:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.720	4.623	4.823	491.620	500.000	-1.7
Aroclor-1016-2	4.737	4.641	4.841	482.580	500.000	-3.5
Aroclor-1016-3	4.912	4.816	5.016	504.890	500.000	1.0
Aroclor-1016-4	4.954	4.858	5.058	490.290	500.000	-1.9
Aroclor-1016-5	5.166	5.070	5.270	519.370	500.000	3.9
Aroclor-1260-1	6.192	6.096	6.296	443.310	500.000	-11.3
Aroclor-1260-2	6.380	6.284	6.484	448.550	500.000	-10.3
Aroclor-1260-3	6.530	6.435	6.635	405.930	500.000	-18.8
Aroclor-1260-4	6.999	6.903	7.103	421.210	500.000	-15.8
Aroclor-1260-5	7.242	7.146	7.346	418.450	500.000	-16.3
Decachlorobiphenyl	8.609	8.514	8.714	47.100	50.000	-5.8
Tetrachloro-m-xylene	3.650	3.551	3.751	51.830	50.000	3.7