

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO091525

Review By	yogesh	Review On	9/15/2025 12:11:18 PM
Supervise By	mohammad	Supervise On	9/17/2025 2:07:54 AM
SubDirectory	PO091525	HP Acquire Method	HP Processing Method PO091525
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO113718.D	15 Sep 2025 08:58	YP/AJ	Ok
2	I.BLK	PO113719.D	15 Sep 2025 09:16	YP/AJ	Ok
3	AR1660ICC1000	PO113720.D	15 Sep 2025 09:34	YP/AJ	Ok
4	AR1660ICC750	PO113721.D	15 Sep 2025 09:53	YP/AJ	Ok
5	AR1660ICC500	PO113722.D	15 Sep 2025 10:11	YP/AJ	Ok
6	AR1660ICC250	PO113723.D	15 Sep 2025 10:30	YP/AJ	Ok,M
7	AR1660ICC050	PO113724.D	15 Sep 2025 10:48	YP/AJ	Ok,M
8	AR1221ICC500	PO113725.D	15 Sep 2025 11:06	YP/AJ	Ok
9	AR1232ICC500	PO113726.D	15 Sep 2025 11:25	YP/AJ	Ok
10	AR1242ICC1000	PO113727.D	15 Sep 2025 11:43	YP/AJ	Ok
11	AR1242ICC750	PO113728.D	15 Sep 2025 12:02	YP/AJ	Ok
12	AR1242ICC500	PO113729.D	15 Sep 2025 12:20	YP/AJ	Ok
13	AR1242ICC250	PO113730.D	15 Sep 2025 12:38	YP/AJ	Ok
14	AR1242ICC050	PO113731.D	15 Sep 2025 13:34	YP/AJ	Ok,M
15	AR1248ICC1000	PO113732.D	15 Sep 2025 13:54	YP/AJ	Ok
16	AR1248ICC750	PO113733.D	15 Sep 2025 14:13	YP/AJ	Ok
17	AR1248ICC500	PO113734.D	15 Sep 2025 14:31	YP/AJ	Ok
18	AR1248ICC250	PO113735.D	15 Sep 2025 14:49	YP/AJ	Ok,M
19	AR1248ICC050	PO113736.D	15 Sep 2025 15:26	YP/AJ	Ok
20	AR1254ICC1000	PO113737.D	15 Sep 2025 15:45	YP/AJ	Ok
21	AR1254ICC750	PO113738.D	15 Sep 2025 16:02	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO091525

Review By	yogesh	Review On	9/15/2025 12:11:18 PM
Supervise By	mohammad	Supervise On	9/17/2025 2:07:54 AM
SubDirectory	PO091525	HP Acquire Method	HP Processing Method PO091525
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC			
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PO113739.D	15 Sep 2025 16:20	YP/AJ	Ok
23	AR1254ICC250	PO113740.D	15 Sep 2025 16:38	YP/AJ	Ok
24	AR1254ICC050	PO113741.D	15 Sep 2025 17:14	YP/AJ	Ok,M
25	AR1262ICC500	PO113742.D	15 Sep 2025 17:33	YP/AJ	Ok
26	AR1268ICC1000	PO113743.D	15 Sep 2025 17:51	YP/AJ	Ok
27	AR1268ICC750	PO113744.D	15 Sep 2025 18:10	YP/AJ	Ok
28	AR1268ICC500	PO113745.D	15 Sep 2025 18:27	YP/AJ	Ok
29	AR1268ICC250	PO113746.D	15 Sep 2025 18:45	YP/AJ	Ok
30	AR1268ICC050	PO113747.D	15 Sep 2025 19:21	YP/AJ	Ok,M
31	PO091525ICV500	PO113748.D	15 Sep 2025 19:40	YP/AJ	Ok,M
32	AR1242ICV500	PO113749.D	15 Sep 2025 20:16	YP/AJ	Ok
33	AR1248ICV500	PO113750.D	15 Sep 2025 21:11	YP/AJ	Ok
34	AR1254ICV500	PO113751.D	15 Sep 2025 22:05	YP/AJ	Ok
35	AR1268ICV500	PO113752.D	15 Sep 2025 23:00	YP/AJ	Ok
36	DDT ANALOG	PO113753.D	15 Sep 2025 23:37	YP/AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO092325

Review By	yogesh	Review On	9/23/2025 11:26:49 AM
Supervise By	mohammad	Supervise On	9/25/2025 2:25:31 AM
SubDirectory	PO092325	HP Acquire Method	HP Processing Method PO091925
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO113959.D	23 Sep 2025 10:09	YP/AJ	Ok
2	AR1660CCC500	PO113960.D	23 Sep 2025 10:28	YP/AJ	Ok,M
3	AR1242CCC500	PO113961.D	23 Sep 2025 10:46	YP/AJ	Ok
4	AR1248CCC500	PO113962.D	23 Sep 2025 11:05	YP/AJ	Ok
5	AR1254CCC500	PO113963.D	23 Sep 2025 11:22	YP/AJ	Ok
6	I.BLK	PO113964.D	23 Sep 2025 11:40	YP/AJ	Ok
7	PB169759BL	PO113965.D	23 Sep 2025 11:58	YP/AJ	Ok
8	PB169759BS	PO113966.D	23 Sep 2025 12:16	YP/AJ	Ok
9	PB169795BL	PO113967.D	23 Sep 2025 12:34	YP/AJ	Ok
10	PB169795BS	PO113968.D	23 Sep 2025 12:52	YP/AJ	Not Ok
11	Q3163-01	PO113969.D	23 Sep 2025 13:10	YP/AJ	Ok,M
12	Q3163-02	PO113970.D	23 Sep 2025 13:29	YP/AJ	Ok,M
13	Q3136-05DL	PO113971.D	23 Sep 2025 14:24	YP/AJ	Ok,M
14	AR1660CCC500	PO113972.D	23 Sep 2025 15:11	YP/AJ	Ok
15	AR1242CCC500	PO113973.D	23 Sep 2025 15:29	YP/AJ	Ok
16	AR1248CCC500	PO113974.D	23 Sep 2025 15:47	YP/AJ	Ok
17	AR1254CCC500	PO113975.D	23 Sep 2025 16:04	YP/AJ	Ok
18	I.BLK	PO113976.D	23 Sep 2025 16:22	YP/AJ	Ok
19	DDT ANALOG	PO113977.D	23 Sep 2025 16:40	YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO092425

Review By	yogesh	Review On	9/24/2025 11:26:58 AM
Supervise By		Supervise On	
SubDirectory	PO092425	HP Acquire Method	HP Processing Method PO091925
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO113978.D	24 Sep 2025 08:38	YP/AJ	Ok
2	AR1660CCC500	PO113979.D	24 Sep 2025 09:36	YP/AJ	Ok,NS
3	AR1242CCC500	PO113980.D	24 Sep 2025 09:54	YP/AJ	Ok
4	AR1248CCC500	PO113981.D	24 Sep 2025 10:12	YP/AJ	Ok
5	AR1254CCC500	PO113982.D	24 Sep 2025 10:31	YP/AJ	Ok
6	I.BLK	PO113983.D	24 Sep 2025 10:49	YP/AJ	Ok
7	PB169795BS	PO113984.D	24 Sep 2025 11:08	YP/AJ	Ok
8	PB169805BL	PO113985.D	24 Sep 2025 12:49	YP/AJ	Ok
9	PB169805BS	PO113986.D	24 Sep 2025 13:07	YP/AJ	Not Ok
10	Q3160-01	PO113987.D	24 Sep 2025 13:24	YP/AJ	Ok,NS
11	Q3159-03	PO113988.D	24 Sep 2025 13:43	YP/AJ	Ok,NS
12	Q3159-03MS	PO113989.D	24 Sep 2025 14:01	YP/AJ	Ok,NS
13	Q3159-03MSD	PO113990.D	24 Sep 2025 14:20	YP/AJ	Ok,NS
14	Q3168-03	PO113991.D	24 Sep 2025 14:37	YP/AJ	Ok,NS
15	Q3167-01	PO113992.D	24 Sep 2025 14:56	YP/AJ	Ok,NS
16	Q3169-05	PO113993.D	24 Sep 2025 15:14	YP/AJ	Dilution
17	AR1660CCC500	PO113994.D	24 Sep 2025 16:34	YP/AJ	Ok
18	AR1242CCC500	PO113995.D	24 Sep 2025 16:52	YP/AJ	Ok
19	AR1248CCC500	PO113996.D	24 Sep 2025 17:09	YP/AJ	Ok
20	AR1254CCC500	PO113997.D	24 Sep 2025 17:28	YP/AJ	Ok
21	I.BLK	PO113998.D	24 Sep 2025 17:46	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QCBatch ID # PO092425

Review By	yogesh	Review On	9/24/2025 11:26:58 AM
Supervise By		Supervise On	
SubDirectory	PO092425	HP Acquire Method	HP Processing Method PO091925
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PB169805BS	PO113999.D	24 Sep 2025 18:04	YP/AJ	Ok
23	Q3169-06	PO114000.D	24 Sep 2025 18:21	YP/AJ	Dilution
24	Q3169-07	PO114001.D	24 Sep 2025 18:40	YP/AJ	Dilution
25	Q3169-08	PO114002.D	24 Sep 2025 18:58	YP/AJ	Dilution
26	Q3169-05DL	PO114003.D	24 Sep 2025 19:16	YP/AJ	Ok,NS
27	Q3169-06DL	PO114004.D	24 Sep 2025 19:34	YP/AJ	Ok,NS
28	Q3169-07DL	PO114005.D	24 Sep 2025 19:52	YP/AJ	Ok,NS
29	Q3169-08DL	PO114006.D	24 Sep 2025 20:10	YP/AJ	Ok,NS
30	AR1660CCC500	PO114007.D	24 Sep 2025 21:10	YP/AJ	Ok
31	AR1242CCC500	PO114008.D	24 Sep 2025 22:03	YP/AJ	Ok
32	AR1248CCC500	PO114009.D	24 Sep 2025 22:21	YP/AJ	Ok,NS
33	AR1254CCC500	PO114010.D	24 Sep 2025 22:39	YP/AJ	Ok
34	I.BLK	PO114011.D	24 Sep 2025 22:56	YP/AJ	Ok
35	DDT ANALOG	PO114012.D	24 Sep 2025 23:15	YP/AJ	Ok

M : Manual Integration