



A

B

C

SAMPLE DATA

Report of Analysis

Client:	ATG-GREENVILLE AEC	Date Collected:	09/18/25 09:18
Project:	AEC-2025-0013 CSC 4151 Coram	Date Received:	09/22/25
Client Sample ID:	DA-1	SDG No.:	Q3164
Lab Sample ID:	Q3164-01	Matrix:	Water
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
COD	190	D	5	7.50	50.0	mg/L		09/25/25 15:47	SM 5220 D-11
Oil and Grease	18.4		1	0.29	5.00	mg/L		09/26/25 09:00	1664A

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	ATG-GREENVILLE AEC	Date Collected:	09/18/25 09:30
Project:	AEC-2025-0013 CSC 4151 Coram	Date Received:	09/22/25
Client Sample ID:	DA-2	SDG No.:	Q3164
Lab Sample ID:	Q3164-02	Matrix:	Water
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
COD	160	D	5	7.50	50.0	mg/L		09/25/25 15:49	SM 5220 D-11
Oil and Grease	11.9		1	0.29	5.00	mg/L		09/26/25 09:00	1664A

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	ATG-GREENVILLE AEC	Date Collected:	09/18/25 09:52
Project:	AEC-2025-0013 CSC 4151 Coram	Date Received:	09/22/25
Client Sample ID:	DA-3	SDG No.:	Q3164
Lab Sample ID:	Q3164-03	Matrix:	Water
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
COD	190	D	5	7.50	50.0	mg/L		09/25/25 15:50	SM 5220 D-11
Oil and Grease	133		1	0.29	5.00	mg/L		09/26/25 09:00	1664A

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

LAB CHRONICLE

OrderID:	Q3164	OrderDate:	9/22/2025 2:37:00 PM
Client:	ATG-GREENVILLE AEC	Project:	AEC-2025-0013 CSC 4151 Coram
Contact:	Staci Bruce	Location:	J22,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3164-01	DA-1	Water			09/18/25 09:18			09/22/25
			COD	SM5220 D				
			Oil and Grease	1664A			09/25/25 15:47 09/26/25 09:00	
Q3164-02	DA-2	Water			09/18/25 09:30			09/22/25
			COD	SM5220 D				
			Oil and Grease	1664A			09/25/25 15:49 09/26/25 09:00	
Q3164-03	DA-3	Water			09/18/25 09:52			09/22/25
			COD	SM5220 D				
			Oil and Grease	1664A			09/25/25 15:50 09/26/25 09:00	