

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q3164
Project:	AEC-2025-0013 CSC 4151 Coram	Sample ID:	Q3164-01
Client ID:	DA-1MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	236		190	D	50.0	5	92		09/25/2025

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q3164
Project:	AEC-2025-0013 CSC 4151 Coram	Sample ID:	Q3164-01
Client ID:	DA-1MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	231		190	D	50.0	5	82		09/25/2025

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q3164
Project:	AEC-2025-0013 CSC 4151 Coram	Sample ID:	Q3184-01
Client ID:	Q3184-1MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	45.7		25.6		20.0	1	101		09/26/2025

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q3164
Project:	AEC-2025-0013 CSC 4151 Coram	Sample ID:	Q3184-01
Client ID:	Q3184-1MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	45.7		25.6		20.0	1	101		09/26/2025