

Laboratory Control Sample Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q3164

Project: AEC-2025-0013 CSC 4151 Coram

Run No.: LB137315

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137315BS							
COD	mg/L	50	51.2		102	1	90-110	09/25/2025

Laboratory Control Sample Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q3164

Project: AEC-2025-0013 CSC 4151 Coram

Run No.: LB137326

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137326BS							
Oil and Grease	mg/L	20.0	17.1		86	1	78-114	09/26/2025