

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>ATGG01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3164</u>
Instrument ID:	<u>MSVOA_V</u>	Calibration Date/Time:	<u>09/25/2025</u> <u>09:59</u>
Lab File ID:	<u>VV039168.D</u>	Init. Calib. Date(s):	<u>09/22/2025</u> <u>09/22/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>12:26</u> <u>16:35</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Benzene	1.979	1.926		-2.68	20
Toluene	1.147	1.204		4.97	20
Ethyl Benzene	2.108	2.320		10.06	20
m/p-Xylenes	0.814	0.936		14.99	20
o-Xylene	0.720	0.841		16.81	20
1,2-Dichloroethane-d4	0.724	0.609		-15.88	20
Dibromofluoromethane	0.402	0.384		-4.48	20
Toluene-d8	1.181	1.121		-5.08	20
4-Bromofluorobenzene	0.486	0.508		4.53	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.