284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_V

Daily Analysis Runlog For Sequence/QCBatch ID # VV092225

Review By	Mahesh Dadoda	Review On	9/26/2025 1:39:32 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	9/26/2025 1:56:56 AM		
SubDirectory	VV092225	HP Acquire Method	MSVOA_V	HP Processing Method	82v092225w.m

STD. NAME	STD REF.#
Tune/Reschk	VP135580
Initial Calibration Stds	VP135581,VP135582,VP135583,VP135584,VP135585,VP135586
CCC	
Internal Standard/PEM	VP133935
ICV/I.BLK	VP135587
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VV039134.D	22 Sep 2025 10:15		SY/MD	Ok
2	VSTDICC001	VSTDICC001	VV039135.D	22 Sep 2025 10:59	Not use	SY/MD	Not Ok
3	VSTDICC005	VSTDICC005	VV039136.D	22 Sep 2025 11:30	Not use	SY/MD	Not Ok
4	VSTDICC010	VSTDICC010	VV039137.D	22 Sep 2025 12:26		SY/MD	Ok
5	VSTDICC020	VSTDICC020	VV039138.D	22 Sep 2025 12:48	QR- 06,16,20,70,92	SY/MD	Ok
6	VSTDICCC050	VSTDICCC050	VV039139.D	22 Sep 2025 13:26	LR- 41,58,59	SY/MD	Ok
7	VSTDICC100	VSTDICC100	VV039140.D	22 Sep 2025 13:49		SY/MD	Ok
8	VIBLK	VIBLK	VV039141.D	22 Sep 2025 14:21		SY/MD	Ok
9	VSTDICC005	VSTDICC005	VV039142.D	22 Sep 2025 15:37		SY/MD	Ok,M
10	VSTDICC001	VSTDICC001	VV039143.D	22 Sep 2025 16:35		SY/MD	Ok,M
11	VSTDICV050	ICVVV092225	VV039144.D	22 Sep 2025 16:58	Goof for Non-DOD	SY/MD	Ok,M

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_V

Daily Analysis Runlog For Sequence/QCBatch ID # VV092425

Review By	Mahesh Dadoda	Review On	9/26/2025 1:45:06 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	9/26/2025 2:03:33 AM		
SubDirectory	VV092425	HP Acquire Method	MSVOA_V	HP Processing Method	

STD. NAME	STD REF.#
Tune/Reschk	VP135622
Initial Calibration Stds	
CCC	VP135623,VP135624
Internal Standard/PEM	VP133935
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VV039145.D	24 Sep 2025 09:52		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VV039146.D	24 Sep 2025 10:49	pH#lot#v14222	SY/MD	Ok
3	VV0924MBL01	VV0924MBL01	VV039147.D	24 Sep 2025 12:08		SY/MD	Ok
4	VV0924WBL01	VV0924WBL01	VV039148.D	24 Sep 2025 13:09		SY/MD	Ok
5	VV0924WBS01	VV0924WBS01	VV039149.D	24 Sep 2025 13:37	Recovery fail	SY/MD	Not Ok
6	VV0924WBS02	VV0924WBS02	VV039150.D	24 Sep 2025 14:01		SY/MD	Ok
7	Q3173-02	Field Blank	VV039151.D	24 Sep 2025 14:46	pH<2.0 A,FB	SY/MD	Ok
8	Q3173-01	OBS1	VV039152.D	24 Sep 2025 15:08	pH<2.0 A	SY/MD	Ok
9	Q3175-01	MW2	VV039153.D	24 Sep 2025 15:31	pH<2.0 A ,Need 20X	SY/MD	Dilution
10	Q3175-02	MW4	VV039154.D	24 Sep 2025 15:53	pH<2.0 A ,Need 20X	SY/MD	Dilution
11	VIBLK	VIBLK	VV039155.D	24 Sep 2025 16:16		SY/MD	Ok
12	VIBLK	VIBLK	VV039156.D	24 Sep 2025 16:38		SY/MD	Ok
13	Q3172-01	FW7D	VV039157.D	24 Sep 2025 17:00	pH<2.0 A ,Surrogate Fail;Need 100X	SY/MD	Dilution
14	Q3172-02	FW6D	VV039158.D	24 Sep 2025 17:22	pH<2.0 A ,Need 4X	SY/MD	Dilution
15	Q3172-03	GB3W	VV039159.D	24 Sep 2025 17:45	pH<2.0 A ,E flag of previous sample	SY/MD	ReRun
16	Q3164-01	DA-1	VV039160.D	24 Sep 2025 18:07	pH<2.0 A	SY/MD	Ok
17	Q3164-02	DA-2	VV039161.D	24 Sep 2025 18:30	pH<2.0 A ,Confirm Concentration	SY/MD	Not Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_V

Daily Analysis Runlog For Sequence/QCBatch ID # VV092425

Review By	Mahesh Dadoda	Review On	9/26/2025 1:45:06 AM	
Supervise By	Semsettin Yesilyurt	Supervise On	9/26/2025 2:03:33 AM	
SubDirectory	VV092425	HP Acquire Method	MSVOA_V	HP Processing Method

STD. NAME	STD REF.#
Tune/Reschk	VP135622
Initial Calibration Stds	
CCC	VP135623,VP135624
Internal Standard/PEM	VP133935
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	Q3164-03	DA-3	VV039162.D	24 Sep 2025 18:52	pH<2.0 A	SY/MD	Ok
19	VIBLK	VIBLK	VV039163.D	24 Sep 2025 19:15		SY/MD	Ok
20	VIBLK	VIBLK	VV039164.D	24 Sep 2025 19:37		SY/MD	Ok,M
21	VV0924WBSD02	VV0924WBSD02	VV039165.D	24 Sep 2025 19:59		SY/MD	Ok
22	VSTDCCC050	VSTDCCC050EC	VV039166.D	24 Sep 2025 20:22		SY/MD	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_V

Daily Analysis Runlog For Sequence/QCBatch ID # VV092525

Review By	Mahesh Dadoda	Review On	9/29/2025 12:56:22 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	9/29/2025 1:08:36 AM		
SubDirectory	VV092525	HP Acquire Method	8260_V	HP Processing Method	82v092225w.m

STD. NAME	STD REF.#
Tune/Reschk	VP135649
Initial Calibration Stds	
CCC	VP135650,VP135651
Internal Standard/PEM	VP133935
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VV039167.D	25 Sep 2025 09:14		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VV039168.D	25 Sep 2025 09:59	pH#lot#v14222	SY/MD	Ok
3	VV0925MBL01	VV0925MBL01	VV039169.D	25 Sep 2025 10:31		SY/MD	Ok
4	VV0925WBL01	VV0925WBL01	VV039170.D	25 Sep 2025 11:01		SY/MD	Ok
5	VV0925WBS01	VV0925WBS01	VV039171.D	25 Sep 2025 11:24		SY/MD	Ok
6	VV0925WBSD01	VV0925WBSD01	VV039172.D	25 Sep 2025 11:48		SY/MD	Ok
7	Q3164-02	DA-2	VV039173.D	25 Sep 2025 12:10	pH<2.0 B	SY/MD	Ok
8	Q3175-01DL	MW2DL	VV039174.D	25 Sep 2025 12:33	Need 100X	SY/MD	Dilution
9	Q3175-02DL	MW4DL	VV039175.D	25 Sep 2025 12:55		SY/MD	Ok,M
10	VIBLK	VIBLK	VV039176.D	25 Sep 2025 13:18		SY/MD	Ok,M
11	Q3168-05	TOTES COMP#1	VV039177.D	25 Sep 2025 13:40	pH<2.0 A, Surrogate fail	SY/MD	ReRun
12	Q3168-06	TOTES COMP#2	VV039178.D	25 Sep 2025 14:03	pH<2.0 A, Surrogate fail	SY/MD	ReRun
13	Q3175-01DL2	MW2DL2	VV039179.D	25 Sep 2025 14:25		SY/MD	Ok,M
14	Q3172-02DL	FW6DDL	VV039180.D	25 Sep 2025 14:48		SY/MD	Ok
15	VIBLK	VIBLK	VV039181.D	25 Sep 2025 15:10		SY/MD	Ok
16	Q3172-01DL	FW7DDL	VV039182.D	25 Sep 2025 15:33	Surrogate fail	SY/MD	Ok
17	Q3168-05RE	TOTES COMP#1RE	VV039183.D	25 Sep 2025 15:55	pH<2.0 B,Surrogate fail	SY/MD	Confirms
18	Q3168-06RE	TOTES COMP#2RE	VV039184.D	25 Sep 2025 16:18	pH<2.0 B,Surrogate fail	SY/MD	Confirms

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_V

Daily Analysis Runlog For Sequence/QCBatch ID # VV092525

Review By	Mahesh Dadoda	Review On	9/29/2025 12:56:22 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	9/29/2025 1:08:36 AM		
SubDirectory	VV092525	HP Acquire Method	8260_V	HP Processing Method	82v092225w.m

STD. NAME	STD REF.#
Tune/Reschk	VP135649
Initial Calibration Stds	
CCC	VP135650,VP135651
Internal Standard/PEM	VP133935
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	Q3172-03	GB3W	VV039185.D	25 Sep 2025 16:40	pH<2.0 A	SY/MD	Ok
20	Q3195-22	RW-1-092425	VV039186.D	25 Sep 2025 17:03	pH<2.0 A	SY/MD	Ok
21	Q3195-23	EB-01-091925	VV039187.D	25 Sep 2025 17:25	pH<2.0 A ,EB	SY/MD	Ok
22	Q3195-24	EB-02-092225	VV039188.D	25 Sep 2025 17:47	pH<2.0 A ,Internal Standard Fail	SY/MD	ReRun
23	Q3195-25	FB-01-091925	VV039189.D	25 Sep 2025 18:10	pH<2.0 A ,FB	SY/MD	Ok
24	Q3195-26	FB-02-092225	VV039190.D	25 Sep 2025 18:32	pH<2.0 A FB	SY/MD	Ok
25	Q3195-27	TB-01-091525	VV039191.D	25 Sep 2025 18:55	pH<2.0 A,TB	SY/MD	Ok
26	Q3195-02	MW-1-091725	VV039192.D	25 Sep 2025 19:17	pH<2.0 A	SY/MD	Ok
27	Q3195-09	MW-22-091825	VV039193.D	25 Sep 2025 19:40	pH<2.0 A	SY/MD	Ok
28	Q3195-10	MW-23-092325	VV039194.D	25 Sep 2025 20:02	pH<2.0 A	SY/MD	Ok
29	Q3195-11	MW-24-091925	VV039195.D	25 Sep 2025 20:25	pH<2.0 A	SY/MD	Ok
30	VSTDCCC050	VSTDCCC050EC	VV039196.D	25 Sep 2025 20:47		SY/MD	Ok

M : Manual Integration