

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3164</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>ATG-GREENVILLE AEC</u>	Datafile :	<u>VV039171.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VV0925WBS01	Benzene	20	19.9	ug/L	100			82	109	
	Toluene	20	21.5	ug/L	108			82	110	
	Ethyl Benzene	20	20.0	ug/L	100			83	109	
	m/p-Xylenes	40	43.1	ug/L	108			82	110	
	o-Xylene	20	21.0	ug/L	105			83	109	