

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>ATGG01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3164</u>
Lab File ID:	<u>VV039146.D</u>	Date Analyzed:	<u>09/24/2025</u>
Instrument ID:	<u>MSVOA_V</u>	Time Analyzed:	<u>10:49</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	551620	11.172				
UPPER LIMIT	1103240	11.672				
LOWER LIMIT	275810	10.672				
EPA SAMPLE NO.						
DA-1	589916	11.17				
DA-3	607726	11.17				
VV0924WBL01	379013	11.18				
VV0924WBS02	468522	11.17				
VV0924WBSD02	510716	11.17				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.