

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: ATGG01

Lab Code: ACE

SDG NO.: Q3164

Lab File ID: VV039145.D

BFB Injection Date: 09/24/2025

Instrument ID: MSVOA_V

BFB Injection Time: 09:52

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.6
75	30.0 - 60.0% of mass 95	51.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	1.4 (1.9) 1
174	50.0 - 100.0% of mass 95	77.1
175	5.0 - 9.0% of mass 174	5.2 (6.7) 1
176	95.0 - 101.0% of mass 174	73.5 (95.3) 1
177	5.0 - 9.0% of mass 176	4.9 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VV039146.D	09/24/2025	10:49
VV0924WBL01	VV0924WBL01	VV039148.D	09/24/2025	13:09
VV0924WBS02	VV0924WBS02	VV039150.D	09/24/2025	14:01
DA-1	Q3164-01	VV039160.D	09/24/2025	18:07
DA-3	Q3164-03	VV039162.D	09/24/2025	18:52
VV0924WBSD02	VV0924WBSD02	VV039165.D	09/24/2025	19:59