

ANALYTICAL RESULTS SUMMARY

SEMI-VOLATILE ORGANICS
VOLATILE ORGANICS

PROJECT NAME : NWIRP BETHPAGE 112G08005-WE13

TETRA TECH NUS, INC.

661 Andersen Drive

Suite 200

Pittsburgh, PA - 15220-2745

Phone No: 412-921-7090

ORDER ID : Q3165

ATTENTION : Ernie Wu



Laboratory Certification ID # 20012



1) Signature Page	3
2) Case Narrative	4
2.1) VOC-TCLVOA-10- Case Narrative	4
2.2) SVOC-SIMGroup1- Case Narrative	7
3) Qualifier Page	9
4) QA Checklist	10
5) VOC-TCLVOA-10 Data	11
6) SVOC-SIMGroup1 Data	71
7) Shipping Document	95
7.1) CHAIN OF CUSTODY	96
7.2) Lab Certificate	97
7.3) Internal COC	98

1
2
3
4
5
6
7

Cover Page

Order ID : Q3165

Project ID : NWIRP Bethpage 112G08005-WE13

Client : Tetra Tech NUS, Inc.

Lab Sample Number

Q3165-01
Q3165-02
Q3165-03
Q3165-04
Q3165-05
Q3165-06
Q3165-07
Q3165-08
Q3165-09
Q3165-10

Client Sample Number

BP-VPB-184-TB-20250919
VPB184-HYD-20250919
BP-VPB-184-GW-240-242
BP-VPB-184-GW-260-262
BP-VPB-184-GW-280-282
BP-VPB-184-DUP-20250922
BP-VPB-184-GW-300-302
BP-VPB-184-GW-320-322
BP-VPB-184-GW-320-322MS
BP-VPB-184-GW-320-322MSD

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature :

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 10:31 am, Oct 06, 2025

Date: 10/4/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

Tetra Tech NUS, Inc.

Project Name: NWIRP Bethpage 112G08005-WE13

Project Manager: Ernie Wu

Order ID # Q3165

Test Name: VOC-TCLVOA-10

A. Number of Samples and Date of Receipt:

10 Water samples were received on 09/22/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: VOC-TCLVOA-10. This data package contains results for VOC-TCLVOA-10.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_X were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI. The analysis of VOC-TCLVOA-10 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis except for BP-VPB-184-GW-320-322MSD [1,2-Dichloroethane-d4 - 124%] but original sample and MS passing for surrogate therefore no corrective action taken.

The Internal Standards Areas were met for all analysis.

The Retention Times were met for all analysis.

The MS recoveries met the requirements for all compounds.

The MSD {Q3165-10MSD} with File ID: VX047822.D recoveries met the requirements for all compounds except for 1,1,2,2-Tetrachloroethane[153%], 1,1,2-Trichloroethane[129%], 1,1-Dichloroethane[130%], 1,2-Dibromo-3-Chloropropane[137%], 1,2-Dibromoethane[127%], 2-Butanone[144%], 4-Methyl-2-Pentanone[140%], Bromochloromethane[133%], Carbon Tetrachloride[71%], Chloroform[139%], cis-1,2-Dichloroethene[127%], Methyl tert-butyl Ether[136%], Methylcyclohexane[54%] and Methylene Chloride[134%] due to matrix interference.

The sample # BP-VPB-184-GW-320-322MS and BP-VPB-184-GW-320-322MSD is failing for some compounds and the original sample(BP-VPB-184-GW-320-322) is reported with M flag for these compounds.

The RPD for {Q3165-10MSD} with File ID: VX047822.D met criteria except for 1,1,2,2-Tetrachloroethane[45%], 1,1,2-Trichloroethane[23%], 1,2-Dibromo-3-Chloropropane[39%], 1,2-Dichlorobenzene[23%], 1,2-Dichloroethane[21%], 2-Butanone[29%], 2-Hexanone[27%], 4-Methyl-2-Pentanone[22%], Acetone[23%], Bromomethane[23%], Carbon Tetrachloride[26%], Chloroethane[22%], Chloroform[25%], Chloromethane[24%], Cyclohexane[22%], Methyl Acetate[30%], Methyl tert-butyl Ether[26%], Methylcyclohexane[39%] and Methylene Chloride[24%] due to difference in results of MS-MSD.

The Blank Spike met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration File ID VX047795.D met the requirements except for Dichlorodifluoromethane and Methyl Acetate failing high but no positive hit in associated samples therefore no corrective action taken.

The Tuning criteria met requirements.

E. Additional Comments:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is)."

The Sample #BP-VPB-184-GW-260-262 have the concentration of target compound below Method detection limits, therefore it is not reported as Hit in Form1.

The not QT review data is reported in the Miscellaneous.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 10:31 am, Oct 06, 2025

Signature_____

CASE NARRATIVE

Tetra Tech NUS, Inc.

Project Name: NWIRP Bethpage 112G08005-WE13

Project Manager : Ernie Wu

Order ID # Q3165

Test Name: SVOC-SIMGroup1

A. Number of Samples and Date of Receipt:

3 Water samples were received on 09/22/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1. This data package contains results for SVOC-SIMGroup1.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_N using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGA. The analysis of SVOC-SIMGroup1 was based on method 8270-Modified and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis except for, VPB184-HYD-20250919 [2-Methylnaphthalene-d10 - 1%, Fluoranthene-d10 - 16%]. Surrogates are failed due to matrix interference as seen in the chromatogram. Therefor no corrective action was taken.

The Internal Standards Areas were met for all analysis except for, BP-VPB-184-GW-260-262, BP-VPB-184-GW-300-302. Failed internal standard are not associated with reporting list, therefor no further corrective action was taken. The Retention Times were met for all analysis.

The RPD were met for all analysis.

The Blank Spike met requirements for all compounds.

The Blank Spike Duplicate met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

E. Additional Comments:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is).”

The Sample VPB184-HYD-20250919 has the concentration of target compound below Method detection limit; therefore it is not reported as Hit in Form1.

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 10:31 am, Oct 06, 2025

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q3165

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 10/04/2025

LAB CHRONICLE

OrderID:	Q3165	OrderDate:	9/22/2025 4:07:00 PM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	J43,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3165-01	BP-VPB-184-TB-2025 0919	Water			09/19/25			09/22/25
			VOC-TCLVOA-10	8260-Low			09/25/25	
Q3165-02	VPB184-HYD-202509 19	Water			09/19/25			09/22/25
			VOC-TCLVOA-10	8260-Low			09/25/25	
Q3165-03	BP-VPB-184-GW-240- 242	Water			09/19/25			09/22/25
			VOC-TCLVOA-10	8260-Low			09/25/25	
Q3165-04	BP-VPB-184-GW-260- 262	Water			09/19/25			09/22/25
			VOC-TCLVOA-10	8260-Low			09/25/25	
Q3165-05	BP-VPB-184-GW-280- 282	Water			09/22/25			09/22/25
			VOC-TCLVOA-10	8260-Low			09/25/25	
Q3165-06	BP-VPB-184-DUP-202 50922	Water			09/22/25			09/22/25
			VOC-TCLVOA-10	8260-Low			09/25/25	
Q3165-07	BP-VPB-184-GW-300- 302	Water			09/22/25			09/22/25
			VOC-TCLVOA-10	8260-Low			09/25/25	
Q3165-08	BP-VPB-184-GW-320- 322	Water			09/22/25			09/22/25
			VOC-TCLVOA-10	8260-Low			09/25/25	

Hit Summary Sheet
SW-846

SDG No.: Q3165
Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID: BP-VPB-184-GW-240-242									
Q3165-03	BP-VPB-184-GW-2 Water		Chloromethane	0.67	J	0.32	0.50	1.00	ug/L
Q3165-03	BP-VPB-184-GW-2 Water		Acetone	10.8		1.50	3.80	5.00	ug/L
Q3165-03	BP-VPB-184-GW-2 Water		Carbon Disulfide	0.60	J	0.21	0.75	1.00	ug/L
			Total Voc :	12.1					
			Total Concentration:	12.1					
Client ID: BP-VPB-184-GW-260-262									
Q3165-04	BP-VPB-184-GW-2 Water		Acetone	8.30		1.50	3.80	5.00	ug/L
Q3165-04	BP-VPB-184-GW-2 Water		Carbon Disulfide	0.33	J	0.21	0.75	1.00	ug/L
			Total Voc :	8.63					
			Total Concentration:	8.63					
Client ID: BP-VPB-184-GW-280-282									
Q3165-05	BP-VPB-184-GW-2 Water		Chloromethane	0.57	J	0.32	0.50	1.00	ug/L
Q3165-05	BP-VPB-184-GW-2 Water		Acetone	9.40		1.50	3.80	5.00	ug/L
Q3165-05	BP-VPB-184-GW-2 Water		Carbon Disulfide	0.74	J	0.21	0.75	1.00	ug/L
Q3165-05	BP-VPB-184-GW-2 Water		2-Butanone	2.00	J	0.98	2.50	5.00	ug/L
			Total Voc :	12.7					
			Total Concentration:	12.7					
Client ID: BP-VPB-184-DUP-20250922									
Q3165-06	BP-VPB-184-DUP- Water		Chloromethane	0.63	J	0.32	0.50	1.00	ug/L
Q3165-06	BP-VPB-184-DUP- Water		Acetone	9.50		1.50	3.80	5.00	ug/L
Q3165-06	BP-VPB-184-DUP- Water		Carbon Disulfide	0.80	J	0.21	0.75	1.00	ug/L
			Total Voc :	10.9					
			Total Concentration:	10.9					
Client ID: BP-VPB-184-GW-300-302									
Q3165-07	BP-VPB-184-GW-3 Water		Chloromethane	0.74	J	0.32	0.50	1.00	ug/L
Q3165-07	BP-VPB-184-GW-3 Water		Acetone	16.2		1.50	3.80	5.00	ug/L
Q3165-07	BP-VPB-184-GW-3 Water		Carbon Disulfide	0.71	J	0.21	0.75	1.00	ug/L
Q3165-07	BP-VPB-184-GW-3 Water		2-Butanone	3.50	J	0.98	2.50	5.00	ug/L
			Total Voc :	21.1					
			Total Concentration:	21.1					
Client ID: BP-VPB-184-GW-320-322									
Q3165-08	BP-VPB-184-GW-3 Water		Chloromethane	0.46	J	0.32	0.50	1.00	ug/L
Q3165-08	BP-VPB-184-GW-3 Water		Acetone	8.70		1.50	3.80	5.00	ug/L
Q3165-08	BP-VPB-184-GW-3 Water		Carbon Disulfide	0.40	J	0.21	0.75	1.00	ug/L
			Total Voc :	9.56					
			Total Concentration:	9.56					



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/19/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25	
Client Sample ID:	BP-VPB-184-TB-20250919		SDG No.:	Q3165	
Lab Sample ID:	Q3165-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047801.D	1	09/25/25 11:24	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/19/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25	
Client Sample ID:	BP-VPB-184-TB-20250919		SDG No.:	Q3165	
Lab Sample ID:	Q3165-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047801.D	1	09/25/25 11:24	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	50.5		81 - 118		101%	SPK: 50
1868-53-7	Dibromofluoromethane	47.1		80 - 119		94%	SPK: 50
2037-26-5	Toluene-d8	47.9		89 - 112		96%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.4		85 - 114		105%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	120000	5.544				
540-36-3	1,4-Difluorobenzene	255000	6.745				
3114-55-4	Chlorobenzene-d5	257000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	128000	12.006				
TENTATIVE IDENTIFIED COMPOUNDS							

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/19/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25
Client Sample ID:	BP-VPB-184-TB-20250919		SDG No.:	Q3165
Lab Sample ID:	Q3165-01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047801.D	1	09/25/25 11:24	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
75-43-4	Dichlorofluoromethane	N.D					

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/19/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25	
Client Sample ID:	VPB184-HYD-20250919		SDG No.:	Q3165	
Lab Sample ID:	Q3165-02		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047800.D	1	09/25/25 11:03	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/19/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25	
Client Sample ID:	VPB184-HYD-20250919		SDG No.:	Q3165	
Lab Sample ID:	Q3165-02		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047800.D	1	09/25/25 11:03	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	49.7		81 - 118		99%	SPK: 50
1868-53-7	Dibromofluoromethane	47.1		80 - 119		94%	SPK: 50
2037-26-5	Toluene-d8	47.8		89 - 112		96%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.8		85 - 114		106%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	120000	5.537				
540-36-3	1,4-Difluorobenzene	251000	6.745				
3114-55-4	Chlorobenzene-d5	253000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	125000	12.006				
TENTATIVE IDENTIFIED COMPOUNDS							

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/19/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25
Client Sample ID:	VPB184-HYD-20250919		SDG No.:	Q3165
Lab Sample ID:	Q3165-02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047800.D	1	09/25/25 11:03	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
75-43-4	Dichlorofluoromethane	N.D					

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	09/19/25	
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	09/22/25	
Client Sample ID:	BP-VPB-184-GW-240-242			SDG No.:	Q3165	
Lab Sample ID:	Q3165-03			Matrix:	Water	
Analytical Method:	8260D			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID :	0.18	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047802.D	1	09/25/25 11:46	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.67	J	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	10.8		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.60	J	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	09/19/25	
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	09/22/25	
Client Sample ID:	BP-VPB-184-GW-240-242			SDG No.:	Q3165	
Lab Sample ID:	Q3165-03			Matrix:	Water	
Analytical Method:	8260D			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID :	0.18	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047802.D	1	09/25/25 11:46	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	49.8		81 - 118		100%	SPK: 50
1868-53-7	Dibromofluoromethane	47.1		80 - 119		94%	SPK: 50
2037-26-5	Toluene-d8	47.4		89 - 112		95%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.6		85 - 114		103%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	124000	5.544				
540-36-3	1,4-Difluorobenzene	260000	6.751				
3114-55-4	Chlorobenzene-d5	265000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	130000	12.006				
TENTATIVE IDENTIFIED COMPOUNDS							

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/19/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25	
Client Sample ID:	BP-VPB-184-GW-240-242		SDG No.:	Q3165	
Lab Sample ID:	Q3165-03		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047802.D	1	09/25/25 11:46	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
75-43-4	Dichlorofluoromethane	N.D					

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/19/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25	
Client Sample ID:	BP-VPB-184-GW-260-262		SDG No.:	Q3165	
Lab Sample ID:	Q3165-04		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047803.D	1	09/25/25 12:07	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	8.30		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.33	J	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/19/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25	
Client Sample ID:	BP-VPB-184-GW-260-262		SDG No.:	Q3165	
Lab Sample ID:	Q3165-04		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047803.D	1	09/25/25 12:07	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	54.7		81 - 118		109%	SPK: 50
1868-53-7	Dibromofluoromethane	49.8		80 - 119		100%	SPK: 50
2037-26-5	Toluene-d8	48.7		89 - 112		97%	SPK: 50
460-00-4	4-Bromofluorobenzene	56.7		85 - 114		113%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	129000	5.544				
540-36-3	1,4-Difluorobenzene	276000	6.751				
3114-55-4	Chlorobenzene-d5	289000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	152000	12.006				
TENTATIVE IDENTIFIED COMPOUNDS							

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/19/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25
Client Sample ID:	BP-VPB-184-GW-260-262		SDG No.:	Q3165
Lab Sample ID:	Q3165-04		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047803.D	1	09/25/25 12:07	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
75-43-4	Dichlorofluoromethane	N.D					

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/22/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25	
Client Sample ID:	BP-VPB-184-GW-280-282		SDG No.:	Q3165	
Lab Sample ID:	Q3165-05		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047804.D	1	09/25/25 12:28	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.57	J	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	9.40		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.74	J	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.00	J	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/22/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25	
Client Sample ID:	BP-VPB-184-GW-280-282		SDG No.:	Q3165	
Lab Sample ID:	Q3165-05		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047804.D	1	09/25/25 12:28	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	51.7		81 - 118		103%	SPK: 50
1868-53-7	Dibromofluoromethane	48.6		80 - 119		97%	SPK: 50
2037-26-5	Toluene-d8	48.2		89 - 112		96%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.3		85 - 114		109%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	122000	5.544				
540-36-3	1,4-Difluorobenzene	261000	6.751				
3114-55-4	Chlorobenzene-d5	267000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	135000	12.006				
TENTATIVE IDENTIFIED COMPOUNDS							

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/22/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25	
Client Sample ID:	BP-VPB-184-GW-280-282		SDG No.:	Q3165	
Lab Sample ID:	Q3165-05		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047804.D	1	09/25/25 12:28	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
75-43-4	Dichlorofluoromethane	N.D					

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/22/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25	
Client Sample ID:	BP-VPB-184-DUP-20250922		SDG No.:	Q3165	
Lab Sample ID:	Q3165-06		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047805.D	1	09/25/25 12:49	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.63	J	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	9.50		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.80	J	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/22/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25	
Client Sample ID:	BP-VPB-184-DUP-20250922		SDG No.:	Q3165	
Lab Sample ID:	Q3165-06		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047805.D	1	09/25/25 12:49	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	54.4		81 - 118		109%	SPK: 50
1868-53-7	Dibromofluoromethane	49.9		80 - 119		100%	SPK: 50
2037-26-5	Toluene-d8	49.2		89 - 112		98%	SPK: 50
460-00-4	4-Bromofluorobenzene	55.5		85 - 114		111%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	118000	5.543				
540-36-3	1,4-Difluorobenzene	253000	6.744				
3114-55-4	Chlorobenzene-d5	262000	10.036				
3855-82-1	1,4-Dichlorobenzene-d4	135000	12.006				
TENTATIVE IDENTIFIED COMPOUNDS							

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/22/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25	
Client Sample ID:	BP-VPB-184-DUP-20250922		SDG No.:	Q3165	
Lab Sample ID:	Q3165-06		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047805.D	1	09/25/25 12:49	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
75-43-4	Dichlorofluoromethane	N.D					

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/22/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25	
Client Sample ID:	BP-VPB-184-GW-300-302		SDG No.:	Q3165	
Lab Sample ID:	Q3165-07		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047806.D	1	09/25/25 13:10	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.74	J	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	16.2		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.71	J	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	3.50	J	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/22/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25	
Client Sample ID:	BP-VPB-184-GW-300-302		SDG No.:	Q3165	
Lab Sample ID:	Q3165-07		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047806.D	1	09/25/25 13:10	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.9		81 - 118		106%	SPK: 50
1868-53-7	Dibromofluoromethane	49.7		80 - 119		99%	SPK: 50
2037-26-5	Toluene-d8	49.6		89 - 112		99%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.3		85 - 114		109%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	119000	5.544				
540-36-3	1,4-Difluorobenzene	254000	6.745				
3114-55-4	Chlorobenzene-d5	262000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	127000	12.006				
TENTATIVE IDENTIFIED COMPOUNDS							

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/22/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25
Client Sample ID:	BP-VPB-184-GW-300-302		SDG No.:	Q3165
Lab Sample ID:	Q3165-07		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047806.D	1	09/25/25 13:10	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
75-43-4	Dichlorofluoromethane	N.D					

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	09/22/25
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	09/22/25
Client Sample ID:	BP-VPB-184-GW-320-322			SDG No.:	Q3165
Lab Sample ID:	Q3165-08			Matrix:	Water
Analytical Method:	8260D			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID :	0.18	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047820.D	1	09/25/25 18:06	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.46	J	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	8.70		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.40	J	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	UM	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	UM	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	UM	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	UM	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	UM	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	UM	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	UM	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	UM	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	UM	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	UM	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/22/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25	
Client Sample ID:	BP-VPB-184-GW-320-322		SDG No.:	Q3165	
Lab Sample ID:	Q3165-08		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047820.D	1	09/25/25 18:06	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	UM	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	UM	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	UM	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	UM	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.9		81 - 118		108%	SPK: 50
1868-53-7	Dibromofluoromethane	50.4		80 - 119		101%	SPK: 50
2037-26-5	Toluene-d8	49.5		89 - 112		99%	SPK: 50
460-00-4	4-Bromofluorobenzene	55.6		85 - 114		111%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	130000	5.538				
540-36-3	1,4-Difluorobenzene	276000	6.751				
3114-55-4	Chlorobenzene-d5	286000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	146000	12.006				
TENTATIVE IDENTIFIED COMPOUNDS							

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/22/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25
Client Sample ID:	BP-VPB-184-GW-320-322		SDG No.:	Q3165
Lab Sample ID:	Q3165-08		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047820.D	1	09/25/25 18:06	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
75-43-4	Dichlorofluoromethane	N.D					

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: **Q3165**

Client: **Tetra Tech NUS, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3165-01	BP-VPB-184-TB-20250919	1,2-Dichloroethane-d4	50	50.5	101		81	118
		Dibromofluoromethane	50	47.1	94		80	119
		Toluene-d8	50	47.9	96		89	112
		4-Bromofluorobenzene	50	52.4	105		85	114
Q3165-02	VPB184-HYD-20250919	1,2-Dichloroethane-d4	50	49.7	99		81	118
		Dibromofluoromethane	50	47.1	94		80	119
		Toluene-d8	50	47.8	96		89	112
		4-Bromofluorobenzene	50	52.8	106		85	114
Q3165-03	BP-VPB-184-GW-240-242	1,2-Dichloroethane-d4	50	49.8	100		81	118
		Dibromofluoromethane	50	47.1	94		80	119
		Toluene-d8	50	47.4	95		89	112
		4-Bromofluorobenzene	50	51.6	103		85	114
Q3165-04	BP-VPB-184-GW-260-262	1,2-Dichloroethane-d4	50	54.7	109		81	118
		Dibromofluoromethane	50	49.8	100		80	119
		Toluene-d8	50	48.7	97		89	112
		4-Bromofluorobenzene	50	56.7	113		85	114
Q3165-05	BP-VPB-184-GW-280-282	1,2-Dichloroethane-d4	50	51.7	103		81	118
		Dibromofluoromethane	50	48.6	97		80	119
		Toluene-d8	50	48.2	96		89	112
		4-Bromofluorobenzene	50	54.3	109		85	114
Q3165-06	BP-VPB-184-DUP-20250922	1,2-Dichloroethane-d4	50	54.4	109		81	118
		Dibromofluoromethane	50	49.9	100		80	119
		Toluene-d8	50	49.2	98		89	112
		4-Bromofluorobenzene	50	55.5	111		85	114
Q3165-07	BP-VPB-184-GW-300-302	1,2-Dichloroethane-d4	50	52.9	106		81	118
		Dibromofluoromethane	50	49.7	99		80	119
		Toluene-d8	50	49.6	99		89	112
		4-Bromofluorobenzene	50	54.3	109		85	114
Q3165-08	BP-VPB-184-GW-320-322	1,2-Dichloroethane-d4	50	53.9	108		81	118
		Dibromofluoromethane	50	50.4	101		80	119
		Toluene-d8	50	49.5	99		89	112
		4-Bromofluorobenzene	50	55.6	111		85	114
Q3165-09MS	BP-VPB-184-GW-320-322MS	1,2-Dichloroethane-d4	50	57.0	114		81	118
		Dibromofluoromethane	50	53.2	106		80	119
		Toluene-d8	50	50.7	101		89	112
		4-Bromofluorobenzene	50	53.3	107		85	114
Q3165-10MSD	BP-VPB-184-GW-320-322MSD	1,2-Dichloroethane-d4	50	61.8	124	*	81	118
		Dibromofluoromethane	50	56.6	113		80	119
		Toluene-d8	50	49.9	100		89	112
		4-Bromofluorobenzene	50	46.0	92		85	114
VX0925WBL01	VX0925WBL01	1,2-Dichloroethane-d4	50	51.1	102		81	118
		Dibromofluoromethane	50	48.4	97		80	119
		Toluene-d8	50	48.3	97		89	112
		4-Bromofluorobenzene	50	53.3	107		85	114
VX0925WBS01	VX0925WBS01	1,2-Dichloroethane-d4	50	53.3	107		81	118
		Dibromofluoromethane	50	52.0	104		80	119
		Toluene-d8	50	50.2	100		89	112
		4-Bromofluorobenzene	50	53.1	106		85	114

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3165

Analytical Method: SW8260-Low

Client: Tetra Tech NUS, Inc.

Datafile : VX047821.D

										Limits		
Parameter	Spike	Sample	Result	Units	Rec		RPD	RPD		Low	High	RPD
		Result			Qual	Qual						
Lab Sample ID : Q3165-09MS		Client Sample ID :		BP-VPB-184-GW-320-322MS								
Dichlorodifluoromethane	50	0	54.6	ug/L	109					32	152	
Chloromethane	50	0.46	53.5	ug/L	106					50	139	
Vinyl chloride	50	0	54.3	ug/L	109					58	137	
Bromomethane	50	0	51.5	ug/L	103					53	141	
Chloroethane	50	0	52.9	ug/L	106					60	138	
Trichlorofluoromethane	50	0	51.7	ug/L	103					65	141	
1,1,2-Trichlorotrifluoroethane	50	0	48.6	ug/L	97					70	136	
1,1-Dichloroethene	50	0	51.8	ug/L	104					71	131	
Acetone	250	8.70	240	ug/L	93					39	160	
Carbon disulfide	50	0.40	51.7	ug/L	103					64	133	
Methyl tert-butyl Ether	50	0	52.5	ug/L	105					71	124	
Methyl Acetate	50	0	41.6	ug/L	83					56	136	
Methylene Chloride	50	0	52.6	ug/L	105					74	124	
trans-1,2-Dichloroethene	50	0	53.5	ug/L	107					75	124	
1,1-Dichloroethane	50	0	53.4	ug/L	107					77	125	
Cyclohexane	50	0	46.7	ug/L	93					71	130	
2-Butanone	250	0	270	ug/L	108					56	143	
Carbon Tetrachloride	50	0	46.4	ug/L	93					72	136	
cis-1,2-Dichloroethene	50	0	52.3	ug/L	105					78	123	
Bromochloromethane	50	0	56.6	ug/L	113					78	123	
Chloroform	50	0	53.8	ug/L	108					79	124	
1,1,1-Trichloroethane	50	0	50.8	ug/L	102					74	131	
Methylcyclohexane	50	0	40.4	ug/L	81					72	132	
Benzene	50	0	49.7	ug/L	99					79	120	
1,2-Dichloroethane	50	0	48.4	ug/L	97					73	128	
Trichloroethene	50	0	48.2	ug/L	96					79	123	
1,2-Dichloropropane	50	0	50.2	ug/L	100					78	122	
Bromodichloromethane	50	0	50.0	ug/L	100					79	125	
4-Methyl-2-Pentanone	250	0	280	ug/L	112					67	130	
Toluene	50	0	48.9	ug/L	98					80	121	
t-1,3-Dichloropropene	50	0	40.7	ug/L	81					73	127	
cis-1,3-Dichloropropene	50	0	41.0	ug/L	82					75	124	
1,1,2-Trichloroethane	50	0	51.5	ug/L	103					80	119	
2-Hexanone	250	0	260	ug/L	104					57	139	
Dibromochloromethane	50	0	51.3	ug/L	103					74	126	
1,2-Dibromoethane	50	0	53.0	ug/L	106					77	121	
Tetrachloroethene	50	0	42.4	ug/L	85					74	129	
Chlorobenzene	50	0	45.7	ug/L	91					82	118	
Ethyl Benzene	50	0	44.9	ug/L	90					79	121	
m/p-Xylenes	100	0	90.5	ug/L	91					80	121	
o-Xylene	50	0	45.5	ug/L	91					78	122	
Styrene	50	0	42.5	ug/L	85					78	123	
Bromoform	50	0	47.8	ug/L	96					66	130	
Isopropylbenzene	50	0	42.3	ug/L	85					72	131	
1,1,2,2-Tetrachloroethane	50	0	48.3	ug/L	97					71	121	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3165

Analytical Method: SW8260-Low

Client: Tetra Tech NUS, Inc.

Datafile : VX047821.D

Parameter	Spike	Sample	Result	Units	Rec			RPD	Limits		
		Result			Rec	Qual	RPD	Qual	Low	High	RPD
1,3-Dichlorobenzene	50	0	42.3	ug/L	85				80	119	
1,4-Dichlorobenzene	50	0	42.3	ug/L	85				79	118	
1,2-Dichlorobenzene	50	0	44.3	ug/L	89				80	119	
1,2-Dibromo-3-Chloropropane	50	0	46.4	ug/L	93				62	128	
1,2,4-Trichlorobenzene	50	0	39.1	ug/L	78				69	130	
1,2,3-Trichlorobenzene	50	0	41.1	ug/L	82				69	129	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3165

Analytical Method: SW8260-Low

Client: Tetra Tech NUS, Inc.

Datafile : VX047822.D

Parameter	Spike	Sample	Result	Units	Rec		RPD		Limits		
		Result			Qual	RPD	Qual	Low	High	RPD	
Lab Sample ID :	Q3165-10MSD	Client Sample ID :	BP-VPB-184-GW-320-322MSD								
Dichlorodifluoromethane	50	0	53.8	ug/L	108		1		32	152	20
Chloromethane	50	0.46	68.2	ug/L	135		24	*	50	139	20
Vinyl chloride	50	0	63.8	ug/L	128		16		58	137	20
Bromomethane	50	0	65.0	ug/L	130		23	*	53	141	20
Chloroethane	50	0	66.2	ug/L	132		22	*	60	138	20
Trichlorofluoromethane	50	0	52.9	ug/L	106		2		65	141	20
1,1,2-Trichlorotrifluoroethane	50	0	41.7	ug/L	83		15		70	136	20
1,1-Dichloroethene	50	0	61.8	ug/L	124		18		71	131	20
Acetone	250	8.70	300	ug/L	117		23	*	39	160	20
Carbon disulfide	50	0.40	44.3	ug/L	88		16		64	133	20
Methyl tert-butyl Ether	50	0	68.1	ug/L	136	*	26	*	71	124	20
Methyl Acetate	50	0	56.1	ug/L	112		30	*	56	136	20
Methylene Chloride	50	0	67.0	ug/L	134	*	24	*	74	124	20
trans-1,2-Dichloroethene	50	0	57.7	ug/L	115		8		75	124	20
1,1-Dichloroethane	50	0	65.2	ug/L	130	*	20		77	125	20
Cyclohexane	50	0	37.5	ug/L	75		22	*	71	130	20
2-Butanone	250	0	360	ug/L	144	*	29	*	56	143	20
Carbon Tetrachloride	50	0	35.6	ug/L	71	*	26	*	72	136	20
cis-1,2-Dichloroethene	50	0	63.6	ug/L	127	*	19		78	123	20
Bromochloromethane	50	0	66.6	ug/L	133	*	16		78	123	20
Chloroform	50	0	69.3	ug/L	139	*	25	*	79	124	20
1,1,1-Trichloroethane	50	0	58.7	ug/L	117		14		74	131	20
Methylcyclohexane	50	0	27.2	ug/L	54	*	39	*	72	132	20
Benzene	50	0	58.2	ug/L	116		16		79	120	20
1,2-Dichloroethane	50	0	59.7	ug/L	119		21	*	73	128	20
Trichloroethene	50	0	56.2	ug/L	112		15		79	123	20
1,2-Dichloropropane	50	0	61.1	ug/L	122		20		78	122	20
Bromodichloromethane	50	0	44.8	ug/L	90		11		79	125	20
4-Methyl-2-Pentanone	250	0	350	ug/L	140	*	22	*	67	130	20
Toluene	50	0	53.2	ug/L	106		8		80	121	20
t-1,3-Dichloropropene	50	0	42.1	ug/L	84		3		73	127	20
cis-1,3-Dichloropropene	50	0	42.4	ug/L	85		3		75	124	20
1,1,2-Trichloroethane	50	0	64.7	ug/L	129	*	23	*	80	119	20
2-Hexanone	250	0	340	ug/L	136		27	*	57	139	20
Dibromochloromethane	50	0	44.0	ug/L	88		15		74	126	20
1,2-Dibromoethane	50	0	63.4	ug/L	127	*	18		77	121	20
Tetrachloroethene	50	0	49.7	ug/L	99		16		74	129	20
Chlorobenzene	50	0	53.9	ug/L	108		16		82	118	20
Ethyl Benzene	50	0	49.0	ug/L	98		9		79	121	20
m/p-Xylenes	100	0	97.5	ug/L	98		7		80	121	20
o-Xylene	50	0	51.7	ug/L	103		13		78	122	20
Styrene	50	0	42.3	ug/L	85		0		78	123	20
Bromoform	50	0	43.1	ug/L	86		10		66	130	20
Isopropylbenzene	50	0	50.2	ug/L	100		17		72	131	20
1,1,2,2-Tetrachloroethane	50	0	76.6	ug/L	153	*	45	*	71	121	20

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3165

Analytical Method: SW8260-Low

Client: Tetra Tech NUS, Inc.

Datafile : VX047822.D

Parameter	Spike	Sample Result	Result	Units	Rec		RPD		Limits		RPD
					Rec	Qual	RPD	Qual	Low	High	
1,3-Dichlorobenzene	50	0	51.1	ug/L	102		19		80	119	20
1,4-Dichlorobenzene	50	0	50.8	ug/L	102		18		79	118	20
1,2-Dichlorobenzene	50	0	55.9	ug/L	112		23	*	80	119	20
1,2-Dibromo-3-Chloropropane	50	0	68.7	ug/L	137	*	39	*	62	128	20
1,2,4-Trichlorobenzene	50	0	42.8	ug/L	86		9		69	130	20
1,2,3-Trichlorobenzene	50	0	48.3	ug/L	97		16		69	129	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3165</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Tetra Tech NUS, Inc.</u>	Datafile :	<u>VX047798.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0925WBS01	Dichlorodifluoromethane	20	22.1	ug/L	111			32	152	
	Chloromethane	20	20.6	ug/L	103			50	139	
	Vinyl chloride	20	20.6	ug/L	103			58	137	
	Bromomethane	20	20.8	ug/L	104			53	141	
	Chloroethane	20	20.2	ug/L	101			60	138	
	Trichlorofluoromethane	20	19.8	ug/L	99			65	141	
	1,1,2-Trichlorotrifluoroethane	20	20.3	ug/L	102			70	136	
	1,1-Dichloroethene	20	19.3	ug/L	97			71	131	
	Acetone	100	96.4	ug/L	96			39	160	
	Carbon disulfide	20	19.6	ug/L	98			64	133	
	Methyl tert-butyl Ether	20	20.0	ug/L	100			71	124	
	Methyl Acetate	20	22.6	ug/L	113			56	136	
	Methylene Chloride	20	20.6	ug/L	103			74	124	
	trans-1,2-Dichloroethene	20	20.2	ug/L	101			75	124	
	1,1-Dichloroethane	20	20.4	ug/L	102			77	125	
	Cyclohexane	20	19.2	ug/L	96			71	130	
	2-Butanone	100	110	ug/L	110			56	143	
	Carbon Tetrachloride	20	18.2	ug/L	91			72	136	
	cis-1,2-Dichloroethene	20	20.3	ug/L	102			78	123	
	Bromochloromethane	20	20.0	ug/L	100			78	123	
	Chloroform	20	20.7	ug/L	104			79	124	
	1,1,1-Trichloroethane	20	19.8	ug/L	99			74	131	
	Methylcyclohexane	20	17.8	ug/L	89			72	132	
	Benzene	20	19.9	ug/L	100			79	120	
	1,2-Dichloroethane	20	19.5	ug/L	98			73	128	
	Trichloroethene	20	19.7	ug/L	99			79	123	
	1,2-Dichloropropane	20	20.1	ug/L	101			78	122	
	Bromodichloromethane	20	19.5	ug/L	98			79	125	
	4-Methyl-2-Pentanone	100	110	ug/L	110			67	130	
	Toluene	20	20.0	ug/L	100			80	121	
	t-1,3-Dichloropropene	20	18.9	ug/L	95			73	127	
	cis-1,3-Dichloropropene	20	19.3	ug/L	97			75	124	
	1,1,2-Trichloroethane	20	21.0	ug/L	105			80	119	
	2-Hexanone	100	110	ug/L	110			57	139	
	Dibromochloromethane	20	20.5	ug/L	103			74	126	
	1,2-Dibromoethane	20	20.8	ug/L	104			77	121	
	Tetrachloroethene	20	18.7	ug/L	94			74	129	
	Chlorobenzene	20	19.0	ug/L	95			82	118	
	Ethyl Benzene	20	18.1	ug/L	91			79	121	
	m/p-Xylenes	40	37.6	ug/L	94			80	121	
	o-Xylene	20	18.5	ug/L	93			78	122	
	Styrene	20	18.9	ug/L	95			78	123	
	Bromoform	20	19.6	ug/L	98			66	130	
	Isopropylbenzene	20	16.8	ug/L	84			72	131	
	1,1,2,2-Tetrachloroethane	20	18.9	ug/L	95			71	121	
	1,3-Dichlorobenzene	20	17.5	ug/L	88			80	119	
	1,4-Dichlorobenzene	20	17.8	ug/L	89			79	118	
	1,2-Dichlorobenzene	20	17.8	ug/L	89			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3165</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Tetra Tech NUS, Inc.</u>	Datafile :	<u>VX047798.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0925WBS01	1,2-Dibromo-3-Chloropropane	20	17.6	ug/L	88			62	128	
	1,2,4-Trichlorobenzene	20	16.2	ug/L	81			69	130	
	1,2,3-Trichlorobenzene	20	16.9	ug/L	85			69	129	

VOLATILE METHOD BLANK SUMMARY

Client ID

VX0925WBL01

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q3165

Lab File ID: VX047797.D

Lab Sample ID: VX0925WBL01

Date Analyzed: 09/25/2025

Time Analyzed: 09:46

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0925WBS01	VX0925WBS01	VX047798.D	09/25/2025
VPB184-HYD-20250919	Q3165-02	VX047800.D	09/25/2025
BP-VPB-184-TB-20250919	Q3165-01	VX047801.D	09/25/2025
BP-VPB-184-GW-240-242	Q3165-03	VX047802.D	09/25/2025
BP-VPB-184-GW-260-262	Q3165-04	VX047803.D	09/25/2025
BP-VPB-184-GW-280-282	Q3165-05	VX047804.D	09/25/2025
BP-VPB-184-DUP-20250922	Q3165-06	VX047805.D	09/25/2025
BP-VPB-184-GW-300-302	Q3165-07	VX047806.D	09/25/2025
BP-VPB-184-GW-320-322	Q3165-08	VX047820.D	09/25/2025
BP-VPB-184-GW-320-322MS	Q3165-09MS	VX047821.D	09/25/2025
BP-VPB-184-GW-320-322MSD	Q3165-10MSD	VX047822.D	09/25/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: VX047584.D
Instrument ID: MSVOA_X
GC Column: DB-624UI ID: 0.18 (mm)

Contract: TETR06
SDG NO.: Q3165
BFB Injection Date: 09/16/2025
BFB Injection Time: 08:56
Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20
75	30.0 - 60.0% of mass 95	54
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7
173	Less than 2.0% of mass 174	0.7 (1) 1
174	50.0 - 100.0% of mass 95	67.3
175	5.0 - 9.0% of mass 174	5.4 (8.1) 1
176	95.0 - 101.0% of mass 174	66.3 (98.6) 1
177	5.0 - 9.0% of mass 176	4.3 (6.4) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX047585.D	09/16/2025	09:23
VSTDIC005	VSTDIC005	VX047586.D	09/16/2025	10:16
VSTDIC020	VSTDIC020	VX047587.D	09/16/2025	10:37
VSTDIC050	VSTDIC050	VX047588.D	09/16/2025	10:58
VSTDIC100	VSTDIC100	VX047589.D	09/16/2025	11:40
VSTDIC150	VSTDIC150	VX047590.D	09/16/2025	12:01

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: VX047794.D
Instrument ID: MSVOA_X
GC Column: DB-624UI ID: 0.18 (mm)

Contract: TETR06
SDG NO.: Q3165
BFB Injection Date: 09/25/2025
BFB Injection Time: 08:24
Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20.1
75	30.0 - 60.0% of mass 95	52
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7
173	Less than 2.0% of mass 174	0.7 (1) 1
174	50.0 - 100.0% of mass 95	69.9
175	5.0 - 9.0% of mass 174	5.4 (7.8) 1
176	95.0 - 101.0% of mass 174	67 (95.7) 1
177	5.0 - 9.0% of mass 176	4.6 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX047795.D	09/25/2025	08:56
VX0925WBL01	VX0925WBL01	VX047797.D	09/25/2025	09:46
VX0925WBS01	VX0925WBS01	VX047798.D	09/25/2025	10:13
VPB184-HYD-20250919	Q3165-02	VX047800.D	09/25/2025	11:03
BP-VPB-184-TB-20250919	Q3165-01	VX047801.D	09/25/2025	11:24
BP-VPB-184-GW-240-242	Q3165-03	VX047802.D	09/25/2025	11:46
BP-VPB-184-GW-260-262	Q3165-04	VX047803.D	09/25/2025	12:07
BP-VPB-184-GW-280-282	Q3165-05	VX047804.D	09/25/2025	12:28
BP-VPB-184-DUP-20250922	Q3165-06	VX047805.D	09/25/2025	12:49
BP-VPB-184-GW-300-302	Q3165-07	VX047806.D	09/25/2025	13:10
BP-VPB-184-GW-320-322	Q3165-08	VX047820.D	09/25/2025	18:06
BP-VPB-184-GW-320-322MS	Q3165-09MS	VX047821.D	09/25/2025	18:27
BP-VPB-184-GW-320-322MSD	Q3165-10MSD	VX047822.D	09/25/2025	18:48
VSTDCCC050EC	VSTDCCC050	VX047823.D	09/25/2025	19:09

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TETR06
Lab Code: ACE SDG NO.: Q3165
Lab File ID: VX047795.D Date Analyzed: 09/25/2025
Instrument ID: MSVOA_X Time Analyzed: 08:56
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	136107	5.53	245940	6.75	220262	10.04
UPPER LIMIT	272214	6.032	491880	7.245	440524	10.537
LOWER LIMIT	68053.5	5.032	122970	6.245	110131	9.537
EPA SAMPLE NO.						
BP-VPB-184-TB-20250919	119517	5.54	255275	6.75	257309	10.04
VPB184-HYD-20250919	120300	5.54	251100	6.75	252620	10.04
BP-VPB-184-GW-240-242	123778	5.54	260434	6.75	265025	10.04
BP-VPB-184-GW-260-262	129350	5.54	276238	6.75	289210	10.04
BP-VPB-184-GW-280-282	122004	5.54	260982	6.75	266894	10.04
BP-VPB-184-DUP-20250922	118043	5.54	252631	6.74	261796	10.04
BP-VPB-184-GW-300-302	119390	5.54	254220	6.75	261705	10.04
BP-VPB-184-GW-320-322	130467	5.54	276215	6.75	285610	10.04
BP-VPB-184-GW-320-322MS	135494	5.54	257414	6.75	243581	10.04
BP-VPB-184-GW-320-322MSD	107908	5.54	210706	6.75	179962	10.04
VX0925WBL01	117291	5.54	245721	6.75	251657	10.04
VX0925WBS01	146604	5.54	268605	6.75	252515	10.04

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3165</u>
Lab File ID:	<u>VX047795.D</u>	Date Analyzed:	<u>09/25/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>08:56</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	107538	12.00€				
UPPER LIMIT	215076	12.50€				
LOWER LIMIT	53769	11.50€				
EPA SAMPLE NO.						
BP-VPB-184-TB-20250919	127824	12.01				
VPB184-HYD-20250919	124631	12.01				
BP-VPB-184-GW-240-242	130114	12.01				
BP-VPB-184-GW-260-262	151799	12.01				
BP-VPB-184-GW-280-282	134705	12.01				
BP-VPB-184-DUP-20250922	135087	12.01				
BP-VPB-184-GW-300-302	127218	12.01				
BP-VPB-184-GW-320-322	145772	12.01				
BP-VPB-184-GW-320-322MS	121955	12.01				
BP-VPB-184-GW-320-322MSD	77374	12.01				
VX0925WBL01	126490	12.01				
VX0925WBS01	131521	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0925WBL01		SDG No.:	Q3165
Lab Sample ID:	VX0925WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047797.D	1	09/25/25 09:46	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0925WBL01		SDG No.:	Q3165
Lab Sample ID:	VX0925WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047797.D	1	09/25/25 09:46	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	51.1		81 - 118		102%	SPK: 50
1868-53-7	Dibromofluoromethane	48.4		80 - 119		97%	SPK: 50
2037-26-5	Toluene-d8	48.3		89 - 112		97%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.3		85 - 114		107%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	117000	5.538				
540-36-3	1,4-Difluorobenzene	246000	6.745				
3114-55-4	Chlorobenzene-d5	252000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	126000	12.006				

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0925WBL01		SDG No.:	Q3165
Lab Sample ID:	VX0925WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047797.D	1	09/25/25 09:46	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0925WBS01		SDG No.:	Q3165
Lab Sample ID:	VX0925WBS01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047798.D	1	09/25/25 10:13	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	22.1		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	20.6		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	20.6		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	20.8		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	20.2		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.8		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	20.3		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.3		0.23	0.75	1.00	ug/L
67-64-1	Acetone	96.4		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	19.6		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	20.0		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	22.6		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	20.6		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	20.2		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	20.4		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	19.2		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	110		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	18.2		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	20.3		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	20.0		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	20.7		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.8		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	17.8		0.16	0.50	1.00	ug/L
71-43-2	Benzene	19.9		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	19.5		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	19.7		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	20.1		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	19.5		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	110		0.68	2.50	5.00	ug/L
108-88-3	Toluene	20.0		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0925WBS01		SDG No.:	Q3165
Lab Sample ID:	VX0925WBS01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047798.D	1	09/25/25 10:13	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	18.9		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	19.3		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	21.0		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	110		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	20.5		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	20.8		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	18.7		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	19.0		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	18.1		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	37.6		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	18.5		0.12	0.50	1.00	ug/L
100-42-5	Styrene	18.9		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	19.6		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	16.8		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	18.9		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	17.5		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	17.8		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	17.8		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	17.6		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	16.2		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	16.9		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.3		81 - 118		107%	SPK: 50
1868-53-7	Dibromofluoromethane	52.0		80 - 119		104%	SPK: 50
2037-26-5	Toluene-d8	50.2		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.1		85 - 114		106%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	147000	5.538				
540-36-3	1,4-Difluorobenzene	269000	6.745				
3114-55-4	Chlorobenzene-d5	253000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	132000	12.006				

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0925WBS01		SDG No.:	Q3165
Lab Sample ID:	VX0925WBS01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047798.D	1	09/25/25 10:13	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/22/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25	
Client Sample ID:	BP-VPB-184-GW-320-322MS		SDG No.:	Q3165	
Lab Sample ID:	Q3165-09MS		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047821.D	1	09/25/25 18:27	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	54.6		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	53.5		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	54.3		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	51.5		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	52.9		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	51.7		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	48.6		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	51.8		0.23	0.75	1.00	ug/L
67-64-1	Acetone	240		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	51.7		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	52.5		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	41.6		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	52.6		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	53.5		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	53.4		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	46.7		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	270		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	46.4		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	52.3		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	56.6		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	53.8		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	50.8		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	40.4		0.16	0.50	1.00	ug/L
71-43-2	Benzene	49.7		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	48.4		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	48.2		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	50.2		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	50.0		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	280		0.68	2.50	5.00	ug/L
108-88-3	Toluene	48.9		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/22/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25	
Client Sample ID:	BP-VPB-184-GW-320-322MS		SDG No.:	Q3165	
Lab Sample ID:	Q3165-09MS		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047821.D	1	09/25/25 18:27	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	40.7		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	41.0		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	51.5		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	260		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	51.3		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	53.0		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	42.4		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	45.7		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	44.9		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	90.5		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	45.5		0.12	0.50	1.00	ug/L
100-42-5	Styrene	42.5		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	47.8		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	42.3		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	48.3		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	42.3		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	42.3		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	44.3		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	46.4		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	39.1		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	41.1		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	57.0		81 - 118		114%	SPK: 50
1868-53-7	Dibromofluoromethane	53.2		80 - 119		106%	SPK: 50
2037-26-5	Toluene-d8	50.7		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.3		85 - 114		107%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	135000	5.538				
540-36-3	1,4-Difluorobenzene	257000	6.751				
3114-55-4	Chlorobenzene-d5	244000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	122000	12.006				

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/22/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25	
Client Sample ID:	BP-VPB-184-GW-320-322MS		SDG No.:	Q3165	
Lab Sample ID:	Q3165-09MS		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047821.D	1	09/25/25 18:27	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/22/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25	
Client Sample ID:	BP-VPB-184-GW-320-322MSD		SDG No.:	Q3165	
Lab Sample ID:	Q3165-10MSD		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047822.D	1	09/25/25 18:48	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	53.8		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	68.2		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	63.8		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	65.0		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	66.2		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	52.9		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	41.7		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	61.8		0.23	0.75	1.00	ug/L
67-64-1	Acetone	300		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	44.3		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	68.1		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	56.1		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	67.0		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	57.7		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	65.2		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	37.5		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	360		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	35.6		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	63.6		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	66.6		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	69.3		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	58.7		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	27.2		0.16	0.50	1.00	ug/L
71-43-2	Benzene	58.2		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	59.7		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	56.2		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	61.1		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	44.8		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	350		0.68	2.50	5.00	ug/L
108-88-3	Toluene	53.2		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/22/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25	
Client Sample ID:	BP-VPB-184-GW-320-322MSD		SDG No.:	Q3165	
Lab Sample ID:	Q3165-10MSD		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047822.D	1	09/25/25 18:48	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	42.1		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	42.4		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	64.7		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	340		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	44.0		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	63.4		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	49.7		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	53.9		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	49.0		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	97.5		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	51.7		0.12	0.50	1.00	ug/L
100-42-5	Styrene	42.3		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	43.1		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	50.2		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	76.6		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	51.1		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	50.8		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	55.9		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	68.7		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	42.8		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	48.3		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	61.8	*	81 - 118		124%	SPK: 50
1868-53-7	Dibromofluoromethane	56.6		80 - 119		113%	SPK: 50
2037-26-5	Toluene-d8	49.9		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	46.0		85 - 114		92%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	108000	5.544				
540-36-3	1,4-Difluorobenzene	211000	6.751				
3114-55-4	Chlorobenzene-d5	180000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	77400	12.006				

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/22/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/22/25
Client Sample ID:	BP-VPB-184-GW-320-322MSD		SDG No.:	Q3165
Lab Sample ID:	Q3165-10MSD		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047822.D	1	09/25/25 18:48	VX092525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: TETR06
Lab Code: ACE SDG No.: Q3165
Instrument ID: MSVOA_X Calibration Date(s): 09/16/2025 09/16/2025
Heated Purge: (Y/N) N Calibration Time(s): 09:23 12:01
GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:	RRF001 = VX047585.D	RRF005 = VX047586.D	RRF020 = VX047587.D					
	RRF050 = VX047588.D	RRF100 = VX047589.D	RRF150 = VX047590.D					
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Dichlorodifluoromethane	0.447	0.575	0.582	0.501	0.510	0.492	0.518	10
Chloromethane	0.610	0.749	0.739	0.661	0.677	0.647	0.680	7.9
Vinyl Chloride	0.572	0.713	0.716	0.661	0.677	0.658	0.666	7.9
Bromomethane		0.493	0.451	0.424	0.423	0.321	0.422	15
Chloroethane	0.433	0.450	0.469	0.443	0.450	0.426	0.445	3.4
Trichlorofluoromethane	0.865	1.033	1.038	0.995	1.018	0.986	0.989	6.5
1,1,2-Trichlorotrifluoroethane	0.519	0.591	0.603	0.583	0.590	0.583	0.578	5.2
1,1-Dichloroethene	0.526	0.612	0.618	0.604	0.609	0.596	0.594	5.7
Acetone	0.343	0.416	0.400	0.317	0.302	0.299	0.346	14.7
Carbon Disulfide	1.555	1.735	1.758	1.566	1.589	1.555	1.626	5.8
Methyl tert-butyl Ether	1.884	2.238	2.221	2.247	2.212	2.209	2.169	6.5
Methyl Acetate	0.612	0.673	0.714	0.876	0.865	0.881	0.770	15.4
Methylene Chloride	0.718	0.772	0.754	0.730	0.716	0.705	0.732	3.5
trans-1,2-Dichloroethene	0.543	0.703	0.665	0.649	0.643	0.635	0.640	8.3
1,1-Dichloroethane	1.090	1.325	1.299	1.297	1.287	1.278	1.263	6.8
Cyclohexane		1.161	1.092	1.030	0.984	0.993	1.052	7.1
2-Butanone	0.370	0.470	0.473	0.439	0.414	0.424	0.432	9
Carbon Tetrachloride	0.482	0.574	0.555	0.539	0.526	0.518	0.532	6
cis-1,2-Dichloroethene	0.698	0.806	0.803	0.808	0.787	0.797	0.783	5.4
Bromochloromethane	0.457	0.607	0.450	0.577	0.587	0.630	0.551	14.2
Chloroform	1.034	1.364	1.366	1.323	1.282	1.288	1.276	9.7
1,1,1-Trichloroethane	0.901	1.123	1.141	1.127	1.100	1.103	1.082	8.3
Methylcyclohexane	0.490	0.589	0.591	0.565	0.548	0.554	0.556	6.6
Benzene	1.330	1.577	1.566	1.523	1.473	1.459	1.488	6.1
1,2-Dichloroethane	0.513	0.595	0.600	0.582	0.554	0.553	0.566	5.8
Trichloroethene	0.296	0.382	0.382	0.370	0.364	0.363	0.360	8.9
1,2-Dichloropropane	0.326	0.403	0.394	0.397	0.382	0.384	0.381	7.4
Bromodichloromethane	0.461	0.599	0.601	0.607	0.579	0.586	0.572	9.7
4-Methyl-2-Pentanone	0.390	0.493	0.503	0.511	0.479	0.487	0.477	9.3
Toluene	0.841	0.967	0.960	0.940	0.897	0.897	0.917	5.2

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG No.: Q3165

Instrument ID: MSVOA_X

Calibration Date(s): 09/16/2025 09/16/2025

Heated Purge: (Y/N) N

Calibration Time(s): 09:23 12:01

GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF001 = VX047585.D		RRF005 = VX047586.D		RRF020 = VX047587.D			
		RRF050 = VX047588.D		RRF100 = VX047589.D		RRF150 = VX047590.D			
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
t-1,3-Dichloropropene	0.487	0.588	0.602	0.619	0.595	0.611	0.584	8.3	
cis-1,3-Dichloropropene	0.513	0.652	0.644	0.653	0.637	0.644	0.624	8.7	
1,1,2-Trichloroethane	0.323	0.365	0.371	0.367	0.344	0.351	0.354	5.1	
2-Hexanone	0.274	0.374	0.372	0.360	0.333	0.343	0.343	10.8	
Dibromochloromethane	0.321	0.414	0.431	0.435	0.421	0.422	0.407	10.5	
1,2-Dibromoethane	0.287	0.379	0.384	0.379	0.364	0.367	0.360	10.1	
Tetrachloroethene	0.299	0.354	0.342	0.326	0.320	0.314	0.326	6	
Chlorobenzene	0.998	1.182	1.177	1.178	1.139	1.142	1.136	6.2	
Ethyl Benzene	1.625	2.077	2.049	2.039	1.996	1.973	1.960	8.6	
m/p-Xylenes	0.604	0.774	0.765	0.765	0.739	0.730	0.729	8.8	
o-Xylene	0.576	0.734	0.734	0.749	0.721	0.719	0.706	9.1	
Styrene	1.019	1.288	1.290	1.299	1.257	1.265	1.236	8.7	
Bromoform	0.221	0.293	0.301	0.313	0.312	0.317	0.293	12.4	
Isopropylbenzene	3.124	3.907	3.986	4.047	3.907	3.839	3.801	8.9	
1,1,2,2-Tetrachloroethane	0.965	1.182	1.203	1.225	1.152	1.173	1.150	8.2	
1,3-Dichlorobenzene	1.462	1.800	1.829	1.817	1.768	1.758	1.739	8	
1,4-Dichlorobenzene	1.570	1.890	1.868	1.832	1.767	1.783	1.785	6.5	
1,2-Dichlorobenzene	1.359	1.691	1.757	1.746	1.696	1.692	1.657	9	
1,2-Dibromo-3-Chloropropane	0.173	0.218	0.242	0.254	0.250	0.260	0.233	14.1	
1,2,4-Trichlorobenzene	0.831	1.099	1.107	1.156	1.140	1.126	1.077	11.3	
1,2,3-Trichlorobenzene	0.753	0.986	1.031	1.088	1.078	1.073	1.002	12.7	
1,2-Dichloroethane-d4		0.884	0.647	0.770	0.815	0.863	0.796	11.8	
Dibromofluoromethane		0.356	0.266	0.331	0.355	0.368	0.335	12.3	
Toluene-d8		1.237	0.943	1.147	1.211	1.263	1.160	11.1	
4-Bromofluorobenzene		0.478	0.360	0.427	0.452	0.484	0.440	11.4	

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3165</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/25/2025 08:56</u>
Lab File ID:	<u>VX047795.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.627		21.04	20
Chloromethane	0.680	0.789	0.1	16.03	20
Vinyl Chloride	0.666	0.779		16.97	20
Bromomethane	0.422	0.471		11.61	20
Chloroethane	0.445	0.495		11.24	20
Trichlorofluoromethane	0.989	1.083		9.51	20
1,1,2-Trichlorotrifluoroethane	0.578	0.618		6.92	20
1,1-Dichloroethene	0.594	0.663		11.62	20
Acetone	0.346	0.336		-2.89	20
Carbon Disulfide	1.626	1.765		8.55	20
Methyl tert-butyl Ether	2.169	2.414		11.3	20
Methyl Acetate	0.770	0.985		27.92	20
Methylene Chloride	0.732	0.817		11.61	20
trans-1,2-Dichloroethene	0.640	0.703		9.84	20
1,1-Dichloroethane	1.263	1.370	0.1	8.47	20
Cyclohexane	1.052	1.056		0.38	20
2-Butanone	0.432	0.496		14.81	20
Carbon Tetrachloride	0.532	0.538		1.13	20
cis-1,2-Dichloroethene	0.783	0.854		9.07	20
Bromochloromethane	0.551	0.621		12.7	20
Chloroform	1.276	1.421		11.36	20
1,1,1-Trichloroethane	1.082	1.142		5.55	20
Methylcyclohexane	0.556	0.530		-4.68	20
Benzene	1.488	1.598		7.39	20
1,2-Dichloroethane	0.566	0.594		4.95	20
Trichloroethene	0.360	0.380		5.56	20
1,2-Dichloropropane	0.381	0.418		9.71	20
Bromodichloromethane	0.572	0.613		7.17	20
4-Methyl-2-Pentanone	0.477	0.566		18.66	20
Toluene	0.917	0.962		4.91	20
t-1,3-Dichloropropene	0.584	0.544		-6.85	20
cis-1,3-Dichloropropene	0.624	0.601		-3.69	20
1,1,2-Trichloroethane	0.354	0.390		10.17	20
2-Hexanone	0.343	0.392		14.29	20
Dibromochloromethane	0.407	0.449		10.32	20
1,2-Dibromoethane	0.360	0.409		13.61	20
Tetrachloroethene	0.326	0.353		8.28	20
Chlorobenzene	1.136	1.192	0.3	4.93	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3165</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/25/2025 08:56</u>
Lab File ID:	<u>VX047795.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	2.012		2.65	20
m/p-Xylenes	0.729	0.757		3.84	20
o-Xylene	0.706	0.737		4.39	20
Styrene	1.236	1.303		5.42	20
Bromoform	0.293	0.318	0.1	8.53	20
Isopropylbenzene	3.801	3.828		0.71	20
1,1,2,2-Tetrachloroethane	1.150	1.273	0.3	10.7	20
1,3-Dichlorobenzene	1.739	1.748		0.52	20
1,4-Dichlorobenzene	1.785	1.750		-1.96	20
1,2-Dichlorobenzene	1.657	1.689		1.93	20
1,2-Dibromo-3-Chloropropane	0.233	0.250		7.3	20
1,2,4-Trichlorobenzene	1.077	1.039		-3.53	20
1,2,3-Trichlorobenzene	1.002	1.016		1.4	20
1,2-Dichloroethane-d4	0.796	0.817		2.64	20
Dibromofluoromethane	0.335	0.351		4.78	20
Toluene-d8	1.160	1.182		1.9	20
4-Bromofluorobenzene	0.440	0.443		0.68	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETRO6</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3165</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/25/2025</u> <u>19:09</u>
Lab File ID:	<u>VX047823.D</u>	Init. Calib. Date(s):	<u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23</u> <u>12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.570		10.04	50
Chloromethane	0.680	0.706	0.1	3.82	50
Vinyl Chloride	0.666	0.715		7.36	50
Bromomethane	0.422	0.476		12.8	50
Chloroethane	0.445	0.468		5.17	50
Trichlorofluoromethane	0.989	1.012		2.33	50
1,1,2-Trichlorotrifluoroethane	0.578	0.582		0.69	50
1,1-Dichloroethene	0.594	0.611		2.86	50
Acetone	0.346	0.295		-14.74	50
Carbon Disulfide	1.626	1.666		2.46	50
Methyl tert-butyl Ether	2.169	2.267		4.52	50
Methyl Acetate	0.770	0.897		16.49	50
Methylene Chloride	0.732	0.760		3.83	50
trans-1,2-Dichloroethene	0.640	0.669		4.53	50
1,1-Dichloroethane	1.263	1.314	0.1	4.04	50
Cyclohexane	1.052	1.017		-3.33	50
2-Butanone	0.432	0.457		5.79	50
Carbon Tetrachloride	0.532	0.519		-2.44	50
cis-1,2-Dichloroethene	0.783	0.823		5.11	50
Bromochloromethane	0.551	0.605		9.8	50
Chloroform	1.276	1.355		6.19	50
1,1,1-Trichloroethane	1.082	1.111		2.68	50
Methylcyclohexane	0.556	0.527		-5.22	50
Benzene	1.488	1.537		3.29	50
1,2-Dichloroethane	0.566	0.566		0	50
Trichloroethene	0.360	0.370		2.78	50
1,2-Dichloropropane	0.381	0.403		5.77	50
Bromodichloromethane	0.572	0.597		4.37	50
4-Methyl-2-Pentanone	0.477	0.541		13.42	50
Toluene	0.917	0.955		4.14	50
t-1,3-Dichloropropene	0.584	0.574		-1.71	50
cis-1,3-Dichloropropene	0.624	0.627		0.48	50
1,1,2-Trichloroethane	0.354	0.385		8.76	50
2-Hexanone	0.343	0.377		9.91	50
Dibromochloromethane	0.407	0.444		9.09	50
1,2-Dibromoethane	0.360	0.402		11.67	50
Tetrachloroethene	0.326	0.327		0.31	50
Chlorobenzene	1.136	1.165	0.3	2.55	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3165</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/25/2025 19:09</u>
Lab File ID:	<u>VX047823.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	1.999		1.99	50
m/p-Xylenes	0.729	0.748		2.61	50
o-Xylene	0.706	0.733		3.82	50
Styrene	1.236	1.280		3.56	50
Bromoform	0.293	0.319	0.1	8.87	50
Isopropylbenzene	3.801	3.759		-1.11	50
1,1,2,2-Tetrachloroethane	1.150	1.240	0.3	7.83	50
1,3-Dichlorobenzene	1.739	1.729		-0.57	50
1,4-Dichlorobenzene	1.785	1.757		-1.57	50
1,2-Dichlorobenzene	1.657	1.693		2.17	50
1,2-Dibromo-3-Chloropropane	0.233	0.237		1.72	50
1,2,4-Trichlorobenzene	1.077	1.045		-2.97	50
1,2,3-Trichlorobenzene	1.002	1.018		1.6	50
1,2-Dichloroethane-d4	0.796	0.769		-3.39	50
Dibromofluoromethane	0.335	0.335		0	50
Toluene-d8	1.160	1.142		-1.55	50
4-Bromofluorobenzene	0.440	0.431		-2.05	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

LAB CHRONICLE

OrderID:	Q3165	OrderDate:	9/22/2025 4:07:00 PM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	J43,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3165-02	VPB184-HYD-202509 19	Water			09/19/25			09/22/25
			SVOC-SIMGroup1	8270-Modified		09/25/25	09/29/25	
Q3165-04	BP-VPB-184-GW-260- 262	Water			09/19/25			09/22/25
			SVOC-SIMGroup1	8270-Modified		09/25/25	09/29/25	
Q3165-07	BP-VPB-184-GW-300- 302	Water			09/22/25			09/22/25
			SVOC-SIMGroup1	8270-Modified		09/25/25	09/29/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet SW-846

SDG No.: Q3165

Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID : BP-VPB-184-GW-260-262								
Q3165-04	BP-VPB-184-GW-260-262 WATER	1,4-Dioxane	0.430		0.08	0.24	0.24	ug/L
		Total Svoc :		0.43				
		Total Concentration:		0.43				
Client ID : BP-VPB-184-GW-300-302								
Q3165-07	BP-VPB-184-GW-300-302 WATER	1,4-Dioxane	0.210	J	0.09	0.26	0.26	ug/L
		Total Svoc :		0.21				
		Total Concentration:		0.21				



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	09/19/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	09/22/25
Client Sample ID:	VPB184-HYD-20250919	SDG No.:	Q3165
Lab Sample ID:	Q3165-02	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	900 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037935.D	1	09/25/25 11:00	09/29/25 13:55	PB169839

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.22	U	0.070	0.22	0.22	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.0060	*	30 - 150		1%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.065	*	30 - 150		16%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.38		55 - 111		94%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.36		53 - 106		91%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.46		58 - 132		114%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	8470	7.818				
1146-65-2	Naphthalene-d8	24200	10.616				
15067-26-2	Acenaphthene-d10	12400	14.473				
1517-22-2	Phenanthrene-d10	25700	17.223				
1719-03-5	Chrysene-d12	20800	21.402				
1520-96-3	Perylene-d12	17200	23.724				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	09/19/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	09/22/25
Client Sample ID:	BP-VPB-184-GW-260-262	SDG No.:	Q3165
Lab Sample ID:	Q3165-04	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	840 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037936.D	1	09/25/25 11:00	09/29/25 14:32	PB169839

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.43		0.080	0.24	0.24	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.34		30 - 150		85%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.46		30 - 150		115%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.32		55 - 111		81%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.35		53 - 106		87%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.43		58 - 132		108%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	9860	7.818				
1146-65-2	Naphthalene-d8	30100	10.616				
15067-26-2	Acenaphthene-d10	16700	14.473				
1517-22-2	Phenanthrene-d10	35400	17.211				
1719-03-5	Chrysene-d12	36100	21.393				
1520-96-3	Perylene-d12	36300	23.721				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	09/22/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	09/22/25
Client Sample ID:	BP-VPB-184-GW-300-302	SDG No.:	Q3165
Lab Sample ID:	Q3165-07	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	760 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037937.D	1	09/25/25 11:00	09/29/25 15:08	PB169839

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.21	J	0.090	0.26	0.26	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.34		30 - 150		85%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.45		30 - 150		113%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.34		55 - 111		84%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.34		53 - 106		84%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.42		58 - 132		105%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	8450	7.818				
1146-65-2	Naphthalene-d8	25700	10.616				
15067-26-2	Acenaphthene-d10	14600	14.473				
1517-22-2	Phenanthrene-d10	31200	17.211				
1719-03-5	Chrysene-d12	32100	21.402				
1520-96-3	Perylene-d12	35200	23.724				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q3165

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169839BL	PB169839BL	2-Methylnaphthalene-d10	0.4	0.35	88		30	150
		Fluoranthene-d10	0.4	0.36	90		30	150
		Nitrobenzene-d5	0.4	0.37	93		55	111
		2-Fluorobiphenyl	0.4	0.38	94		53	106
		Terphenyl-d14	0.4	0.43	106		58	132
PB169839BS	PB169839BS	2-Methylnaphthalene-d10	0.4	0.37	91		30	150
		Fluoranthene-d10	0.4	0.37	91		30	150
		Nitrobenzene-d5	0.4	0.37	92		55	111
		2-Fluorobiphenyl	0.4	0.37	92		53	106
		Terphenyl-d14	0.4	0.38	95		58	132
PB169839BSD	PB169839BSD	2-Methylnaphthalene-d10	0.4	0.36	91		30	150
		Fluoranthene-d10	0.4	0.37	91		30	150
		Nitrobenzene-d5	0.4	0.37	93		55	111
		2-Fluorobiphenyl	0.4	0.37	93		53	106
		Terphenyl-d14	0.4	0.39	97		58	132
Q3165-02	VPB184-HYD-20250919	2-Methylnaphthalene-d10	0.4	0.0060	1	*	30	150
		Fluoranthene-d10	0.4	0.065	16	*	30	150
		Nitrobenzene-d5	0.4	0.38	94		55	111
		2-Fluorobiphenyl	0.4	0.36	91		53	106
		Terphenyl-d14	0.4	0.46	114		58	132
Q3165-04	BP-VPB-184-GW-260-262	2-Methylnaphthalene-d10	0.4	0.34	85		30	150
		Fluoranthene-d10	0.4	0.46	115		30	150
		Nitrobenzene-d5	0.4	0.32	81		55	111
		2-Fluorobiphenyl	0.4	0.35	87		53	106
		Terphenyl-d14	0.4	0.43	108		58	132
Q3165-07	BP-VPB-184-GW-300-302	2-Methylnaphthalene-d10	0.4	0.34	85		30	150
		Fluoranthene-d10	0.4	0.45	113		30	150
		Nitrobenzene-d5	0.4	0.34	84		55	111
		2-Fluorobiphenyl	0.4	0.34	84		53	106
		Terphenyl-d14	0.4	0.42	105		58	132

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3165

Analytical Method: 8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN037933.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB169839BS	1,4-Dioxane	0.4	0.31	ug/L	78				70	130	

A

B

C

D

E

F

G

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3165

Analytical Method: 8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN037934.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB169839BSD	1,4-Dioxane	0.4	0.31	ug/L	78	3			70	130		20

A
B
C
D
E
F
G

4B

SEMIVOLATILE METHOD BLANK SUMMARY

Client ID

PB169839BL

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q3165

Lab File ID: BN037932.D

Lab Sample ID: PB169839BL

Instrument ID: BNA_N

Date Extracted: 09/25/2025

Matrix: (soil/water) Water

Date Analyzed: 09/29/2025

Level: (low/med) LOW

Time Analyzed: 12:06

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB169839BS	PB169839BS	BN037933.D	09/29/2025
PB169839BSD	PB169839BSD	BN037934.D	09/29/2025
VPB184-HYD-20250919	Q3165-02	BN037935.D	09/29/2025
BP-VPB-184-GW-260-262	Q3165-04	BN037936.D	09/29/2025
BP-VPB-184-GW-300-302	Q3165-07	BN037937.D	09/29/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037860.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q3165
DFTPP Injection Date: 09/23/2025
DFTPP Injection Time: 11:33

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
365	Greater than 1% of mass 198	4.2
441	Present, but less than mass 443	72.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	17.7 (20.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN037861.D	09/23/2025	12:13
SSTDICC0.2	SSTDICC0.2	BN037862.D	09/23/2025	12:49
SSTDICCC0.4	SSTDICCC0.4	BN037863.D	09/23/2025	13:25
SSTDICC0.8	SSTDICC0.8	BN037864.D	09/23/2025	14:02
SSTDICC1.6	SSTDICC1.6	BN037865.D	09/23/2025	14:38
SSTDICC3.2	SSTDICC3.2	BN037866.D	09/23/2025	15:15
SSTDICC5.0	SSTDICC5.0	BN037867.D	09/23/2025	15:51

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037930.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q3165
DFTPP Injection Date: 09/29/2025
DFTPP Injection Time: 10:15

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.6 (1.8) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	81.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	17.1 (20) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037931.D	09/29/2025	11:30
PB169839BL	PB169839BL	BN037932.D	09/29/2025	12:06
PB169839BS	PB169839BS	BN037933.D	09/29/2025	12:43
PB169839BSD	PB169839BSD	BN037934.D	09/29/2025	13:19
VPB184-HYD-20250919	Q3165-02	BN037935.D	09/29/2025	13:55
BP-VPB-184-GW-260-262	Q3165-04	BN037936.D	09/29/2025	14:32
BP-VPB-184-GW-300-302	Q3165-07	BN037937.D	09/29/2025	15:08
SSTDCCC0.4EC	SSTDCCC0.4	BN037941.D	09/29/2025	18:46

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3165

Client ID : SSTDCCC0.4

Date Analyzed: 09/29/2025

Lab File ID: BN037931.D

Time Analyzed: 11:30

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	9694	7.818	28657	10.62	15093	14.47
UPPER LIMIT	19388	8.318	57314	11.116	30186	14.973
LOWER LIMIT	4847	7.318	14328.5	10.116	7546.5	13.973
EPA SAMPLE NO.						
01 PB169839BL	8651	7.82	24367	10.62	11932	14.47
02 PB169839BS	7496	7.82	21013	10.62	10987	14.47
03 PB169839BSD	7869	7.82	22045	10.62	11433	14.47
04 VPB184-HYD-20250919	8469	7.82	24220	10.62	12389	14.47
05 BP-VPB-184-GW-260-262	9861	7.82	30069	10.62	16737	14.47
06 BP-VPB-184-GW-300-302	8452	7.82	25701	10.62	14574	14.47

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3165

Client ID: SSTDCCC0.4

Date Analyzed: 09/29/2025

Lab File ID: BN037931.D

Time Analyzed: 11:30

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	28466	17.211	20163	21.402	15946	23.724
UPPER LIMIT	56932	17.711	40326	21.902	31892	24.224
LOWER LIMIT	14233	16.711	10081.5	20.902	7973	23.224
EPA SAMPLE NO.						
01 PB169839BL	23393	17.22	17581	21.40	14270	23.72
02 PB169839BS	21519	17.21	19399	21.40	17959	23.73
03 PB169839BSD	22296	17.22	19503	21.40	17544	23.73
04 VPB184-HYD-20250919	25703	17.22	20793	21.40	17157	23.72
05 BP-VPB-184-GW-260-262	35434	17.21	36101	21.39	36307 *	23.72
06 BP-VPB-184-GW-300-302	31224	17.21	32073	21.40	35207 *	23.72

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:		
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:		
Client Sample ID:	PB169839BL		SDG No.:	Q3165	
Lab Sample ID:	PB169839BL		Matrix:	Water	
Analytical Method:	SW8270ESIM		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037932.D	1	09/25/25 11:00	09/29/25 12:06	PB169839

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.35		30 - 150		88%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.36		30 - 150		90%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.37		55 - 111		93%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.38		53 - 106		94%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.43		58 - 132		106%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	8650	7.818				
1146-65-2	Naphthalene-d8	24400	10.616				
15067-26-2	Acenaphthene-d10	11900	14.473				
1517-22-2	Phenanthrene-d10	23400	17.223				
1719-03-5	Chrysene-d12	17600	21.402				
1520-96-3	Perylene-d12	14300	23.724				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB169839BS		SDG No.:	Q3165
Lab Sample ID:	PB169839BS		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037933.D	1	09/25/25 11:00	09/29/25 12:43	PB169839

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.31		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.37		30 - 150		91%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.37		30 - 150		91%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.37		55 - 111		92%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.37		53 - 106		92%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.38		58 - 132		95%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	7500	7.818				
1146-65-2	Naphthalene-d8	21000	10.616				
15067-26-2	Acenaphthene-d10	11000	14.473				
1517-22-2	Phenanthrene-d10	21500	17.211				
1719-03-5	Chrysene-d12	19400	21.402				
1520-96-3	Perylene-d12	18000	23.727				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	
Client Sample ID:	PB169839BSD	SDG No.:	Q3165
Lab Sample ID:	PB169839BSD	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037934.D	1	09/25/25 11:00	09/29/25 13:19	PB169839

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.31		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.36		30 - 150		91%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.37		30 - 150		91%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.37		55 - 111		93%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.37		53 - 106		93%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.39		58 - 132		97%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	7870	7.818				
1146-65-2	Naphthalene-d8	22000	10.616				
15067-26-2	Acenaphthene-d10	11400	14.473				
1517-22-2	Phenanthrene-d10	22300	17.223				
1719-03-5	Chrysene-d12	19500	21.402				
1520-96-3	Perylene-d12	17500	23.73				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN092325.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Sep 24 11:00:01 2025
 Response Via : Initial Calibration

Calibration Files

0.1 =BN037861.D 0.2 =BN037862.D 0.4 =BN037863.D 0.8 =BN037864.D 1.6 =BN037865.D 3.2 =BN037866.D 5 =BN037867.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.413	0.410	0.403	0.425	0.410	0.376	0.406	4.04	
3)	n-Nitrosodimet...	0.541	0.517	0.526	0.561	0.549	0.547	0.540	3.03	
4) S	2-Fluorophenol	0.944	0.921	0.856	0.839	0.908	0.923	0.998	0.913	5.84
5) S	Phenol-d6	1.190	1.200	1.118	1.088	1.211	1.236	1.353	1.199	7.15
6)	bis(2-Chloroet...	1.133	1.096	1.082	1.066	1.150	1.124	1.143	1.113	2.89
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.299	0.289	0.289	0.290	0.310	0.320	0.338	0.305	6.14
9)	Naphthalene	1.105	1.089	1.048	1.051	1.123	1.137	1.160	1.102	3.84
10)	Hexachlorobuta...	0.198	0.184	0.182	0.183	0.192	0.188	0.188	0.188	2.94
11) SURR2	Methylnaphth...	0.515	0.518	0.499	0.504	0.553	0.587	0.714	0.556	13.74
12)	2-Methylnaphth...	0.678	0.681	0.668	0.682	0.754	0.775	0.802	0.720	7.68
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.148	0.142	0.135	0.139	0.163	0.184	0.152	12.15	
15) S	2-Fluorobiphenyl	1.660	1.631	1.559	1.547	1.654	1.650	1.762	1.638	4.36
16)	Acenaphthylene	1.698	1.721	1.663	1.680	1.907	1.951	2.094	1.816	9.23
17)	Acenaphthene	1.262	1.230	1.198	1.208	1.351	1.371	1.437	1.294	7.16
18)	Fluorene	1.455	1.455	1.452	1.466	1.664	1.717	1.748	1.565	8.77
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.041	0.044	0.046	0.058	0.071	0.084	0.057	29.75	
21)	4-Bromophenyl-...	0.252	0.247	0.244	0.244	0.272	0.275	0.291	0.261	7.07
22)	Hexachlorobenzene	0.317	0.312	0.305	0.300	0.324	0.323	0.332	0.316	3.53
23)	Atrazine	0.180	0.181	0.172	0.178	0.205	0.224	0.249	0.199	14.70
24)	Pentachlorophenol	0.101	0.092	0.097	0.116	0.136	0.108	16.50		
25)	Phenanthrene	1.301	1.271	1.230	1.224	1.358	1.379	1.452	1.316	6.37
26)	Anthracene	1.034	1.035	1.023	1.057	1.209	1.275	1.361	1.142	12.10
27) SURR	Fluoranthene-d10	0.880	0.918	0.887	0.900	1.004	1.085	1.370	1.006	17.59
28)	Fluoranthene	1.278	1.315	1.311	1.358	1.555	1.617	1.713	1.450	12.07
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.576	1.600	1.521	1.476	1.607	1.595	1.679	1.579	4.12
31) S	Terphenyl-d14	0.799	0.791	0.753	0.735	0.798	0.809	0.893	0.797	6.34
32)	Benzo(a)anthra...	1.376	1.359	1.294	1.285	1.423	1.483	1.567	1.398	7.27
33)	Chrysene	1.508	1.526	1.462	1.446	1.559	1.512	1.569	1.512	3.03
34)	Bis(2-ethylhex...	0.592	0.487	0.476	0.518	0.569	0.676	0.553	13.57	
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN092325.M

36)	Indeno(1,2,3-c...	1.715	1.672	1.658	1.707	1.887	1.996	2.159	1.828	10.52
37)	Benzo(b)fluora...	1.599	1.606	1.641	1.590	1.810	1.848	1.981	1.725	8.95
38)	Benzo(k)fluora...	1.620	1.586	1.597	1.628	1.875	1.917	1.926	1.736	9.26
39) C	Benzo(a)pyrene	1.254	1.296	1.231	1.254	1.413	1.492	1.610	1.364	10.61
40)	Dibenzo(a,h)an...	1.285	1.283	1.303	1.333	1.495	1.587	1.705	1.427	11.87
41)	Benzo(g,h,i)pe...	1.407	1.381	1.404	1.397	1.521	1.613	1.773	1.499	9.85

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG No.: Q3165
 Instrument ID: BNA_N Calibration Date/Time: 09/29/2025 11:30
 Lab File ID: BN037931.D Init. Calib. Date(s): 09/23/2025 09/23/2025
 EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 12:13 15:51
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.556	0.515		-7.4	20.0
Fluoranthene-d10	1.006	0.853		-15.2	20.0
2-Fluorophenol	0.913	0.894		-2.1	20.0
Phenol-d6	1.199	1.159		-3.3	20.0
Nitrobenzene-d5	0.305	0.277		-9.2	20.0
2-Fluorobiphenyl	1.638	1.554		-5.1	20.0
2,4,6-Tribromophenol	0.152	0.134		-11.8	20.0
Terphenyl-d14	0.797	0.840		5.4	20.0
1,4-Dioxane	0.406	0.433		6.7	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG No.: Q3165
 Instrument ID: BNA_N Calibration Date/Time: 09/29/2025 18:46
 Lab File ID: BN037941.D Init. Calib. Date(s): 09/23/2025 09/23/2025
 EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 12:13 15:51
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.556	0.490		-11.9	50.0
Fluoranthene-d10	1.006	0.914		-9.1	50.0
2-Fluorophenol	0.913	0.907		-0.7	50.0
Phenol-d6	1.199	1.110		-7.4	50.0
Nitrobenzene-d5	0.305	0.282		-7.5	50.0
2-Fluorobiphenyl	1.638	1.513		-7.6	50.0
2,4,6-Tribromophenol	0.152	0.142		-6.6	50.0
Terphenyl-d14	0.797	0.716		-10.2	50.0
1,4-Dioxane	0.406	0.382		-5.9	50.0

All other compounds must meet a minimum RRF of 0.010.



SHIPPING DOCUMENTS

CHEMTECH

CHAIN OF CUSTODY RECORD

284 Sheffield Street, Mountainside, NJ 07092

(908) 789-8900 Fax: (908) 78-8922

www.chemtech.net

Chemtech Project Number:

Q 3165

COC Number:

Q 3166

CLIENT INFORMATION

COMPANY: Tetra Tech

ADDRESS: 4433 Corporation Lane Suite 300

CITY: Virginia Beach

STATE: VA

ZIP: 23462

ATTENTION: Ernie Wu

PHONE: 757-466-4901

FAX: 757-461-4148

PROJECT INFORMATION

PROJECT NAME: NWIRP Bethpage

PROJECT #: 112G08005-WE13 LOCATION: VPB-184

PROJECT MANAGER: Ernie Wu

E-MAIL: ernie.wu@tetrattech.com

PHONE: 757-466-4901

FAX: 757-461-4148

BILLING INFORMATION

BILL TO: SEE CONTRACT

PO#

ADDRESS:

CITY:

STATE: ZIP:

ATTENTION:

PHONE:

DATA TURNAROUND INFORMATION

FAX: 2 & 10 DAYS*

HARD COPY: 2 & 10 DAYS*

EDD: 2 & 10 DAYS*

* TO BE APPROVED BY CHEMTECH

STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

- ☐ RESEULTS ONLY ☐ USEPA CLP
☐ RESULTS + QC ☐ New York State ASP "B"
☐ New Jersey REDUCED ☐ New York State ASP "A"
☐ New Jersey CLP ☐ Other _____
☐ EDD Format _____

ANALYSIS

VOC(SW846-8260B)
1,4 Dioxane (8270 SIM)

1 2 3 4 5 6 7 8 9

PRESERVATIVES

COMMENTS

<-- Specify Preservatives
A-HCl B-HNO3
C-H2SO4 D-NaOH
E-ICE F-Other

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# of Bottles	A										
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9		
1.	BP-VPB-184-TB-20250919	QA		X	9/19/25	9:00	2	2										
2.	VPB184-HYD-20250919	AQ		X	9/19/25	13:15	5	2	1	2								
3.	BP-VPB-184-GW-240-242	AQ		X	9/19/25	10:47	3	3										
4.	BP-VPB-184-GW-260-262	AQ		X	9/19/25	12:42	3	2	1									
5.	BP-VPB-184-GW-280-282	AQ		X	9/22/25	10:22	3	3										
6.	BP-VPB-184-DUP-20250922	QA		X	9/22/25	12:00	3	3										
7.	BP-VPB-184-GW-300-302	AQ		X	9/22/25	12:22	3	2	1									
8.	BP-VPB-184-GW-320-322	AQ		X	9/22/25	14:16	9	9										
9.																		
10.																		

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE PROSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER	DATE/TIME	RECEIVED BY	1550
1. <i>[Signature]</i>	9/22/25	1. <i>[Signature]</i>	4-22-25
RELINQUISHED BY	DATE/TIME	RECEIVED BY	
2.		2.	
RELINQUISHED BY	DATE/TIME	RECEIVED FOR LAB BY	
3. <i>[Signature]</i>	9/22/25	3.	

Conditions of bottles or coolers at receipt: ☐ Compliant ☐ Non Compliant ☐ Cooler Temp 2.0
MeOH extraction requires an additional 4oz. Jar for percent solid ☐ Ice in
Cooler?: _____
Comments: 5 Day TAT - For VOC's see worksheet #15 of SAP 2018 for VPB program VOC list
10-DAY TAT - For 1,4 Dioxane (8270 SIM)

Page 1 of 1

SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight
CHEMTECH: ☐ Picked Up ☐ Overnight

Shipment Complete
☐ YES ☐ NO

WHITE - CHEMTECH COPY FOR RETURN TO CLIENT

YELLOW - CHEMTECH COPY

PINK - SAMPLER COPY

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q3165	TETR06	Order Date : 9/22/2025 4:07:00 PM	Project Mgr :
Client Name : Tetra Tech NUS, Inc.		Project Name : NWIRP Bethpage 112G080	Report Type : Level 4
Client Contact : Ernie Wu		Receive DateTime : 9/23/2025 6:00:00 PM	EDD Type : ADAPT
Invoice Name : Tetra Tech NUS, Inc.		Purchase Order :	Hard Copy Date :
Invoice Contact : Ernie Wu			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q3165-01	BP-VPB-184-TB-20250919	Water	09/19/2025	09:00	VOC-TCLVOA-10		8260-Low	5 Bus. Days	
Q3165-02	VPB184-HYD-20250919	Water	09/19/2025	13:15	VOC-TCLVOA-10		8260-Low	5 Bus. Days	
Q3165-03	BP-VPB-184-GW-240-242	Water	09/19/2025	10:47	VOC-TCLVOA-10		8260-Low	5 Bus. Days	
Q3165-04	BP-VPB-184-GW-260-262	Water	09/19/2025	12:42	VOC-TCLVOA-10		8260-Low	5 Bus. Days	
Q3165-05	BP-VPB-184-GW-280-282	Water	09/22/2025	10:22	VOC-TCLVOA-10		8260-Low	5 Bus. Days	
Q3165-06	BP-VPB-184-DUP-20250922	Water	09/22/2025	12:00	VOC-TCLVOA-10		8260-Low	5 Bus. Days	
Q3165-07	BP-VPB-184-GW-300-302	Water	09/22/2025	12:22	VOC-TCLVOA-10		8260-Low	5 Bus. Days	
Q3165-08	BP-VPB-184-GW-320-322	Water	09/22/2025	14:16					

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q3165	TETR06	Order Date : 9/22/2025 4:07:00 PM	Project Mgr :
Client Name : Tetra Tech NUS, Inc.		Project Name : NWIRP Bethpage 112G080	Report Type : Level 4
Client Contact : Ernie Wu		Receive DateTime : 9/23/2025 6:00:00 PM	EDD Type : ADAPT
Invoice Name : Tetra Tech NUS, Inc.		Purchase Order :	Hard Copy Date :
Invoice Contact : Ernie Wu			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q3165-09	Q3165-8MS	Water	09/22/2025	14:16	VOC-TCLVOA-10		8260-Low	5 Bus. Days	
Q3165-10	Q3165-8MSD	Water	09/22/2025	14:16	VOC-TCLVOA-10		8260-Low	5 Bus. Days	
					VOC-TCLVOA-10		8260-Low	5 Bus. Days	

Relinquished By :

Date / Time : 9/23/25 1035

Received By :

Date / Time :

Storage Area : VOA Refridgerator Room