

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q3165

MATRIX: Water

METHOD: 8260D

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications BFB Meet Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements. The Initial Calibration met the requirements. The Continuous Calibration File ID VX047795.D met the requirements except for Dichlorodifluoromethane and Methyl Acetate failing high but no positive hit in associated samples therefore no corrective action taken.			✓
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges. The Surrogate recoveries were met for all analysis except for BP-VPB-184-GW-320- 322MSD [1,2-Dichloroethane-d4 - 124%] but original sample and MS passing for surrogate therefore no corrective action taken.		✓	

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

	NA	NO	YES
8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria		✓	
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
The MS recoveries met the requirements for all compounds.			
The MSD {Q3165-10MSD} with File ID: VX047822.D recoveries met the requirements for all compounds except for 1,1,2,2-Tetrachloroethane[153%], 1,1,2-Trichloroethane[129%], 1,1-Dichloroethane[130%], 1,2-Dibromo-3-Chloropropane[137%], 1,2-Dibromoethane[127%], 2-Butanone[144%], 4-Methyl-2-Pentanone[140%], Bromochloromethane[133%], Carbon Tetrachloride[71%], Chloroform[139%], cis-1,2-Dichloroethene[127%], Methyl tert-butyl Ether[136%], Methylcyclohexane[54%] and Methylene Chloride[134%] due to matrix interference.			
The sample # BP-VPB-184-GW-320-322MS and BP-VPB-184-GW-320-322MSD is failing for some compounds and the original sample(BP-VPB-184-GW-320-322) is reported with M flag for these compounds.			
The RPD for {Q3165-10MSD} with File ID: VX047822.D met criteria except for 1,1,2,2-Tetrachloroethane[45%], 1,1,2-Trichloroethane[23%], 1,2-Dibromo-3-Chloropropane[39%], 1,2-Dichlorobenzene[23%], 1,2-Dichloroethane[21%], 2-Butanone[29%], 2-Hexanone[27%], 4-Methyl-2-Pentanone[22%], Acetone[23%], Bromomethane[23%], Carbon Tetrachloride[26%], Chloroethane[22%], Chloroform[25%], Chloromethane[24%], Cyclohexane[22%], Methyl Acetate[30%], Methyl tert-butyl Ether[26%], Methylcyclohexane[39%] and Methylene Chloride[24%] due to difference in results of MS-MSD.			
9. Internal Standard Area/Retention Time Shift Meet Criteria			✓
Comments:			
10. Analysis Holding Time Met			✓
If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is)."

The Sample #BP-VPB-184-GW-260-262 have the concentration of target compound below Method detection limits, therefore it is not reported as Hit in Form1.

The not QT review data is reported in the Miscellaneous.

QA REVIEW

Date

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES