

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN092325

Review By	Rahul	Review On	9/24/2025 8:51:58 AM		
Supervise By	Jagrut	Supervise On	9/24/2025 9:44:29 AM		
SubDirectory	BN092325	HP Acquire Method	BNA_N	HP Processing Method	bn092325

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6842, SP6843, SP6844, SP6845, SP6846, SP6847, SP6848
CCC	SP6846
Internal Standard/PEM	SP6830, 1ul/100ul sample
ICV/I.BLK	SP6854
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN037860.D	23 Sep 2025 11:33		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN037861.D	23 Sep 2025 12:13		RC/JU	Ok
3	SSTDICC0.2	SSTDICC0.2	BN037862.D	23 Sep 2025 12:49		RC/JU	Ok
4	SSTDICCC0.4	SSTDICCC0.4	BN037863.D	23 Sep 2025 13:25	Compound #20 kept on QR	RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN037864.D	23 Sep 2025 14:02		RC/JU	Ok
6	SSTDICC1.6	SSTDICC1.6	BN037865.D	23 Sep 2025 14:38		RC/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN037866.D	23 Sep 2025 15:15		RC/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN037867.D	23 Sep 2025 15:51	Compound#14,24 removed from 5.0 ppm	RC/JU	Ok
9	SSTDICV0.4	ICVBN092325	BN037868.D	23 Sep 2025 16:28		RC/JU	Ok
10	PB169743BL	PB169743BL	BN037869.D	23 Sep 2025 17:42	END CCC missing, Analyzed for Contamination check.	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN092925

Review By	rahul	Review On	9/30/2025 8:38:19 AM
Supervise By	Jagrut	Supervise On	9/30/2025 10:51:24 AM
SubDirectory	BN092925	HP Acquire Method	BNA_N, 8270_HP Processing Method bn092325

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6842,SP6843,SP6844,SP6845,SP6846,SP6847,SP6848
CCC	SP6846
Internal Standard/PEM	SP6830,1ul/100ul sample
ICV/I.BLK	SP6854
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN037930.D	29 Sep 2025 10:15		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN037931.D	29 Sep 2025 11:30		RC/JU	Ok
3	PB169839BL	PB169839BL	BN037932.D	29 Sep 2025 12:06		RC/JU	Ok
4	PB169839BS	PB169839BS	BN037933.D	29 Sep 2025 12:43		RC/JU	Ok
5	PB169839BSD	PB169839BSD	BN037934.D	29 Sep 2025 13:19		RC/JU	Ok
6	Q3165-02	VPB184-HYD-2025091	BN037935.D	29 Sep 2025 13:55	Surrogate Fail	RC/JU	Ok
7	Q3165-04	BP-VPB-184-GW-260-2	BN037936.D	29 Sep 2025 14:32		RC/JU	Ok
8	Q3165-07	BP-VPB-184-GW-300-3	BN037937.D	29 Sep 2025 15:08		RC/JU	Ok
9	Q3156-05DL	RE132D2-20250918DL	BN037938.D	29 Sep 2025 16:57		RC/JU	Ok
10	Q3156-06DL	RE132D4-20250918DL	BN037939.D	29 Sep 2025 17:33	Internal standard fail	RC/JU	ReRun
11	Q3156-07DL	TT189D2-20250918DL	BN037940.D	29 Sep 2025 18:09		RC/JU	Ok
12	SSTDCCC0.4	SSTDCCC0.4EC	BN037941.D	29 Sep 2025 18:46		RC/JU	Ok

M : Manual Integration