

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3165

Client ID : SSTDCCC0.4

Date Analyzed: 09/29/2025

Lab File ID: BN037931.D

Time Analyzed: 11:30

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	9694	7.818	28657	10.62	15093	14.47
UPPER LIMIT	19388	8.318	57314	11.116	30186	14.973
LOWER LIMIT	4847	7.318	14328.5	10.116	7546.5	13.973
EPA SAMPLE NO.						
01 PB169839BL	8651	7.82	24367	10.62	11932	14.47
02 PB169839BS	7496	7.82	21013	10.62	10987	14.47
03 PB169839BSD	7869	7.82	22045	10.62	11433	14.47
04 VPB184-HYD-20250919	8469	7.82	24220	10.62	12389	14.47
05 BP-VPB-184-GW-260-262	9861	7.82	30069	10.62	16737	14.47
06 BP-VPB-184-GW-300-302	8452	7.82	25701	10.62	14574	14.47

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.