



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
Data File : VX047794.D
Acq On : 25 Sep 2025 08:24
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Title : SW846 8260
Last Update : Wed Sep 17 06:39:58 2025



AutoFind: Scans 1635, 1636, 1637; Background Corrected with Scan 1628

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	20.1	10851	PASS
75	95	30	60	52.0	28035	PASS
95	95	100	100	100.0	53944	PASS
96	95	5	9	7.0	3775	PASS
173	174	0.00	2	1.0	365	PASS
174	95	50	100	69.9	37728	PASS
175	174	5	9	7.8	2936	PASS
176	174	95	101	95.7	36117	PASS
177	176	5	9	6.9	2482	PASS