

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3165</u>
Lab File ID:	<u>VX047795.D</u>	Date Analyzed:	<u>09/25/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>08:56</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	107538	12.00€				
UPPER LIMIT	215076	12.50€				
LOWER LIMIT	53769	11.50€				
EPA SAMPLE NO.						
BP-VPB-184-TB-20250919	127824	12.01				
VPB184-HYD-20250919	124631	12.01				
BP-VPB-184-GW-240-242	130114	12.01				
BP-VPB-184-GW-260-262	151799	12.01				
BP-VPB-184-GW-280-282	134705	12.01				
BP-VPB-184-DUP-20250922	135087	12.01				
BP-VPB-184-GW-300-302	127218	12.01				
BP-VPB-184-GW-320-322	145772	12.01				
BP-VPB-184-GW-320-322MS	121955	12.01				
BP-VPB-184-GW-320-322MSD	77374	12.01				
VX0925WBL01	126490	12.01				
VX0925WBS01	131521	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.