

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: G Environmental **Level:** LOW **SDG No.:** Q3172
Contract: GENV01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q3168-05 **Client ID:** TOTES COMP#1DUP
Percent Solids for Sample: NA **Duplicate ID** Q3168-05DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	628		576		9		P
Antimony	ug/L	20	25.0	U	25.0	U			P
Arsenic	ug/L	20	10.0	U	10.0	U			P
Barium	ug/L	20	15.5	J	19.1	J	21		P
Beryllium	ug/L	20	3.00	U	3.00	U			P
Cadmium	ug/L	20	3.00	U	3.00	U			P
Calcium	ug/L	20	27600		31500		13		P
Chromium	ug/L	20	1.92	J	1.86	J	3		P
Cobalt	ug/L	20	15.0	U	15.0	U			P
Copper	ug/L	20	12.4		12.1		2		P
Iron	ug/L	20	1670		1740		4		P
Lead	ug/L	20	12.9		13.7		6		P
Magnesium	ug/L	20	1000		1130		12		P
Manganese	ug/L	20	46.1		52.1		12		P
Nickel	ug/L	20	1.67	J	1.63	J	2		P
Potassium	ug/L	20	1940		2190		12		P
Selenium	ug/L	20	10.0	U	10.0	U			P
Silver	ug/L	20	5.00	U	5.00	U			P
Sodium	ug/L	20	3870		4520		15		P
Thallium	ug/L	20	20.0	U	20.0	U			P
Vanadium	ug/L	20	20.0	U	20.0	U			P
Zinc	ug/L	20	355		396		11		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: G Environmental **Level:** LOW **SDG No.:** Q3172
Contract: GENV01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q3168-05MS **Client ID:** TOTES COMP#1MSD
Percent Solids for Sample: NA **Duplicate ID** Q3168-05MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	1520		1480		3		P
Antimony	ug/L	20	385		395		3		P
Arsenic	ug/L	20	387		377		3		P
Barium	ug/L	20	108		108		0		P
Beryllium	ug/L	20	91.7		89.8		2		P
Cadmium	ug/L	20	91.9		89.0		3		P
Calcium	ug/L	20	31900		30300		5		P
Chromium	ug/L	20	203		196		4		P
Cobalt	ug/L	20	94.3		91.5		3		P
Copper	ug/L	20	157		152		3		P
Iron	ug/L	20	3190		3020		5		P
Lead	ug/L	20	466		451		3		P
Magnesium	ug/L	20	2020		1970		3		P
Manganese	ug/L	20	143		141		1		P
Nickel	ug/L	20	236		228		3		P
Potassium	ug/L	20	7220		6900		5		P
Selenium	ug/L	20	960		935		3		P
Silver	ug/L	20	33.6		31.7		6		P
Sodium	ug/L	20	5960		5840		2		P
Thallium	ug/L	20	988		970		2		P
Vanadium	ug/L	20	144		145		1		P
Zinc	ug/L	20	503		468		7		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>G Environmental</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3172</u>
Contract:	<u>GENV01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3205-01</u>	Client ID:	<u>TRE-25-0106DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3205-01DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	0.085	J	0.20	U	200.0		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>G Environmental</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3172</u>
Contract:	<u>GENV01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3205-01MS</u>	Client ID:	<u>TRE-25-0106MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3205-01MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	3.87		4.02		4		CV