

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental

SDG No.: Q3172

Contract: GENV01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169812BS							
Aluminum	ug/L	1000	933		93	80 - 120	P
Antimony	ug/L	400	391		98	80 - 120	P
Arsenic	ug/L	400	376		94	80 - 120	P
Barium	ug/L	100	87.3		87	80 - 120	P
Beryllium	ug/L	100	94.6		95	80 - 120	P
Cadmium	ug/L	100	91.9		92	80 - 120	P
Calcium	ug/L	500	469	J	94	80 - 120	P
Chromium	ug/L	200	193		96	80 - 120	P
Cobalt	ug/L	100	92.3		92	80 - 120	P
Copper	ug/L	150	146		97	80 - 120	P
Iron	ug/L	1500	1440		96	80 - 120	P
Lead	ug/L	500	453		91	80 - 120	P
Magnesium	ug/L	1000	931	J	93	80 - 120	P
Manganese	ug/L	100	91.9		92	80 - 120	P
Nickel	ug/L	250	232		93	80 - 120	P
Potassium	ug/L	5000	4960		99	80 - 120	P
Selenium	ug/L	1000	967		97	80 - 120	P
Silver	ug/L	37.5	34.4		92	80 - 120	P
Sodium	ug/L	1500	1400		93	80 - 120	P
Thallium	ug/L	1000	954		95	80 - 120	P
Vanadium	ug/L	150	139		93	80 - 120	P
Zinc	ug/L	100	98.6		99	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental

SDG No.: Q3172

Contract: GENV01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169872BS Mercury	ug/L	4.0	3.59		90	80 - 120	CV