

Surrogate Summary

SDG No.: **Q3172**

Client: **G Environmental**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3172-01	FW7D	1,2-Dichloroethane-d4	50	50.8	102		74	125
		Dibromofluoromethane	50	39.0	78		75	124
		Toluene-d8	50	41.6	83	*	86	113
		4-Bromofluorobenzene	50	41.8	83		77	121
Q3172-01DL	FW7DDL	1,2-Dichloroethane-d4	50	47.2	94		74	125
		Dibromofluoromethane	50	42.7	85		75	124
		Toluene-d8	50	39.9	80	*	86	113
		4-Bromofluorobenzene	50	39.3	79		77	121
Q3172-02	FW6D	1,2-Dichloroethane-d4	50	55.0	110		74	125
		Dibromofluoromethane	50	43.2	86		75	124
		Toluene-d8	50	44.7	89		86	113
		4-Bromofluorobenzene	50	42.2	84		77	121
Q3172-02DL	FW6DDL	1,2-Dichloroethane-d4	50	56.6	113		74	125
		Dibromofluoromethane	50	49.1	98		75	124
		Toluene-d8	50	45.3	91		86	113
		4-Bromofluorobenzene	50	44.8	90		77	121
Q3172-03	GB3W	1,2-Dichloroethane-d4	50	55.1	110		74	125
		Dibromofluoromethane	50	50.7	101		75	124
		Toluene-d8	50	46.5	93		86	113
		4-Bromofluorobenzene	50	46.9	94		77	121
VV0924WBL01	VV0924WBL01	1,2-Dichloroethane-d4	50	62.2	124		74	125
		Dibromofluoromethane	50	54.6	109		75	124
		Toluene-d8	50	44.1	88		86	113
		4-Bromofluorobenzene	50	48.0	96		77	121
VV0924WBS02	VV0924WBS02	1,2-Dichloroethane-d4	50	49.0	98		74	125
		Dibromofluoromethane	50	50.0	100		75	124
		Toluene-d8	50	49.5	99		86	113
		4-Bromofluorobenzene	50	52.0	104		77	121
VV0924WBSD0	VV0924WBSD02	1,2-Dichloroethane-d4	50	49.5	99		74	125
		Dibromofluoromethane	50	50.1	100		75	124
		Toluene-d8	50	50.7	101		86	113
		4-Bromofluorobenzene	50	55.0	110		77	121
VV0925WBL01	VV0925WBL01	1,2-Dichloroethane-d4	50	60.1	120		74	125
		Dibromofluoromethane	50	53.3	107		75	124
		Toluene-d8	50	47.3	95		86	113
		4-Bromofluorobenzene	50	44.1	88		77	121
VV0925WBS01	VV0925WBS01	1,2-Dichloroethane-d4	50	45.1	90		74	125
		Dibromofluoromethane	50	47.8	96		75	124
		Toluene-d8	50	47.9	96		86	113
		4-Bromofluorobenzene	50	51.2	102		77	121