

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3172</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>G Environmental</u>	Datafile :	<u>VV039165.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VV0924WBSD02	Dichlorodifluoromethane	20	18.1	ug/L	91	5		69	116	19
	Chloromethane	20	18.2	ug/L	91	5		65	116	21
	Vinyl chloride	20	18.1	ug/L	91	6		65	117	19
	Bromomethane	20	14.7	ug/L	74	34	*	58	125	20
	Chloroethane	20	20.0	ug/L	100	6		56	128	20
	Trichlorofluoromethane	20	18.9	ug/L	95	5		73	115	16
	1,1,2-Trichlorotrifluoroethane	20	19.0	ug/L	95	4		80	112	15
	Tert butyl alcohol	100	110	ug/L	110	0		48	142	30
	1,1-Dichloroethene	20	19.4	ug/L	97	3		74	110	20
	Acetone	100	79.7	ug/L	80	32	*	60	125	20
	Carbon disulfide	20	18.3	ug/L	92	6		64	112	20
	Methyl tert-butyl Ether	20	19.3	ug/L	97	7		78	114	20
	Methyl Acetate	20	20.1	ug/L	101	10		67	125	20
	Methylene Chloride	20	20.9	ug/L	104	9		72	114	20
	trans-1,2-Dichloroethene	20	18.6	ug/L	93	6		75	108	16
	1,1-Dichloroethane	20	18.1	ug/L	91	7		78	112	20
	Cyclohexane	20	18.4	ug/L	92	2		75	110	20
	2-Butanone	100	92.0	ug/L	92	18		65	122	26
	Carbon Tetrachloride	20	18.2	ug/L	91	8		77	113	15
	cis-1,2-Dichloroethene	20	18.8	ug/L	94	4		77	110	20
	Bromochloromethane	20	19.2	ug/L	96	6		70	124	20
	Chloroform	20	18.4	ug/L	92	8		79	113	20
	1,1,1-Trichloroethane	20	18.4	ug/L	92	8		80	108	20
	Methylcyclohexane	20	18.2	ug/L	91	7		72	115	20
	Benzene	20	18.8	ug/L	94	7		82	109	15
	1,2-Dichloroethane	20	18.5	ug/L	93	11		80	115	20
	Trichloroethene	20	19.2	ug/L	96	6		77	113	15
	1,2-Dichloropropane	20	19.4	ug/L	97	7		83	111	16
	Bromodichloromethane	20	18.3	ug/L	92	11		83	110	16
	4-Methyl-2-Pentanone	100	100	ug/L	100	18		74	118	25
	Toluene	20	20.0	ug/L	100	9		82	110	16
	t-1,3-Dichloropropene	20	18.5	ug/L	93	15		79	110	20
	cis-1,3-Dichloropropene	20	19.2	ug/L	96	12		82	110	16
	1,1,2-Trichloroethane	20	18.5	ug/L	93	12		83	112	20
	2-Hexanone	100	91.4	ug/L	91	19		73	117	25
	Dibromochloromethane	20	18.5	ug/L	93	10		82	110	20
	1,2-Dibromoethane	20	19.3	ug/L	97	11		81	110	20
	Tetrachloroethene	20	19.7	ug/L	99	2		67	123	15
	Chlorobenzene	20	19.9	ug/L	100	1		82	109	15
	Ethyl Benzene	20	19.2	ug/L	96	4		83	109	16
	m/p-Xylenes	40	40.4	ug/L	101	7		82	110	15
	o-Xylene	20	19.8	ug/L	99	3		83	109	20
	Styrene	20	17.9	ug/L	90	5		80	111	17
	Bromoform	20	18.8	ug/L	94	13		79	109	20
	Isopropylbenzene	20	19.6	ug/L	98	1		83	112	29
	1,1,2,2-Tetrachloroethane	20	17.9	ug/L	90	6		76	118	20
	1,3-Dichlorobenzene	20	18.7	ug/L	94	4		82	108	20
	1,4-Dichlorobenzene	20	17.9	ug/L	90	6		82	107	15

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3172</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>G Environmental</u>	Datafile :	<u>VV039165.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VV0924WBSD02	1,2-Dichlorobenzene	20	18.5	ug/L	93	2		82	109	20
	1,2-Dibromo-3-Chloropropane	20	19.7	ug/L	99	7		68	112	20
	1,2,4-Trichlorobenzene	20	18.4	ug/L	92	2		75	113	29
	1,2,3-Trichlorobenzene	20	18.7	ug/L	94	2		76	114	29