

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
Lab Code: ACE SDG NO.: Q3172
Lab File ID: VV039146.D Date Analyzed: 09/24/2025
Instrument ID: MSVOA_V Time Analyzed: 10:49
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	558154	4.60	907128	5.53	923608	8.78
UPPER LIMIT	1116310	5.1	1814260	6.032	1847220	9.277
LOWER LIMIT	279077	4.1	453564	5.032	461804	8.277
EPA SAMPLE NO.						
FW7D	784952	4.60	1453021	5.53	1186446	8.81
FW6D	463506	4.60	853323	5.54	809313	8.78
VV0924WBL01	463770	4.60	858892	5.53	831675	8.77
VV0924WBS02	492318	4.60	795602	5.53	784865	8.78
VV0924WBSD02	545364	4.60	903558	5.53	885424	8.78

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.