

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3172
 Lab File ID: VV039168.D Date Analyzed: 09/25/2025
 Instrument ID: MSVOA_V Time Analyzed: 09:59
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	605297	4.60	933338	5.53	901987	8.77
UPPER LIMIT	1210590	5.096	1866680	6.029	1803970	9.273
LOWER LIMIT	302649	4.096	466669	5.029	450994	8.273
EPA SAMPLE NO.						
FW7DDL	686688	4.60	1217519	5.54	1189450	8.78
FW6DDL	642311	4.60	1161576	5.54	1114736	8.78
GB3W	368346	4.60	653076	5.54	662159	8.78
VV0925WBL01	480047	4.60	864912	5.53	870930	8.78
VV0925WBS01	557017	4.60	898902	5.54	871912	8.78

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.