

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3173
 Lab File ID: VV039146.D Date Analyzed: 09/24/2025
 Instrument ID: MSVOA_V Time Analyzed: 10:49
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	558154	4.60	907128	5.53	923608	8.78
UPPER LIMIT	1116310	5.1	1814260	6.032	1847220	9.277
LOWER LIMIT	279077	4.1	453564	5.032	461804	8.277
EPA SAMPLE NO.						
OBS1	483819	4.60	891080	5.54	856052	8.78
Field Blank	552493	4.60	1029152	5.53	962323	8.78
VV0924WBL01	463770	4.60	858892	5.53	831675	8.77
VV0924WBS02	492318	4.60	795602	5.53	784865	8.78
VV0924WBSD02	545364	4.60	903558	5.53	885424	8.78

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.