

Laboratory Control Sample Summary

Client: G Environmental

SDG No.: Q3174

Project: Buffington

Run No.: LB137344

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137344BSS							
Bromide	mg/Kg	200	202		101	1	90-110	09/26/2025
Chloride	mg/Kg	60	59.2		99	1	90-110	09/26/2025
Fluoride	mg/Kg	40	40.6		102	1	90-110	09/26/2025
Nitrite	mg/Kg	60	60.3		100	1	90-110	09/26/2025
Nitrate	mg/Kg	50	48.8		98	1	90-110	09/26/2025
Sulfate	mg/Kg	300	296		99	1	90-110	09/26/2025
Orthophosphate as P	mg/Kg	100	98.5		98	1	90-110	09/26/2025