



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090825\
Data File : VY023302.D
Acq On : 08 Sep 2025 09:07
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090825S.M
Title : SW846 8260
Last Update : Tue Sep 09 02:06:08 2025



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1744

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.3	30416	PASS
75	95	30	60	49.4	82160	PASS
95	95	100	100	100.0	166443	PASS
96	95	5	9	6.2	10390	PASS
173	174	0.00	2	1.2	1857	PASS
174	95	50	100	90.6	150859	PASS
175	174	5	9	7.6	11485	PASS
176	174	95	101	98.7	148856	PASS
177	176	5	9	6.6	9762	PASS