

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3174
 Lab File ID: VY023363.D Date Analyzed: 09/25/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:02
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	631796	7.71	940041	8.62	821307	11.41
UPPER LIMIT	1263590	8.207	1880080	9.116	1642610	11.914
LOWER LIMIT	315898	7.207	470021	8.116	410654	10.914
EPA SAMPLE NO.						
RPXY1	636134	7.71	1166329	8.62	1147896	11.41
SBF1	616227	7.71	1138197	8.62	1103007	11.41
VY0925SBL01	701667	7.71	1240929	8.62	1153810	11.41
VY0925SBS01	598680	7.71	900176	8.62	766959	11.41
VY0925SBSD01	592925	7.71	889198	8.62	770961	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.