

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3174
 Lab File ID: VY023363.D Date Analyzed: 09/25/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:02
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	424523	13.347				
UPPER LIMIT	849046	13.847				
LOWER LIMIT	212262	12.847				
EPA SAMPLE NO.						
RPXY1	537673	13.35				
SBF1	505673	13.35				
VY0925SBL01	541851	13.35				
VY0925SBS01	395796	13.35				
VY0925SBSD01	390902	13.34				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.