

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3175</u>
Instrument ID:	<u>MSVOA_V</u>	Calibration Date/Time:	<u>09/24/2025 10:49</u>
Lab File ID:	<u>VV039146.D</u>	Init. Calib. Date(s):	<u>09/22/2025 09/22/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>12:26 16:35</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.792	0.627		-20.83	20
Chloromethane	0.850	0.730	0.1	-14.12	20
Vinyl Chloride	0.910	0.742		-18.46	20
Bromomethane	0.573	0.504		-12.04	20
Chloroethane	0.631	0.488		-22.66	20
Trichlorofluoromethane	1.317	1.061		-19.44	20
1,1,2-Trichlorotrifluoroethane	0.784	0.628		-19.9	20
Tert butyl alcohol	0.153	0.157		2.61	20
1,1-Dichloroethene	0.754	0.615		-18.43	20
Acetone	0.523	0.448		-14.34	20
Carbon Disulfide	2.306	1.849		-19.82	20
Methyl tert-butyl Ether	2.402	2.294		-4.5	20
Methyl Acetate	1.072	1.090		1.68	20
Methylene Chloride	1.066	0.777		-27.11	20
trans-1,2-Dichloroethene	0.803	0.683		-14.94	20
1,1-Dichloroethane	1.576	1.323	0.1	-16.05	20
Cyclohexane	1.291	1.047		-18.9	20
2-Butanone	0.579	0.608		5.01	20
Carbon Tetrachloride	0.743	0.609		-18.03	20
cis-1,2-Dichloroethene	0.885	0.808		-8.7	20
Bromochloromethane	0.696	0.683		-1.87	20
Chloroform	1.645	1.449		-11.91	20
1,1,1-Trichloroethane	1.424	1.169		-17.91	20
Methylcyclohexane	0.734	0.667		-9.13	20
Benzene	1.979	1.817		-8.19	20
1,2-Dichloroethane	0.693	0.655		-5.48	20
Trichloroethene	0.457	0.405		-11.38	20
1,2-Dichloropropane	0.492	0.476		-3.25	20
Bromodichloromethane	0.778	0.727		-6.55	20
4-Methyl-2-Pentanone	0.661	0.751		13.62	20
Toluene	1.147	1.120		-2.35	20
t-1,3-Dichloropropene	0.678	0.704		3.84	20
cis-1,3-Dichloropropene	0.749	0.769		2.67	20
1,1,2-Trichloroethane	0.514	0.500		-2.72	20
2-Hexanone	0.463	0.557		20.3	20
Dibromochloromethane	0.586	0.559		-4.61	20
1,2-Dibromoethane	0.526	0.530		0.76	20
Tetrachloroethene	0.419	0.347		-17.18	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3175</u>
Instrument ID:	<u>MSVOA_V</u>	Calibration Date/Time:	<u>09/24/2025</u> <u>10:49</u>
Lab File ID:	<u>VV039146.D</u>	Init. Calib. Date(s):	<u>09/22/2025</u> <u>09/22/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>12:26</u> <u>16:35</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chlorobenzene	1.393	1.221	0.3	-12.35	20
Ethyl Benzene	2.108	1.997		-5.27	20
m/p-Xylenes	0.814	0.814		0	20
o-Xylene	0.720	0.732		1.67	20
Styrene	1.263	1.382		9.42	20
Bromoform	0.393	0.378	0.1	-3.82	20
Isopropylbenzene	3.627	3.308		-8.8	20
1,1,2,2-Tetrachloroethane	1.589	1.354	0.3	-14.79	20
1,3-Dichlorobenzene	1.904	1.649		-13.39	20
1,4-Dichlorobenzene	2.089	1.730		-17.18	20
1,2-Dichlorobenzene	1.779	1.564		-12.09	20
1,2-Dibromo-3-Chloropropane	0.330	0.288		-12.73	20
1,2,4-Trichlorobenzene	1.020	0.948		-7.06	20
1,2,3-Trichlorobenzene	1.043	1.000		-4.12	20
1,2-Dichloroethane-d4	0.724	0.661		-8.7	20
Dibromofluoromethane	0.402	0.366		-8.95	20
Toluene-d8	1.181	1.053		-10.84	20
4-Bromofluorobenzene	0.486	0.489		0.62	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.