

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG No.: Q3175
 Instrument ID: MSVOA_V Calibration Date/Time: 09/24/2025 10:49
 Lab File ID: VV039146.D Init. Calib. Date(s): 09/22/2025 09/22/2025
 Heated Purge: (Y/N) N Init. Calib. Time(s): 12:26 16:35
 GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Chloroethane	50.000	45.23	9.540001	20
Acetone	250.000	250.30	-0.1179993	20
Methylene Chloride	50.000	49.66	0.6800003	20
2-Hexanone	250.000	260.90	-4.359204	20
Styrene	50.000	45.39	9.213997	20
1,2-Dibromo-3-Chloropropane	50.000	50.35	-0.6900024	20

All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.