

## Cover Page

**Order ID :** Q3175

**Project ID :** Summer

**Client :** G Environmental

**Lab Sample Number**

Q3175-01

Q3175-02

**Client Sample Number**

MW2

MW4

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : \_\_\_\_\_

Date: 10/6/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

## **CASE NARRATIVE**

### **G Environmental**

**Project Name: Summer**

**Project # N/A**

**Order ID # Q3175**

**Test Name: VOCMS Group1**

#### **A. Number of Samples and Date of Receipt:**

2 Water samples were received on 09/23/2025.

#### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: VOCMS Group1. This data package contains results for VOCMS Group1.

#### **C. Analytical Techniques:**

The analysis performed on instrument MSVOA\_V were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI. The analysis of VOCMS Group1 was based on method 8260D.

#### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Internal Standards Areas were met for all analysis.

The Retention Times were met for all analysis.

The RPD for {VV0924WBSD02} with File ID: VV039165.D met criteria except for Acetone[32%], Bromomethane[34%] due to difference in results of BS-BSD.

The Blank Spike for {VV0924WBS02} with File ID: VV039150.D met requirements for all compounds except for 4-Methyl-2-Pentanone[120%]failing high but no positive hit in associated samples therefore no corrective action taken.

The Blank analysis did not indicate the presence of lab contamination.

The %RSD is greater than 20% in the Initial Calibration method (82V092225S.M) for 2-Hexanone passing on Linear regression and Chloroethane, Acetone ,Methylene Chloride,Styrene,1,2-Dibromo-3-Chloropropane are passing on Quadratic Regression.

The Continuous Calibration File ID VV039146.D met the requirements except for Dichlorodifluoromethane failing marginally low therefore no corrective action taken.

The Continuous Calibration File ID VV039168.D met the requirements except for Bromomethane failing marginally low therefore no corrective action taken.



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

The Tuning criteria met requirements.

Samples MW2, MW2DL and MW4 were diluted due to high concentrations.

**E. Additional Comments:**

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

Trip Blank was not provided with this set of samples.

**F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_

## DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

|       |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                           |
|-------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Value | If the result is a value greater than or equal to the detection limit, report the value                                                                                                                                                                                                                                                                                                                                                                                                                                                   |
| U     | Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.                                                                                                                                                                                                                                 |
| ND    | Indicates the analyte was analyzed for, but not detected                                                                                                                                                                                                                                                                                                                                                                                                                                                                                  |
| J     | Indicates an estimated value. This flag is used:<br>(1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.)<br>(2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others. |
| B     | Indicates the analyte was found in the blank as well as the sample report as “12 B”.                                                                                                                                                                                                                                                                                                                                                                                                                                                      |
| E     | Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.                                                                                                                                                                                                                                                                                                                                                                                                                         |
| D     | This flag identifies all compounds identified in an analysis at a secondary dilution factor.                                                                                                                                                                                                                                                                                                                                                                                                                                              |
| P     | This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.                                                                                                                                                                                                                                                                                                                        |
| N     | This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.                                                                                                                                                                                                                  |
| A     | This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.                                                                                                                                                                                                                                                                                                                                                                                                                                     |
| Q     | Indicates the LCS did not meet the control limits requirements                                                                                                                                                                                                                                                                                                                                                                                                                                                                            |

## APPENDIX A

### QA REVIEW GENERAL DOCUMENTATION

Project #: Q3175

Completed

For thorough review, the report must have the following:

#### GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

#### COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

#### CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

#### ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: PATEL VAISHALI

Date: 10/06/2025

## LAB CHRONICLE

|                 |                 |                   |                      |
|-----------------|-----------------|-------------------|----------------------|
| <b>OrderID:</b> | Q3175           | <b>OrderDate:</b> | 9/24/2025 9:56:00 AM |
| <b>Client:</b>  | G Environmental | <b>Project:</b>   | Summer               |
| <b>Contact:</b> | Gary Landis     | <b>Location:</b>  | VOA Lab              |

| LabID                   | ClientID      | Matrix       | Test         | Method   | Sample Date     | Prep Date | Anal Date | Received        |
|-------------------------|---------------|--------------|--------------|----------|-----------------|-----------|-----------|-----------------|
| <b>Q3175-01</b>         | <b>MW2</b>    | <b>Water</b> | VOCMS Group1 | 8260-Low | <b>09/23/25</b> |           | 09/24/25  | <b>09/23/25</b> |
| <b>Q3175-01DL</b>       | <b>MW2DL</b>  | <b>Water</b> | VOCMS Group1 | 8260-Low | <b>09/23/25</b> |           | 09/25/25  | <b>09/23/25</b> |
| <b>Q3175-01DL<br/>2</b> | <b>MW2DL2</b> | <b>Water</b> | VOCMS Group1 | 8260-Low | <b>09/23/25</b> |           | 09/25/25  | <b>09/23/25</b> |
| <b>Q3175-02</b>         | <b>MW4</b>    | <b>Water</b> | VOCMS Group1 | 8260-Low | <b>09/23/25</b> |           | 09/24/25  | <b>09/23/25</b> |
| <b>Q3175-02DL</b>       | <b>MW4DL</b>  | <b>Water</b> | VOCMS Group1 | 8260-Low | <b>09/23/25</b> |           | 09/25/25  | <b>09/23/25</b> |

**Hit Summary Sheet**  
**SW-846**

**SDG No.:** Q3175  
**Client:** G Environmental

| Sample ID         | Client ID    | Matrix | Parameter                      | Concentration | C | MDL   | RDL  | Units |
|-------------------|--------------|--------|--------------------------------|---------------|---|-------|------|-------|
| <b>Client ID:</b> | <b>MW2</b>   |        |                                |               |   |       |      |       |
| Q3175-01          | MW2          | Water  | Carbon Disulfide               | 1.50          |   | 0.21  | 1.00 | ug/L  |
| Q3175-01          | MW2          | Water  | Cyclohexane                    | 410           | E | 1.50  | 5.00 | ug/L  |
| Q3175-01          | MW2          | Water  | Methylcyclohexane              | 270           | E | 0.16  | 1.00 | ug/L  |
| Q3175-01          | MW2          | Water  | Benzene                        | 460           | E | 0.15  | 1.00 | ug/L  |
| Q3175-01          | MW2          | Water  | Toluene                        | 100           | E | 0.14  | 1.00 | ug/L  |
| Q3175-01          | MW2          | Water  | Ethyl Benzene                  | 1300          | E | 0.13  | 1.00 | ug/L  |
| Q3175-01          | MW2          | Water  | m/p-Xylenes                    | 3100          | E | 0.24  | 2.00 | ug/L  |
| Q3175-01          | MW2          | Water  | o-Xylene                       | 120           | E | 0.12  | 1.00 | ug/L  |
| Q3175-01          | MW2          | Water  | Isopropylbenzene               | 430           | E | 0.12  | 1.00 | ug/L  |
|                   |              |        | <b>Total Voc :</b>             |               |   | 6190  |      |       |
| Q3175-01          | MW2          | Water  | Naphthalene, 1-methyl-         | * 370         | J | 0     | 0    | ug/L  |
| Q3175-01          | MW2          | Water  | Naphthalene, 2-methyl-         | * 180         | J | 0     | 0    | ug/L  |
| Q3175-01          | MW2          | Water  | Benzene, 1,2,4,5-tetramethyl-  | * 340         | J | 0     | 0    | ug/L  |
| Q3175-01          | MW2          | Water  | Benzene, 1,2,3,4-tetramethyl-  | * 250         | J | 0     | 0    | ug/L  |
| Q3175-01          | MW2          | Water  | Benzene, 1,2,3-trimethyl-      | * 1400        | J | 0     | 0    | ug/L  |
| Q3175-01          | MW2          | Water  | o-Cymene                       | * 210         | J | 0     | 0    | ug/L  |
| Q3175-01          | MW2          | Water  | Benzene, 1-ethyl-2-methyl-     | * 1100        | J | 0     | 0    | ug/L  |
| Q3175-01          | MW2          | Water  | Benzene, 1-ethyl-4-methyl-     | * 440         | J | 0     | 0    | ug/L  |
| Q3175-01          | MW2          | Water  | 1H-Indene, 2,3-dihydro-4-meth  | * 250         | J | 0     | 0    | ug/L  |
| Q3175-01          | MW2          | Water  | Benzene, 4-ethyl-1,2-dimethyl- | * 490         | J | 0     | 0    | ug/L  |
| Q3175-01          | MW2          | Water  | Benzene, 1-methyl-4-propyl-    | * 120         | J | 0     | 0    | ug/L  |
| Q3175-01          | MW2          | Water  | Benzene, 2-ethyl-1,4-dimethyl- | * 270         | J | 0     | 0    | ug/L  |
| Q3175-01          | MW2          | Water  | Benzene, 2-ethenyl-1,4-dimethy | * 560         | J | 0     | 0    | ug/L  |
| Q3175-01          | MW2          | Water  | Benzene, 1-ethenyl-3-ethyl-    | * 230         | J | 0     | 0    | ug/L  |
| Q3175-01          | MW2          | Water  | n-propylbenzene                | * 650         | J | 0.13  | 1.00 | ug/L  |
| Q3175-01          | MW2          | Water  | 1,3,5-Trimethylbenzene         | * 1000        | J | 0.15  | 1.00 | ug/L  |
| Q3175-01          | MW2          | Water  | 1,2,4-Trimethylbenzene         | * 1600        | J | 0.14  | 1.00 | ug/L  |
| Q3175-01          | MW2          | Water  | sec-Butylbenzene               | * 48.9        | J | 0.13  | 1.00 | ug/L  |
| Q3175-01          | MW2          | Water  | p-Isopropyltoluene             | * 34.0        | J | 0.13  | 1.00 | ug/L  |
| Q3175-01          | MW2          | Water  | n-Butylbenzene                 | * 140         | J | 0.15  | 1.00 | ug/L  |
| Q3175-01          | MW2          | Water  | Naphthalene                    | * 1100        | J | 0.20  | 1.00 | ug/L  |
|                   |              |        | <b>Total Tics :</b>            |               |   | 10800 |      |       |
|                   |              |        | <b>Total Concentration:</b>    |               |   | 17000 |      |       |
| <b>Client ID:</b> | <b>MW2DL</b> |        |                                |               |   |       |      |       |
| Q3175-01DL        | MW2DL        | Water  | Cyclohexane                    | 220           | D | 29.0  | 100  | ug/L  |
| Q3175-01DL        | MW2DL        | Water  | Methylcyclohexane              | 150           | D | 3.20  | 20.0 | ug/L  |

**Hit Summary Sheet**  
**SW-846**

**SDG No.:** Q3175

**Client:** G Environmental

| Sample ID            | Client ID | Matrix | Parameter                       | Concentration | C  | MDL  | RDL  | Units |
|----------------------|-----------|--------|---------------------------------|---------------|----|------|------|-------|
| Q3175-01DL           | MW2DL     | Water  | Benzene                         | 270           | D  | 3.00 | 20.0 | ug/L  |
| Q3175-01DL           | MW2DL     | Water  | Toluene                         | 57.2          | D  | 2.80 | 20.0 | ug/L  |
| Q3175-01DL           | MW2DL     | Water  | Ethyl Benzene                   | 4900          | ED | 2.60 | 20.0 | ug/L  |
| Q3175-01DL           | MW2DL     | Water  | m/p-Xylenes                     | 5900          | ED | 4.80 | 40.0 | ug/L  |
| Q3175-01DL           | MW2DL     | Water  | o-Xylene                        | 55.3          | D  | 2.40 | 20.0 | ug/L  |
| Q3175-01DL           | MW2DL     | Water  | Isopropylbenzene                | 160           | D  | 2.40 | 20.0 | ug/L  |
| Total Voc :          |           |        |                                 | 11700         |    |      |      |       |
| Total Concentration: |           |        |                                 | 11700         |    |      |      |       |
| Client ID:           | MW2DL2    |        |                                 |               |    |      |      |       |
| Q3175-01DL2          | MW2DL2    | Water  | Cyclohexane                     | 310           | JD | 150  | 500  | ug/L  |
| Q3175-01DL2          | MW2DL2    | Water  | Methylcyclohexane               | 160           | D  | 16.0 | 100  | ug/L  |
| Q3175-01DL2          | MW2DL2    | Water  | Benzene                         | 360           | D  | 15.0 | 100  | ug/L  |
| Q3175-01DL2          | MW2DL2    | Water  | Toluene                         | 70.2          | JD | 14.0 | 100  | ug/L  |
| Q3175-01DL2          | MW2DL2    | Water  | Ethyl Benzene                   | 5500          | D  | 13.0 | 100  | ug/L  |
| Q3175-01DL2          | MW2DL2    | Water  | m/p-Xylenes                     | 7100          | D  | 24.0 | 200  | ug/L  |
| Q3175-01DL2          | MW2DL2    | Water  | o-Xylene                        | 76.1          | JD | 12.0 | 100  | ug/L  |
| Q3175-01DL2          | MW2DL2    | Water  | Isopropylbenzene                | 220           | D  | 12.0 | 100  | ug/L  |
| Total Voc :          |           |        |                                 | 13800         |    |      |      |       |
| Total Concentration: |           |        |                                 | 13800         |    |      |      |       |
| Client ID:           | MW4       |        |                                 |               |    |      |      |       |
| Q3175-02             | MW4       | Water  | Carbon Disulfide                | 0.57          | J  | 0.21 | 1.00 | ug/L  |
| Q3175-02             | MW4       | Water  | Cyclohexane                     | 160           | E  | 1.50 | 5.00 | ug/L  |
| Q3175-02             | MW4       | Water  | Methylcyclohexane               | 230           | E  | 0.16 | 1.00 | ug/L  |
| Q3175-02             | MW4       | Water  | Benzene                         | 760           | E  | 0.15 | 1.00 | ug/L  |
| Q3175-02             | MW4       | Water  | Toluene                         | 71.3          |    | 0.14 | 1.00 | ug/L  |
| Q3175-02             | MW4       | Water  | Ethyl Benzene                   | 480           | E  | 0.13 | 1.00 | ug/L  |
| Q3175-02             | MW4       | Water  | m/p-Xylenes                     | 1300          | E  | 0.24 | 2.00 | ug/L  |
| Q3175-02             | MW4       | Water  | o-Xylene                        | 230           | E  | 0.12 | 1.00 | ug/L  |
| Q3175-02             | MW4       | Water  | Isopropylbenzene                | 110           | E  | 0.12 | 1.00 | ug/L  |
| Total Voc :          |           |        |                                 | 3340          |    |      |      |       |
| Q3175-02             | MW4       | Water  | Naphthalene, 2-methyl-          | * 120         | J  | 0    | 0    | ug/L  |
| Q3175-02             | MW4       | Water  | Benzene, 1,2,4,5-tetramethyl-   | * 140         | J  | 0    | 0    | ug/L  |
| Q3175-02             | MW4       | Water  | Naphthalene, 1,2,3,4-tetrahydrc | * 75.1        | J  | 0    | 0    | ug/L  |
| Q3175-02             | MW4       | Water  | Benzene, 1,2,3,4-tetramethyl-   | * 120         | J  | 0    | 0    | ug/L  |
| Q3175-02             | MW4       | Water  | Indane                          | * 390         | J  | 0    | 0    | ug/L  |
| Q3175-02             | MW4       | Water  | Benzene, 1,2,3-trimethyl-       | * 370         | J  | 0    | 0    | ug/L  |
| Q3175-02             | MW4       | Water  | Benzene, 1-ethyl-2-methyl-      | * 430         | J  | 0    | 0    | ug/L  |

### Hit Summary Sheet SW-846

SDG No.: Q3175

Client: G Environmental

| Sample ID            | Client ID | Matrix | Parameter                        | Concentration | C | MDL  | RDL  | Units |
|----------------------|-----------|--------|----------------------------------|---------------|---|------|------|-------|
| Q3175-02             | MW4       | Water  | Benzene, 1-ethyl-4-methyl-       | * 190         | J | 0    | 0    | ug/L  |
| Q3175-02             | MW4       | Water  | Indan, 1-methyl-                 | * 120         | J | 0    | 0    | ug/L  |
| Q3175-02             | MW4       | Water  | Cyclopentane, 1,2-dimethyl-, tr  | * 80.5        | J | 0    | 0    | ug/L  |
| Q3175-02             | MW4       | Water  | Benzene, 4-ethyl-1,2-dimethyl-   | * 230         | J | 0    | 0    | ug/L  |
| Q3175-02             | MW4       | Water  | Benzene, 2-ethyl-1,4-dimethyl-   | * 100         | J | 0    | 0    | ug/L  |
| Q3175-02             | MW4       | Water  | Benzene, 2-ethenyl-1,4-dimethyl- | * 280         | J | 0    | 0    | ug/L  |
| Q3175-02             | MW4       | Water  | Benzene, 1-ethenyl-3-ethyl-      | * 130         | J | 0    | 0    | ug/L  |
| Q3175-02             | MW4       | Water  | Cyclopentene, 1,5-dimethyl-      | * 120         | J | 0    | 0    | ug/L  |
| Q3175-02             | MW4       | Water  | n-propylbenzene                  | * 250         | J | 0.13 | 1.00 | ug/L  |
| Q3175-02             | MW4       | Water  | 1,3,5-Trimethylbenzene           | * 390         | J | 0.15 | 1.00 | ug/L  |
| Q3175-02             | MW4       | Water  | 1,2,4-Trimethylbenzene           | * 670         | J | 0.14 | 1.00 | ug/L  |
| Q3175-02             | MW4       | Water  | sec-Butylbenzene                 | * 18.4        | J | 0.13 | 1.00 | ug/L  |
| Q3175-02             | MW4       | Water  | p-Isopropyltoluene               | * 14.2        | J | 0.13 | 1.00 | ug/L  |
| Q3175-02             | MW4       | Water  | n-Butylbenzene                   | * 47.9        | J | 0.15 | 1.00 | ug/L  |
| Q3175-02             | MW4       | Water  | Naphthalene                      | * 300         | J | 0.20 | 1.00 | ug/L  |
| Total Tics :         |           |        |                                  | 4590          |   |      |      |       |
| Total Concentration: |           |        |                                  | 7930          |   |      |      |       |
| Client ID:           | MW4DL     |        |                                  |               |   |      |      |       |
| Q3175-02DL           | MW4DL     | Water  | Cyclohexane                      | 120           | D | 29.0 | 100  | ug/L  |
| Q3175-02DL           | MW4DL     | Water  | Methylcyclohexane                | 140           | D | 3.20 | 20.0 | ug/L  |
| Q3175-02DL           | MW4DL     | Water  | Benzene                          | 1100          | D | 3.00 | 20.0 | ug/L  |
| Q3175-02DL           | MW4DL     | Water  | Toluene                          | 48.3          | D | 2.80 | 20.0 | ug/L  |
| Q3175-02DL           | MW4DL     | Water  | Ethyl Benzene                    | 530           | D | 2.60 | 20.0 | ug/L  |
| Q3175-02DL           | MW4DL     | Water  | m/p-Xylenes                      | 1600          | D | 4.80 | 40.0 | ug/L  |
| Q3175-02DL           | MW4DL     | Water  | o-Xylene                         | 120           | D | 2.40 | 20.0 | ug/L  |
| Q3175-02DL           | MW4DL     | Water  | Isopropylbenzene                 | 51.6          | D | 2.40 | 20.0 | ug/L  |
| Total Voc :          |           |        |                                  | 3710          |   |      |      |       |
| Total Concentration: |           |        |                                  | 3710          |   |      |      |       |



# QC SUMMARY

## Surrogate Summary

SDG No.: **Q3175**

Client: **G Environmental**

Analytical Method: **SW8260-Low**

| Lab Sample ID | Client ID    | Parameter             | Spike | Result | Recovery (%) | Qual | Limits (%) |      |
|---------------|--------------|-----------------------|-------|--------|--------------|------|------------|------|
|               |              |                       |       |        |              |      | Low        | High |
| Q3175-01      | MW2          | 1,2-Dichloroethane-d4 | 50    | 44.7   | 89           |      | 74         | 125  |
|               |              | Dibromofluoromethane  | 50    | 38.9   | 78           |      | 75         | 124  |
|               |              | Toluene-d8            | 50    | 47.6   | 95           |      | 86         | 113  |
|               |              | 4-Bromofluorobenzene  | 50    | 45.2   | 90           |      | 77         | 121  |
| Q3175-01DL    | MW2DL        | 1,2-Dichloroethane-d4 | 50    | 52.4   | 105          |      | 74         | 125  |
|               |              | Dibromofluoromethane  | 50    | 48.9   | 98           |      | 75         | 124  |
|               |              | Toluene-d8            | 50    | 48.3   | 97           |      | 86         | 113  |
|               |              | 4-Bromofluorobenzene  | 50    | 50.2   | 100          |      | 77         | 121  |
| Q3175-01DL2   | MW2DL2       | 1,2-Dichloroethane-d4 | 50    | 55.6   | 111          |      | 74         | 125  |
|               |              | Dibromofluoromethane  | 50    | 49.9   | 100          |      | 75         | 124  |
|               |              | Toluene-d8            | 50    | 45.3   | 91           |      | 86         | 113  |
|               |              | 4-Bromofluorobenzene  | 50    | 46.4   | 93           |      | 77         | 121  |
| Q3175-02      | MW4          | 1,2-Dichloroethane-d4 | 50    | 46.6   | 93           |      | 74         | 125  |
|               |              | Dibromofluoromethane  | 50    | 38.8   | 78           |      | 75         | 124  |
|               |              | Toluene-d8            | 50    | 47.7   | 95           |      | 86         | 113  |
|               |              | 4-Bromofluorobenzene  | 50    | 48.4   | 97           |      | 77         | 121  |
| Q3175-02DL    | MW4DL        | 1,2-Dichloroethane-d4 | 50    | 54.6   | 109          |      | 74         | 125  |
|               |              | Dibromofluoromethane  | 50    | 48.8   | 98           |      | 75         | 124  |
|               |              | Toluene-d8            | 50    | 47.6   | 95           |      | 86         | 113  |
|               |              | 4-Bromofluorobenzene  | 50    | 48.5   | 97           |      | 77         | 121  |
| VV0924WBL01   | VV0924WBL01  | 1,2-Dichloroethane-d4 | 50    | 62.2   | 124          |      | 74         | 125  |
|               |              | Dibromofluoromethane  | 50    | 54.6   | 109          |      | 75         | 124  |
|               |              | Toluene-d8            | 50    | 44.1   | 88           |      | 86         | 113  |
|               |              | 4-Bromofluorobenzene  | 50    | 48.0   | 96           |      | 77         | 121  |
| VV0924WBS02   | VV0924WBS02  | 1,2-Dichloroethane-d4 | 50    | 49.0   | 98           |      | 74         | 125  |
|               |              | Dibromofluoromethane  | 50    | 50.0   | 100          |      | 75         | 124  |
|               |              | Toluene-d8            | 50    | 49.5   | 99           |      | 86         | 113  |
|               |              | 4-Bromofluorobenzene  | 50    | 52.0   | 104          |      | 77         | 121  |
| VV0924WBSD0   | VV0924WBSD02 | 1,2-Dichloroethane-d4 | 50    | 49.5   | 99           |      | 74         | 125  |
|               |              | Dibromofluoromethane  | 50    | 50.1   | 100          |      | 75         | 124  |
|               |              | Toluene-d8            | 50    | 50.7   | 101          |      | 86         | 113  |
|               |              | 4-Bromofluorobenzene  | 50    | 55.0   | 110          |      | 77         | 121  |
| VV0925WBL01   | VV0925WBL01  | 1,2-Dichloroethane-d4 | 50    | 60.1   | 120          |      | 74         | 125  |
|               |              | Dibromofluoromethane  | 50    | 53.3   | 107          |      | 75         | 124  |
|               |              | Toluene-d8            | 50    | 47.3   | 95           |      | 86         | 113  |
|               |              | 4-Bromofluorobenzene  | 50    | 44.1   | 88           |      | 77         | 121  |
| VV0925WBS01   | VV0925WBS01  | 1,2-Dichloroethane-d4 | 50    | 45.1   | 90           |      | 74         | 125  |
|               |              | Dibromofluoromethane  | 50    | 47.8   | 96           |      | 75         | 124  |
|               |              | Toluene-d8            | 50    | 47.9   | 96           |      | 86         | 113  |
|               |              | 4-Bromofluorobenzene  | 50    | 51.2   | 102          |      | 77         | 121  |

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

|          |                        |                    |                   |
|----------|------------------------|--------------------|-------------------|
| SDG No.: | <u>Q3175</u>           | Analytical Method: | <u>SW8260-Low</u> |
| Client:  | <u>G Environmental</u> | Datafile :         | <u>VV039150.D</u> |

| Lab Sample ID | Parameter                      | Spike | Result | Unit | Rec | RPD | Qual | Low | Limits<br>High | RPD |
|---------------|--------------------------------|-------|--------|------|-----|-----|------|-----|----------------|-----|
| VV0924WBS02   | Dichlorodifluoromethane        | 20    | 19.1   | ug/L | 96  |     |      | 69  | 116            |     |
|               | Chloromethane                  | 20    | 19.1   | ug/L | 96  |     |      | 65  | 116            |     |
|               | Vinyl chloride                 | 20    | 19.4   | ug/L | 97  |     |      | 65  | 117            |     |
|               | Bromomethane                   | 20    | 20.9   | ug/L | 104 |     |      | 58  | 125            |     |
|               | Chloroethane                   | 20    | 21.1   | ug/L | 106 |     |      | 56  | 128            |     |
|               | Trichlorofluoromethane         | 20    | 20.0   | ug/L | 100 |     |      | 73  | 115            |     |
|               | 1,1,2-Trichlorotrifluoroethane | 20    | 19.7   | ug/L | 99  |     |      | 80  | 112            |     |
|               | Tert butyl alcohol             | 100   | 110    | ug/L | 110 |     |      | 48  | 142            |     |
|               | 1,1-Dichloroethene             | 20    | 20.0   | ug/L | 100 |     |      | 74  | 110            |     |
|               | Acetone                        | 100   | 110    | ug/L | 110 |     |      | 60  | 125            |     |
|               | Carbon disulfide               | 20    | 19.6   | ug/L | 98  |     |      | 64  | 112            |     |
|               | Methyl tert-butyl Ether        | 20    | 20.9   | ug/L | 104 |     |      | 78  | 114            |     |
|               | Methyl Acetate                 | 20    | 22.4   | ug/L | 112 |     |      | 67  | 125            |     |
|               | Methylene Chloride             | 20    | 22.7   | ug/L | 114 |     |      | 72  | 114            |     |
|               | trans-1,2-Dichloroethene       | 20    | 19.7   | ug/L | 99  |     |      | 75  | 108            |     |
|               | 1,1-Dichloroethane             | 20    | 19.6   | ug/L | 98  |     |      | 78  | 112            |     |
|               | Cyclohexane                    | 20    | 18.7   | ug/L | 94  |     |      | 75  | 110            |     |
|               | 2-Butanone                     | 100   | 110    | ug/L | 110 |     |      | 65  | 122            |     |
|               | Carbon Tetrachloride           | 20    | 19.7   | ug/L | 99  |     |      | 77  | 113            |     |
|               | cis-1,2-Dichloroethene         | 20    | 19.5   | ug/L | 98  |     |      | 77  | 110            |     |
|               | Bromochloromethane             | 20    | 20.3   | ug/L | 102 |     |      | 70  | 124            |     |
|               | Chloroform                     | 20    | 19.9   | ug/L | 100 |     |      | 79  | 113            |     |
|               | 1,1,1-Trichloroethane          | 20    | 20.0   | ug/L | 100 |     |      | 80  | 108            |     |
|               | Methylcyclohexane              | 20    | 19.6   | ug/L | 98  |     |      | 72  | 115            |     |
|               | Benzene                        | 20    | 20.1   | ug/L | 101 |     |      | 82  | 109            |     |
|               | 1,2-Dichloroethane             | 20    | 20.7   | ug/L | 104 |     |      | 80  | 115            |     |
|               | Trichloroethene                | 20    | 20.3   | ug/L | 102 |     |      | 77  | 113            |     |
|               | 1,2-Dichloropropane            | 20    | 20.9   | ug/L | 104 |     |      | 83  | 111            |     |
|               | Bromodichloromethane           | 20    | 20.6   | ug/L | 103 |     |      | 83  | 110            |     |
|               | 4-Methyl-2-Pentanone           | 100   | 120    | ug/L | 120 |     | *    | 74  | 118            |     |
|               | Toluene                        | 20    | 21.8   | ug/L | 109 |     |      | 82  | 110            |     |
|               | t-1,3-Dichloropropene          | 20    | 21.6   | ug/L | 108 |     |      | 79  | 110            |     |
|               | cis-1,3-Dichloropropene        | 20    | 21.5   | ug/L | 108 |     |      | 82  | 110            |     |
|               | 1,1,2-Trichloroethane          | 20    | 21.0   | ug/L | 105 |     |      | 83  | 112            |     |
|               | 2-Hexanone                     | 100   | 110    | ug/L | 110 |     |      | 73  | 117            |     |
|               | Dibromochloromethane           | 20    | 20.5   | ug/L | 103 |     |      | 82  | 110            |     |
|               | 1,2-Dibromoethane              | 20    | 21.5   | ug/L | 108 |     |      | 81  | 110            |     |
|               | Tetrachloroethene              | 20    | 20.2   | ug/L | 101 |     |      | 67  | 123            |     |
|               | Chlorobenzene                  | 20    | 19.8   | ug/L | 99  |     |      | 82  | 109            |     |
|               | Ethyl Benzene                  | 20    | 19.9   | ug/L | 100 |     |      | 83  | 109            |     |
|               | m/p-Xylenes                    | 40    | 43.2   | ug/L | 108 |     |      | 82  | 110            |     |
|               | o-Xylene                       | 20    | 20.4   | ug/L | 102 |     |      | 83  | 109            |     |
|               | Styrene                        | 20    | 18.9   | ug/L | 95  |     |      | 80  | 111            |     |
|               | Bromoform                      | 20    | 21.4   | ug/L | 107 |     |      | 79  | 109            |     |
|               | Isopropylbenzene               | 20    | 19.8   | ug/L | 99  |     |      | 83  | 112            |     |
|               | 1,1,2,2-Tetrachloroethane      | 20    | 19.2   | ug/L | 96  |     |      | 76  | 118            |     |
|               | 1,3-Dichlorobenzene            | 20    | 19.6   | ug/L | 98  |     |      | 82  | 108            |     |
|               | 1,4-Dichlorobenzene            | 20    | 19.2   | ug/L | 96  |     |      | 82  | 107            |     |

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

|                 |                               |                           |                          |
|-----------------|-------------------------------|---------------------------|--------------------------|
| <b>SDG No.:</b> | <u><b>Q3175</b></u>           | <b>Analytical Method:</b> | <u><b>SW8260-Low</b></u> |
| <b>Client:</b>  | <u><b>G Environmental</b></u> | <b>Datafile :</b>         | <u><b>VV039150.D</b></u> |

| Lab Sample ID      | Parameter                   | Spike | Result | Unit | Rec | RPD | Qual | Low | Limits<br>High | RPD |
|--------------------|-----------------------------|-------|--------|------|-----|-----|------|-----|----------------|-----|
| <b>VV0924WBS02</b> | 1,2-Dichlorobenzene         | 20    | 19.0   | ug/L | 95  |     |      | 82  | 109            |     |
|                    | 1,2-Dibromo-3-Chloropropane | 20    | 21.3   | ug/L | 106 |     |      | 68  | 112            |     |
|                    | 1,2,4-Trichlorobenzene      | 20    | 18.7   | ug/L | 94  |     |      | 75  | 113            |     |
|                    | 1,2,3-Trichlorobenzene      | 20    | 19.1   | ug/L | 96  |     |      | 76  | 114            |     |

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

|          |                        |                    |                   |
|----------|------------------------|--------------------|-------------------|
| SDG No.: | <u>Q3175</u>           | Analytical Method: | <u>SW8260-Low</u> |
| Client:  | <u>G Environmental</u> | Datafile :         | <u>VV039165.D</u> |

| Lab Sample ID | Parameter                      | Spike | Result | Unit | Rec | RPD | Qual | Low | Limits High | RPD |
|---------------|--------------------------------|-------|--------|------|-----|-----|------|-----|-------------|-----|
| VV0924WBSD02  | Dichlorodifluoromethane        | 20    | 18.1   | ug/L | 91  | 5   |      | 69  | 116         | 19  |
|               | Chloromethane                  | 20    | 18.2   | ug/L | 91  | 5   |      | 65  | 116         | 21  |
|               | Vinyl chloride                 | 20    | 18.1   | ug/L | 91  | 6   |      | 65  | 117         | 19  |
|               | Bromomethane                   | 20    | 14.7   | ug/L | 74  | 34  | *    | 58  | 125         | 20  |
|               | Chloroethane                   | 20    | 20.0   | ug/L | 100 | 6   |      | 56  | 128         | 20  |
|               | Trichlorofluoromethane         | 20    | 18.9   | ug/L | 95  | 5   |      | 73  | 115         | 16  |
|               | 1,1,2-Trichlorotrifluoroethane | 20    | 19.0   | ug/L | 95  | 4   |      | 80  | 112         | 15  |
|               | Tert butyl alcohol             | 100   | 110    | ug/L | 110 | 0   |      | 48  | 142         | 30  |
|               | 1,1-Dichloroethene             | 20    | 19.4   | ug/L | 97  | 3   |      | 74  | 110         | 20  |
|               | Acetone                        | 100   | 79.7   | ug/L | 80  | 32  | *    | 60  | 125         | 20  |
|               | Carbon disulfide               | 20    | 18.3   | ug/L | 92  | 6   |      | 64  | 112         | 20  |
|               | Methyl tert-butyl Ether        | 20    | 19.3   | ug/L | 97  | 7   |      | 78  | 114         | 20  |
|               | Methyl Acetate                 | 20    | 20.1   | ug/L | 101 | 10  |      | 67  | 125         | 20  |
|               | Methylene Chloride             | 20    | 20.9   | ug/L | 104 | 9   |      | 72  | 114         | 20  |
|               | trans-1,2-Dichloroethene       | 20    | 18.6   | ug/L | 93  | 6   |      | 75  | 108         | 16  |
|               | 1,1-Dichloroethane             | 20    | 18.1   | ug/L | 91  | 7   |      | 78  | 112         | 20  |
|               | Cyclohexane                    | 20    | 18.4   | ug/L | 92  | 2   |      | 75  | 110         | 20  |
|               | 2-Butanone                     | 100   | 92.0   | ug/L | 92  | 18  |      | 65  | 122         | 26  |
|               | Carbon Tetrachloride           | 20    | 18.2   | ug/L | 91  | 8   |      | 77  | 113         | 15  |
|               | cis-1,2-Dichloroethene         | 20    | 18.8   | ug/L | 94  | 4   |      | 77  | 110         | 20  |
|               | Bromochloromethane             | 20    | 19.2   | ug/L | 96  | 6   |      | 70  | 124         | 20  |
|               | Chloroform                     | 20    | 18.4   | ug/L | 92  | 8   |      | 79  | 113         | 20  |
|               | 1,1,1-Trichloroethane          | 20    | 18.4   | ug/L | 92  | 8   |      | 80  | 108         | 20  |
|               | Methylcyclohexane              | 20    | 18.2   | ug/L | 91  | 7   |      | 72  | 115         | 20  |
|               | Benzene                        | 20    | 18.8   | ug/L | 94  | 7   |      | 82  | 109         | 15  |
|               | 1,2-Dichloroethane             | 20    | 18.5   | ug/L | 93  | 11  |      | 80  | 115         | 20  |
|               | Trichloroethene                | 20    | 19.2   | ug/L | 96  | 6   |      | 77  | 113         | 15  |
|               | 1,2-Dichloropropane            | 20    | 19.4   | ug/L | 97  | 7   |      | 83  | 111         | 16  |
|               | Bromodichloromethane           | 20    | 18.3   | ug/L | 92  | 11  |      | 83  | 110         | 16  |
|               | 4-Methyl-2-Pentanone           | 100   | 100    | ug/L | 100 | 18  |      | 74  | 118         | 25  |
|               | Toluene                        | 20    | 20.0   | ug/L | 100 | 9   |      | 82  | 110         | 16  |
|               | t-1,3-Dichloropropene          | 20    | 18.5   | ug/L | 93  | 15  |      | 79  | 110         | 20  |
|               | cis-1,3-Dichloropropene        | 20    | 19.2   | ug/L | 96  | 12  |      | 82  | 110         | 16  |
|               | 1,1,2-Trichloroethane          | 20    | 18.5   | ug/L | 93  | 12  |      | 83  | 112         | 20  |
|               | 2-Hexanone                     | 100   | 91.4   | ug/L | 91  | 19  |      | 73  | 117         | 25  |
|               | Dibromochloromethane           | 20    | 18.5   | ug/L | 93  | 10  |      | 82  | 110         | 20  |
|               | 1,2-Dibromoethane              | 20    | 19.3   | ug/L | 97  | 11  |      | 81  | 110         | 20  |
|               | Tetrachloroethene              | 20    | 19.7   | ug/L | 99  | 2   |      | 67  | 123         | 15  |
|               | Chlorobenzene                  | 20    | 19.9   | ug/L | 100 | 1   |      | 82  | 109         | 15  |
|               | Ethyl Benzene                  | 20    | 19.2   | ug/L | 96  | 4   |      | 83  | 109         | 16  |
|               | m/p-Xylenes                    | 40    | 40.4   | ug/L | 101 | 7   |      | 82  | 110         | 15  |
|               | o-Xylene                       | 20    | 19.8   | ug/L | 99  | 3   |      | 83  | 109         | 20  |
|               | Styrene                        | 20    | 17.9   | ug/L | 90  | 5   |      | 80  | 111         | 17  |
|               | Bromoform                      | 20    | 18.8   | ug/L | 94  | 13  |      | 79  | 109         | 20  |
|               | Isopropylbenzene               | 20    | 19.6   | ug/L | 98  | 1   |      | 83  | 112         | 29  |
|               | 1,1,2,2-Tetrachloroethane      | 20    | 17.9   | ug/L | 90  | 6   |      | 76  | 118         | 20  |
|               | 1,3-Dichlorobenzene            | 20    | 18.7   | ug/L | 94  | 4   |      | 82  | 108         | 20  |
|               | 1,4-Dichlorobenzene            | 20    | 17.9   | ug/L | 90  | 6   |      | 82  | 107         | 15  |

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

|                 |                               |                           |                          |
|-----------------|-------------------------------|---------------------------|--------------------------|
| <b>SDG No.:</b> | <u><b>Q3175</b></u>           | <b>Analytical Method:</b> | <u><b>SW8260-Low</b></u> |
| <b>Client:</b>  | <u><b>G Environmental</b></u> | <b>Datafile :</b>         | <u><b>VV039165.D</b></u> |

| Lab Sample ID       | Parameter                   | Spike | Result | Unit | Rec | RPD | Qual | Low | Limits<br>High | RPD |
|---------------------|-----------------------------|-------|--------|------|-----|-----|------|-----|----------------|-----|
| <b>VV0924WBSD02</b> | 1,2-Dichlorobenzene         | 20    | 18.5   | ug/L | 93  | 2   |      | 82  | 109            | 20  |
|                     | 1,2-Dibromo-3-Chloropropane | 20    | 19.7   | ug/L | 99  | 7   |      | 68  | 112            | 20  |
|                     | 1,2,4-Trichlorobenzene      | 20    | 18.4   | ug/L | 92  | 2   |      | 75  | 113            | 29  |
|                     | 1,2,3-Trichlorobenzene      | 20    | 18.7   | ug/L | 94  | 2   |      | 76  | 114            | 29  |

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

|                 |                        |                           |                   |
|-----------------|------------------------|---------------------------|-------------------|
| <b>SDG No.:</b> | <b>Q3175</b>           | <b>Analytical Method:</b> | <b>SW8260-Low</b> |
| <b>Client:</b>  | <b>G Environmental</b> | <b>Datafile :</b>         | <b>VV039171.D</b> |

| Lab Sample ID      | Parameter                      | Spike | Result | Unit | Rec | RPD | Qual | Low | Limits<br>High | RPD |
|--------------------|--------------------------------|-------|--------|------|-----|-----|------|-----|----------------|-----|
| <b>VV0925WBS01</b> | Dichlorodifluoromethane        | 20    | 18.2   | ug/L | 91  |     |      | 69  | 116            |     |
|                    | Chloromethane                  | 20    | 19.1   | ug/L | 96  |     |      | 65  | 116            |     |
|                    | Vinyl chloride                 | 20    | 19.4   | ug/L | 97  |     |      | 65  | 117            |     |
|                    | Bromomethane                   | 20    | 18.7   | ug/L | 94  |     |      | 58  | 125            |     |
|                    | Chloroethane                   | 20    | 19.3   | ug/L | 97  |     |      | 56  | 128            |     |
|                    | Trichlorofluoromethane         | 20    | 19.4   | ug/L | 97  |     |      | 73  | 115            |     |
|                    | 1,1,2-Trichlorotrifluoroethane | 20    | 20.0   | ug/L | 100 |     |      | 80  | 112            |     |
|                    | Tert butyl alcohol             | 100   | 110    | ug/L | 110 |     |      | 48  | 142            |     |
|                    | 1,1-Dichloroethene             | 20    | 19.9   | ug/L | 100 |     |      | 74  | 110            |     |
|                    | Acetone                        | 100   | 89.0   | ug/L | 89  |     |      | 60  | 125            |     |
|                    | Carbon disulfide               | 20    | 19.0   | ug/L | 95  |     |      | 64  | 112            |     |
|                    | Methyl tert-butyl Ether        | 20    | 20.0   | ug/L | 100 |     |      | 78  | 114            |     |
|                    | Methyl Acetate                 | 20    | 21.2   | ug/L | 106 |     |      | 67  | 125            |     |
|                    | Methylene Chloride             | 20    | 20.6   | ug/L | 103 |     |      | 72  | 114            |     |
|                    | trans-1,2-Dichloroethene       | 20    | 19.7   | ug/L | 99  |     |      | 75  | 108            |     |
|                    | 1,1-Dichloroethane             | 20    | 19.3   | ug/L | 97  |     |      | 78  | 112            |     |
|                    | Cyclohexane                    | 20    | 18.2   | ug/L | 91  |     |      | 75  | 110            |     |
|                    | 2-Butanone                     | 100   | 97.1   | ug/L | 97  |     |      | 65  | 122            |     |
|                    | Carbon Tetrachloride           | 20    | 19.4   | ug/L | 97  |     |      | 77  | 113            |     |
|                    | cis-1,2-Dichloroethene         | 20    | 19.5   | ug/L | 98  |     |      | 77  | 110            |     |
|                    | Bromochloromethane             | 20    | 20.5   | ug/L | 103 |     |      | 70  | 124            |     |
|                    | Chloroform                     | 20    | 19.2   | ug/L | 96  |     |      | 79  | 113            |     |
|                    | 1,1,1-Trichloroethane          | 20    | 19.0   | ug/L | 95  |     |      | 80  | 108            |     |
|                    | Methylcyclohexane              | 20    | 19.2   | ug/L | 96  |     |      | 72  | 115            |     |
|                    | Benzene                        | 20    | 19.9   | ug/L | 100 |     |      | 82  | 109            |     |
|                    | 1,2-Dichloroethane             | 20    | 19.6   | ug/L | 98  |     |      | 80  | 115            |     |
|                    | Trichloroethene                | 20    | 20.4   | ug/L | 102 |     |      | 77  | 113            |     |
|                    | 1,2-Dichloropropane            | 20    | 20.5   | ug/L | 103 |     |      | 83  | 111            |     |
|                    | Bromodichloromethane           | 20    | 19.3   | ug/L | 97  |     |      | 83  | 110            |     |
|                    | 4-Methyl-2-Pentanone           | 100   | 110    | ug/L | 110 |     |      | 74  | 118            |     |
|                    | Toluene                        | 20    | 21.5   | ug/L | 108 |     |      | 82  | 110            |     |
|                    | t-1,3-Dichloropropene          | 20    | 20.1   | ug/L | 101 |     |      | 79  | 110            |     |
|                    | cis-1,3-Dichloropropene        | 20    | 20.2   | ug/L | 101 |     |      | 82  | 110            |     |
|                    | 1,1,2-Trichloroethane          | 20    | 20.0   | ug/L | 100 |     |      | 83  | 112            |     |
|                    | 2-Hexanone                     | 100   | 99.6   | ug/L | 100 |     |      | 73  | 117            |     |
|                    | Dibromochloromethane           | 20    | 20.2   | ug/L | 101 |     |      | 82  | 110            |     |
|                    | 1,2-Dibromoethane              | 20    | 20.6   | ug/L | 103 |     |      | 81  | 110            |     |
|                    | Tetrachloroethene              | 20    | 20.2   | ug/L | 101 |     |      | 67  | 123            |     |
|                    | Chlorobenzene                  | 20    | 20.1   | ug/L | 101 |     |      | 82  | 109            |     |
|                    | Ethyl Benzene                  | 20    | 20.0   | ug/L | 100 |     |      | 83  | 109            |     |
|                    | m/p-Xylenes                    | 40    | 43.1   | ug/L | 108 |     |      | 82  | 110            |     |
|                    | o-Xylene                       | 20    | 21.0   | ug/L | 105 |     |      | 83  | 109            |     |
|                    | Styrene                        | 20    | 19.3   | ug/L | 97  |     |      | 80  | 111            |     |
|                    | Bromoform                      | 20    | 20.5   | ug/L | 103 |     |      | 79  | 109            |     |
|                    | Isopropylbenzene               | 20    | 20.6   | ug/L | 103 |     |      | 83  | 112            |     |
|                    | 1,1,2,2-Tetrachloroethane      | 20    | 19.4   | ug/L | 97  |     |      | 76  | 118            |     |
|                    | 1,3-Dichlorobenzene            | 20    | 19.6   | ug/L | 98  |     |      | 82  | 108            |     |
|                    | 1,4-Dichlorobenzene            | 20    | 18.9   | ug/L | 95  |     |      | 82  | 107            |     |

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

|                 |                               |                           |                          |
|-----------------|-------------------------------|---------------------------|--------------------------|
| <b>SDG No.:</b> | <u><b>Q3175</b></u>           | <b>Analytical Method:</b> | <u><b>SW8260-Low</b></u> |
| <b>Client:</b>  | <u><b>G Environmental</b></u> | <b>Datafile :</b>         | <u><b>VV039171.D</b></u> |

| Lab Sample ID | Parameter                   | Spike | Result | Unit | Rec | RPD | Qual | Low | Limits | RPD |
|---------------|-----------------------------|-------|--------|------|-----|-----|------|-----|--------|-----|
|               |                             |       |        |      |     |     |      |     | High   |     |
| VV0925WBS01   | 1,2-Dichlorobenzene         | 20    | 19.8   | ug/L | 99  |     |      | 82  | 109    |     |
|               | 1,2-Dibromo-3-Chloropropane | 20    | 22.1   | ug/L | 111 |     |      | 68  | 112    |     |
|               | 1,2,4-Trichlorobenzene      | 20    | 19.2   | ug/L | 96  |     |      | 75  | 113    |     |
|               | 1,2,3-Trichlorobenzene      | 20    | 19.9   | ug/L | 100 |     |      | 76  | 114    |     |

VOLATILE METHOD BLANK SUMMARY

Client ID

VV0924WBL01

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG NO.: Q3175

Lab File ID: VV039148.D

Lab Sample ID: VV0924WBL01

Date Analyzed: 09/24/2025

Time Analyzed: 13:09

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA\_V

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

| EPA<br>SAMPLE NO. | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED |
|-------------------|------------------|----------------|------------------|
| VV0924WBS02       | VV0924WBS02      | VV039150.D     | 09/24/2025       |
| MW2               | Q3175-01         | VV039153.D     | 09/24/2025       |
| MW4               | Q3175-02         | VV039154.D     | 09/24/2025       |
| VV0924WBSD02      | VV0924WBSD02     | VV039165.D     | 09/24/2025       |

COMMENTS: \_\_\_\_\_  
\_\_\_\_\_

VOLATILE METHOD BLANK SUMMARY

Client ID

VV0925WBL01

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG NO.: Q3175

Lab File ID: VV039170.D

Lab Sample ID: VV0925WBL01

Date Analyzed: 09/25/2025

Time Analyzed: 11:01

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA\_V

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

| EPA<br>SAMPLE NO. | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED |
|-------------------|------------------|----------------|------------------|
| VV0925WBS01       | VV0925WBS01      | VV039171.D     | 09/25/2025       |
| MW2DL             | Q3175-01DL       | VV039174.D     | 09/25/2025       |
| MW4DL             | Q3175-02DL       | VV039175.D     | 09/25/2025       |
| MW2DL2            | Q3175-01DL2      | VV039179.D     | 09/25/2025       |

COMMENTS: \_\_\_\_\_  
\_\_\_\_\_



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG NO.: Q3175

Lab File ID: VV039134.D

BFB Injection Date: 09/22/2025

Instrument ID: MSVOA\_V

BFB Injection Time: 10:15

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

| m/e | ION ABUNDANCE CRITERIA             | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 50  | 15.0 - 40.0% of mass 95            | 19.6                 |
| 75  | 30.0 - 60.0% of mass 95            | 51.9                 |
| 95  | Base Peak, 100% relative abundance | 100                  |
| 96  | 5.0 - 9.0% of mass 95              | 7.1                  |
| 173 | Less than 2.0% of mass 174         | 1.1 ( 1.6 ) 1        |
| 174 | 50.0 - 100.0% of mass 95           | 73.3                 |
| 175 | 5.0 - 9.0% of mass 174             | 5.5 ( 7.5 ) 1        |
| 176 | 95.0 - 101.0% of mass 174          | 70.4 ( 96 ) 1        |
| 177 | 5.0 - 9.0% of mass 176             | 4.4 ( 6.3 ) 2        |

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| CLIENT ID | LAB SAMPLE ID | LAB FILE ID | DATE ANALYZED | TIME ANALYZED |
|-----------|---------------|-------------|---------------|---------------|
| VSTDIC010 | VSTDIC010     | VV039137.D  | 09/22/2025    | 12:26         |
| VSTDIC020 | VSTDIC020     | VV039138.D  | 09/22/2025    | 12:48         |
| VSTDIC050 | VSTDIC050     | VV039139.D  | 09/22/2025    | 13:26         |
| VSTDIC100 | VSTDIC100     | VV039140.D  | 09/22/2025    | 13:49         |
| VSTDIC005 | VSTDIC005     | VV039142.D  | 09/22/2025    | 15:37         |
| VSTDIC001 | VSTDIC001     | VV039143.D  | 09/22/2025    | 16:35         |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG NO.: Q3175

Lab File ID: VV039145.D

BFB Injection Date: 09/24/2025

Instrument ID: MSVOA\_V

BFB Injection Time: 09:52

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

| m/e | ION ABUNDANCE CRITERIA             | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 50  | 15.0 - 40.0% of mass 95            | 19.6                 |
| 75  | 30.0 - 60.0% of mass 95            | 51.7                 |
| 95  | Base Peak, 100% relative abundance | 100                  |
| 96  | 5.0 - 9.0% of mass 95              | 6.7                  |
| 173 | Less than 2.0% of mass 174         | 1.4 ( 1.9 ) 1        |
| 174 | 50.0 - 100.0% of mass 95           | 77.1                 |
| 175 | 5.0 - 9.0% of mass 174             | 5.2 ( 6.7 ) 1        |
| 176 | 95.0 - 101.0% of mass 174          | 73.5 ( 95.3 ) 1      |
| 177 | 5.0 - 9.0% of mass 176             | 4.9 ( 6.6 ) 2        |

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| CLIENT ID    | LAB SAMPLE ID | LAB FILE ID | DATE ANALYZED | TIME ANALYZED |
|--------------|---------------|-------------|---------------|---------------|
| VSTDCCC050   | VSTDCCC050    | VV039146.D  | 09/24/2025    | 10:49         |
| VV0924WBL01  | VV0924WBL01   | VV039148.D  | 09/24/2025    | 13:09         |
| VV0924WBS02  | VV0924WBS02   | VV039150.D  | 09/24/2025    | 14:01         |
| MW2          | Q3175-01      | VV039153.D  | 09/24/2025    | 15:31         |
| MW4          | Q3175-02      | VV039154.D  | 09/24/2025    | 15:53         |
| VV0924WBSD02 | VV0924WBSD02  | VV039165.D  | 09/24/2025    | 19:59         |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG NO.: Q3175

Lab File ID: VV039167.D

BFB Injection Date: 09/25/2025

Instrument ID: MSVOA\_V

BFB Injection Time: 09:14

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

| m/e | ION ABUNDANCE CRITERIA             | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 50  | 15.0 - 40.0% of mass 95            | 16.2                 |
| 75  | 30.0 - 60.0% of mass 95            | 49.9                 |
| 95  | Base Peak, 100% relative abundance | 100                  |
| 96  | 5.0 - 9.0% of mass 95              | 6.3                  |
| 173 | Less than 2.0% of mass 174         | 1.3 ( 1.7 ) 1        |
| 174 | 50.0 - 100.0% of mass 95           | 72.3                 |
| 175 | 5.0 - 9.0% of mass 174             | 5.7 ( 7.9 ) 1        |
| 176 | 95.0 - 101.0% of mass 174          | 69.9 ( 96.6 ) 1      |
| 177 | 5.0 - 9.0% of mass 176             | 4.4 ( 6.3 ) 2        |

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| CLIENT ID   | LAB SAMPLE ID | LAB FILE ID | DATE ANALYZED | TIME ANALYZED |
|-------------|---------------|-------------|---------------|---------------|
| VSTDCCC050  | VSTDCCC050    | VV039168.D  | 09/25/2025    | 09:59         |
| VV0925WBL01 | VV0925WBL01   | VV039170.D  | 09/25/2025    | 11:01         |
| VV0925WBS01 | VV0925WBS01   | VV039171.D  | 09/25/2025    | 11:24         |
| MW2DL       | Q3175-01DL    | VV039174.D  | 09/25/2025    | 12:33         |
| MW4DL       | Q3175-02DL    | VV039175.D  | 09/25/2025    | 12:55         |
| MW2DL2      | Q3175-01DL2   | VV039179.D  | 09/25/2025    | 14:25         |

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01  
Lab Code: ACE SDG NO.: Q3175  
Lab File ID: VV039146.D Date Analyzed: 09/24/2025  
Instrument ID: MSVOA\_V Time Analyzed: 10:49  
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                | IS1<br>AREA # | RT # | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #  |
|----------------|---------------|------|---------------|-------|---------------|-------|
| 12 HOUR STD    | 558154        | 4.60 | 907128        | 5.53  | 923608        | 8.78  |
| UPPER LIMIT    | 1116310       | 5.1  | 1814260       | 6.032 | 1847220       | 9.277 |
| LOWER LIMIT    | 279077        | 4.1  | 453564        | 5.032 | 461804        | 8.277 |
| EPA SAMPLE NO. |               |      |               |       |               |       |
| MW2            | 378275        | 4.60 | 655768        | 5.54  | 600147        | 8.78  |
| MW4            | 627637        | 4.60 | 1091934       | 5.54  | 1002921       | 8.78  |
| VV0924WBL01    | 463770        | 4.60 | 858892        | 5.53  | 831675        | 8.77  |
| VV0924WBS02    | 492318        | 4.60 | 795602        | 5.53  | 784865        | 8.78  |
| VV0924WBSD02   | 545364        | 4.60 | 903558        | 5.53  | 885424        | 8.78  |

IS1 = Pentafluorobenzene  
IS2 = 1,4-Difluorobenzene  
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
AREA LOWER LIMIT = -50% of internal standard area  
RT UPPER LIMIT = +0.50 minutes of internal standard RT  
RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
\* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

|                |                   |                |                   |
|----------------|-------------------|----------------|-------------------|
| Lab Name:      | <u>Alliance</u>   | Contract:      | <u>GENV01</u>     |
| Lab Code:      | <u>ACE</u>        | SDG NO.:       | <u>Q3175</u>      |
| Lab File ID:   | <u>VV039146.D</u> | Date Analyzed: | <u>09/24/2025</u> |
| Instrument ID: | <u>MSVOA_V</u>    | Time Analyzed: | <u>10:49</u>      |
| GC Column:     | <u>DB-624UI</u>   | ID:            | <u>0.18</u> (mm)  |
|                |                   | Heated Purge:  | (Y/N) <u>N</u>    |

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 551620        | 11.172 |  |  |  |  |
| UPPER LIMIT    | 1103240       | 11.672 |  |  |  |  |
| LOWER LIMIT    | 275810        | 10.672 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| MW2            | 318916        | 11.18  |  |  |  |  |
| MW4            | 492612        | 11.18  |  |  |  |  |
| VV0924WBL01    | 379013        | 11.18  |  |  |  |  |
| VV0924WBS02    | 468522        | 11.17  |  |  |  |  |
| VV0924WBSD02   | 510716        | 11.17  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01  
 Lab Code: ACE SDG NO.: Q3175  
 Lab File ID: VV039168.D Date Analyzed: 09/25/2025  
 Instrument ID: MSVOA\_V Time Analyzed: 09:59  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #  |
|----------------|---------------|-------|---------------|-------|---------------|-------|
| 12 HOUR STD    | 605297        | 4.60  | 933338        | 5.53  | 901987        | 8.77  |
| UPPER LIMIT    | 1210590       | 5.096 | 1866680       | 6.029 | 1803970       | 9.273 |
| LOWER LIMIT    | 302649        | 4.096 | 466669        | 5.029 | 450994        | 8.273 |
| EPA SAMPLE NO. |               |       |               |       |               |       |
| MW2DL          | 563178        | 4.60  | 979530        | 5.54  | 954674        | 8.78  |
| MW2DL2         | 457907        | 4.60  | 814896        | 5.54  | 779302        | 8.78  |
| MW4DL          | 560555        | 4.60  | 1013068       | 5.54  | 987441        | 8.78  |
| VV0925WBL01    | 480047        | 4.60  | 864912        | 5.53  | 870930        | 8.78  |
| VV0925WBS01    | 557017        | 4.60  | 898902        | 5.54  | 871912        | 8.78  |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

|                |                   |                |                   |
|----------------|-------------------|----------------|-------------------|
| Lab Name:      | <u>Alliance</u>   | Contract:      | <u>GENV01</u>     |
| Lab Code:      | <u>ACE</u>        | SDG NO.:       | <u>Q3175</u>      |
| Lab File ID:   | <u>VV039168.D</u> | Date Analyzed: | <u>09/25/2025</u> |
| Instrument ID: | <u>MSVOA_V</u>    | Time Analyzed: | <u>09:59</u>      |
| GC Column:     | <u>DB-624UI</u>   | ID:            | <u>0.18</u> (mm)  |
|                |                   | Heated Purge:  | (Y/N) <u>N</u>    |

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 535710        | 11.172 |  |  |  |  |
| UPPER LIMIT    | 1071420       | 11.672 |  |  |  |  |
| LOWER LIMIT    | 267855        | 10.672 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| MW2DL          | 558535        | 11.17  |  |  |  |  |
| MW2DL2         | 406825        | 11.18  |  |  |  |  |
| MW4DL          | 548334        | 11.17  |  |  |  |  |
| VV0925WBL01    | 413592        | 11.18  |  |  |  |  |
| VV0925WBS01    | 508947        | 11.17  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.



# SAMPLE DATA

### Report of Analysis

|                    |                 |                 |              |
|--------------------|-----------------|-----------------|--------------|
| Client:            | G Environmental | Date Collected: | 09/23/25     |
| Project:           | Summer          | Date Received:  | 09/23/25     |
| Client Sample ID:  | MW2             | SDG No.:        | Q3175        |
| Lab Sample ID:     | Q3175-01        | Matrix:         | Water        |
| Analytical Method: | 8260D           | % Solid:        | 0            |
| Sample Wt/Vol:     | 5               | Units:          | mL           |
| Soil Aliquot Vol:  |                 |                 | uL           |
| GC Column:         | DB-624UI        | ID :            | 0.18         |
| Prep Method :      |                 | Level :         | LOW          |
|                    |                 | Final Vol:      | 5000         |
|                    |                 | Test:           | VOCMS Group1 |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039153.D        | 1         | 09/24/25 15:31 | VV092425      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL   | LOQ / CRQL | Units |
|----------------|--------------------------------|-------|-----------|-------|------------|-------|
| <b>TARGETS</b> |                                |       |           |       |            |       |
| 75-71-8        | Dichlorodifluoromethane        | 0.22  | U         | 0.22  | 1.00       | ug/L  |
| 74-87-3        | Chloromethane                  | 0.32  | U         | 0.32  | 1.00       | ug/L  |
| 75-01-4        | Vinyl Chloride                 | 0.26  | U         | 0.26  | 1.00       | ug/L  |
| 74-83-9        | Bromomethane                   | 1.40  | U         | 1.40  | 5.00       | ug/L  |
| 75-00-3        | Chloroethane                   | 0.47  | U         | 0.47  | 1.00       | ug/L  |
| 75-69-4        | Trichlorofluoromethane         | 0.33  | U         | 0.33  | 1.00       | ug/L  |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 0.25  | U         | 0.25  | 1.00       | ug/L  |
| 75-65-0        | Tert butyl alcohol             | 5.50  | U         | 5.50  | 25.0       | ug/L  |
| 75-35-4        | 1,1-Dichloroethene             | 0.23  | U         | 0.23  | 1.00       | ug/L  |
| 67-64-1        | Acetone                        | 1.50  | U         | 1.50  | 5.00       | ug/L  |
| 75-15-0        | Carbon Disulfide               | 1.50  |           | 0.21  | 1.00       | ug/L  |
| 1634-04-4      | Methyl tert-butyl Ether        | 0.16  | U         | 0.16  | 1.00       | ug/L  |
| 79-20-9        | Methyl Acetate                 | 0.27  | U         | 0.27  | 1.00       | ug/L  |
| 75-09-2        | Methylene Chloride             | 0.28  | U         | 0.28  | 1.00       | ug/L  |
| 156-60-5       | trans-1,2-Dichloroethene       | 0.23  | U         | 0.23  | 1.00       | ug/L  |
| 75-34-3        | 1,1-Dichloroethane             | 0.23  | U         | 0.23  | 1.00       | ug/L  |
| 110-82-7       | Cyclohexane                    | 410   | E         | 1.50  | 5.00       | ug/L  |
| 78-93-3        | 2-Butanone                     | 0.98  | U         | 0.98  | 5.00       | ug/L  |
| 56-23-5        | Carbon Tetrachloride           | 0.25  | U         | 0.25  | 1.00       | ug/L  |
| 156-59-2       | cis-1,2-Dichloroethene         | 0.19  | U         | 0.19  | 1.00       | ug/L  |
| 74-97-5        | Bromochloromethane             | 0.22  | U         | 0.22  | 1.00       | ug/L  |
| 67-66-3        | Chloroform                     | 0.25  | U         | 0.25  | 1.00       | ug/L  |
| 71-55-6        | 1,1,1-Trichloroethane          | 0.20  | U         | 0.20  | 1.00       | ug/L  |
| 108-87-2       | Methylcyclohexane              | 270   | E         | 0.16  | 1.00       | ug/L  |
| 71-43-2        | Benzene                        | 460   | E         | 0.15  | 1.00       | ug/L  |
| 107-06-2       | 1,2-Dichloroethane             | 0.22  | U         | 0.22  | 1.00       | ug/L  |
| 79-01-6        | Trichloroethene                | 0.090 | U         | 0.090 | 1.00       | ug/L  |
| 78-87-5        | 1,2-Dichloropropane            | 0.20  | U         | 0.20  | 1.00       | ug/L  |
| 75-27-4        | Bromodichloromethane           | 0.22  | U         | 0.22  | 1.00       | ug/L  |
| 108-10-1       | 4-Methyl-2-Pentanone           | 0.68  | UQ        | 0.68  | 5.00       | ug/L  |

### Report of Analysis

|                    |                 |           |                 |              |    |
|--------------------|-----------------|-----------|-----------------|--------------|----|
| Client:            | G Environmental |           | Date Collected: | 09/23/25     |    |
| Project:           | Summer          |           | Date Received:  | 09/23/25     |    |
| Client Sample ID:  | MW2             |           | SDG No.:        | Q3175        |    |
| Lab Sample ID:     | Q3175-01        |           | Matrix:         | Water        |    |
| Analytical Method: | 8260D           |           | % Solid:        | 0            |    |
| Sample Wt/Vol:     | 5               | Units: mL | Final Vol:      | 5000         | uL |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOCMS Group1 |    |
| GC Column:         | DB-624UI        | ID : 0.18 | Level :         | LOW          |    |
| Prep Method :      |                 |           |                 |              |    |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039153.D        | 1         | 09/24/25 15:31 | VV092425      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units   |
|---------------------------|-----------------------------|--------|-----------|----------|------------|---------|
| 108-88-3                  | Toluene                     | 100    | E         | 0.14     | 1.00       | ug/L    |
| 10061-02-6                | t-1,3-Dichloropropene       | 0.17   | U         | 0.17     | 1.00       | ug/L    |
| 10061-01-5                | cis-1,3-Dichloropropene     | 0.16   | U         | 0.16     | 1.00       | ug/L    |
| 79-00-5                   | 1,1,2-Trichloroethane       | 0.21   | U         | 0.21     | 1.00       | ug/L    |
| 591-78-6                  | 2-Hexanone                  | 0.89   | U         | 0.89     | 5.00       | ug/L    |
| 124-48-1                  | Dibromochloromethane        | 0.18   | U         | 0.18     | 1.00       | ug/L    |
| 106-93-4                  | 1,2-Dibromoethane           | 0.15   | U         | 0.15     | 1.00       | ug/L    |
| 127-18-4                  | Tetrachloroethene           | 0.23   | U         | 0.23     | 1.00       | ug/L    |
| 108-90-7                  | Chlorobenzene               | 0.12   | U         | 0.12     | 1.00       | ug/L    |
| 100-41-4                  | Ethyl Benzene               | 1300   | E         | 0.13     | 1.00       | ug/L    |
| 179601-23-1               | m/p-Xylenes                 | 3100   | E         | 0.24     | 2.00       | ug/L    |
| 95-47-6                   | o-Xylene                    | 120    | E         | 0.12     | 1.00       | ug/L    |
| 100-42-5                  | Styrene                     | 0.15   | U         | 0.15     | 1.00       | ug/L    |
| 75-25-2                   | Bromoform                   | 0.19   | U         | 0.19     | 1.00       | ug/L    |
| 98-82-8                   | Isopropylbenzene            | 430    | E         | 0.12     | 1.00       | ug/L    |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 0.26   | U         | 0.26     | 1.00       | ug/L    |
| 541-73-1                  | 1,3-Dichlorobenzene         | 0.16   | U         | 0.16     | 1.00       | ug/L    |
| 106-46-7                  | 1,4-Dichlorobenzene         | 0.19   | U         | 0.19     | 1.00       | ug/L    |
| 95-50-1                   | 1,2-Dichlorobenzene         | 0.16   | U         | 0.16     | 1.00       | ug/L    |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 0.53   | U         | 0.53     | 1.00       | ug/L    |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 0.20   | U         | 0.20     | 1.00       | ug/L    |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 0.20   | U         | 0.20     | 1.00       | ug/L    |
| <b>SURROGATES</b>         |                             |        |           |          |            |         |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 44.7   |           | 74 - 125 | 89%        | SPK: 50 |
| 1868-53-7                 | Dibromofluoromethane        | 38.9   |           | 75 - 124 | 78%        | SPK: 50 |
| 2037-26-5                 | Toluene-d8                  | 47.6   |           | 86 - 113 | 95%        | SPK: 50 |
| 460-00-4                  | 4-Bromofluorobenzene        | 45.2   |           | 77 - 121 | 90%        | SPK: 50 |
| <b>INTERNAL STANDARDS</b> |                             |        |           |          |            |         |
| 363-72-4                  | Pentafluorobenzene          | 378000 | 4.6       |          |            |         |
| 540-36-3                  | 1,4-Difluorobenzene         | 656000 | 5.535     |          |            |         |
| 3114-55-4                 | Chlorobenzene-d5            | 600000 | 8.779     |          |            |         |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 319000 | 11.178    |          |            |         |

## Report of Analysis

|                    |                            |                 |              |
|--------------------|----------------------------|-----------------|--------------|
| Client:            | G Environmental            | Date Collected: | 09/23/25     |
| Project:           | Summer                     | Date Received:  | 09/23/25     |
| Client Sample ID:  | MW2                        | SDG No.:        | Q3175        |
| Lab Sample ID:     | Q3175-01                   | Matrix:         | Water        |
| Analytical Method: | 8260D                      | % Solid:        | 0            |
| Sample Wt/Vol:     | 5      Units:    mL        | Final Vol:      | 5000      uL |
| Soil Aliquot Vol:  | uL                         | Test:           | VOCMS Group1 |
| GC Column:         | DB-624UI      ID :    0.18 | Level :         | LOW          |
| Prep Method :      |                            |                 |              |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039153.D        | 1         | 09/24/25 15:31 | VV092425      |

| CAS Number                            | Parameter                        | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|---------------------------------------|----------------------------------|-------|-----------|-----|------------|-------|
| <b>TENTATIVE IDENTIFIED COMPOUNDS</b> |                                  |       |           |     |            |       |
| 103-65-1                              | n-propylbenzene                  | 650   | J         |     | 10.3       | ug/L  |
| 000611-14-3                           | Benzene, 1-ethyl-2-methyl-       | 1100  | J         |     | 10.4       | ug/L  |
| 108-67-8                              | 1,3,5-Trimethylbenzene           | 1000  | J         |     | 10.5       | ug/L  |
| 000622-96-8                           | Benzene, 1-ethyl-4-methyl-       | 440   | J         |     | 10.7       | ug/L  |
| 95-63-6                               | 1,2,4-Trimethylbenzene           | 1600  | J         |     | 10.8       | ug/L  |
| 135-98-8                              | sec-Butylbenzene                 | 48.9  | J         |     | 11.0       | ug/L  |
| 99-87-6                               | p-Isopropyltoluene               | 34.0  | J         |     | 11.2       | ug/L  |
| 000526-73-8                           | Benzene, 1,2,3-trimethyl-        | 1400  | J         |     | 11.3       | ug/L  |
| 104-51-8                              | n-Butylbenzene                   | 140   | J         |     | 11.6       | ug/L  |
| 001074-55-1                           | Benzene, 1-methyl-4-propyl-      | 120   | J         |     | 11.7       | ug/L  |
| 001758-88-9                           | Benzene, 2-ethyl-1,4-dimethyl-   | 270   | J         |     | 11.8       | ug/L  |
| 000527-84-4                           | o-Cymene                         | 210   | J         |     | 11.9       | ug/L  |
| 000934-80-5                           | Benzene, 4-ethyl-1,2-dimethyl-   | 490   | J         |     | 11.9       | ug/L  |
| 007525-62-4                           | Benzene, 1-ethenyl-3-ethyl-      | 230   | J         |     | 12.0       | ug/L  |
| 000488-23-3                           | Benzene, 1,2,3,4-tetramethyl-    | 250   | J         |     | 12.3       | ug/L  |
| 000095-93-2                           | Benzene, 1,2,4,5-tetramethyl-    | 340   | J         |     | 12.4       | ug/L  |
| 000824-22-6                           | 1H-Indene, 2,3-dihydro-4-methyl- | 250   | J         |     | 12.7       | ug/L  |
| 002039-89-6                           | Benzene, 2-ethenyl-1,4-dimethyl- | 560   | J         |     | 12.8       | ug/L  |
| 91-20-3                               | Naphthalene                      | 1100  | J         |     | 13.4       | ug/L  |
| 000090-12-0                           | Naphthalene, 1-methyl-           | 370   | J         |     | 14.6       | ug/L  |
| 000091-57-6                           | Naphthalene, 2-methyl-           | 180   | J         |     | 14.7       | ug/L  |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
 Data File : VV039153.D  
 Acq On : 24 Sep 2025 15:31  
 Operator : SY/MD  
 Sample : Q3175-01  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 MW2

Manual Integrations  
 APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025  
 Supervised By :Semsettin Yesilyurt 09/26/2025

Quant Time: Sep 25 04:48:07 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

| Compound                    | R.T.           | QIon | Response   | Conc     | Units  | Dev(Min) |
|-----------------------------|----------------|------|------------|----------|--------|----------|
| -----                       |                |      |            |          |        |          |
| Internal Standards          |                |      |            |          |        |          |
| 1) Pentafluorobenzene       | 4.600          | 168  | 378275     | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 5.535          | 114  | 655768     | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 8.779          | 117  | 600147     | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 11.178         | 152  | 318916     | 50.000   | ug/l   | 0.00     |
|                             |                |      |            |          |        |          |
| System Monitoring Compounds |                |      |            |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 4.944          | 65   | 244683     | 44.693   | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 81 - 118 |      | Recovery = | 89.380%  |        |          |
| 35) Dibromofluoromethane    | 4.503          | 113  | 205255     | 38.934   | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 80 - 119 |      | Recovery = | 77.860%# |        |          |
| 50) Toluene-d8              | 7.239          | 98   | 737311     | 47.619   | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 89 - 112 |      | Recovery = | 95.240%  |        |          |
| 62) 4-Bromofluorobenzene    | 10.005         | 95   | 288262     | 45.224   | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 85 - 114 |      | Recovery = | 90.440%  |        |          |
|                             |                |      |            |          |        |          |
| Target Compounds            |                |      |            |          | Qvalue |          |
| 17) Carbon Disulfide        | 2.262          | 76   | 26662      | 1.529    | ug/l # | 93       |
| 31) Cyclohexane             | 4.583          | 56   | 3994761    | 409.101  | ug/l   | 96       |
| 39) Methylcyclohexane       | 6.046          | 83   | 2582048    | 268.253  | ug/l   | 99       |
| 40) Benzene                 | 5.008          | 78   | 11848259   | 456.496  | ug/l   | 98       |
| 52) Toluene                 | 7.307          | 92   | 1554151    | 103.275  | ug/l   | 99       |
| 67) Ethyl Benzene           | 8.924          | 91   | 32943870m  | 1301.902 | ug/l   |          |
| 68) m/p-Xylenes             | 9.059          | 106  | 30412638m  | 3113.968 | ug/l   |          |
| 69) o-Xylene                | 9.471          | 106  | 1043936    | 120.719  | ug/l   | 99       |
| 73) Isopropylbenzene        | 9.857          | 105  | 9939721    | 429.711  | ug/l   | 99       |
| 78) n-propylbenzene         | 10.268         | 91   | 18384305m  | 652.664  | ug/l   |          |
| 80) 1,3,5-Trimethylbenzene  | 10.458         | 105  | 19445969m  | 1011.427 | ug/l   |          |
| 84) 1,2,4-Trimethylbenzene  | 10.831         | 105  | 29248488m  | 1566.967 | ug/l   |          |
| 85) sec-Butylbenzene        | 11.014         | 105  | 1244555    | 48.943   | ug/l   | 100      |
| 86) p-Isopropyltoluene      | 11.172         | 119  | 718828m    | 33.964   | ug/l   |          |
| 89) n-Butylbenzene          | 11.574         | 91   | 2920675m   | 144.495  | ug/l   |          |
| 95) Naphthalene             | 13.413         | 128  | 22556658m  | 1072.874 | ug/l   |          |
| -----                       |                |      |            |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039153.D  
Acq On : 24 Sep 2025 15:31  
Operator : SY/MD  
Sample : Q3175-01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 10 Sample Multiplier: 1


Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2

Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025  
Supervised By :Semsettin Yesilyurt 09/26/2025

Quant Time: Sep 25 04:48:07 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration









#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 4.600 min Scan# 1106  
Delta R.T. 0.003 min  
Lab File: VV039153.D  
Acq: 24 Sep 2025 15:31

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2

Tgt Ion:168 Resp: 37827  
Ion Ratio Lower Upper  
168 100  
99 61.1 49.6 74.4




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 09/26/2025  
Supervised By :Semsettin Yesilyurt 09/26/2025



#17  
Carbon Disulfide  
Concen: 1.529 ug/l  
RT: 2.262 min Scan# 380  
Delta R.T. 0.009 min  
Lab File: VV039153.D  
Acq: 24 Sep 2025 15:31


Tgt Ion: 76 Resp: 26662  
Ion Ratio Lower Upper  
76 100  
78 6.5 7.4 11.0#





#31  
Cyclohexane  
Concen: 409.101 ug/l  
RT: 4.583 min Scan# 1101  
Delta R.T. 0.003 min  
Lab File: VV039153.D  
Acq: 24 Sep 2025 15:31




Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2



Tgt Ion: 56 Resp: 399476  
Ion Ratio Lower Upper  
56 100  
69 29.2 26.2 39.2  
84 85.5 70.3 105.5




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 09/26/2025  
Supervised By :Semsettin Yesilyurt 09/26/2025



#33  
1,2-Dichloroethane-d4  
Concen: 44.693 ug/l  
RT: 4.944 min Scan# 1214  
Delta R.T. 0.006 min  
Lab File: VV039153.D  
Acq: 24 Sep 2025 15:31

Tgt Ion: 65 Resp: 244683  
Ion Ratio Lower Upper  
65 100  
67 51.0 0.0 107.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 5.535 min Scan# 11

Delta R.T. 0.006 min

Lab File: VV039153.D

Acq: 24 Sep 2025 15:31

Instrument :

MSVOA\_V

ClientSampleId :

MW2

Tgt Ion:114 Resp: 65576

Ion Ratio Lower Upper

114 100

63 20.9 0.0 42.6





88 16.5 0.0 30.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025

Supervised By :Semsettin Yesilyurt 09/26/2025



#35

Dibromofluoromethane

Concen: 38.934 ug/l

RT: 4.503 min Scan# 1077

Delta R.T. 0.006 min

Lab File: VV039153.D

Acq: 24 Sep 2025 15:31




Tgt Ion:113 Resp: 205255


Ion Ratio Lower Upper

113 100

111 102.5 82.4 123.6

192 17.3 13.8 20.8





#39

Methylcyclohexane

Concen: 268.253 ug/l

RT: 6.046 min Scan# 1

Delta R.T. 0.003 min

Lab File: VV039153.D

Acq: 24 Sep 2025 15:31

Instrument :

MSVOA\_V

ClientSampleId :

MW2

Tgt Ion: 83 Resp: 258204

Ion Ratio Lower Upper

83 100

55 77.2 60.5 90.7





98 47.4 37.8 56.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025

Supervised By :Semsettin Yesilyurt 09/26/2025



#40

Benzene

Concen: 456.496 ug/l

RT: 5.008 min Scan# 1234

Delta R.T. 0.003 min

Lab File: VV039153.D




Acq: 24 Sep 2025 15:31


Tgt Ion: 78 Resp:11848259

Ion Ratio Lower Upper

78 100

77 24.4 18.7 28.1





#50

Toluene-d8

Concen: 47.619 ug/l

RT: 7.239 min Scan# 1928

Delta R.T. 0.006 min

Lab File: VV039153.D

Acq: 24 Sep 2025 15:31

Instrument :

MSVOA\_V

ClientSampleId :

MW2

Tgt Ion: 98 Resp: 73731

Ion Ratio Lower Upper





98 100

100 63.9 52.2 78.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025  
Supervised By :Semsettin Yesilyurt 09/26/2025



#52

Toluene

Concen: 103.275 ug/l

RT: 7.307 min Scan# 1949

Delta R.T. 0.003 min

Lab File: VV039153.D




Acq: 24 Sep 2025 15:31


Tgt Ion: 92 Resp: 1554151

Ion Ratio Lower Upper

92 100


91 172.2 139.2 208.8





#62  
4-Bromofluorobenzene  
Concen: 45.224 ug/l  
RT: 10.005 min Scan# 21  
Delta R.T. 0.006 min  
Lab File: VV039153.D  
Acq: 24 Sep 2025 15:31




Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2



Tgt Ion: 95 Resp: 28826  
Ion Ratio Lower Upper  
95 100  
174 72.3 0.0 150.4  
176 70.2 0.0 144.2




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 09/26/2025  
Supervised By :Semsettin Yesilyurt 09/26/2025



#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 8.779 min Scan# 2407  
Delta R.T. 0.006 min  
Lab File: VV039153.D  
Acq: 24 Sep 2025 15:31

Tgt Ion:117 Resp: 600147  
Ion Ratio Lower Upper  
117 100  
82 55.0 45.4 68.2  
119 31.8 25.6 38.4





#67

Ethyl Benzene

Concen: 1301.902 ug/l m

RT: 8.924 min Scan# 2455

Delta R.T. -0.010 min

Lab File: VV039153.D

Acq: 24 Sep 2025 15:31

Instrument :

MSVOA\_V

ClientSampleId :

MW2

Tgt Ion: 91 Resp:32943870

Ion Ratio Lower Upper





91 100

106 47.0 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025  
 Supervised By :Semsettin Yesilyurt 09/26/2025



#68

m/p-Xylenes

Concen: 3113.968 ug/l m

RT: 9.059 min Scan# 2494

Delta R.T. 0.000 min

Lab File: VV039153.D




Acq: 24 Sep 2025 15:31


Tgt Ion:106 Resp:30412638

Ion Ratio Lower Upper

106 100

91 0.0 162.5 243.7#









#69  
o-Xylene  
Concen: 120.719 ug/l  
RT: 9.471 min Scan# 2621  
Delta R.T. 0.003 min  
Lab File: VV039153.D  
Acq: 24 Sep 2025 15:31

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2

Tgt Ion:106 Resp: 1043930  
Ion Ratio Lower Upper  
106 100  
91 215.9 109.1 327.3




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 09/26/2025  
Supervised By :Semsettin Yesilyurt 09/26/2025



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 11.178 min Scan# 3153  
Delta R.T. 0.006 min  
Lab File: VV039153.D  
Acq: 24 Sep 2025 15:31

Tgt Ion:152 Resp: 318916  
Ion Ratio Lower Upper  
152 100  
115 77.3 44.8 134.4  
150 157.7 0.0 353.8





#73

Isopropylbenzene

Concen: 429.711 ug/l

RT: 9.857 min Scan# 21

Delta R.T. 0.003 min

Lab File: VV039153.D

Acq: 24 Sep 2025 15:31

Instrument :

MSVOA\_V

ClientSampleId :

MW2

Tgt Ion:105 Resp: 993972

Ion Ratio Lower Upper

105 100





120 26.7 13.1 39.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025

Supervised By :Semsettin Yesilyurt 09/26/2025



#78

n-propylbenzene

Concen: 652.664 ug/l m

RT: 10.268 min Scan# 2870

Delta R.T. -0.007 min

Lab File: VV039153.D




Acq: 24 Sep 2025 15:31


Tgt Ion: 91 Resp:18384305

Ion Ratio Lower Upper

91 100


120 59.9 11.1 33.3#





#80  
1,3,5-Trimethylbenzene  
Concen: 1011.427 ug/l m  
RT: 10.458 min Scan# 2931  
Delta R.T. -0.007 min  
Lab File: VV039153.D  
Acq: 24 Sep 2025 15:31




Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2



Tgt Ion:105 Resp:19445969  
Ion Ratio Lower Upper  
105 100  
120 21.1 24.3 72.8




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 09/26/2025  
Supervised By :Semsettin Yesilyurt 09/26/2025



#84  
1,2,4-Trimethylbenzene  
Concen: 1566.967 ug/l m  
RT: 10.831 min Scan# 3045  
Delta R.T. -0.007 min  
Lab File: VV039153.D  
Acq: 24 Sep 2025 15:31


Tgt Ion:105 Resp:29248488  
Ion Ratio Lower Upper  
105 100  
120 14.0 22.4 67.2#





#85  
 sec-Butylbenzene  
 Concen: 48.943 ug/l  
 RT: 11.014 min Scan# 3101  
 Delta R.T. 0.003 min  
 Lab File: VV039153.D  
 Acq: 24 Sep 2025 15:31




Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 MW2



Tgt Ion:105 Resp: 124455  
 Ion Ratio Lower Upper  
 105 100  
 134 19.4 9.6 28.8




Manual Integrations  
 APPROVED


Reviewed By :Mahesh Dadoda 09/26/2025  
 Supervised By :Semsettin Yesilyurt 09/26/2025



#86  
 p-Isopropyltoluene  
 Concen: 33.964 ug/l m  
 RT: 11.172 min Scan# 3151  
 Delta R.T. 0.006 min  
 Lab File: VV039153.D  
 Acq: 24 Sep 2025 15:31

Tgt Ion:119 Resp: 718828  
 Ion Ratio Lower Upper  
 119 100  
 134 61.8 12.7 38.0#  
 91 55.4 12.0 36.1#





#89

n-Butylbenzene

Concen: 144.495 ug/l m

RT: 11.574 min Scan# 31

Delta R.T. -0.003 min

Lab File: VV039153.D

Acq: 24 Sep 2025 15:31

Instrument :

MSVOA\_V

ClientSampleId :

MW2

Tgt Ion: 91 Resp: 2920679

Ion Ratio Lower Upper

91 100

92 38.4 26.6 79.8





134 144.8 12.1 36.3

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025

Supervised By :Semsettin Yesilyurt 09/26/2025



#95

Naphthalene

Concen: 1072.874 ug/l m

RT: 13.413 min Scan# 3848

Delta R.T. -0.010 min

Lab File: VV039153.D

Acq: 24 Sep 2025 15:31




Tgt Ion:128 Resp:22556658

Ion Ratio Lower Upper

128 100

127 0.0 10.3 15.5#

129 0.0 8.6 13.0#



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
 Data File : VV039153.D  
 Acq On : 24 Sep 2025 15:31  
 Operator : SY/MD  
 Sample : Q3175-01  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 MW2

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Title : SW846 8260

Signal : TIC: VV039153.D\data.ms

| peak<br># | R.T.<br>min | first<br>scan | max<br>scan | last<br>scan | PK<br>TY | peak<br>height | corr.<br>area | corr.<br>% max. | % of<br>total |
|-----------|-------------|---------------|-------------|--------------|----------|----------------|---------------|-----------------|---------------|
| 1         | 1.294       | 71            | 79          | 107          | rVB      | 4555692        | 6354746       | 1.95%           | 0.308%        |
| 2         | 1.609       | 166           | 177         | 214          | rBV      | 19751438       | 32152253      | 9.87%           | 1.558%        |
| 3         | 1.780       | 220           | 230         | 251          | rBV3     | 9432300        | 17128104      | 5.26%           | 0.830%        |
| 4         | 1.870       | 251           | 258         | 274          | rVV      | 2090420        | 3462107       | 1.06%           | 0.168%        |
| 5         | 1.992       | 288           | 296         | 319          | rVV      | 4986792        | 8804315       | 2.70%           | 0.427%        |
| 6         | 2.481       | 408           | 448         | 455          | rBV2     | 6283373        | 19515166      | 5.99%           | 0.946%        |
| 7         | 2.526       | 456           | 462         | 503          | rVB2     | 6112025        | 14167384      | 4.35%           | 0.687%        |
| 8         | 2.706       | 503           | 518         | 558          | rVB      | 4280423        | 9899501       | 3.04%           | 0.480%        |
| 9         | 2.976       | 589           | 602         | 623          | rBV      | 2417480        | 5301174       | 1.63%           | 0.257%        |
| 10        | 3.198       | 661           | 671         | 687          | rVB      | 2282894        | 4927401       | 1.51%           | 0.239%        |
| 11        | 3.497       | 752           | 764         | 779          | rBV      | 1651064        | 3863844       | 1.19%           | 0.187%        |
| 12        | 3.674       | 803           | 819         | 883          | rVB      | 13346898       | 39427711      | 12.11%          | 1.911%        |
| 13        | 4.362       | 996           | 1033        | 1057         | rBV      | 2542542        | 7334427       | 2.25%           | 0.356%        |
| 14        | 4.583       | 1085          | 1102        | 1118         | rBV2     | 6923575        | 19324545      | 5.93%           | 0.937%        |
| 15        | 4.670       | 1119          | 1129        | 1160         | rVB      | 2987275        | 9000307       | 2.76%           | 0.436%        |
| 16        | 4.818       | 1161          | 1175        | 1204         | rVB2     | 1361447        | 4311867       | 1.32%           | 0.209%        |
| 17        | 5.008       | 1220          | 1234        | 1250         | rBV      | 10324769       | 24093318      | 7.40%           | 1.168%        |
| 18        | 5.146       | 1268          | 1277        | 1294         | rVB5     | 2382293        | 6874519       | 2.11%           | 0.333%        |
| 19        | 5.574       | 1388          | 1410        | 1437         | rBV7     | 1851007        | 7940463       | 2.44%           | 0.385%        |
| 20        | 6.046       | 1533          | 1557        | 1573         | rBV3     | 5686951        | 14344687      | 4.40%           | 0.695%        |
| 21        | 6.722       | 1742          | 1767        | 1776         | rBV3     | 2046267        | 6560523       | 2.01%           | 0.318%        |
| 22        | 7.088       | 1869          | 1881        | 1899         | rVB      | 1758948        | 3564941       | 1.09%           | 0.173%        |
| 23        | 7.307       | 1940          | 1949        | 1962         | rBV      | 3426570        | 6010301       | 1.85%           | 0.291%        |
| 24        | 8.956       | 2439          | 2462        | 2483         | rBV3     | 115757530      | 278788675     | 85.61%          | 13.513%       |
| 25        | 9.091       | 2484          | 2504        | 2533         | rVB5     | 124179454      | 325646250     | 100.00%         | 15.784%       |
| 26        | 9.468       | 2611          | 2621        | 2646         | rVB      | 4049258        | 6250556       | 1.92%           | 0.303%        |
| 27        | 9.857       | 2729          | 2742        | 2762         | rBV      | 16100601       | 24336867      | 7.47%           | 1.180%        |
| 28        | 10.281      | 2860          | 2874        | 2892         | rBV2     | 34163572       | 53862483      | 16.54%          | 2.611%        |
| 29        | 10.400      | 2892          | 2911        | 2922         | rVV2     | 42138944       | 86247752      | 26.49%          | 4.180%        |
| 30        | 10.471      | 2922          | 2933        | 2961         | rVB3     | 58427723       | 91527328      | 28.11%          | 4.436%        |
| 31        | 10.664      | 2981          | 2993        | 3021         | rBV      | 22645947       | 33775587      | 10.37%          | 1.637%        |
| 32        | 10.860      | 3036          | 3054        | 3069         | rBV3     | 121134522      | 263661832     | 80.97%          | 12.780%       |
| 33        | 11.123      | 3126          | 3136        | 3144         | rBV      | 2980129        | 4378544       | 1.34%           | 0.212%        |
| 34        | 11.175      | 3144          | 3152        | 3166         | rVB2     | 2490063        | 3803654       | 1.17%           | 0.184%        |
| 35        | 11.268      | 3167          | 3181        | 3194         | rBV3     | 64262855       | 103920655     | 31.91%          | 5.037%        |
| 36        | 11.458      | 3226          | 3240        | 3250         | rBV4     | 57650138       | 101907087     | 31.29%          | 4.939%        |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039153.D  
Acq On : 24 Sep 2025 15:31  
Operator : SY/MD  
Sample : Q3175-01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Title : SW846 8260

|    |        |      |      |      |      |          |          |        |        |
|----|--------|------|------|------|------|----------|----------|--------|--------|
| 37 | 11.570 | 3262 | 3275 | 3296 | rVB3 | 19499545 | 39004420 | 11.98% | 1.891% |
| 38 | 11.670 | 3296 | 3306 | 3318 | rVV  | 3163047  | 4643009  | 1.43%  | 0.225% |
| 39 | 11.744 | 3319 | 3329 | 3346 | rVB  | 6025232  | 8853223  | 2.72%  | 0.429% |
| 40 | 11.837 | 3347 | 3358 | 3363 | rBV  | 13418058 | 20175833 | 6.20%  | 0.978% |
| 41 | 11.866 | 3363 | 3367 | 3379 | rVV  | 11392449 | 16020669 | 4.92%  | 0.777% |
| 42 | 11.943 | 3379 | 3391 | 3411 | rVV  | 22581837 | 37249716 | 11.44% | 1.806% |
| 43 | 12.040 | 3411 | 3421 | 3453 | rVB2 | 8749135  | 17726383 | 5.44%  | 0.859% |
| 44 | 12.239 | 3473 | 3483 | 3496 | rVB  | 5830558  | 8707350  | 2.67%  | 0.422% |
| 45 | 12.339 | 3497 | 3514 | 3522 | rBV  | 12959061 | 19041462 | 5.85%  | 0.923% |
| 46 | 12.393 | 3522 | 3531 | 3547 | rVB  | 17665753 | 26002021 | 7.98%  | 1.260% |
| 47 | 12.651 | 3601 | 3611 | 3626 | rVB  | 12699755 | 19069019 | 5.86%  | 0.924% |
| 48 | 12.808 | 3641 | 3660 | 3671 | rBV2 | 26564524 | 42643927 | 13.10% | 2.067% |
| 49 | 12.873 | 3673 | 3680 | 3689 | rVB3 | 1983822  | 3302463  | 1.01%  | 0.160% |
| 50 | 12.972 | 3703 | 3711 | 3735 | rVB3 | 4136718  | 7291009  | 2.24%  | 0.353% |
| 51 | 13.197 | 3772 | 3781 | 3787 | rBV2 | 2748067  | 4152955  | 1.28%  | 0.201% |
| 52 | 13.429 | 3831 | 3853 | 3874 | rBV2 | 54637140 | 95072163 | 29.19% | 4.608% |
| 53 | 14.554 | 4191 | 4203 | 4231 | rVB  | 18117572 | 27821376 | 8.54%  | 1.349% |
| 54 | 14.734 | 4249 | 4259 | 4274 | rBV  | 9031814  | 13861563 | 4.26%  | 0.672% |




Sum of corrected areas: 2063109435

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039153.D  
Acq On : 24 Sep 2025 15:31  
Operator : SY/MD  
Sample : Q3175-01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039153.D  
Acq On : 24 Sep 2025 15:31  
Operator : SY/MD  
Sample : Q3175-01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2










Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*  
Peak Number 1 Benzene, 1-ethyl-2-methyl- Concentration Rank 3

| R.T.   | EstConc      | Area     | Relative to ISTD       | R.T.   |
|--------|--------------|----------|------------------------|--------|
| 10.400 | 1133.75 ug/l | 86247800 | 1,4-Dichlorobenzene-d4 | 11.178 |

| Hit# | of 5 | Tentative ID                       | MW  | MolForm | CAS#        | Qual |
|------|------|------------------------------------|-----|---------|-------------|------|
| 1    |      | Benzene, 1-ethyl-2-methyl-         | 120 | C9H12   | 000611-14-3 | 87   |
| 2    |      | 2,3-Heptadien-5-yne, 2,4-dimethyl- | 120 | C9H12   | 041898-89-9 | 83   |
| 3    |      | Benzene, 1,2,3-trimethyl-          | 120 | C9H12   | 000526-73-8 | 81   |
| 4    |      | Benzene, 1-ethyl-4-methyl-         | 120 | C9H12   | 000622-96-8 | 81   |
| 5    |      | Benzene, 1-ethyl-3-methyl-         | 120 | C9H12   | 000620-14-4 | 74   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039153.D  
Acq On : 24 Sep 2025 15:31  
Operator : SY/MD  
Sample : Q3175-01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2










Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*  
Peak Number 2 Benzene, 1-ethyl-4-methyl- Concentration Rank 6

| R.T.   | EstConc     | Area     | Relative to ISTD       | R.T.   |
|--------|-------------|----------|------------------------|--------|
| 10.664 | 443.99 ug/l | 33775600 | 1,4-Dichlorobenzene-d4 | 11.178 |

| Hit# | of | 5 | Tentative ID               | MW  | MolForm | CAS#        | Qual |
|------|----|---|----------------------------|-----|---------|-------------|------|
| 1    |    |   | Benzene, 1-ethyl-2-methyl- | 120 | C9H12   | 000611-14-3 | 94   |
| 2    |    |   | Benzene, 1-ethyl-4-methyl- | 120 | C9H12   | 000622-96-8 | 93   |
| 3    |    |   | Benzene, 1-ethyl-3-methyl- | 120 | C9H12   | 000620-14-4 | 91   |
| 4    |    |   | Mesitylene                 | 120 | C9H12   | 000108-67-8 | 91   |
| 5    |    |   | Benzene, 1,2,3-trimethyl-  | 120 | C9H12   | 000526-73-8 | 91   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039153.D  
Acq On : 24 Sep 2025 15:31  
Operator : SY/MD  
Sample : Q3175-01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2










Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*  
Peak Number 3 Benzene, 1,2,3-trimethyl- Concentration Rank 1

| R.T.   | EstConc      | Area      | Relative to ISTD       | R.T.   |
|--------|--------------|-----------|------------------------|--------|
| 11.268 | 1366.06 ug/l | 103921000 | 1,4-Dichlorobenzene-d4 | 11.178 |

| Hit# | of | 5 | Tentative ID               | MW  | MolForm | CAS#        | Qual |
|------|----|---|----------------------------|-----|---------|-------------|------|
| 1    |    |   | Benzene, 1,2,3-trimethyl-  | 120 | C9H12   | 000526-73-8 | 93   |
| 2    |    |   | Mesitylene                 | 120 | C9H12   | 000108-67-8 | 93   |
| 3    |    |   | Benzene, 1-ethyl-4-methyl- | 120 | C9H12   | 000622-96-8 | 87   |
| 4    |    |   | Benzene, 1-ethyl-3-methyl- | 120 | C9H12   | 000620-14-4 | 87   |
| 5    |    |   | Benzene, 1,2,4-trimethyl-  | 120 | C9H12   | 000095-63-6 | 81   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039153.D  
Acq On : 24 Sep 2025 15:31  
Operator : SY/MD  
Sample : Q3175-01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260










TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*

Peak Number 5 Benzene, 1-methyl-4-propyl- Concentration Rank 15

| R.T.   | EstConc     | Area    | Relative to ISTD       | R.T.   |
|--------|-------------|---------|------------------------|--------|
| 11.744 | 116.38 ug/l | 8853220 | 1,4-Dichlorobenzene-d4 | 11.178 |

| Hit# | of | 5 | Tentative ID                        | MW  | MolForm | CAS#        | Qual |
|------|----|---|-------------------------------------|-----|---------|-------------|------|
| 1    |    |   | Benzene, 1-methyl-4-propyl-         | 134 | C10H14  | 001074-55-1 | 95   |
| 2    |    |   | Benzene, 1-methyl-2-propyl-         | 134 | C10H14  | 001074-17-5 | 94   |
| 3    |    |   | Benzene, (1-methylpropyl)-          | 134 | C10H14  | 000135-98-8 | 91   |
| 4    |    |   | Benzene, 1-methyl-3-propyl-         | 134 | C10H14  | 001074-43-7 | 86   |
| 5    |    |   | Benzeneacetaldehyde, .alpha.-met... | 134 | C9H10O  | 000093-53-8 | 86   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039153.D  
Acq On : 24 Sep 2025 15:31  
Operator : SY/MD  
Sample : Q3175-01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2










Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*  
Peak Number 6 Benzene, 2-ethyl-1,4-dimethyl- Concentration Rank 9

| R.T.   | EstConc     | Area     | Relative to ISTD       | R.T.   |
|--------|-------------|----------|------------------------|--------|
| 11.837 | 265.22 ug/l | 20175800 | 1,4-Dichlorobenzene-d4 | 11.178 |

| Hit# | of | 5 | Tentative ID                        | MW  | MolForm | CAS#        | Qual |
|------|----|---|-------------------------------------|-----|---------|-------------|------|
| 1    |    |   | Benzene, 2-ethyl-1,4-dimethyl-      | 134 | C10H14  | 001758-88-9 | 97   |
| 2    |    |   | o-Cymene                            | 134 | C10H14  | 000527-84-4 | 95   |
| 3    |    |   | Benzene, 4-ethyl-1,2-dimethyl-      | 134 | C10H14  | 000934-80-5 | 95   |
| 4    |    |   | Benzene, 1-ethyl-2,3-dimethyl-      | 134 | C10H14  | 000933-98-2 | 95   |
| 5    |    |   | Benzene, 1-methyl-3-(1-methyleth... | 134 | C10H14  | 000535-77-3 | 93   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039153.D  
Acq On : 24 Sep 2025 15:31  
Operator : SY/MD  
Sample : Q3175-01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2










Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*

Peak Number 7 o-Cymene Concentration Rank 13

| R.T.      | EstConc                        | Area     | Relative to ISTD       | R.T.        |      |
|-----------|--------------------------------|----------|------------------------|-------------|------|
| 11.866    | 210.60 ug/l                    | 16020700 | 1,4-Dichlorobenzene-d4 | 11.178      |      |
| Hit# of 5 | Tentative ID                   | MW       | MolForm                | CAS#        | Qual |
| 1         | o-Cymene                       | 134      | C10H14                 | 000527-84-4 | 97   |
| 2         | p-Cymene                       | 134      | C10H14                 | 000099-87-6 | 95   |
| 3         | Benzene, 1,2,4,5-tetramethyl-  | 134      | C10H14                 | 000095-93-2 | 91   |
| 4         | Benzene, 1-ethyl-2,4-dimethyl- | 134      | C10H14                 | 000874-41-9 | 91   |
| 5         | Benzene, 1,2,3,4-tetramethyl-  | 134      | C10H14                 | 000488-23-3 | 91   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039153.D  
Acq On : 24 Sep 2025 15:31  
Operator : SY/MD  
Sample : Q3175-01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260










TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*

Peak Number 8 Benzene, 4-ethyl-1,2-dimethyl- Concentration Rank 5

| R.T.   | EstConc     | Area     | Relative to ISTD       | R.T.   |
|--------|-------------|----------|------------------------|--------|
| 11.943 | 489.66 ug/l | 37249700 | 1,4-Dichlorobenzene-d4 | 11.178 |

| Hit# | of | 5 | Tentative ID                        | MW  | MolForm | CAS#        | Qual |
|------|----|---|-------------------------------------|-----|---------|-------------|------|
| 1    |    |   | Benzene, 4-ethyl-1,2-dimethyl-      | 134 | C10H14  | 000934-80-5 | 96   |
| 2    |    |   | Benzene, 2-ethyl-1,4-dimethyl-      | 134 | C10H14  | 001758-88-9 | 95   |
| 3    |    |   | o-Cymene                            | 134 | C10H14  | 000527-84-4 | 95   |
| 4    |    |   | Benzene, 1-methyl-3-(1-methyleth... | 134 | C10H14  | 000535-77-3 | 95   |
| 5    |    |   | Benzene, 1-ethyl-2,3-dimethyl-      | 134 | C10H14  | 000933-98-2 | 95   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039153.D  
Acq On : 24 Sep 2025 15:31  
Operator : SY/MD  
Sample : Q3175-01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2










Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*  
Peak Number 9 Benzene, 1-ethenyl-3-ethyl- Concentration Rank 12

| R.T.   | EstConc     | Area     | Relative to ISTD       | R.T.   |
|--------|-------------|----------|------------------------|--------|
| 12.040 | 233.02 ug/l | 17726400 | 1,4-Dichlorobenzene-d4 | 11.178 |

| Hit# | of | 5 | Tentative ID                    | MW  | MolForm | CAS#        | Qual |
|------|----|---|---------------------------------|-----|---------|-------------|------|
| 1    |    |   | Benzene, 1-ethenyl-3-ethyl-     | 132 | C10H12  | 007525-62-4 | 91   |
| 2    |    |   | Indan, 1-methyl-                | 132 | C10H12  | 000767-58-8 | 90   |
| 3    |    |   | Benzene, (2-methyl-1-propenyl)- | 132 | C10H12  | 000768-49-0 | 90   |
| 4    |    |   | Benzene, 1-ethenyl-4-ethyl-     | 132 | C10H12  | 003454-07-7 | 90   |
| 5    |    |   | 1-Methyl-2-phenylcyclopropane   | 132 | C10H12  | 003145-76-4 | 86   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039153.D  
Acq On : 24 Sep 2025 15:31  
Operator : SY/MD  
Sample : Q3175-01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260










TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*

Peak Number 10 Benzene, 1,2,3,4-tetramethyl- Concentration Rank 11

| R.T.   | EstConc     | Area     | Relative to ISTD       | R.T.   |
|--------|-------------|----------|------------------------|--------|
| 12.339 | 250.31 ug/l | 19041500 | 1,4-Dichlorobenzene-d4 | 11.178 |

| Hit# | of | 5 | Tentative ID                   | MW  | MolForm | CAS#        | Qual |
|------|----|---|--------------------------------|-----|---------|-------------|------|
| 1    |    |   | Benzene, 1,2,3,4-tetramethyl-  | 134 | C10H14  | 000488-23-3 | 97   |
| 2    |    |   | Benzene, 1,2,3,5-tetramethyl-  | 134 | C10H14  | 000527-53-7 | 96   |
| 3    |    |   | Benzene, 1,2,4,5-tetramethyl-  | 134 | C10H14  | 000095-93-2 | 96   |
| 4    |    |   | o-Cymene                       | 134 | C10H14  | 000527-84-4 | 95   |
| 5    |    |   | Benzene, 1-ethyl-2,4-dimethyl- | 134 | C10H14  | 000874-41-9 | 94   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039153.D  
Acq On : 24 Sep 2025 15:31  
Operator : SY/MD  
Sample : Q3175-01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P










\*\*\*\*\*

Peak Number 11 Benzene, 1,2,4,5-tetramethyl- Concentration Rank 8

| R.T.   | EstConc     | Area     | Relative to ISTD       | R.T.   |
|--------|-------------|----------|------------------------|--------|
| 12.394 | 341.80 ug/l | 26002000 | 1,4-Dichlorobenzene-d4 | 11.178 |

| Hit# | of | 5 | Tentative ID                   | MW  | MolForm | CAS#        | Qual |
|------|----|---|--------------------------------|-----|---------|-------------|------|
| 1    |    |   | Benzene, 1,2,3,4-tetramethyl-  | 134 | C10H14  | 000488-23-3 | 97   |
| 2    |    |   | Benzene, 1,2,4,5-tetramethyl-  | 134 | C10H14  | 000095-93-2 | 97   |
| 3    |    |   | Benzene, 1,2,3,5-tetramethyl-  | 134 | C10H14  | 000527-53-7 | 96   |
| 4    |    |   | Benzene, 1-ethyl-2,4-dimethyl- | 134 | C10H14  | 000874-41-9 | 95   |
| 5    |    |   | o-Cymene                       | 134 | C10H14  | 000527-84-4 | 95   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039153.D  
Acq On : 24 Sep 2025 15:31  
Operator : SY/MD  
Sample : Q3175-01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260










TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*

Peak Number 12 1H-Indene, 2,3-dihydro-4-me... Concentration Rank 10

| R.T.   | EstConc     | Area     | Relative to ISTD       | R.T.   |
|--------|-------------|----------|------------------------|--------|
| 12.651 | 250.67 ug/l | 19069000 | 1,4-Dichlorobenzene-d4 | 11.178 |

| Hit# | of 5 | Tentative ID                     | MW  | MolForm | CAS#        | Qual |
|------|------|----------------------------------|-----|---------|-------------|------|
| 1    |      | 1H-Indene, 2,3-dihydro-4-methyl- | 132 | C10H12  | 000824-22-6 | 95   |
| 2    |      | Benzene, 1-ethenyl-4-ethyl-      | 132 | C10H12  | 003454-07-7 | 94   |
| 3    |      | 3-Phenylbut-1-ene                | 132 | C10H12  | 000934-10-1 | 94   |
| 4    |      | Benzene, 2-ethenyl-1,4-dimethyl- | 132 | C10H12  | 002039-89-6 | 92   |
| 5    |      | 1H-Indene, 2,3-dihydro-5-methyl- | 132 | C10H12  | 000874-35-1 | 91   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039153.D  
Acq On : 24 Sep 2025 15:31  
Operator : SY/MD  
Sample : Q3175-01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260










TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*

Peak Number 13 Benzene, 2-ethenyl-1,4-dime... Concentration Rank 4

| R.T.   | EstConc     | Area     | Relative to ISTD       | R.T.   |
|--------|-------------|----------|------------------------|--------|
| 12.808 | 560.57 ug/l | 42643900 | 1,4-Dichlorobenzene-d4 | 11.178 |

| Hit# | of | 5 | Tentative ID                     | MW  | MolForm | CAS#        | Qual |
|------|----|---|----------------------------------|-----|---------|-------------|------|
| 1    |    |   | Benzene, 2-ethenyl-1,4-dimethyl- | 132 | C10H12  | 002039-89-6 | 95   |
| 2    |    |   | 1H-Indene, 2,3-dihydro-5-methyl- | 132 | C10H12  | 000874-35-1 | 93   |
| 3    |    |   | 2,4-Dimethylstyrene              | 132 | C10H12  | 002234-20-0 | 87   |
| 4    |    |   | Benzene, 2-ethenyl-1,3-dimethyl- | 132 | C10H12  | 002039-90-9 | 76   |
| 5    |    |   | Benzene, 2-butenyl-              | 132 | C10H12  | 001560-06-1 | 76   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039153.D  
Acq On : 24 Sep 2025 15:31  
Operator : SY/MD  
Sample : Q3175-01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2










Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*  
Peak Number 14 Naphthalene, 1-methyl- Concentration Rank 7

| R.T.   | EstConc     | Area     | Relative to ISTD       | R.T.   |
|--------|-------------|----------|------------------------|--------|
| 14.554 | 365.72 ug/l | 27821400 | 1,4-Dichlorobenzene-d4 | 11.178 |

| Hit# | of | 5 | Tentative ID                        | MW  | MolForm | CAS#        | Qual |
|------|----|---|-------------------------------------|-----|---------|-------------|------|
| 1    |    |   | Naphthalene, 1-methyl-              | 142 | C11H10  | 000090-12-0 | 96   |
| 2    |    |   | Naphthalene, 2-methyl-              | 142 | C11H10  | 000091-57-6 | 96   |
| 3    |    |   | Benzocycloheptatriene               | 142 | C11H10  | 000264-09-5 | 91   |
| 4    |    |   | 1,4-Methanonaphthalene, 1,4-dihy... | 142 | C11H10  | 004453-90-1 | 91   |
| 5    |    |   | 1H-Indene, 1-ethylidene-            | 142 | C11H10  | 002471-83-2 | 72   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039153.D  
Acq On : 24 Sep 2025 15:31  
Operator : SY/MD  
Sample : Q3175-01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260










TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*

Peak Number 15 Naphthalene, 2-methyl- Concentration Rank 14

| R.T.   | EstConc     | Area     | Relative to ISTD       | R.T.   |
|--------|-------------|----------|------------------------|--------|
| 14.734 | 182.21 ug/l | 13861600 | 1,4-Dichlorobenzene-d4 | 11.178 |

| Hit# | of | 5 | Tentative ID                        | MW  | MolForm | CAS#        | Qual |
|------|----|---|-------------------------------------|-----|---------|-------------|------|
| 1    |    |   | Naphthalene, 2-methyl-              | 142 | C11H10  | 000091-57-6 | 96   |
| 2    |    |   | Naphthalene, 1-methyl-              | 142 | C11H10  | 000090-12-0 | 96   |
| 3    |    |   | Benzocycloheptatriene               | 142 | C11H10  | 000264-09-5 | 91   |
| 4    |    |   | 1,4-Methanonaphthalene, 1,4-dihy... | 142 | C11H10  | 004453-90-1 | 91   |
| 5    |    |   | 1H-Indene, 1-ethylidene-            | 142 | C11H10  | 002471-83-2 | 72   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039153.D  
Acq On : 24 Sep 2025 15:31  
Operator : SY/MD  
Sample : Q3175-01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST0.L  
TIC Integration Parameters: LSCINT.P

| TIC Top Hit name   | RT     | EstConc | Units | Response  | --Internal Standard-- |        |         |      |
|--------------------|--------|---------|-------|-----------|-----------------------|--------|---------|------|
|                    |        |         |       |           | #                     | RT     | Resp    | Conc |
| Benzene, 1-ethy... | 10.400 | 1133.8  | ug/l  | 86247800  | 4                     | 11.178 | 3803650 | 50.0 |
| Benzene, 1-ethy... | 10.664 | 444.0   | ug/l  | 33775600  | 4                     | 11.178 | 3803650 | 50.0 |
| Benzene, 1,2,3-... | 11.268 | 1366.1  | ug/l  | 103921000 | 4                     | 11.178 | 3803650 | 50.0 |
| Benzene, 1-meth... | 11.744 | 116.4   | ug/l  | 8853220   | 4                     | 11.178 | 3803650 | 50.0 |
| Benzene, 2-ethy... | 11.837 | 265.2   | ug/l  | 20175800  | 4                     | 11.178 | 3803650 | 50.0 |
| o-Cymene           | 11.866 | 210.6   | ug/l  | 16020700  | 4                     | 11.178 | 3803650 | 50.0 |
| Benzene, 4-ethy... | 11.943 | 489.7   | ug/l  | 37249700  | 4                     | 11.178 | 3803650 | 50.0 |
| Benzene, 1-ethe... | 12.040 | 233.0   | ug/l  | 17726400  | 4                     | 11.178 | 3803650 | 50.0 |
| Benzene, 1,2,3,... | 12.339 | 250.3   | ug/l  | 19041500  | 4                     | 11.178 | 3803650 | 50.0 |
| Benzene, 1,2,4,... | 12.394 | 341.8   | ug/l  | 26002000  | 4                     | 11.178 | 3803650 | 50.0 |
| 1H-Indene, 2,3-... | 12.651 | 250.7   | ug/l  | 19069000  | 4                     | 11.178 | 3803650 | 50.0 |
| Benzene, 2-ethe... | 12.808 | 560.6   | ug/l  | 42643900  | 4                     | 11.178 | 3803650 | 50.0 |
| Naphthalene, 1-... | 14.554 | 365.7   | ug/l  | 27821400  | 4                     | 11.178 | 3803650 | 50.0 |
| Naphthalene, 2-... | 14.734 | 182.2   | ug/l  | 13861600  | 4                     | 11.178 | 3803650 | 50.0 |

### Report of Analysis

|                    |                 |           |                 |              |    |
|--------------------|-----------------|-----------|-----------------|--------------|----|
| Client:            | G Environmental |           | Date Collected: | 09/23/25     |    |
| Project:           | Summer          |           | Date Received:  | 09/23/25     |    |
| Client Sample ID:  | MW2DL           |           | SDG No.:        | Q3175        |    |
| Lab Sample ID:     | Q3175-01DL      |           | Matrix:         | Water        |    |
| Analytical Method: | 8260D           |           | % Solid:        | 0            |    |
| Sample Wt/Vol:     | 5               | Units: mL | Final Vol:      | 5000         | uL |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOCMS Group1 |    |
| GC Column:         | DB-624UI        | ID : 0.18 | Level :         | LOW          |    |
| Prep Method :      |                 |           |                 |              |    |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039174.D        | 20        | 09/25/25 12:33 | VV092525      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units |
|----------------|--------------------------------|-------|-----------|------|------------|-------|
| <b>TARGETS</b> |                                |       |           |      |            |       |
| 75-71-8        | Dichlorodifluoromethane        | 4.40  | UD        | 4.40 | 20.0       | ug/L  |
| 74-87-3        | Chloromethane                  | 6.40  | UD        | 6.40 | 20.0       | ug/L  |
| 75-01-4        | Vinyl Chloride                 | 5.20  | UD        | 5.20 | 20.0       | ug/L  |
| 74-83-9        | Bromomethane                   | 28.8  | UD        | 28.8 | 100        | ug/L  |
| 75-00-3        | Chloroethane                   | 9.40  | UD        | 9.40 | 20.0       | ug/L  |
| 75-69-4        | Trichlorofluoromethane         | 6.60  | UD        | 6.60 | 20.0       | ug/L  |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 5.00  | UD        | 5.00 | 20.0       | ug/L  |
| 75-65-0        | Tert butyl alcohol             | 110   | UD        | 110  | 500        | ug/L  |
| 75-35-4        | 1,1-Dichloroethene             | 4.60  | UD        | 4.60 | 20.0       | ug/L  |
| 67-64-1        | Acetone                        | 30.2  | UD        | 30.2 | 100        | ug/L  |
| 75-15-0        | Carbon Disulfide               | 4.20  | UD        | 4.20 | 20.0       | ug/L  |
| 1634-04-4      | Methyl tert-butyl Ether        | 3.20  | UD        | 3.20 | 20.0       | ug/L  |
| 79-20-9        | Methyl Acetate                 | 5.40  | UD        | 5.40 | 20.0       | ug/L  |
| 75-09-2        | Methylene Chloride             | 5.60  | UD        | 5.60 | 20.0       | ug/L  |
| 156-60-5       | trans-1,2-Dichloroethene       | 4.60  | UD        | 4.60 | 20.0       | ug/L  |
| 75-34-3        | 1,1-Dichloroethane             | 4.60  | UD        | 4.60 | 20.0       | ug/L  |
| 110-82-7       | Cyclohexane                    | 220   | D         | 29.0 | 100        | ug/L  |
| 78-93-3        | 2-Butanone                     | 19.6  | UD        | 19.6 | 100        | ug/L  |
| 56-23-5        | Carbon Tetrachloride           | 5.00  | UD        | 5.00 | 20.0       | ug/L  |
| 156-59-2       | cis-1,2-Dichloroethene         | 3.80  | UD        | 3.80 | 20.0       | ug/L  |
| 74-97-5        | Bromochloromethane             | 4.40  | UD        | 4.40 | 20.0       | ug/L  |
| 67-66-3        | Chloroform                     | 5.00  | UD        | 5.00 | 20.0       | ug/L  |
| 71-55-6        | 1,1,1-Trichloroethane          | 4.00  | UD        | 4.00 | 20.0       | ug/L  |
| 108-87-2       | Methylcyclohexane              | 150   | D         | 3.20 | 20.0       | ug/L  |
| 71-43-2        | Benzene                        | 270   | D         | 3.00 | 20.0       | ug/L  |
| 107-06-2       | 1,2-Dichloroethane             | 4.40  | UD        | 4.40 | 20.0       | ug/L  |
| 79-01-6        | Trichloroethene                | 1.90  | UD        | 1.90 | 20.0       | ug/L  |
| 78-87-5        | 1,2-Dichloropropane            | 4.00  | UD        | 4.00 | 20.0       | ug/L  |
| 75-27-4        | Bromodichloromethane           | 4.40  | UD        | 4.40 | 20.0       | ug/L  |
| 108-10-1       | 4-Methyl-2-Pentanone           | 13.6  | UD        | 13.6 | 100        | ug/L  |

### Report of Analysis

|                    |                 |           |                 |              |    |
|--------------------|-----------------|-----------|-----------------|--------------|----|
| Client:            | G Environmental |           | Date Collected: | 09/23/25     |    |
| Project:           | Summer          |           | Date Received:  | 09/23/25     |    |
| Client Sample ID:  | MW2DL           |           | SDG No.:        | Q3175        |    |
| Lab Sample ID:     | Q3175-01DL      |           | Matrix:         | Water        |    |
| Analytical Method: | 8260D           |           | % Solid:        | 0            |    |
| Sample Wt/Vol:     | 5               | Units: mL | Final Vol:      | 5000         | uL |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOCMS Group1 |    |
| GC Column:         | DB-624UI        | ID : 0.18 | Level :         | LOW          |    |
| Prep Method :      |                 |           |                 |              |    |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039174.D        | 20        | 09/25/25 12:33 | VV092525      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units   |
|---------------------------|-----------------------------|--------|-----------|----------|------------|---------|
| 108-88-3                  | Toluene                     | 57.2   | D         | 2.80     | 20.0       | ug/L    |
| 10061-02-6                | t-1,3-Dichloropropene       | 3.40   | UD        | 3.40     | 20.0       | ug/L    |
| 10061-01-5                | cis-1,3-Dichloropropene     | 3.20   | UD        | 3.20     | 20.0       | ug/L    |
| 79-00-5                   | 1,1,2-Trichloroethane       | 4.20   | UD        | 4.20     | 20.0       | ug/L    |
| 591-78-6                  | 2-Hexanone                  | 17.8   | UD        | 17.8     | 100        | ug/L    |
| 124-48-1                  | Dibromochloromethane        | 3.60   | UD        | 3.60     | 20.0       | ug/L    |
| 106-93-4                  | 1,2-Dibromoethane           | 3.00   | UD        | 3.00     | 20.0       | ug/L    |
| 127-18-4                  | Tetrachloroethene           | 4.60   | UD        | 4.60     | 20.0       | ug/L    |
| 108-90-7                  | Chlorobenzene               | 2.40   | UD        | 2.40     | 20.0       | ug/L    |
| 100-41-4                  | Ethyl Benzene               | 4900   | ED        | 2.60     | 20.0       | ug/L    |
| 179601-23-1               | m/p-Xylenes                 | 5900   | ED        | 4.80     | 40.0       | ug/L    |
| 95-47-6                   | o-Xylene                    | 55.3   | D         | 2.40     | 20.0       | ug/L    |
| 100-42-5                  | Styrene                     | 3.00   | UD        | 3.00     | 20.0       | ug/L    |
| 75-25-2                   | Bromoform                   | 3.80   | UD        | 3.80     | 20.0       | ug/L    |
| 98-82-8                   | Isopropylbenzene            | 160    | D         | 2.40     | 20.0       | ug/L    |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 5.20   | UD        | 5.20     | 20.0       | ug/L    |
| 541-73-1                  | 1,3-Dichlorobenzene         | 3.20   | UD        | 3.20     | 20.0       | ug/L    |
| 106-46-7                  | 1,4-Dichlorobenzene         | 3.80   | UD        | 3.80     | 20.0       | ug/L    |
| 95-50-1                   | 1,2-Dichlorobenzene         | 3.20   | UD        | 3.20     | 20.0       | ug/L    |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 10.6   | UD        | 10.6     | 20.0       | ug/L    |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 4.00   | UD        | 4.00     | 20.0       | ug/L    |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 4.00   | UD        | 4.00     | 20.0       | ug/L    |
| <b>SURROGATES</b>         |                             |        |           |          |            |         |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 52.4   |           | 74 - 125 | 105%       | SPK: 50 |
| 1868-53-7                 | Dibromofluoromethane        | 48.9   |           | 75 - 124 | 98%        | SPK: 50 |
| 2037-26-5                 | Toluene-d8                  | 48.3   |           | 86 - 113 | 97%        | SPK: 50 |
| 460-00-4                  | 4-Bromofluorobenzene        | 50.2   |           | 77 - 121 | 100%       | SPK: 50 |
| <b>INTERNAL STANDARDS</b> |                             |        |           |          |            |         |
| 363-72-4                  | Pentafluorobenzene          | 563000 | 4.6       |          |            |         |
| 540-36-3                  | 1,4-Difluorobenzene         | 980000 | 5.535     |          |            |         |
| 3114-55-4                 | Chlorobenzene-d5            | 955000 | 8.776     |          |            |         |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 559000 | 11.172    |          |            |         |

## Report of Analysis

|                    |                                            |                 |                              |
|--------------------|--------------------------------------------|-----------------|------------------------------|
| Client:            | G Environmental                            | Date Collected: | 09/23/25                     |
| Project:           | Summer                                     | Date Received:  | 09/23/25                     |
| Client Sample ID:  | MW2DL                                      | SDG No.:        | Q3175                        |
| Lab Sample ID:     | Q3175-01DL                                 | Matrix:         | Water                        |
| Analytical Method: | 8260D                                      | % Solid:        | 0                            |
| Sample Wt/Vol:     | 5                      Units:    mL        | Final Vol:      | 5000                      uL |
| Soil Aliquot Vol:  | uL                                         | Test:           | VOCMS Group1                 |
| GC Column:         | DB-624UI                      ID :    0.18 | Level :         | LOW                          |
| Prep Method :      |                                            |                 |                              |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039174.D        | 20        | 09/25/25 12:33 | VV092525      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
 Data File : VV039174.D  
 Acq On : 25 Sep 2025 12:33  
 Operator : SY/MD  
 Sample : Q3175-01DL 20X  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 MW2DL

Manual Integrations  
 APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025  
 Supervised By :Semsettin Yesilyurt 09/29/2025

Quant Time: Sep 26 01:24:38 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

| Compound                    | R.T.           | QIon | Response            | Conc    | Units  | Dev(Min) |
|-----------------------------|----------------|------|---------------------|---------|--------|----------|
| -----                       |                |      |                     |         |        |          |
| Internal Standards          |                |      |                     |         |        |          |
| 1) Pentafluorobenzene       | 4.600          | 168  | 563178              | 50.000  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 5.535          | 114  | 979530              | 50.000  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 8.776          | 117  | 954674              | 50.000  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 11.172         | 152  | 558535              | 50.000  | ug/l   | 0.00     |
|                             |                |      |                     |         |        |          |
| System Monitoring Compounds |                |      |                     |         |        |          |
| 33) 1,2-Dichloroethane-d4   | 4.944          | 65   | 427463              | 52.444  | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 81 - 118 |      | Recovery = 104.880% |         |        |          |
| 35) Dibromofluoromethane    | 4.503          | 113  | 384673              | 48.850  | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 80 - 119 |      | Recovery = 97.700%  |         |        |          |
| 50) Toluene-d8              | 7.236          | 98   | 1117790             | 48.331  | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 89 - 112 |      | Recovery = 96.660%  |         |        |          |
| 62) 4-Bromofluorobenzene    | 10.001         | 95   | 477673              | 50.170  | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 85 - 114 |      | Recovery = 100.340% |         |        |          |
|                             |                |      |                     |         |        |          |
| Target Compounds            |                |      |                     |         | Qvalue |          |
| 31) Cyclohexane             | 4.584          | 56   | 163390              | 11.239  | ug/l   | 93       |
| 39) Methylcyclohexane       | 6.047          | 83   | 104560              | 7.272   | ug/l   | 91       |
| 40) Benzene                 | 5.015          | 78   | 516821              | 13.331  | ug/l   | 98       |
| 52) Toluene                 | 7.317          | 92   | 64289               | 2.860   | ug/l   | 100      |
| 67) Ethyl Benzene           | 8.934          | 91   | 9765537             | 242.606 | ug/l   | 97       |
| 68) m/p-Xylenes             | 9.063          | 106  | 4597374             | 295.919 | ug/l   | 97       |
| 69) o-Xylene                | 9.474          | 106  | 38063               | 2.767   | ug/l   | 89       |
| 73) Isopropylbenzene        | 9.857          | 105  | 331271              | 8.177   | ug/l   | 100      |
| 78) n-propylbenzene         | 10.278         | 91   | 1203830             | 24.402  | ug/l   | 100      |
| 80) 1,3,5-Trimethylbenzene  | 10.464         | 105  | 1628348             | 48.359  | ug/l   | 100      |
| 84) 1,2,4-Trimethylbenzene  | 10.837         | 105  | 6743997             | 206.300 | ug/l   | 98       |
| 85) sec-Butylbenzene        | 11.014         | 105  | 52699m              | 1.183   | ug/l   |          |
| 86) p-Isopropyltoluene      | 11.169         | 119  | 28308m              | 0.764   | ug/l   |          |
| 89) n-Butylbenzene          | 11.574         | 91   | 112240m             | 3.171   | ug/l   |          |
| 95) Naphthalene             | 13.423         | 128  | 2104561             | 57.156  | ug/l   | 100      |
| -----                       |                |      |                     |         |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
Data File : VV039174.D  
Acq On : 25 Sep 2025 12:33  
Operator : SY/MD  
Sample : Q3175-01DL 20X  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 8 Sample Multiplier: 1


Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2DL

Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025


Quant Time: Sep 26 01:24:38 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 4.600 min Scan# 1107  
Delta R.T. 0.003 min  
Lab File: VV039174.D  
Acq: 25 Sep 2025 12:33




Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2DL




Tgt Ion: 168 Resp: 563178  
Ion Ratio Lower Upper  
168 100  
99 62.1 49.6 74.4

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025




#31  
Cyclohexane  
Concen: 11.239 ug/l  
RT: 4.584 min Scan# 1102  
Delta R.T. 0.003 min  
Lab File: VV039174.D  
Acq: 25 Sep 2025 12:33



Tgt Ion: 56 Resp: 163390  
Ion Ratio Lower Upper  
56 100  
69 34.2 26.2 39.2  
84 80.2 70.3 105.5





#33

1,2-Dichloroethane-d4

Concen: 52.444 ug/l

RT: 4.944 min Scan# 1214

Delta R.T. 0.006 min

Lab File: VV039174.D

Acq: 25 Sep 2025 12:33

Instrument :

MSVOA\_V

ClientSampleId :

MW2DL

Tgt Ion: 65 Resp: 42746





Ion Ratio Lower Upper

65 100

67 52.5 0.0 107.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025

#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 5.535 min Scan# 1398

Delta R.T. 0.006 min

Lab File: VV039174.D

Acq: 25 Sep 2025 12:33




Tgt Ion:114 Resp: 979530


Ion Ratio Lower Upper

114 100

63 20.0 0.0 42.6

88 15.4 0.0 30.8





#35

Dibromofluoromethane

Concen: 48.850 ug/l

RT: 4.503 min Scan# 1075

Delta R.T. 0.006 min

Lab File: VV039174.D


Acq: 25 Sep 2025 12:33

Instrument :

MSVOA\_V

ClientSampleId :

MW2DL



Tgt Ion: 113 Resp: 38467

Ion Ratio Lower Upper

113 100




111 105.2 82.4 123.6

192 17.3 13.8 20.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025



#39

Methylcyclohexane

Concen: 7.272 ug/l

RT: 6.047 min Scan# 1557

Delta R.T. 0.004 min

Lab File: VV039174.D

Acq: 25 Sep 2025 12:33




Tgt Ion: 83 Resp: 104560


Ion Ratio Lower Upper

83 100

55 87.3 60.5 90.7

98 48.8 37.8 56.6









#40  
Benzene  
Concen: 13.331 ug/l  
RT: 5.015 min Scan# 1236  
Delta R.T. 0.010 min  
Lab File: VV039174.D  
Acq: 25 Sep 2025 12:33

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2DL

Tgt Ion: 78 Resp: 51682  
Ion Ratio Lower Upper  
78 100  
77 24.5 18.7 28.1




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025



#50  
Toluene-d8  
Concen: 48.331 ug/l  
RT: 7.236 min Scan# 1927  
Delta R.T. 0.003 min  
Lab File: VV039174.D  
Acq: 25 Sep 2025 12:33

Tgt Ion: 98 Resp: 1117790  
Ion Ratio Lower Upper  
98 100  
100 64.9 52.2 78.2





#52

Toluene

Concen: 2.860 ug/l

RT: 7.317 min Scan# 1948

Delta R.T. 0.013 min

Lab File: VV039174.D

Acq: 25 Sep 2025 12:33

Instrument :

MSVOA\_V

ClientSampleId :

MW2DL

Tgt Ion: 92 Resp: 64289

Ion Ratio Lower Upper





92 100

91 174.3 139.2 208.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025



#62

4-Bromofluorobenzene

Concen: 50.170 ug/l

RT: 10.001 min Scan# 2787

Delta R.T. 0.003 min

Lab File: VV039174.D

Acq: 25 Sep 2025 12:33




Tgt Ion: 95 Resp: 477673


Ion Ratio Lower Upper

95 100

174 74.9 0.0 150.4

176 70.6 0.0 144.2





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 8.776 min Scan# 2405

Delta R.T. 0.003 min

Lab File: VV039174.D

Acq: 25 Sep 2025 12:33

Instrument :

MSVOA\_V

ClientSampleId :

MW2DL

Tgt Ion:117 Resp: 954674

Ion Ratio Lower Upper

117 100





82 56.1 45.4 68.2

119 32.0 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025  
 Supervised By :Semsettin Yesilyurt 09/29/2025



#67

Ethyl Benzene

Concen: 242.606 ug/l

RT: 8.934 min Scan# 2455

Delta R.T. -0.000 min

Lab File: VV039174.D




Acq: 25 Sep 2025 12:33


Tgt Ion: 91 Resp: 9765537

Ion Ratio Lower Upper

91 100

106 32.2 24.6 37.0









#68  
m/p-Xylenes  
Concen: 295.919 ug/l  
RT: 9.063 min Scan# 2495  
Delta R.T. 0.004 min  
Lab File: VV039174.D  
Acq: 25 Sep 2025 12:33

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2DL

Tgt Ion:106 Resp: 4597374  
Ion Ratio Lower Upper  
106 100  
91 198.5 162.5 243.7




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025



#69  
o-Xylene  
Concen: 2.767 ug/l  
RT: 9.474 min Scan# 2623  
Delta R.T. 0.006 min  
Lab File: VV039174.D  
Acq: 25 Sep 2025 12:33


Tgt Ion:106 Resp: 38063  
Ion Ratio Lower Upper  
106 100  
91 200.5 109.1 327.3





#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 11.172 min Scan# 3151  
Delta R.T. -0.000 min  
Lab File: VV039174.D  
Acq: 25 Sep 2025 12:33




Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2DL




Tgt Ion:152 Resp: 558538  
Ion Ratio Lower Upper  
152 100  
115 61.1 44.8 134.4  
150 155.7 0.0 353.8

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025




#73  
Isopropylbenzene  
Concen: 8.177 ug/l  
RT: 9.857 min Scan# 2742  
Delta R.T. 0.003 min  
Lab File: VV039174.D  
Acq: 25 Sep 2025 12:33



Tgt Ion:105 Resp: 331271  
Ion Ratio Lower Upper  
105 100  
120 25.9 13.1 39.1









#78  
n-propylbenzene  
Concen: 24.402 ug/l  
RT: 10.278 min Scan# 2873  
Delta R.T. 0.003 min  
Lab File: VV039174.D  
Acq: 25 Sep 2025 12:33

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2DL

Tgt Ion: 91 Resp: 1203830  
Ion Ratio Lower Upper  
91 100  
120 22.1 11.1 33.3




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025



#80  
1,3,5-Trimethylbenzene  
Concen: 48.359 ug/l  
RT: 10.464 min Scan# 2931  
Delta R.T. 0.000 min  
Lab File: VV039174.D  
Acq: 25 Sep 2025 12:33


Tgt Ion:105 Resp: 1628348  
Ion Ratio Lower Upper  
105 100  
120 48.3 24.3 72.8





#84  
1,2,4-Trimethylbenzene  
Concen: 206.300 ug/l  
RT: 10.837 min Scan# 3047  
Delta R.T. 0.000 min  
Lab File: VV039174.D  
Acq: 25 Sep 2025 12:33




Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2DL




Tgt Ion:105 Resp: 674399  
Ion Ratio Lower Upper  
105 100  
120 45.9 22.4 67.2

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025




#85  
sec-Butylbenzene  
Concen: 1.183 ug/l m  
RT: 11.014 min Scan# 3102  
Delta R.T. 0.003 min  
Lab File: VV039174.D  
Acq: 25 Sep 2025 12:33




Tgt Ion:105 Resp: 52699  
Ion Ratio Lower Upper  
105 100  
134 0.0 9.6 28.8#





#86  
p-Isopropyltoluene  
Concen: 0.764 ug/l m  
RT: 11.169 min Scan# 3149  
Delta R.T. 0.003 min  
Lab File: VV039174.D  
Acq: 25 Sep 2025 12:33




Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2DL



Tgt Ion: 119 Resp: 28308  
Ion Ratio Lower Upper  
119 100  
134 51.6 12.7 38.0  
91 49.5 12.0 36.1




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025



#89  
n-Butylbenzene  
Concen: 3.171 ug/l m  
RT: 11.574 min Scan# 3276  
Delta R.T. -0.003 min  
Lab File: VV039174.D  
Acq: 25 Sep 2025 12:33

Tgt Ion: 91 Resp: 112240  
Ion Ratio Lower Upper  
91 100  
92 43.0 26.6 79.8  
134 149.1 12.1 36.3





#95

Naphthalene

Concen: 57.156 ug/l

RT: 13.423 min Scan# 3851

Delta R.T. -0.000 min

Lab File: VV039174.D

Acq: 25 Sep 2025 12:33

Instrument :

MSVOA\_V

ClientSampleId :

MW2DL

Tgt Ion:128 Resp: 210456

Ion Ratio Lower Upper

128 100




127 12.9 10.3 15.5

129 11.1 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025



## Report of Analysis

|                    |                 |                 |          |
|--------------------|-----------------|-----------------|----------|
| Client:            | G Environmental | Date Collected: | 09/23/25 |
| Project:           | Summer          | Date Received:  | 09/23/25 |
| Client Sample ID:  | MW2DL2          | SDG No.:        | Q3175    |
| Lab Sample ID:     | Q3175-01DL2     | Matrix:         | Water    |
| Analytical Method: | 8260D           | % Solid:        | 0        |
| Sample Wt/Vol:     | 5               | Units:          | mL       |
| Soil Aliquot Vol:  |                 |                 | uL       |
| GC Column:         | DB-624UI        | ID :            | 0.18     |
| Prep Method :      |                 | Level :         | LOW      |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039179.D        | 100       | 09/25/25 14:25 | VV092525      |

| CAS Number | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units |
|------------|--------------------------------|-------|-----------|------|------------|-------|
| TARGETS    |                                |       |           |      |            |       |
| 75-71-8    | Dichlorodifluoromethane        | 22.0  | UD        | 22.0 | 100        | ug/L  |
| 74-87-3    | Chloromethane                  | 32.0  | UD        | 32.0 | 100        | ug/L  |
| 75-01-4    | Vinyl Chloride                 | 26.0  | UD        | 26.0 | 100        | ug/L  |
| 74-83-9    | Bromomethane                   | 140   | UD        | 140  | 500        | ug/L  |
| 75-00-3    | Chloroethane                   | 47.0  | UD        | 47.0 | 100        | ug/L  |
| 75-69-4    | Trichlorofluoromethane         | 33.0  | UD        | 33.0 | 100        | ug/L  |
| 76-13-1    | 1,1,2-Trichlorotrifluoroethane | 25.0  | UD        | 25.0 | 100        | ug/L  |
| 75-65-0    | Tert butyl alcohol             | 550   | UD        | 550  | 2500       | ug/L  |
| 75-35-4    | 1,1-Dichloroethene             | 23.0  | UD        | 23.0 | 100        | ug/L  |
| 67-64-1    | Acetone                        | 150   | UD        | 150  | 500        | ug/L  |
| 75-15-0    | Carbon Disulfide               | 21.0  | UD        | 21.0 | 100        | ug/L  |
| 1634-04-4  | Methyl tert-butyl Ether        | 16.0  | UD        | 16.0 | 100        | ug/L  |
| 79-20-9    | Methyl Acetate                 | 27.0  | UD        | 27.0 | 100        | ug/L  |
| 75-09-2    | Methylene Chloride             | 28.0  | UD        | 28.0 | 100        | ug/L  |
| 156-60-5   | trans-1,2-Dichloroethene       | 23.0  | UD        | 23.0 | 100        | ug/L  |
| 75-34-3    | 1,1-Dichloroethane             | 23.0  | UD        | 23.0 | 100        | ug/L  |
| 110-82-7   | Cyclohexane                    | 310   | JD        | 150  | 500        | ug/L  |
| 78-93-3    | 2-Butanone                     | 98.0  | UD        | 98.0 | 500        | ug/L  |
| 56-23-5    | Carbon Tetrachloride           | 25.0  | UD        | 25.0 | 100        | ug/L  |
| 156-59-2   | cis-1,2-Dichloroethene         | 19.0  | UD        | 19.0 | 100        | ug/L  |
| 74-97-5    | Bromochloromethane             | 22.0  | UD        | 22.0 | 100        | ug/L  |
| 67-66-3    | Chloroform                     | 25.0  | UD        | 25.0 | 100        | ug/L  |
| 71-55-6    | 1,1,1-Trichloroethane          | 20.0  | UD        | 20.0 | 100        | ug/L  |
| 108-87-2   | Methylcyclohexane              | 160   | D         | 16.0 | 100        | ug/L  |
| 71-43-2    | Benzene                        | 360   | D         | 15.0 | 100        | ug/L  |
| 107-06-2   | 1,2-Dichloroethane             | 22.0  | UD        | 22.0 | 100        | ug/L  |
| 79-01-6    | Trichloroethene                | 9.30  | UD        | 9.30 | 100        | ug/L  |
| 78-87-5    | 1,2-Dichloropropane            | 20.0  | UD        | 20.0 | 100        | ug/L  |
| 75-27-4    | Bromodichloromethane           | 22.0  | UD        | 22.0 | 100        | ug/L  |
| 108-10-1   | 4-Methyl-2-Pentanone           | 68.0  | UD        | 68.0 | 500        | ug/L  |

### Report of Analysis

|                    |                    |                 |              |
|--------------------|--------------------|-----------------|--------------|
| Client:            | G Environmental    | Date Collected: | 09/23/25     |
| Project:           | Summer             | Date Received:  | 09/23/25     |
| Client Sample ID:  | MW2DL2             | SDG No.:        | Q3175        |
| Lab Sample ID:     | Q3175-01DL2        | Matrix:         | Water        |
| Analytical Method: | 8260D              | % Solid:        | 0            |
| Sample Wt/Vol:     | 5 Units: mL        | Final Vol:      | 5000 uL      |
| Soil Aliquot Vol:  | uL                 | Test:           | VOCMS Group1 |
| GC Column:         | DB-624UI ID : 0.18 | Level :         | LOW          |
| Prep Method :      |                    |                 |              |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039179.D        | 100       | 09/25/25 14:25 | VV092525      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units   |
|---------------------------|-----------------------------|--------|-----------|----------|------------|---------|
| 108-88-3                  | Toluene                     | 70.2   | JD        | 14.0     | 100        | ug/L    |
| 10061-02-6                | t-1,3-Dichloropropene       | 17.0   | UD        | 17.0     | 100        | ug/L    |
| 10061-01-5                | cis-1,3-Dichloropropene     | 16.0   | UD        | 16.0     | 100        | ug/L    |
| 79-00-5                   | 1,1,2-Trichloroethane       | 21.0   | UD        | 21.0     | 100        | ug/L    |
| 591-78-6                  | 2-Hexanone                  | 89.0   | UD        | 89.0     | 500        | ug/L    |
| 124-48-1                  | Dibromochloromethane        | 18.0   | UD        | 18.0     | 100        | ug/L    |
| 106-93-4                  | 1,2-Dibromoethane           | 15.0   | UD        | 15.0     | 100        | ug/L    |
| 127-18-4                  | Tetrachloroethene           | 23.0   | UD        | 23.0     | 100        | ug/L    |
| 108-90-7                  | Chlorobenzene               | 12.0   | UD        | 12.0     | 100        | ug/L    |
| 100-41-4                  | Ethyl Benzene               | 5500   | D         | 13.0     | 100        | ug/L    |
| 179601-23-1               | m/p-Xylenes                 | 7100   | D         | 24.0     | 200        | ug/L    |
| 95-47-6                   | o-Xylene                    | 76.1   | JD        | 12.0     | 100        | ug/L    |
| 100-42-5                  | Styrene                     | 15.0   | UD        | 15.0     | 100        | ug/L    |
| 75-25-2                   | Bromoform                   | 19.0   | UD        | 19.0     | 100        | ug/L    |
| 98-82-8                   | Isopropylbenzene            | 220    | D         | 12.0     | 100        | ug/L    |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 26.0   | UD        | 26.0     | 100        | ug/L    |
| 541-73-1                  | 1,3-Dichlorobenzene         | 16.0   | UD        | 16.0     | 100        | ug/L    |
| 106-46-7                  | 1,4-Dichlorobenzene         | 19.0   | UD        | 19.0     | 100        | ug/L    |
| 95-50-1                   | 1,2-Dichlorobenzene         | 16.0   | UD        | 16.0     | 100        | ug/L    |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 53.0   | UD        | 53.0     | 100        | ug/L    |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 20.0   | UD        | 20.0     | 100        | ug/L    |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 20.0   | UD        | 20.0     | 100        | ug/L    |
| <b>SURROGATES</b>         |                             |        |           |          |            |         |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 55.6   |           | 74 - 125 | 111%       | SPK: 50 |
| 1868-53-7                 | Dibromofluoromethane        | 49.9   |           | 75 - 124 | 100%       | SPK: 50 |
| 2037-26-5                 | Toluene-d8                  | 45.3   |           | 86 - 113 | 91%        | SPK: 50 |
| 460-00-4                  | 4-Bromofluorobenzene        | 46.4   |           | 77 - 121 | 93%        | SPK: 50 |
| <b>INTERNAL STANDARDS</b> |                             |        |           |          |            |         |
| 363-72-4                  | Pentafluorobenzene          | 458000 | 4.603     |          |            |         |
| 540-36-3                  | 1,4-Difluorobenzene         | 815000 | 5.535     |          |            |         |
| 3114-55-4                 | Chlorobenzene-d5            | 779000 | 8.776     |          |            |         |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 407000 | 11.175    |          |            |         |

## Report of Analysis

|                    |                 |                 |          |
|--------------------|-----------------|-----------------|----------|
| Client:            | G Environmental | Date Collected: | 09/23/25 |
| Project:           | Summer          | Date Received:  | 09/23/25 |
| Client Sample ID:  | MW2DL2          | SDG No.:        | Q3175    |
| Lab Sample ID:     | Q3175-01DL2     | Matrix:         | Water    |
| Analytical Method: | 8260D           | % Solid:        | 0        |
| Sample Wt/Vol:     | 5               | Units:          | mL       |
| Soil Aliquot Vol:  |                 |                 | uL       |
| GC Column:         | DB-624UI        | ID :            | 0.18     |
| Prep Method :      |                 | Level :         | LOW      |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039179.D        | 100       | 09/25/25 14:25 | VV092525      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
 Data File : VV039179.D  
 Acq On : 25 Sep 2025 14:25  
 Operator : SY/MD  
 Sample : Q3175-01DL2 100X  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 MW2DL2

Manual Integrations  
 APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025  
 Supervised By :Semsettin Yesilyurt 09/29/2025

Quant Time: Sep 26 01:28:35 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

| Compound                    | R.T.           | QIon | Response            | Conc   | Units  | Dev(Min) |
|-----------------------------|----------------|------|---------------------|--------|--------|----------|
| -----                       |                |      |                     |        |        |          |
| Internal Standards          |                |      |                     |        |        |          |
| 1) Pentafluorobenzene       | 4.603          | 168  | 457907              | 50.000 | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 5.535          | 114  | 814896              | 50.000 | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 8.776          | 117  | 779302              | 50.000 | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 11.175         | 152  | 406825              | 50.000 | ug/l   | 0.00     |
|                             |                |      |                     |        |        |          |
| System Monitoring Compounds |                |      |                     |        |        |          |
| 33) 1,2-Dichloroethane-d4   | 4.944          | 65   | 368510              | 55.605 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 81 - 118 |      | Recovery = 111.220% |        |        |          |
| 35) Dibromofluoromethane    | 4.503          | 113  | 326916              | 49.902 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 80 - 119 |      | Recovery = 99.800%  |        |        |          |
| 50) Toluene-d8              | 7.239          | 98   | 872246              | 45.333 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 89 - 112 |      | Recovery = 90.660%  |        |        |          |
| 62) 4-Bromofluorobenzene    | 10.001         | 95   | 367137              | 46.351 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 85 - 114 |      | Recovery = 92.700%  |        |        |          |
|                             |                |      |                     |        |        |          |
| Target Compounds            |                |      |                     |        |        |          |
|                             |                |      |                     |        | Qvalue |          |
| 31) Cyclohexane             | 4.590          | 56   | 36779               | 3.112  | ug/l # | 79       |
| 39) Methylcyclohexane       | 6.047          | 83   | 19339               | 1.617  | ug/l   | 93       |
| 40) Benzene                 | 5.018          | 78   | 114906              | 3.563  | ug/l   | 97       |
| 52) Toluene                 | 7.326          | 92   | 13136               | 0.702  | ug/l   | 99       |
| 67) Ethyl Benzene           | 8.937          | 91   | 1804165             | 54.908 | ug/l   | 100      |
| 68) m/p-Xylenes             | 9.062          | 106  | 895770              | 70.633 | ug/l   | 98       |
| 69) o-Xylene                | 9.480          | 106  | 8550                | 0.761  | ug/l   | 88       |
| 73) Isopropylbenzene        | 9.860          | 105  | 64305               | 2.179  | ug/l   | 97       |
| 78) n-propylbenzene         | 10.281         | 91   | 211654              | 5.890  | ug/l   | 100      |
| 80) 1,3,5-Trimethylbenzene  | 10.468         | 105  | 267764              | 10.918 | ug/l   | 100      |
| 84) 1,2,4-Trimethylbenzene  | 10.841         | 105  | 1212989             | 50.943 | ug/l   | 99       |
| 85) sec-Butylbenzene        | 11.014         | 105  | 11370m              | 0.351  | ug/l   |          |
| 86) p-Isopropyltoluene      | 11.172         | 119  | 6248m               | 0.231  | ug/l   |          |
| 89) n-Butylbenzene          | 11.580         | 91   | 22134m              | 0.858  | ug/l   |          |
| 95) Naphthalene             | 13.429         | 128  | 343331              | 12.801 | ug/l   | 99       |
| -----                       |                |      |                     |        |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
Data File : VV039179.D  
Acq On : 25 Sep 2025 14:25  
Operator : SY/MD  
Sample : Q3175-01DL2 100X  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 13 Sample Multiplier: 1


Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2DL2

Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025

Quant Time: Sep 26 01:28:35 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration









#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 4.603 min Scan# 1108  
Delta R.T. 0.006 min  
Lab File: VV039179.D  
Acq: 25 Sep 2025 14:25

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2DL2

Tgt Ion:168 Resp: 45790  
Ion Ratio Lower Upper  
168 100  
99 63.4 49.6 74.4




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025



#31  
Cyclohexane  
Concen: 3.112 ug/l  
RT: 4.590 min Scan# 1104  
Delta R.T. 0.010 min  
Lab File: VV039179.D  
Acq: 25 Sep 2025 14:25

Tgt Ion: 56 Resp: 36779  
Ion Ratio Lower Upper  
56 100  
69 67.7 26.2 39.2#  
84 92.6 70.3 105.5





#33

1,2-Dichloroethane-d4

Concen: 55.605 ug/l

RT: 4.944 min Scan# 1214

Delta R.T. 0.006 min

Lab File: VV039179.D

Acq: 25 Sep 2025 14:25

Instrument :

MSVOA\_V

ClientSampleId :

MW2DL2

Tgt Ion: 65 Resp: 368510





Ion Ratio Lower Upper

65 100

67 53.0 0.0 107.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025

#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 5.535 min Scan# 1398

Delta R.T. 0.006 min

Lab File: VV039179.D

Acq: 25 Sep 2025 14:25




Tgt Ion:114 Resp: 814896


Ion Ratio Lower Upper

114 100

63 20.4 0.0 42.6

88 16.2 0.0 30.8





#35

Dibromofluoromethane

Concen: 49.902 ug/l

RT: 4.503 min Scan# 1075

Delta R.T. 0.006 min

Lab File: VV039179.D

Acq: 25 Sep 2025 14:25

Instrument :

MSVOA\_V

ClientSampleId :

MW2DL2

Tgt Ion: 113 Resp: 326910

Ion Ratio Lower Upper

113 100





111 103.8 82.4 123.6

192 17.2 13.8 20.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025  
 Supervised By :Semsettin Yesilyurt 09/29/2025



#39

Methylcyclohexane

Concen: 1.617 ug/l

RT: 6.047 min Scan# 1557

Delta R.T. 0.004 min

Lab File: VV039179.D

Acq: 25 Sep 2025 14:25




Tgt Ion: 83 Resp: 19339


Ion Ratio Lower Upper

83 100

55 83.8 60.5 90.7

98 49.3 37.8 56.6





#40

Benzene

Concen: 3.563 ug/l

RT: 5.018 min Scan# 11

Delta R.T. 0.013 min

Lab File: VV039179.D

Acq: 25 Sep 2025 14:25

Instrument :

MSVOA\_V

ClientSampleId :

MW2DL2

Tgt Ion: 78 Resp: 114900

Ion Ratio Lower Upper





78 100

77 25.0 18.7 28.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025  
 Supervised By :Semsettin Yesilyurt 09/29/2025



#50

Toluene-d8

Concen: 45.333 ug/l

RT: 7.239 min Scan# 1928

Delta R.T. 0.006 min

Lab File: VV039179.D




Acq: 25 Sep 2025 14:25


Tgt Ion: 98 Resp: 872246

Ion Ratio Lower Upper

98 100

100 64.7 52.2 78.2





#52

Toluene

Concen: 0.702 ug/l

RT: 7.326 min Scan# 1948

Delta R.T. 0.022 min

Lab File: VV039179.D

Acq: 25 Sep 2025 14:25

Instrument :

MSVOA\_V

ClientSampleId :

MW2DL2

Tgt Ion: 92 Resp: 13130

Ion Ratio Lower Upper





92 100

91 174.9 139.2 208.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025  
 Supervised By :Semsettin Yesilyurt 09/29/2025



#62

4-Bromofluorobenzene

Concen: 46.351 ug/l

RT: 10.001 min Scan# 2787

Delta R.T. 0.003 min

Lab File: VV039179.D

Acq: 25 Sep 2025 14:25




Tgt Ion: 95 Resp: 367137


Ion Ratio Lower Upper

95 100

174 73.4 0.0 150.4

176 69.4 0.0 144.2





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 8.776 min Scan# 2405

Delta R.T. 0.003 min

Lab File: VV039179.D

Acq: 25 Sep 2025 14:25

Instrument :

MSVOA\_V

ClientSampleId :

MW2DL2

Tgt Ion:117 Resp: 779302

Ion Ratio Lower Upper

117 100





82 55.9 45.4 68.2

119 31.3 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025  
 Supervised By :Semsettin Yesilyurt 09/29/2025



#67

Ethyl Benzene

Concen: 54.908 ug/l

RT: 8.937 min Scan# 2456

Delta R.T. 0.003 min

Lab File: VV039179.D




Acq: 25 Sep 2025 14:25


Tgt Ion: 91 Resp: 1804165

Ion Ratio Lower Upper

91 100

106 30.7 24.6 37.0





#68

m/p-Xylenes

Concen: 70.633 ug/l

RT: 9.062 min Scan# 2494

Delta R.T. 0.003 min

Lab File: VV039179.D

Acq: 25 Sep 2025 14:25

Instrument :

MSVOA\_V

ClientSampleId :

MW2DL2

Tgt Ion:106 Resp: 895770





Ion Ratio Lower Upper

106 100

91 206.7 162.5 243.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025

#69

o-Xylene

Concen: 0.761 ug/l

RT: 9.480 min Scan# 2625

Delta R.T. 0.013 min

Lab File: VV039179.D




Acq: 25 Sep 2025 14:25


Tgt Ion:106 Resp: 8550

Ion Ratio Lower Upper

106 100


91 198.5 109.1 327.3





#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 11.175 min Scan# 3151  
Delta R.T. 0.003 min  
Lab File: VV039179.D  
Acq: 25 Sep 2025 14:25




Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2DL2




Tgt Ion:152 Resp: 406825  
Ion Ratio Lower Upper  
152 100  
115 62.4 44.8 134.4  
150 158.5 0.0 353.8

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025




#73  
Isopropylbenzene  
Concen: 2.179 ug/l  
RT: 9.860 min Scan# 2743  
Delta R.T. 0.006 min  
Lab File: VV039179.D  
Acq: 25 Sep 2025 14:25



Tgt Ion:105 Resp: 64305  
Ion Ratio Lower Upper  
105 100  
120 24.6 13.1 39.1









#78  
n-propylbenzene  
Concen: 5.890 ug/l  
RT: 10.281 min Scan# 211654  
Delta R.T. 0.006 min  
Lab File: VV039179.D  
Acq: 25 Sep 2025 14:25

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2DL2

Tgt Ion: 91 Resp: 211654  
Ion Ratio Lower Upper  
91 100  
120 22.0 11.1 33.3




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025



#80  
1,3,5-Trimethylbenzene  
Concen: 10.918 ug/l  
RT: 10.468 min Scan# 2932  
Delta R.T. 0.003 min  
Lab File: VV039179.D  
Acq: 25 Sep 2025 14:25


Tgt Ion:105 Resp: 267764  
Ion Ratio Lower Upper  
105 100  
120 48.5 24.3 72.8





#84  
1,2,4-Trimethylbenzene  
Concen: 50.943 ug/l  
RT: 10.841 min Scan# 3048  
Delta R.T. 0.003 min  
Lab File: VV039179.D  
Acq: 25 Sep 2025 14:25




Instrument :  
MSVOA\_V  
ClientSampleId :  
MW2DL2



Tgt Ion:105 Resp: 1212989  
Ion Ratio Lower Upper  
105 100  
120 44.3 22.4 67.2




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025



#85  
sec-Butylbenzene  
Concen: 0.351 ug/l m  
RT: 11.014 min Scan# 3102  
Delta R.T. 0.003 min  
Lab File: VV039179.D  
Acq: 25 Sep 2025 14:25

Tgt Ion:105 Resp: 11370  
Ion Ratio Lower Upper  
105 100  
134 3.4 9.6 28.8#





#86

p-Isopropyltoluene

Concen: 0.231 ug/l m

RT: 11.172 min Scan# 3149

Delta R.T. 0.006 min

Lab File: VV039179.D

Acq: 25 Sep 2025 14:25

Instrument :

MSVOA\_V

ClientSampleId :

MW2DL2

Tgt Ion:119 Resp: 6248

Ion Ratio Lower Upper





119 100

134 51.4 12.7 38.0

91 61.4 12.0 36.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025

#89

n-Butylbenzene

Concen: 0.858 ug/l m

RT: 11.580 min Scan# 3278

Delta R.T. 0.003 min

Lab File: VV039179.D

Acq: 25 Sep 2025 14:25




Tgt Ion: 91 Resp: 22134


Ion Ratio Lower Upper

91 100

92 44.1 26.6 79.8

134 135.1 12.1 36.3





#95

Naphthalene

Concen: 12.801 ug/l

RT: 13.429 min Scan# 3853

Delta R.T. 0.006 min

Lab File: VV039179.D


Acq: 25 Sep 2025 14:25

Instrument :

MSVOA\_V

ClientSampleId :

MW2DL2



Tgt Ion:128 Resp: 34333

Ion Ratio Lower Upper

128 100

127 13.9 10.3 15.5



129 10.9 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025

Supervised By :Semsettin Yesilyurt 09/29/2025



## Report of Analysis

|                    |                                            |                 |                              |
|--------------------|--------------------------------------------|-----------------|------------------------------|
| Client:            | G Environmental                            | Date Collected: | 09/23/25                     |
| Project:           | Summer                                     | Date Received:  | 09/23/25                     |
| Client Sample ID:  | MW4                                        | SDG No.:        | Q3175                        |
| Lab Sample ID:     | Q3175-02                                   | Matrix:         | Water                        |
| Analytical Method: | 8260D                                      | % Solid:        | 0                            |
| Sample Wt/Vol:     | 5                      Units:    mL        | Final Vol:      | 5000                      uL |
| Soil Aliquot Vol:  | uL                                         | Test:           | VOCMS Group1                 |
| GC Column:         | DB-624UI                      ID :    0.18 | Level :         | LOW                          |
| Prep Method :      |                                            |                 |                              |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039154.D        | 1         | 09/24/25 15:53 | VV092425      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL   | LOQ / CRQL | Units |
|----------------|--------------------------------|-------|-----------|-------|------------|-------|
| <b>TARGETS</b> |                                |       |           |       |            |       |
| 75-71-8        | Dichlorodifluoromethane        | 0.22  | U         | 0.22  | 1.00       | ug/L  |
| 74-87-3        | Chloromethane                  | 0.32  | U         | 0.32  | 1.00       | ug/L  |
| 75-01-4        | Vinyl Chloride                 | 0.26  | U         | 0.26  | 1.00       | ug/L  |
| 74-83-9        | Bromomethane                   | 1.40  | U         | 1.40  | 5.00       | ug/L  |
| 75-00-3        | Chloroethane                   | 0.47  | U         | 0.47  | 1.00       | ug/L  |
| 75-69-4        | Trichlorofluoromethane         | 0.33  | U         | 0.33  | 1.00       | ug/L  |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 0.25  | U         | 0.25  | 1.00       | ug/L  |
| 75-65-0        | Tert butyl alcohol             | 5.50  | U         | 5.50  | 25.0       | ug/L  |
| 75-35-4        | 1,1-Dichloroethene             | 0.23  | U         | 0.23  | 1.00       | ug/L  |
| 67-64-1        | Acetone                        | 1.50  | U         | 1.50  | 5.00       | ug/L  |
| 75-15-0        | Carbon Disulfide               | 0.57  | J         | 0.21  | 1.00       | ug/L  |
| 1634-04-4      | Methyl tert-butyl Ether        | 0.16  | U         | 0.16  | 1.00       | ug/L  |
| 79-20-9        | Methyl Acetate                 | 0.27  | U         | 0.27  | 1.00       | ug/L  |
| 75-09-2        | Methylene Chloride             | 0.28  | U         | 0.28  | 1.00       | ug/L  |
| 156-60-5       | trans-1,2-Dichloroethene       | 0.23  | U         | 0.23  | 1.00       | ug/L  |
| 75-34-3        | 1,1-Dichloroethane             | 0.23  | U         | 0.23  | 1.00       | ug/L  |
| 110-82-7       | Cyclohexane                    | 160   | E         | 1.50  | 5.00       | ug/L  |
| 78-93-3        | 2-Butanone                     | 0.98  | U         | 0.98  | 5.00       | ug/L  |
| 56-23-5        | Carbon Tetrachloride           | 0.25  | U         | 0.25  | 1.00       | ug/L  |
| 156-59-2       | cis-1,2-Dichloroethene         | 0.19  | U         | 0.19  | 1.00       | ug/L  |
| 74-97-5        | Bromochloromethane             | 0.22  | U         | 0.22  | 1.00       | ug/L  |
| 67-66-3        | Chloroform                     | 0.25  | U         | 0.25  | 1.00       | ug/L  |
| 71-55-6        | 1,1,1-Trichloroethane          | 0.20  | U         | 0.20  | 1.00       | ug/L  |
| 108-87-2       | Methylcyclohexane              | 230   | E         | 0.16  | 1.00       | ug/L  |
| 71-43-2        | Benzene                        | 760   | E         | 0.15  | 1.00       | ug/L  |
| 107-06-2       | 1,2-Dichloroethane             | 0.22  | U         | 0.22  | 1.00       | ug/L  |
| 79-01-6        | Trichloroethene                | 0.090 | U         | 0.090 | 1.00       | ug/L  |
| 78-87-5        | 1,2-Dichloropropane            | 0.20  | U         | 0.20  | 1.00       | ug/L  |
| 75-27-4        | Bromodichloromethane           | 0.22  | U         | 0.22  | 1.00       | ug/L  |
| 108-10-1       | 4-Methyl-2-Pentanone           | 0.68  | UQ        | 0.68  | 5.00       | ug/L  |

### Report of Analysis

|                    |                 |                 |              |
|--------------------|-----------------|-----------------|--------------|
| Client:            | G Environmental | Date Collected: | 09/23/25     |
| Project:           | Summer          | Date Received:  | 09/23/25     |
| Client Sample ID:  | MW4             | SDG No.:        | Q3175        |
| Lab Sample ID:     | Q3175-02        | Matrix:         | Water        |
| Analytical Method: | 8260D           | % Solid:        | 0            |
| Sample Wt/Vol:     | 5               | Units:          | mL           |
| Soil Aliquot Vol:  |                 |                 | uL           |
| GC Column:         | DB-624UI        | ID :            | 0.18         |
| Prep Method :      |                 | Level :         | LOW          |
|                    |                 | Final Vol:      | 5000         |
|                    |                 | Test:           | VOCMS Group1 |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039154.D        | 1         | 09/24/25 15:53 | VV092425      |

| CAS Number                | Parameter                   | Conc.   | Qualifier | MDL      | LOQ / CRQL | Units   |
|---------------------------|-----------------------------|---------|-----------|----------|------------|---------|
| 108-88-3                  | Toluene                     | 71.3    |           | 0.14     | 1.00       | ug/L    |
| 10061-02-6                | t-1,3-Dichloropropene       | 0.17    | U         | 0.17     | 1.00       | ug/L    |
| 10061-01-5                | cis-1,3-Dichloropropene     | 0.16    | U         | 0.16     | 1.00       | ug/L    |
| 79-00-5                   | 1,1,2-Trichloroethane       | 0.21    | U         | 0.21     | 1.00       | ug/L    |
| 591-78-6                  | 2-Hexanone                  | 0.89    | U         | 0.89     | 5.00       | ug/L    |
| 124-48-1                  | Dibromochloromethane        | 0.18    | U         | 0.18     | 1.00       | ug/L    |
| 106-93-4                  | 1,2-Dibromoethane           | 0.15    | U         | 0.15     | 1.00       | ug/L    |
| 127-18-4                  | Tetrachloroethene           | 0.23    | U         | 0.23     | 1.00       | ug/L    |
| 108-90-7                  | Chlorobenzene               | 0.12    | U         | 0.12     | 1.00       | ug/L    |
| 100-41-4                  | Ethyl Benzene               | 480     | E         | 0.13     | 1.00       | ug/L    |
| 179601-23-1               | m/p-Xylenes                 | 1300    | E         | 0.24     | 2.00       | ug/L    |
| 95-47-6                   | o-Xylene                    | 230     | E         | 0.12     | 1.00       | ug/L    |
| 100-42-5                  | Styrene                     | 0.15    | U         | 0.15     | 1.00       | ug/L    |
| 75-25-2                   | Bromoform                   | 0.19    | U         | 0.19     | 1.00       | ug/L    |
| 98-82-8                   | Isopropylbenzene            | 110     | E         | 0.12     | 1.00       | ug/L    |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 0.26    | U         | 0.26     | 1.00       | ug/L    |
| 541-73-1                  | 1,3-Dichlorobenzene         | 0.16    | U         | 0.16     | 1.00       | ug/L    |
| 106-46-7                  | 1,4-Dichlorobenzene         | 0.19    | U         | 0.19     | 1.00       | ug/L    |
| 95-50-1                   | 1,2-Dichlorobenzene         | 0.16    | U         | 0.16     | 1.00       | ug/L    |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 0.53    | U         | 0.53     | 1.00       | ug/L    |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 0.20    | U         | 0.20     | 1.00       | ug/L    |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 0.20    | U         | 0.20     | 1.00       | ug/L    |
| <b>SURROGATES</b>         |                             |         |           |          |            |         |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 46.6    |           | 74 - 125 | 93%        | SPK: 50 |
| 1868-53-7                 | Dibromofluoromethane        | 38.8    |           | 75 - 124 | 78%        | SPK: 50 |
| 2037-26-5                 | Toluene-d8                  | 47.7    |           | 86 - 113 | 95%        | SPK: 50 |
| 460-00-4                  | 4-Bromofluorobenzene        | 48.4    |           | 77 - 121 | 97%        | SPK: 50 |
| <b>INTERNAL STANDARDS</b> |                             |         |           |          |            |         |
| 363-72-4                  | Pentafluorobenzene          | 628000  | 4.603     |          |            |         |
| 540-36-3                  | 1,4-Difluorobenzene         | 1090000 | 5.535     |          |            |         |
| 3114-55-4                 | Chlorobenzene-d5            | 1000000 | 8.776     |          |            |         |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 493000  | 11.175    |          |            |         |

## Report of Analysis

|                    |                            |                 |              |
|--------------------|----------------------------|-----------------|--------------|
| Client:            | G Environmental            | Date Collected: | 09/23/25     |
| Project:           | Summer                     | Date Received:  | 09/23/25     |
| Client Sample ID:  | MW4                        | SDG No.:        | Q3175        |
| Lab Sample ID:     | Q3175-02                   | Matrix:         | Water        |
| Analytical Method: | 8260D                      | % Solid:        | 0            |
| Sample Wt/Vol:     | 5      Units:    mL        | Final Vol:      | 5000      uL |
| Soil Aliquot Vol:  | uL                         | Test:           | VOCMS Group1 |
| GC Column:         | DB-624UI      ID :    0.18 | Level :         | LOW          |
| Prep Method :      |                            |                 |              |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039154.D        | 1         | 09/24/25 15:53 | VV092425      |

| CAS Number                            | Parameter                          | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|---------------------------------------|------------------------------------|-------|-----------|-----|------------|-------|
| <b>TENTATIVE IDENTIFIED COMPOUNDS</b> |                                    |       |           |     |            |       |
| 000822-50-4                           | Cyclopentane, 1,2-dimethyl-, trans | 80.5  | J         |     | 5.23       | ug/L  |
| 016491-15-9                           | Cyclopentene, 1,5-dimethyl-        | 120   | J         |     | 6.72       | ug/L  |
| 103-65-1                              | n-propylbenzene                    | 250   | J         |     | 10.3       | ug/L  |
| 000611-14-3                           | Benzene, 1-ethyl-2-methyl-         | 430   | J         |     | 10.4       | ug/L  |
| 108-67-8                              | 1,3,5-Trimethylbenzene             | 390   | J         |     | 10.5       | ug/L  |
| 000622-96-8                           | Benzene, 1-ethyl-4-methyl-         | 190   | J         |     | 10.7       | ug/L  |
| 95-63-6                               | 1,2,4-Trimethylbenzene             | 670   | J         |     | 10.8       | ug/L  |
| 135-98-8                              | sec-Butylbenzene                   | 18.4  | J         |     | 11.0       | ug/L  |
| 99-87-6                               | p-Isopropyltoluene                 | 14.2  | J         |     | 11.2       | ug/L  |
| 000526-73-8                           | Benzene, 1,2,3-trimethyl-          | 370   | J         |     | 11.3       | ug/L  |
| 000496-11-7                           | Indane                             | 390   | J         |     | 11.5       | ug/L  |
| 104-51-8                              | n-Butylbenzene                     | 47.9  | J         |     | 11.6       | ug/L  |
| 001758-88-9                           | Benzene, 2-ethyl-1,4-dimethyl-     | 100   | J         |     | 11.8       | ug/L  |
| 000934-80-5                           | Benzene, 4-ethyl-1,2-dimethyl-     | 230   | J         |     | 11.9       | ug/L  |
| 007525-62-4                           | Benzene, 1-ethenyl-3-ethyl-        | 130   | J         |     | 12.0       | ug/L  |
| 000488-23-3                           | Benzene, 1,2,3,4-tetramethyl-      | 120   | J         |     | 12.3       | ug/L  |
| 000095-93-2                           | Benzene, 1,2,4,5-tetramethyl-      | 140   | J         |     | 12.4       | ug/L  |
| 000767-58-8                           | Indan, 1-methyl-                   | 120   | J         |     | 12.6       | ug/L  |
| 002039-89-6                           | Benzene, 2-ethenyl-1,4-dimethyl-   | 280   | J         |     | 12.8       | ug/L  |
| 000119-64-2                           | Naphthalene, 1,2,3,4-tetrahydro-   | 75.1  | J         |     | 13.0       | ug/L  |
| 91-20-3                               | Naphthalene                        | 300   | J         |     | 13.4       | ug/L  |
| 000091-57-6                           | Naphthalene, 2-methyl-             | 120   | J         |     | 14.6       | ug/L  |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
 Data File : VV039154.D  
 Acq On : 24 Sep 2025 15:53  
 Operator : SY/MD  
 Sample : Q3175-02  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 MW4

Manual Integrations  
 APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025  
 Supervised By :Semsettin Yesilyurt 09/26/2025

Quant Time: Sep 25 04:48:59 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

| Compound                    | R.T.           | QIon | Response   | Conc     | Units  | Dev(Min) |
|-----------------------------|----------------|------|------------|----------|--------|----------|
| -----                       |                |      |            |          |        |          |
| Internal Standards          |                |      |            |          |        |          |
| 1) Pentafluorobenzene       | 4.603          | 168  | 627637     | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 5.535          | 114  | 1091934    | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 8.776          | 117  | 1002921    | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 11.175         | 152  | 492612     | 50.000   | ug/l   | 0.00     |
|                             |                |      |            |          |        |          |
| System Monitoring Compounds |                |      |            |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 4.944          | 65   | 423649     | 46.638   | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 81 - 118 |      | Recovery = | 93.280%  |        |          |
| 35) Dibromofluoromethane    | 4.503          | 113  | 340181     | 38.753   | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 80 - 119 |      | Recovery = | 77.500%# |        |          |
| 50) Toluene-d8              | 7.236          | 98   | 1230675    | 47.734   | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 89 - 112 |      | Recovery = | 95.460%  |        |          |
| 62) 4-Bromofluorobenzene    | 10.001         | 95   | 513340     | 48.366   | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 85 - 114 |      | Recovery = | 96.740%  |        |          |
|                             |                |      |            |          |        |          |
| Target Compounds            |                |      |            |          | Qvalue |          |
| 17) Carbon Disulfide        | 2.262          | 76   | 16408      | 0.567    | ug/l # | 91       |
| 31) Cyclohexane             | 4.583          | 56   | 2558050    | 157.888  | ug/l   | 98       |
| 39) Methylcyclohexane       | 6.046          | 83   | 3663235    | 228.559  | ug/l   | 99       |
| 40) Benzene                 | 4.992          | 78   | 32865540m  | 760.461  | ug/l   |          |
| 52) Toluene                 | 7.307          | 92   | 1785393    | 71.251   | ug/l   | 98       |
| 67) Ethyl Benzene           | 8.924          | 91   | 20451028m  | 483.626  | ug/l   |          |
| 68) m/p-Xylenes             | 9.056          | 106  | 21385207m  | 1310.281 | ug/l   |          |
| 69) o-Xylene                | 9.468          | 106  | 3284665    | 227.292  | ug/l   | 99       |
| 73) Isopropylbenzene        | 9.857          | 105  | 3830198    | 107.200  | ug/l   | 100      |
| 78) n-propylbenzene         | 10.278         | 91   | 10733473   | 246.691  | ug/l   | 98       |
| 80) 1,3,5-Trimethylbenzene  | 10.464         | 105  | 11542912   | 388.679  | ug/l   | 99       |
| 84) 1,2,4-Trimethylbenzene  | 10.831         | 105  | 19453064m  | 674.708  | ug/l   |          |
| 85) sec-Butylbenzene        | 11.014         | 105  | 722406     | 18.392   | ug/l   | 98       |
| 86) p-Isopropyltoluene      | 11.168         | 119  | 463357     | 14.174   | ug/l   | 97       |
| 89) n-Butylbenzene          | 11.574         | 91   | 1495426m   | 47.897   | ug/l   |          |
| 95) Naphthalene             | 13.422         | 128  | 9872434    | 303.997  | ug/l   | 99       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039154.D  
Acq On : 24 Sep 2025 15:53  
Operator : SY/MD  
Sample : Q3175-02  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 11 Sample Multiplier: 1


Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4

Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025  
Supervised By :Semsettin Yesilyurt 09/26/2025

Quant Time: Sep 25 04:48:59 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration









#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 4.603 min Scan# 1106  
Delta R.T. 0.006 min  
Lab File: VV039154.D  
Acq: 24 Sep 2025 15:53

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4

Tgt Ion: 168 Resp: 62763  
Ion Ratio Lower Upper  
168 100  
99 64.5 49.6 74.4




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 09/26/2025  
Supervised By :Semsettin Yesilyurt 09/26/2025



#17  
Carbon Disulfide  
Concen: 0.567 ug/l  
RT: 2.262 min Scan# 380  
Delta R.T. 0.009 min  
Lab File: VV039154.D  
Acq: 24 Sep 2025 15:53

Tgt Ion: 76 Resp: 16408  
Ion Ratio Lower Upper  
76 100  
78 6.0 7.4 11.0#





#31

Cyclohexane

Concen: 157.888 ug/l

RT: 4.583 min Scan# 1101

Delta R.T. 0.003 min

Lab File: VV039154.D

Acq: 24 Sep 2025 15:53

Instrument :

MSVOA\_V

ClientSampleId :

MW4

Tgt Ion: 56 Resp: 255805

Ion Ratio Lower Upper





56 100

69 29.8 26.2 39.2

84 88.4 70.3 105.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025  
Supervised By :Semsettin Yesilyurt 09/26/2025

#33

1,2-Dichloroethane-d4

Concen: 46.638 ug/l

RT: 4.944 min Scan# 1214

Delta R.T. 0.006 min

Lab File: VV039154.D




Acq: 24 Sep 2025 15:53


Tgt Ion: 65 Resp: 423649

Ion Ratio Lower Upper

65 100

67 49.5 0.0 107.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 5.535 min Scan# 111

Delta R.T. 0.006 min

Lab File: VV039154.D

Acq: 24 Sep 2025 15:53

Instrument :

MSVOA\_V

ClientSampleId :

MW4

Tgt Ion:114 Resp: 1091934

Ion Ratio Lower Upper

114 100

63 20.1 0.0 42.6





88 16.0 0.0 30.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025

Supervised By :Semsettin Yesilyurt 09/26/2025



#35

Dibromofluoromethane

Concen: 38.753 ug/l

RT: 4.503 min Scan# 1077

Delta R.T. 0.006 min

Lab File: VV039154.D

Acq: 24 Sep 2025 15:53




Tgt Ion:113 Resp: 340181


Ion Ratio Lower Upper

113 100

111 102.3 82.4 123.6

192 17.1 13.8 20.8





#39

Methylcyclohexane

Concen: 228.559 ug/l

RT: 6.046 min Scan# 1

Delta R.T. 0.003 min

Lab File: VV039154.D

Acq: 24 Sep 2025 15:53

Instrument :

MSVOA\_V

ClientSampleId :

MW4

Tgt Ion: 83 Resp: 366323

Ion Ratio Lower Upper





83 100

55 75.6 60.5 90.7

98 48.3 37.8 56.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025  
Supervised By :Semsettin Yesilyurt 09/26/2025

#40

Benzene

Concen: 760.461 ug/l m

RT: 4.992 min Scan# 1229

Delta R.T. -0.013 min

Lab File: VV039154.D




Acq: 24 Sep 2025 15:53


Tgt Ion: 78 Resp:32865540

Ion Ratio Lower Upper

78 100

77 33.9 18.7 28.1#





#50

Toluene-d8

Concen: 47.734 ug/l

RT: 7.236 min Scan# 1926

Delta R.T. 0.003 min

Lab File: VV039154.D

Acq: 24 Sep 2025 15:53

Instrument :

MSVOA\_V

ClientSampleId :

MW4

Tgt Ion: 98 Resp: 123067





Ion Ratio Lower Upper

98 100

100 64.9 52.2 78.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025  
Supervised By :Semsettin Yesilyurt 09/26/2025

#52

Toluene

Concen: 71.251 ug/l

RT: 7.307 min Scan# 1949

Delta R.T. 0.003 min

Lab File: VV039154.D




Acq: 24 Sep 2025 15:53


Tgt Ion: 92 Resp: 1785393

Ion Ratio Lower Upper

92 100

91 171.3 139.2 208.8





#62

4-Bromofluorobenzene

Concen: 48.366 ug/l

RT: 10.001 min Scan# 21

Delta R.T. 0.003 min

Lab File: VV039154.D

Acq: 24 Sep 2025 15:53

Instrument :

MSVOA\_V

ClientSampleId :

MW4

Tgt Ion: 95 Resp: 513340

Ion Ratio Lower Upper

95 100

174 68.6 0.0 150.4





176 65.8 0.0 144.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025

Supervised By :Semsettin Yesilyurt 09/26/2025



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 8.776 min Scan# 2406

Delta R.T. 0.003 min

Lab File: VV039154.D

Acq: 24 Sep 2025 15:53




Tgt Ion:117 Resp: 1002921


Ion Ratio Lower Upper

117 100

82 56.2 45.4 68.2

119 32.4 25.6 38.4





#67

Ethyl Benzene

Concen: 483.626 ug/l m

RT: 8.924 min Scan# 2455

Delta R.T. -0.010 min

Lab File: VV039154.D

Acq: 24 Sep 2025 15:53

Instrument :

MSVOA\_V

ClientSampleId :

MW4

Tgt Ion: 91 Resp:20451028

Ion Ratio Lower Upper





91 100

106 34.2 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025  
 Supervised By :Semsettin Yesilyurt 09/26/2025



#68

m/p-Xylenes

Concen: 1310.281 ug/l m

RT: 9.056 min Scan# 2493

Delta R.T. -0.003 min

Lab File: VV039154.D




Acq: 24 Sep 2025 15:53


Tgt Ion:106 Resp:21385207

Ion Ratio Lower Upper

106 100


91 0.0 162.5 243.7#





#69  
o-Xylene  
Concen: 227.292 ug/l  
RT: 9.468 min Scan# 2621  
Delta R.T. -0.000 min  
Lab File: VV039154.D  
Acq: 24 Sep 2025 15:53




Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4




Tgt Ion:106 Resp: 328466  
Ion Ratio Lower Upper  
106 100  
91 216.1 109.1 327.3

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 09/26/2025  
Supervised By :Semsettin Yesilyurt 09/26/2025




#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 11.175 min Scan# 3152  
Delta R.T. 0.003 min  
Lab File: VV039154.D  
Acq: 24 Sep 2025 15:53



Tgt Ion:152 Resp: 492612  
Ion Ratio Lower Upper  
152 100  
115 70.6 44.8 134.4  
150 156.9 0.0 353.8





#73

Isopropylbenzene

Concen: 107.200 ug/l

RT: 9.857 min Scan# 21

Delta R.T. 0.003 min

Lab File: VV039154.D

Acq: 24 Sep 2025 15:53

Instrument :

MSVOA\_V

ClientSampleId :

MW4

Tgt Ion:105 Resp: 3830198

Ion Ratio Lower Upper





105 100

120 26.0 13.1 39.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025  
 Supervised By :Semsettin Yesilyurt 09/26/2025



#78

n-propylbenzene

Concen: 246.691 ug/l

RT: 10.278 min Scan# 2873

Delta R.T. 0.003 min

Lab File: VV039154.D




Acq: 24 Sep 2025 15:53


Tgt Ion: 91 Resp:10733473

Ion Ratio Lower Upper

91 100

120 22.9 11.1 33.3





#80

1,3,5-Trimethylbenzene

Concen: 388.679 ug/l

RT: 10.464 min Scan# 2931

Delta R.T. -0.000 min

Lab File: VV039154.D

Acq: 24 Sep 2025 15:53

Instrument :

MSVOA\_V

ClientSampleId :

MW4

Tgt Ion:105 Resp:11542912

Ion Ratio Lower Upper





105 100

120 49.2 24.3 72.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025  
Supervised By :Semsettin Yesilyurt 09/26/2025



#84

1,2,4-Trimethylbenzene

Concen: 674.708 ug/l m

RT: 10.831 min Scan# 3045

Delta R.T. -0.007 min

Lab File: VV039154.D




Acq: 24 Sep 2025 15:53


Tgt Ion:105 Resp:19453064

Ion Ratio Lower Upper

105 100

120 9.7 22.4 67.2#





#85

sec-Butylbenzene

Concen: 18.392 ug/l

RT: 11.014 min Scan# 3101

Delta R.T. 0.003 min

Lab File: VV039154.D

Acq: 24 Sep 2025 15:53

Instrument :

MSVOA\_V

ClientSampleId :

MW4

Tgt Ion:105 Resp: 722400

Ion Ratio Lower Upper





105 100

134 18.5 9.6 28.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025  
 Supervised By :Semsettin Yesilyurt 09/26/2025



#86

p-Isopropyltoluene

Concen: 14.174 ug/l

RT: 11.168 min Scan# 3150

Delta R.T. 0.003 min

Lab File: VV039154.D

Acq: 24 Sep 2025 15:53




Tgt Ion:119 Resp: 463357


Ion Ratio Lower Upper

119 100

134 25.7 12.7 38.0

91 26.9 12.0 36.1





#89

n-Butylbenzene

Concen: 47.897 ug/l m

RT: 11.574 min Scan# 3277

Delta R.T. -0.003 min

Lab File: VV039154.D

Acq: 24 Sep 2025 15:53

Instrument :

MSVOA\_V

ClientSampleId :

MW4

Tgt Ion: 91 Resp: 1495420

Ion Ratio Lower Upper





91 100

92 38.2 26.6 79.8

134 130.1 12.1 36.3

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025  
Supervised By :Semsettin Yesilyurt 09/26/2025

#95

Naphthalene

Concen: 303.997 ug/l

RT: 13.422 min Scan# 3851

Delta R.T. -0.001 min

Lab File: VV039154.D

Acq: 24 Sep 2025 15:53




Tgt Ion:128 Resp: 9872434

Ion Ratio Lower Upper

128 100

127 13.2 10.3 15.5

129 11.5 8.6 13.0



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039154.D  
Acq On : 24 Sep 2025 15:53  
Operator : SY/MD  
Sample : Q3175-02  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Title : SW846 8260

Signal : TIC: VV039154.D\data.ms

| peak<br># | R.T.<br>min | first<br>scan | max<br>scan | last<br>scan | PK<br>TY | peak<br>height | corr.<br>area | corr.<br>% max. | % of<br>total |
|-----------|-------------|---------------|-------------|--------------|----------|----------------|---------------|-----------------|---------------|
| 1         | 1.204       | 42            | 51          | 70           | rBV2     | 1017587        | 2416686       | 1.46%           | 0.251%        |
| 2         | 1.294       | 72            | 79          | 89           | rVV      | 2140183        | 2654546       | 1.60%           | 0.275%        |
| 3         | 1.609       | 167           | 177         | 212          | rBV      | 6652502        | 10835140      | 6.52%           | 1.124%        |
| 4         | 1.780       | 221           | 230         | 251          | rBV3     | 2579472        | 4697637       | 2.83%           | 0.487%        |
| 5         | 1.992       | 288           | 296         | 318          | rVB      | 1557653        | 2663073       | 1.60%           | 0.276%        |
| 6         | 2.449       | 412           | 438         | 442          | rBV3     | 1761921        | 4128700       | 2.49%           | 0.428%        |
| 7         | 2.706       | 504           | 518         | 555          | rVB2     | 1950193        | 4614467       | 2.78%           | 0.479%        |
| 8         | 2.976       | 589           | 602         | 625          | rBV      | 1141967        | 2571913       | 1.55%           | 0.267%        |
| 9         | 3.674       | 803           | 819         | 888          | rVB      | 4921140        | 14966090      | 9.01%           | 1.553%        |
| 10        | 4.583       | 1086          | 1102        | 1119         | rVV3     | 4725931        | 13839376      | 8.33%           | 1.436%        |
| 11        | 4.670       | 1121          | 1129        | 1160         | rVB      | 1474380        | 4321156       | 2.60%           | 0.448%        |
| 12        | 4.821       | 1161          | 1176        | 1204         | rBV4     | 825206         | 3007219       | 1.81%           | 0.312%        |
| 13        | 5.011       | 1218          | 1235        | 1255         | rBV2     | 44295281       | 112420596     | 67.69%          | 11.663%       |
| 14        | 5.233       | 1294          | 1304        | 1344         | rVB      | 1276597        | 3563982       | 2.15%           | 0.370%        |
| 15        | 5.407       | 1345          | 1358        | 1387         | rBV2     | 775863         | 1803166       | 1.09%           | 0.187%        |
| 16        | 5.535       | 1387          | 1398        | 1403         | rBV      | 1156407        | 2214586       | 1.33%           | 0.230%        |
| 17        | 6.046       | 1534          | 1557        | 1575         | rBV2     | 7739371        | 18672361      | 11.24%          | 1.937%        |
| 18        | 6.571       | 1706          | 1720        | 1742         | rVB3     | 902166         | 2416262       | 1.45%           | 0.251%        |
| 19        | 6.722       | 1743          | 1767        | 1777         | rBV2     | 1705868        | 5240201       | 3.16%           | 0.544%        |
| 20        | 7.088       | 1869          | 1881        | 1898         | rVB3     | 1396993        | 2884091       | 1.74%           | 0.299%        |
| 21        | 7.185       | 1898          | 1911        | 1918         | rBV2     | 1186525        | 2410369       | 1.45%           | 0.250%        |
| 22        | 7.236       | 1919          | 1927        | 1939         | rVB      | 1754302        | 3334176       | 2.01%           | 0.346%        |
| 23        | 7.307       | 1939          | 1949        | 1961         | rBV      | 3931275        | 6736526       | 4.06%           | 0.699%        |
| 24        | 8.095       | 2176          | 2194        | 2203         | rBV2     | 1078253        | 2298444       | 1.38%           | 0.238%        |
| 25        | 8.776       | 2397          | 2406        | 2416         | rBV      | 1711551        | 2707532       | 1.63%           | 0.281%        |
| 26        | 8.937       | 2444          | 2456        | 2479         | rVB2     | 43703650       | 72858250      | 43.87%          | 7.559%        |
| 27        | 9.072       | 2481          | 2498        | 2512         | rVV3     | 86999967       | 166081247     | 100.00%         | 17.231%       |
| 28        | 9.468       | 2610          | 2621        | 2646         | rVB      | 12428326       | 19625417      | 11.82%          | 2.036%        |
| 29        | 9.857       | 2730          | 2742        | 2759         | rBV      | 5994216        | 9396139       | 5.66%           | 0.975%        |
| 30        | 10.001      | 2778          | 2787        | 2797         | rBV      | 1386700        | 2247513       | 1.35%           | 0.233%        |
| 31        | 10.278      | 2861          | 2873        | 2891         | rBV      | 14549636       | 21724884      | 13.08%          | 2.254%        |
| 32        | 10.374      | 2891          | 2903        | 2908         | rBV      | 21538720       | 35407876      | 21.32%          | 3.673%        |
| 33        | 10.464      | 2921          | 2931        | 2964         | rVB      | 22234958       | 33446589      | 20.14%          | 3.470%        |
| 34        | 10.660      | 2981          | 2992        | 3018         | rBV      | 10749077       | 16010670      | 9.64%           | 1.661%        |
| 35        | 10.844      | 3036          | 3049        | 3066         | rVB2     | 58234275       | 90419172      | 54.44%          | 9.381%        |
| 36        | 11.120      | 3126          | 3135        | 3143         | rBV      | 1426738        | 2059606       | 1.24%           | 0.214%        |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039154.D  
Acq On : 24 Sep 2025 15:53  
Operator : SY/MD  
Sample : Q3175-02  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Title : SW846 8260

|    |        |      |      |      |      |          |          |        |        |
|----|--------|------|------|------|------|----------|----------|--------|--------|
| 37 | 11.172 | 3143 | 3151 | 3166 | rVB2 | 2733046  | 4148042  | 2.50%  | 0.430% |
| 38 | 11.262 | 3167 | 3179 | 3194 | rBV  | 20641315 | 31078759 | 18.71% | 3.224% |
| 39 | 11.451 | 3226 | 3238 | 3249 | rBV  | 18174355 | 32743829 | 19.72% | 3.397% |
| 40 | 11.500 | 3250 | 3253 | 3261 | rVV  | 3966928  | 4764714  | 2.87%  | 0.494% |
| 41 | 11.567 | 3261 | 3274 | 3291 | rVB3 | 9099548  | 18251840 | 10.99% | 1.894% |
| 42 | 11.670 | 3296 | 3306 | 3317 | rVV  | 1351566  | 2118502  | 1.28%  | 0.220% |
| 43 | 11.744 | 3319 | 3329 | 3346 | rVB  | 2388890  | 3590737  | 2.16%  | 0.373% |
| 44 | 11.837 | 3347 | 3358 | 3363 | rBV  | 5643222  | 8518899  | 5.13%  | 0.884% |
| 45 | 11.866 | 3363 | 3367 | 3378 | rVV  | 4584244  | 6017807  | 3.62%  | 0.624% |
| 46 | 11.940 | 3379 | 3390 | 3410 | rVV  | 11137748 | 18945147 | 11.41% | 1.966% |
| 47 | 12.040 | 3410 | 3421 | 3453 | rVB2 | 5476630  | 11192756 | 6.74%  | 1.161% |
| 48 | 12.239 | 3472 | 3483 | 3495 | rVB2 | 2159748  | 3517549  | 2.12%  | 0.365% |
| 49 | 12.339 | 3495 | 3514 | 3522 | rBV  | 6565019  | 9901260  | 5.96%  | 1.027% |
| 50 | 12.390 | 3522 | 3530 | 3546 | rVB  | 7914579  | 11627651 | 7.00%  | 1.206% |
| 51 | 12.647 | 3601 | 3610 | 3626 | rVB  | 6436224  | 9744643  | 5.87%  | 1.011% |
| 52 | 12.805 | 3640 | 3659 | 3671 | rBV2 | 14429742 | 23293637 | 14.03% | 2.417% |
| 53 | 12.969 | 3702 | 3710 | 3732 | rVB4 | 3723837  | 6227533  | 3.75%  | 0.646% |
| 54 | 13.139 | 3755 | 3763 | 3772 | rVB  | 1441457  | 2084806  | 1.26%  | 0.216% |
| 55 | 13.194 | 3772 | 3780 | 3787 | rBV2 | 2042198  | 3085434  | 1.86%  | 0.320% |
| 56 | 13.422 | 3834 | 3851 | 3862 | rBV  | 13390847 | 20527847 | 12.36% | 2.130% |
| 57 | 13.631 | 3906 | 3916 | 3928 | rVV4 | 973922   | 1830962  | 1.10%  | 0.190% |
| 58 | 13.834 | 3969 | 3979 | 3994 | rBV  | 1279698  | 2004078  | 1.21%  | 0.208% |
| 59 | 14.027 | 4025 | 4039 | 4050 | rBV3 | 2412208  | 4756085  | 2.86%  | 0.493% |
| 60 | 14.352 | 4130 | 4140 | 4154 | rBV2 | 1281902  | 2113589  | 1.27%  | 0.219% |
| 61 | 14.551 | 4190 | 4202 | 4214 | rBV  | 6554015  | 10277203 | 6.19%  | 1.066% |
| 62 | 14.734 | 4250 | 4259 | 4273 | rVB  | 3150156  | 4819790  | 2.90%  | 0.500% |




Sum of corrected areas: 963878757

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039154.D  
Acq On : 24 Sep 2025 15:53  
Operator : SY/MD  
Sample : Q3175-02  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039154.D  
Acq On : 24 Sep 2025 15:53  
Operator : SY/MD  
Sample : Q3175-02  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260










TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*

Peak Number 1 Cyclopentane, 1,2-dimethyl-... Concentration Rank 14

| R.T.  | EstConc    | Area    | Relative to ISTD    | R.T.  |
|-------|------------|---------|---------------------|-------|
| 5.233 | 80.47 ug/l | 3563980 | 1,4-Difluorobenzene | 5.535 |

| Hit# | of | 5 | Tentative ID                        | MW | MolForm | CAS#        | Qual |
|------|----|---|-------------------------------------|----|---------|-------------|------|
| 1    |    |   | Cyclopentane, 1,2-dimethyl-, trans- | 98 | C7H14   | 000822-50-4 | 87   |
| 2    |    |   | Isopropylcyclobutane                | 98 | C7H14   | 000872-56-0 | 87   |
| 3    |    |   | Cyclopentane, 1,3-dimethyl-, cis-   | 98 | C7H14   | 002532-58-3 | 86   |
| 4    |    |   | Cyclobutanone, 3-ethyl-             | 98 | C6H10O  | 056335-73-0 | 64   |
| 5    |    |   | (Z)-2-Heptene                       | 98 | C7H14   | 006443-92-1 | 58   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039154.D  
Acq On : 24 Sep 2025 15:53  
Operator : SY/MD  
Sample : Q3175-02  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260










TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*

Peak Number 2 Cyclopentene, 1,5-dimethyl- Concentration Rank 11

| R.T.  | EstConc     | Area    | Relative to ISTD    | R.T.  |
|-------|-------------|---------|---------------------|-------|
| 6.722 | 118.31 ug/l | 5240200 | 1,4-Difluorobenzene | 5.535 |

| Hit# | of | 5 | Tentative ID                       | MW | MolForm | CAS#        | Qual |
|------|----|---|------------------------------------|----|---------|-------------|------|
| 1    |    |   | Cyclopentene, 1,5-dimethyl-        | 96 | C7H12   | 016491-15-9 | 90   |
| 2    |    |   | Cyclobutane, (1-methylethylidene)- | 96 | C7H12   | 001528-22-9 | 90   |
| 3    |    |   | 1,4-Pentadiene, 3,3-dimethyl-      | 96 | C7H12   | 001112-35-2 | 72   |
| 4    |    |   | 3,5-Dimethylcyclopentene           | 96 | C7H12   | 007459-71-4 | 64   |
| 5    |    |   | Cyclohexene, 4-methyl-             | 96 | C7H12   | 000591-47-9 | 64   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039154.D  
Acq On : 24 Sep 2025 15:53  
Operator : SY/MD  
Sample : Q3175-02  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260










TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*

Peak Number 3 Benzene, 1-ethyl-2-methyl- Concentration Rank 1

| R.T.   | EstConc     | Area     | Relative to ISTD       | R.T.   |
|--------|-------------|----------|------------------------|--------|
| 10.374 | 426.80 ug/l | 35407900 | 1,4-Dichlorobenzene-d4 | 11.175 |

| Hit# | of | 5 | Tentative ID               | MW  | MolForm | CAS#        | Qual |
|------|----|---|----------------------------|-----|---------|-------------|------|
| 1    |    |   | Benzene, 1-ethyl-2-methyl- | 120 | C9H12   | 000611-14-3 | 95   |
| 2    |    |   | Benzene, 1-ethyl-3-methyl- | 120 | C9H12   | 000620-14-4 | 95   |
| 3    |    |   | Benzene, 1,2,4-trimethyl-  | 120 | C9H12   | 000095-63-6 | 91   |
| 4    |    |   | Benzene, 1-ethyl-4-methyl- | 120 | C9H12   | 000622-96-8 | 91   |
| 5    |    |   | Benzene, 1,2,3-trimethyl-  | 120 | C9H12   | 000526-73-8 | 91   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039154.D  
Acq On : 24 Sep 2025 15:53  
Operator : SY/MD  
Sample : Q3175-02  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4










Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*  
Peak Number 4 Benzene, 1-ethyl-4-methyl- Concentration Rank 6

| R.T.   | EstConc     | Area     | Relative to ISTD       | R.T.   |
|--------|-------------|----------|------------------------|--------|
| 10.660 | 192.99 ug/l | 16010700 | 1,4-Dichlorobenzene-d4 | 11.175 |

| Hit# | of | 5 | Tentative ID               | MW  | MolForm | CAS#        | Qual |
|------|----|---|----------------------------|-----|---------|-------------|------|
| 1    |    |   | Benzene, 1-ethyl-2-methyl- | 120 | C9H12   | 000611-14-3 | 94   |
| 2    |    |   | Mesitylene                 | 120 | C9H12   | 000108-67-8 | 91   |
| 3    |    |   | Benzene, 1-ethyl-4-methyl- | 120 | C9H12   | 000622-96-8 | 91   |
| 4    |    |   | Benzene, 1-ethyl-3-methyl- | 120 | C9H12   | 000620-14-4 | 91   |
| 5    |    |   | Benzene, 1,2,3-trimethyl-  | 120 | C9H12   | 000526-73-8 | 91   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039154.D  
Acq On : 24 Sep 2025 15:53  
Operator : SY/MD  
Sample : Q3175-02  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4










Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*  
Peak Number 5 Benzene, 1,2,3-trimethyl- Concentration Rank 3

| R.T.   | EstConc     | Area     | Relative to ISTD       | R.T.   |
|--------|-------------|----------|------------------------|--------|
| 11.262 | 374.62 ug/l | 31078800 | 1,4-Dichlorobenzene-d4 | 11.175 |

| Hit# | of | 5 | Tentative ID               | MW  | MolForm | CAS#        | Qual |
|------|----|---|----------------------------|-----|---------|-------------|------|
| 1    |    |   | Mesitylene                 | 120 | C9H12   | 000108-67-8 | 97   |
| 2    |    |   | Benzene, 1,2,3-trimethyl-  | 120 | C9H12   | 000526-73-8 | 94   |
| 3    |    |   | Benzene, 1,2,4-trimethyl-  | 120 | C9H12   | 000095-63-6 | 94   |
| 4    |    |   | Benzene, 1-ethyl-3-methyl- | 120 | C9H12   | 000620-14-4 | 91   |
| 5    |    |   | Benzene, 1-ethyl-2-methyl- | 120 | C9H12   | 000611-14-3 | 90   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039154.D  
Acq On : 24 Sep 2025 15:53  
Operator : SY/MD  
Sample : Q3175-02  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4










Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*

Peak Number 6 Indane Concentration Rank 2

| R.T.      | EstConc                      | Area     | Relative to ISTD       | R.T.        |      |
|-----------|------------------------------|----------|------------------------|-------------|------|
| 11.451    | 394.69 ug/l                  | 32743800 | 1,4-Dichlorobenzene-d4 | 11.175      |      |
| Hit# of 5 | Tentative ID                 | MW       | MolForm                | CAS#        | Qual |
| 1         | Indane                       | 118      | C9H10                  | 000496-11-7 | 93   |
| 2         | Benzene, 1-ethenyl-2-methyl- | 118      | C9H10                  | 000611-15-4 | 87   |
| 3         | Benzene, 2-propenyl-         | 118      | C9H10                  | 000300-57-2 | 83   |
| 4         | Benzene, cyclopropyl-        | 118      | C9H10                  | 000873-49-4 | 80   |
| 5         | Deltacyclene                 | 118      | C9H10                  | 007785-10-6 | 72   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039154.D  
Acq On : 24 Sep 2025 15:53  
Operator : SY/MD  
Sample : Q3175-02  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 11 Sample Multiplier: 1

**Instrument :**  
MSVOA\_V  
**ClientSampleId :**  
MW4










Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*

|             |   |                                |               |      |    |
|-------------|---|--------------------------------|---------------|------|----|
| Peak Number | 7 | Benzene, 2-ethyl-1,4-dimethyl- | Concentration | Rank | 13 |
|-------------|---|--------------------------------|---------------|------|----|

| R.T.    | EstConc     | Area                                 | Relative to ISTD       |         |             | R.T.   |
|---------|-------------|--------------------------------------|------------------------|---------|-------------|--------|
| 11.837  | 102.69 ug/l | 8518900                              | 1,4-Dichlorobenzene-d4 |         |             | 11.175 |
| Hit# of | 5           | Tentative ID                         | MW                     | MolForm | CAS#        | Qual   |
| 1       |             | Benzene, 2-ethyl-1,4-dimethyl-       | 134                    | C10H14  | 001758-88-9 | 97     |
| 2       |             | o-Cymene                             | 134                    | C10H14  | 000527-84-4 | 95     |
| 3       |             | Benzene, 4-ethyl-1,2-dimethyl-       | 134                    | C10H14  | 000934-80-5 | 95     |
| 4       |             | Benzene, 1-ethyl-2,3-dimethyl-       | 134                    | C10H14  | 000933-98-2 | 95     |
| 5       |             | Benzene, 1-methyl-3-(1-methylethyl)- | 134                    | C10H14  | 000535-77-3 | 94     |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039154.D  
Acq On : 24 Sep 2025 15:53  
Operator : SY/MD  
Sample : Q3175-02  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260










TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*

Peak Number 8 Benzene, 4-ethyl-1,2-dimethyl- Concentration Rank 5

| R.T.   | EstConc     | Area     | Relative to ISTD       | R.T.   |
|--------|-------------|----------|------------------------|--------|
| 11.940 | 228.36 ug/l | 18945100 | 1,4-Dichlorobenzene-d4 | 11.175 |

| Hit# | of | 5 | Tentative ID                   | MW  | MolForm | CAS#        | Qual |
|------|----|---|--------------------------------|-----|---------|-------------|------|
| 1    |    |   | Benzene, 4-ethyl-1,2-dimethyl- | 134 | C10H14  | 000934-80-5 | 97   |
| 2    |    |   | Benzene, 2-ethyl-1,4-dimethyl- | 134 | C10H14  | 001758-88-9 | 96   |
| 3    |    |   | o-Cymene                       | 134 | C10H14  | 000527-84-4 | 95   |
| 4    |    |   | Benzene, 1-ethyl-2,3-dimethyl- | 134 | C10H14  | 000933-98-2 | 95   |
| 5    |    |   | Benzene, 1-ethyl-2,4-dimethyl- | 134 | C10H14  | 000874-41-9 | 94   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039154.D  
Acq On : 24 Sep 2025 15:53  
Operator : SY/MD  
Sample : Q3175-02  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4










Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*  
Peak Number 9 Benzene, 1-ethenyl-3-ethyl- Concentration Rank 8

| R.T.   | EstConc     | Area     | Relative to ISTD       | R.T.   |
|--------|-------------|----------|------------------------|--------|
| 12.040 | 134.92 ug/l | 11192800 | 1,4-Dichlorobenzene-d4 | 11.175 |

| Hit# | of | 5 | Tentative ID                    | MW  | MolForm | CAS#        | Qual |
|------|----|---|---------------------------------|-----|---------|-------------|------|
| 1    |    |   | Benzene, 1-ethenyl-3-ethyl-     | 132 | C10H12  | 007525-62-4 | 91   |
| 2    |    |   | Indan, 1-methyl-                | 132 | C10H12  | 000767-58-8 | 90   |
| 3    |    |   | Benzene, 1-ethenyl-4-ethyl-     | 132 | C10H12  | 003454-07-7 | 90   |
| 4    |    |   | Benzene, (2-methyl-1-propenyl)- | 132 | C10H12  | 000768-49-0 | 87   |
| 5    |    |   | 1-Methyl-2-phenylcyclopropane   | 132 | C10H12  | 003145-76-4 | 87   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039154.D  
Acq On : 24 Sep 2025 15:53  
Operator : SY/MD  
Sample : Q3175-02  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260










TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*

Peak Number 10 Benzene, 1,2,3,4-tetramethyl- Concentration Rank 10

| R.T.   | EstConc     | Area    | Relative to ISTD       | R.T.   |
|--------|-------------|---------|------------------------|--------|
| 12.339 | 119.35 ug/l | 9901260 | 1,4-Dichlorobenzene-d4 | 11.175 |

| Hit# | of | 5 | Tentative ID                   | MW  | MolForm | CAS#        | Qual |
|------|----|---|--------------------------------|-----|---------|-------------|------|
| 1    |    |   | Benzene, 1,2,3,4-tetramethyl-  | 134 | C10H14  | 000488-23-3 | 97   |
| 2    |    |   | Benzene, 1,2,4,5-tetramethyl-  | 134 | C10H14  | 000095-93-2 | 96   |
| 3    |    |   | Benzene, 1,2,3,5-tetramethyl-  | 134 | C10H14  | 000527-53-7 | 96   |
| 4    |    |   | o-Cymene                       | 134 | C10H14  | 000527-84-4 | 95   |
| 5    |    |   | Benzene, 1-ethyl-2,4-dimethyl- | 134 | C10H14  | 000874-41-9 | 94   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039154.D  
Acq On : 24 Sep 2025 15:53  
Operator : SY/MD  
Sample : Q3175-02  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P










\*\*\*\*\*

Peak Number 11 Benzene, 1,2,4,5-tetramethyl- Concentration Rank 7

| R.T.   | EstConc     | Area     | Relative to ISTD       | R.T.   |
|--------|-------------|----------|------------------------|--------|
| 12.390 | 140.16 ug/l | 11627700 | 1,4-Dichlorobenzene-d4 | 11.175 |

| Hit# | of 5 | Tentative ID                   | MW  | MolForm | CAS#        | Qual |
|------|------|--------------------------------|-----|---------|-------------|------|
| 1    |      | Benzene, 1,2,3,4-tetramethyl-  | 134 | C10H14  | 000488-23-3 | 97   |
| 2    |      | Benzene, 1,2,4,5-tetramethyl-  | 134 | C10H14  | 000095-93-2 | 97   |
| 3    |      | Benzene, 1,2,3,5-tetramethyl-  | 134 | C10H14  | 000527-53-7 | 96   |
| 4    |      | o-Cymene                       | 134 | C10H14  | 000527-84-4 | 95   |
| 5    |      | Benzene, 1-ethyl-2,4-dimethyl- | 134 | C10H14  | 000874-41-9 | 94   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039154.D  
Acq On : 24 Sep 2025 15:53  
Operator : SY/MD  
Sample : Q3175-02  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260










TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*

Peak Number 12 Indan, 1-methyl- Concentration Rank 12

| R.T.   | EstConc     | Area    | Relative to ISTD       | R.T.   |
|--------|-------------|---------|------------------------|--------|
| 12.648 | 117.46 ug/l | 9744640 | 1,4-Dichlorobenzene-d4 | 11.175 |

| Hit# | of | 5 | Tentative ID                     | MW  | MolForm | CAS#        | Qual |
|------|----|---|----------------------------------|-----|---------|-------------|------|
| 1    |    |   | Indan, 1-methyl-                 | 132 | C10H12  | 000767-58-8 | 93   |
| 2    |    |   | Benzene, 2-ethenyl-1,4-dimethyl- | 132 | C10H12  | 002039-89-6 | 92   |
| 3    |    |   | 1H-Indene, 2,3-dihydro-5-methyl- | 132 | C10H12  | 000874-35-1 | 91   |
| 4    |    |   | 1-Phenyl-1-butene                | 132 | C10H12  | 000824-90-8 | 91   |
| 5    |    |   | 3-Phenylbut-1-ene                | 132 | C10H12  | 000934-10-1 | 91   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039154.D  
Acq On : 24 Sep 2025 15:53  
Operator : SY/MD  
Sample : Q3175-02  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260










TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*

Peak Number 13 Benzene, 2-ethenyl-1,4-dime... Concentration Rank 4

| R.T.   | EstConc     | Area     | Relative to ISTD       | R.T.   |
|--------|-------------|----------|------------------------|--------|
| 12.805 | 280.78 ug/l | 23293600 | 1,4-Dichlorobenzene-d4 | 11.175 |

| Hit# | of | 5 | Tentative ID                        | MW  | MolForm | CAS#        | Qual |
|------|----|---|-------------------------------------|-----|---------|-------------|------|
| 1    |    |   | Benzene, 2-ethenyl-1,4-dimethyl-    | 132 | C10H12  | 002039-89-6 | 95   |
| 2    |    |   | Benzene, (1-methyl-1-propenyl)-,... | 132 | C10H12  | 000768-00-3 | 89   |
| 3    |    |   | Benzene, 2-butenyl-                 | 132 | C10H12  | 001560-06-1 | 76   |
| 4    |    |   | 1-Phenyl-1-butene                   | 132 | C10H12  | 000824-90-8 | 76   |
| 5    |    |   | Benzene, 1-ethenyl-4-ethyl-         | 132 | C10H12  | 003454-07-7 | 70   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039154.D  
Acq On : 24 Sep 2025 15:53  
Operator : SY/MD  
Sample : Q3175-02  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260










TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*

Peak Number 14 Naphthalene, 1,2,3,4-tetrahydro... Concentration Rank 15

| R.T.   | EstConc    | Area    | Relative to ISTD       | R.T.   |
|--------|------------|---------|------------------------|--------|
| 12.969 | 75.07 ug/l | 6227530 | 1,4-Dichlorobenzene-d4 | 11.175 |

| Hit# | of | 5 | Tentative ID                        | MW  | MolForm  | CAS#        | Qual |
|------|----|---|-------------------------------------|-----|----------|-------------|------|
| 1    |    |   | Naphthalene, 1,2,3,4-tetrahydro-    | 132 | C10H12   | 000119-64-2 | 95   |
| 2    |    |   | Benzene, cyclobutyl-                | 132 | C10H12   | 004392-30-7 | 50   |
| 3    |    |   | Indan, epoxide                      | 132 | C9H8O    | 000768-22-9 | 46   |
| 4    |    |   | 2H-Inden-2-one, 1,3-dihydro-        | 132 | C9H8O    | 000615-13-4 | 43   |
| 5    |    |   | 2-Propenoic acid, 2-phenylethyl ... | 176 | C11H12O2 | 003530-36-7 | 27   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039154.D  
Acq On : 24 Sep 2025 15:53  
Operator : SY/MD  
Sample : Q3175-02  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4










Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

\*\*\*\*\*  
Peak Number 15 Naphthalene, 2-methyl- Concentration Rank 9

| R.T.   | EstConc     | Area     | Relative to ISTD       | R.T.   |
|--------|-------------|----------|------------------------|--------|
| 14.551 | 123.88 ug/l | 10277200 | 1,4-Dichlorobenzene-d4 | 11.175 |

| Hit# | of | 5 | Tentative ID                        | MW  | MolForm | CAS#        | Qual |
|------|----|---|-------------------------------------|-----|---------|-------------|------|
| 1    |    |   | Naphthalene, 2-methyl-              | 142 | C11H10  | 000091-57-6 | 96   |
| 2    |    |   | Naphthalene, 1-methyl-              | 142 | C11H10  | 000090-12-0 | 96   |
| 3    |    |   | Benzocycloheptatriene               | 142 | C11H10  | 000264-09-5 | 91   |
| 4    |    |   | 1,4-Methanonaphthalene, 1,4-dihy... | 142 | C11H10  | 004453-90-1 | 91   |
| 5    |    |   | 1H-Indene, 1-ethylidene-            | 142 | C11H10  | 002471-83-2 | 64   |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039154.D  
Acq On : 24 Sep 2025 15:53  
Operator : SY/MD  
Sample : Q3175-02  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST0.L  
TIC Integration Parameters: LSCINT.P

| TIC Top Hit name   | RT     | EstConc | Units | Response | --Internal Standard-- |        |         |      |
|--------------------|--------|---------|-------|----------|-----------------------|--------|---------|------|
|                    |        |         |       |          | #                     | RT     | Resp    | Conc |
| Cyclopentane, 1... | 5.233  | 80.5    | ug/l  | 3563980  | 2                     | 5.535  | 2214590 | 50.0 |
| Cyclopentene, 1... | 6.722  | 118.3   | ug/l  | 5240200  | 2                     | 5.535  | 2214590 | 50.0 |
| Benzene, 1-ethy... | 10.374 | 426.8   | ug/l  | 35407900 | 4                     | 11.175 | 4148040 | 50.0 |
| Benzene, 1-ethy... | 10.660 | 193.0   | ug/l  | 16010700 | 4                     | 11.175 | 4148040 | 50.0 |
| Benzene, 1,2,3-... | 11.262 | 374.6   | ug/l  | 31078800 | 4                     | 11.175 | 4148040 | 50.0 |
| Indane             | 11.451 | 394.7   | ug/l  | 32743800 | 4                     | 11.175 | 4148040 | 50.0 |
| Benzene, 2-ethy... | 11.837 | 102.7   | ug/l  | 8518900  | 4                     | 11.175 | 4148040 | 50.0 |
| Benzene, 4-ethy... | 11.940 | 228.4   | ug/l  | 18945100 | 4                     | 11.175 | 4148040 | 50.0 |
| Benzene, 1-ethe... | 12.040 | 134.9   | ug/l  | 11192800 | 4                     | 11.175 | 4148040 | 50.0 |
| Benzene, 1,2,3,... | 12.339 | 119.3   | ug/l  | 9901260  | 4                     | 11.175 | 4148040 | 50.0 |
| Benzene, 1,2,4,... | 12.390 | 140.2   | ug/l  | 11627700 | 4                     | 11.175 | 4148040 | 50.0 |
| Indan, 1-methyl-   | 12.648 | 117.5   | ug/l  | 9744640  | 4                     | 11.175 | 4148040 | 50.0 |
| Benzene, 2-ethe... | 12.805 | 280.8   | ug/l  | 23293600 | 4                     | 11.175 | 4148040 | 50.0 |
| Naphthalene, 1,... | 12.969 | 75.1    | ug/l  | 6227530  | 4                     | 11.175 | 4148040 | 50.0 |
| Naphthalene, 2-... | 14.551 | 123.9   | ug/l  | 10277200 | 4                     | 11.175 | 4148040 | 50.0 |

## Report of Analysis

|                    |                            |                 |              |
|--------------------|----------------------------|-----------------|--------------|
| Client:            | G Environmental            | Date Collected: | 09/23/25     |
| Project:           | Summer                     | Date Received:  | 09/23/25     |
| Client Sample ID:  | MW4DL                      | SDG No.:        | Q3175        |
| Lab Sample ID:     | Q3175-02DL                 | Matrix:         | Water        |
| Analytical Method: | 8260D                      | % Solid:        | 0            |
| Sample Wt/Vol:     | 5      Units:    mL        | Final Vol:      | 5000      uL |
| Soil Aliquot Vol:  | uL                         | Test:           | VOCMS Group1 |
| GC Column:         | DB-624UI      ID :    0.18 | Level :         | LOW          |
| Prep Method :      |                            |                 |              |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039175.D        | 20        | 09/25/25 12:55 | VV092525      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL  | LOQ / CRQL | Units |
|----------------|--------------------------------|-------|-----------|------|------------|-------|
| <b>TARGETS</b> |                                |       |           |      |            |       |
| 75-71-8        | Dichlorodifluoromethane        | 4.40  | UD        | 4.40 | 20.0       | ug/L  |
| 74-87-3        | Chloromethane                  | 6.40  | UD        | 6.40 | 20.0       | ug/L  |
| 75-01-4        | Vinyl Chloride                 | 5.20  | UD        | 5.20 | 20.0       | ug/L  |
| 74-83-9        | Bromomethane                   | 28.8  | UD        | 28.8 | 100        | ug/L  |
| 75-00-3        | Chloroethane                   | 9.40  | UD        | 9.40 | 20.0       | ug/L  |
| 75-69-4        | Trichlorofluoromethane         | 6.60  | UD        | 6.60 | 20.0       | ug/L  |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 5.00  | UD        | 5.00 | 20.0       | ug/L  |
| 75-65-0        | Tert butyl alcohol             | 110   | UD        | 110  | 500        | ug/L  |
| 75-35-4        | 1,1-Dichloroethene             | 4.60  | UD        | 4.60 | 20.0       | ug/L  |
| 67-64-1        | Acetone                        | 30.2  | UD        | 30.2 | 100        | ug/L  |
| 75-15-0        | Carbon Disulfide               | 4.20  | UD        | 4.20 | 20.0       | ug/L  |
| 1634-04-4      | Methyl tert-butyl Ether        | 3.20  | UD        | 3.20 | 20.0       | ug/L  |
| 79-20-9        | Methyl Acetate                 | 5.40  | UD        | 5.40 | 20.0       | ug/L  |
| 75-09-2        | Methylene Chloride             | 5.60  | UD        | 5.60 | 20.0       | ug/L  |
| 156-60-5       | trans-1,2-Dichloroethene       | 4.60  | UD        | 4.60 | 20.0       | ug/L  |
| 75-34-3        | 1,1-Dichloroethane             | 4.60  | UD        | 4.60 | 20.0       | ug/L  |
| 110-82-7       | Cyclohexane                    | 120   | D         | 29.0 | 100        | ug/L  |
| 78-93-3        | 2-Butanone                     | 19.6  | UD        | 19.6 | 100        | ug/L  |
| 56-23-5        | Carbon Tetrachloride           | 5.00  | UD        | 5.00 | 20.0       | ug/L  |
| 156-59-2       | cis-1,2-Dichloroethene         | 3.80  | UD        | 3.80 | 20.0       | ug/L  |
| 74-97-5        | Bromochloromethane             | 4.40  | UD        | 4.40 | 20.0       | ug/L  |
| 67-66-3        | Chloroform                     | 5.00  | UD        | 5.00 | 20.0       | ug/L  |
| 71-55-6        | 1,1,1-Trichloroethane          | 4.00  | UD        | 4.00 | 20.0       | ug/L  |
| 108-87-2       | Methylcyclohexane              | 140   | D         | 3.20 | 20.0       | ug/L  |
| 71-43-2        | Benzene                        | 1100  | D         | 3.00 | 20.0       | ug/L  |
| 107-06-2       | 1,2-Dichloroethane             | 4.40  | UD        | 4.40 | 20.0       | ug/L  |
| 79-01-6        | Trichloroethene                | 1.90  | UD        | 1.90 | 20.0       | ug/L  |
| 78-87-5        | 1,2-Dichloropropane            | 4.00  | UD        | 4.00 | 20.0       | ug/L  |
| 75-27-4        | Bromodichloromethane           | 4.40  | UD        | 4.40 | 20.0       | ug/L  |
| 108-10-1       | 4-Methyl-2-Pentanone           | 13.6  | UD        | 13.6 | 100        | ug/L  |

### Report of Analysis

|                    |                 |                 |              |
|--------------------|-----------------|-----------------|--------------|
| Client:            | G Environmental | Date Collected: | 09/23/25     |
| Project:           | Summer          | Date Received:  | 09/23/25     |
| Client Sample ID:  | MW4DL           | SDG No.:        | Q3175        |
| Lab Sample ID:     | Q3175-02DL      | Matrix:         | Water        |
| Analytical Method: | 8260D           | % Solid:        | 0            |
| Sample Wt/Vol:     | 5               | Units:          | mL           |
| Soil Aliquot Vol:  |                 |                 | uL           |
| GC Column:         | DB-624UI        | ID :            | 0.18         |
| Prep Method :      |                 | Level :         | LOW          |
|                    |                 | Final Vol:      | 5000         |
|                    |                 | Test:           | VOCMS Group1 |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039175.D        | 20        | 09/25/25 12:55 | VV092525      |

| CAS Number                | Parameter                   | Conc.   | Qualifier | MDL      | LOQ / CRQL | Units   |
|---------------------------|-----------------------------|---------|-----------|----------|------------|---------|
| 108-88-3                  | Toluene                     | 48.3    | D         | 2.80     | 20.0       | ug/L    |
| 10061-02-6                | t-1,3-Dichloropropene       | 3.40    | UD        | 3.40     | 20.0       | ug/L    |
| 10061-01-5                | cis-1,3-Dichloropropene     | 3.20    | UD        | 3.20     | 20.0       | ug/L    |
| 79-00-5                   | 1,1,2-Trichloroethane       | 4.20    | UD        | 4.20     | 20.0       | ug/L    |
| 591-78-6                  | 2-Hexanone                  | 17.8    | UD        | 17.8     | 100        | ug/L    |
| 124-48-1                  | Dibromochloromethane        | 3.60    | UD        | 3.60     | 20.0       | ug/L    |
| 106-93-4                  | 1,2-Dibromoethane           | 3.00    | UD        | 3.00     | 20.0       | ug/L    |
| 127-18-4                  | Tetrachloroethene           | 4.60    | UD        | 4.60     | 20.0       | ug/L    |
| 108-90-7                  | Chlorobenzene               | 2.40    | UD        | 2.40     | 20.0       | ug/L    |
| 100-41-4                  | Ethyl Benzene               | 530     | D         | 2.60     | 20.0       | ug/L    |
| 179601-23-1               | m/p-Xylenes                 | 1600    | D         | 4.80     | 40.0       | ug/L    |
| 95-47-6                   | o-Xylene                    | 120     | D         | 2.40     | 20.0       | ug/L    |
| 100-42-5                  | Styrene                     | 3.00    | UD        | 3.00     | 20.0       | ug/L    |
| 75-25-2                   | Bromoform                   | 3.80    | UD        | 3.80     | 20.0       | ug/L    |
| 98-82-8                   | Isopropylbenzene            | 51.6    | D         | 2.40     | 20.0       | ug/L    |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 5.20    | UD        | 5.20     | 20.0       | ug/L    |
| 541-73-1                  | 1,3-Dichlorobenzene         | 3.20    | UD        | 3.20     | 20.0       | ug/L    |
| 106-46-7                  | 1,4-Dichlorobenzene         | 3.80    | UD        | 3.80     | 20.0       | ug/L    |
| 95-50-1                   | 1,2-Dichlorobenzene         | 3.20    | UD        | 3.20     | 20.0       | ug/L    |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 10.6    | UD        | 10.6     | 20.0       | ug/L    |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 4.00    | UD        | 4.00     | 20.0       | ug/L    |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 4.00    | UD        | 4.00     | 20.0       | ug/L    |
| <b>SURROGATES</b>         |                             |         |           |          |            |         |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 54.6    |           | 74 - 125 | 109%       | SPK: 50 |
| 1868-53-7                 | Dibromofluoromethane        | 48.8    |           | 75 - 124 | 98%        | SPK: 50 |
| 2037-26-5                 | Toluene-d8                  | 47.6    |           | 86 - 113 | 95%        | SPK: 50 |
| 460-00-4                  | 4-Bromofluorobenzene        | 48.5    |           | 77 - 121 | 97%        | SPK: 50 |
| <b>INTERNAL STANDARDS</b> |                             |         |           |          |            |         |
| 363-72-4                  | Pentafluorobenzene          | 561000  | 4.6       |          |            |         |
| 540-36-3                  | 1,4-Difluorobenzene         | 1010000 | 5.535     |          |            |         |
| 3114-55-4                 | Chlorobenzene-d5            | 987000  | 8.776     |          |            |         |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 548000  | 11.172    |          |            |         |

## Report of Analysis

|                    |                 |           |                 |              |    |
|--------------------|-----------------|-----------|-----------------|--------------|----|
| Client:            | G Environmental |           | Date Collected: | 09/23/25     |    |
| Project:           | Summer          |           | Date Received:  | 09/23/25     |    |
| Client Sample ID:  | MW4DL           |           | SDG No.:        | Q3175        |    |
| Lab Sample ID:     | Q3175-02DL      |           | Matrix:         | Water        |    |
| Analytical Method: | 8260D           |           | % Solid:        | 0            |    |
| Sample Wt/Vol:     | 5               | Units: mL | Final Vol:      | 5000         | uL |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOCMS Group1 |    |
| GC Column:         | DB-624UI        | ID : 0.18 | Level :         | LOW          |    |
| Prep Method :      |                 |           |                 |              |    |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039175.D        | 20        | 09/25/25 12:55 | VV092525      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
 Data File : VV039175.D  
 Acq On : 25 Sep 2025 12:55  
 Operator : SY/MD  
 Sample : Q3175-02DL 20X  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 MW4DL

Manual Integrations  
 APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025  
 Supervised By :Semsettin Yesilyurt 09/29/2025

Quant Time: Sep 26 01:25:29 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

| Compound                    | R.T.           | QIon | Response            | Conc   | Units  | Dev(Min) |
|-----------------------------|----------------|------|---------------------|--------|--------|----------|
| -----                       |                |      |                     |        |        |          |
| Internal Standards          |                |      |                     |        |        |          |
| 1) Pentafluorobenzene       | 4.600          | 168  | 560555              | 50.000 | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 5.535          | 114  | 1013068             | 50.000 | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 8.776          | 117  | 987441              | 50.000 | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 11.172         | 152  | 548334              | 50.000 | ug/l   | 0.00     |
|                             |                |      |                     |        |        |          |
| System Monitoring Compounds |                |      |                     |        |        |          |
| 33) 1,2-Dichloroethane-d4   | 4.944          | 65   | 442857              | 54.587 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 81 - 118 |      | Recovery = 109.180% |        |        |          |
| 35) Dibromofluoromethane    | 4.503          | 113  | 397640              | 48.825 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 80 - 119 |      | Recovery = 97.640%  |        |        |          |
| 50) Toluene-d8              | 7.236          | 98   | 1138170             | 47.583 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 89 - 112 |      | Recovery = 95.160%  |        |        |          |
| 62) 4-Bromofluorobenzene    | 10.001         | 95   | 477707              | 48.512 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 85 - 114 |      | Recovery = 97.020%  |        |        |          |
|                             |                |      |                     |        |        |          |
| Target Compounds            |                |      |                     |        |        |          |
|                             |                |      |                     |        | Qvalue |          |
| 31) Cyclohexane             | 4.590          | 56   | 88730               | 6.132  | ug/l # | 85       |
| 39) Methylcyclohexane       | 6.050          | 83   | 106117              | 7.136  | ug/l   | 96       |
| 40) Benzene                 | 5.011          | 78   | 2259910             | 56.362 | ug/l   | 99       |
| 52) Toluene                 | 7.320          | 92   | 56174               | 2.416  | ug/l   | 94       |
| 67) Ethyl Benzene           | 8.937          | 91   | 1097863             | 26.369 | ug/l   | 100      |
| 68) m/p-Xylenes             | 9.063          | 106  | 1299190             | 80.850 | ug/l   | 100      |
| 69) o-Xylene                | 9.471          | 106  | 82647               | 5.809  | ug/l   | 99       |
| 73) Isopropylbenzene        | 9.860          | 105  | 102681              | 2.582  | ug/l   | 99       |
| 78) n-propylbenzene         | 10.281         | 91   | 321661              | 6.642  | ug/l   | 99       |
| 80) 1,3,5-Trimethylbenzene  | 10.464         | 105  | 387119              | 11.711 | ug/l   | 99       |
| 84) 1,2,4-Trimethylbenzene  | 10.841         | 105  | 1210229             | 37.710 | ug/l   | 99       |
| 85) sec-Butylbenzene        | 11.017         | 105  | 25782m              | 0.590  | ug/l   |          |
| 86) p-Isopropyltoluene      | 11.172         | 119  | 14766m              | 0.406  | ug/l   |          |
| 89) n-Butylbenzene          | 11.574         | 91   | 38955m              | 1.121  | ug/l   |          |
| 95) Naphthalene             | 13.429         | 128  | 248885              | 6.885  | ug/l   | 100      |
| -----                       |                |      |                     |        |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
Data File : VV039175.D  
Acq On : 25 Sep 2025 12:55  
Operator : SY/MD  
Sample : Q3175-02DL 20X  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 9 Sample Multiplier: 1


Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4DL

Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025

Quant Time: Sep 26 01:25:29 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration









#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 4.600 min Scan# 1107  
Delta R.T. 0.003 min  
Lab File: VV039175.D  
Acq: 25 Sep 2025 12:55

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4DL

Tgt Ion:168 Resp: 56055  
Ion Ratio Lower Upper  
168 100  
99 63.4 49.6 74.4




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025



#31  
Cyclohexane  
Concen: 6.132 ug/l  
RT: 4.590 min Scan# 1104  
Delta R.T. 0.010 min  
Lab File: VV039175.D  
Acq: 25 Sep 2025 12:55

Tgt Ion: 56 Resp: 88730  
Ion Ratio Lower Upper  
56 100  
69 52.9 26.2 39.2#  
84 95.1 70.3 105.5





#33

1,2-Dichloroethane-d4

Concen: 54.587 ug/l

RT: 4.944 min Scan# 1214

Delta R.T. 0.006 min

Lab File: VV039175.D

Acq: 25 Sep 2025 12:55

Instrument :

MSVOA\_V

ClientSampleId :

MW4DL

Tgt Ion: 65 Resp: 44285

Ion Ratio Lower Upper





65 100

67 51.6 0.0 107.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025  
 Supervised By :Semsettin Yesilyurt 09/29/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 5.535 min Scan# 1398

Delta R.T. 0.006 min

Lab File: VV039175.D

Acq: 25 Sep 2025 12:55




Tgt Ion:114 Resp: 1013068


Ion Ratio Lower Upper

114 100

63 20.6 0.0 42.6

88 15.9 0.0 30.8





#35

Dibromofluoromethane

Concen: 48.825 ug/l

RT: 4.503 min Scan# 1075

Delta R.T. 0.006 min

Lab File: VV039175.D

Acq: 25 Sep 2025 12:55

Instrument :

MSVOA\_V

ClientSampleId :

MW4DL

Tgt Ion:113 Resp: 397640

Ion Ratio Lower Upper

113 100





111 101.8 82.4 123.6

192 17.2 13.8 20.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025  
 Supervised By :Semsettin Yesilyurt 09/29/2025



#39

Methylcyclohexane

Concen: 7.136 ug/l

RT: 6.050 min Scan# 1558

Delta R.T. 0.007 min

Lab File: VV039175.D

Acq: 25 Sep 2025 12:55




Tgt Ion: 83 Resp: 106117


Ion Ratio Lower Upper

83 100

55 78.9 60.5 90.7

98 50.2 37.8 56.6









#40  
Benzene  
Concen: 56.362 ug/l  
RT: 5.011 min Scan# 1235  
Delta R.T. 0.006 min  
Lab File: VV039175.D  
Acq: 25 Sep 2025 12:55

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4DL

Tgt Ion: 78 Resp: 2259910  
Ion Ratio Lower Upper  
78 100  
77 24.1 18.7 28.1




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025



#50  
Toluene-d8  
Concen: 47.583 ug/l  
RT: 7.236 min Scan# 1927  
Delta R.T. 0.003 min  
Lab File: VV039175.D  
Acq: 25 Sep 2025 12:55

Tgt Ion: 98 Resp: 1138170  
Ion Ratio Lower Upper  
98 100  
100 64.2 52.2 78.2





#52

Toluene

Concen: 2.416 ug/l

RT: 7.320 min Scan# 1948

Delta R.T. 0.016 min

Lab File: VV039175.D

Acq: 25 Sep 2025 12:55

Instrument :

MSVOA\_V

ClientSampleId :

MW4DL

Tgt Ion: 92 Resp: 56174

Ion Ratio Lower Upper





92 100

91 166.1 139.2 208.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025



#62

4-Bromofluorobenzene

Concen: 48.512 ug/l

RT: 10.001 min Scan# 2787

Delta R.T. 0.003 min

Lab File: VV039175.D

Acq: 25 Sep 2025 12:55




Tgt Ion: 95 Resp: 477707


Ion Ratio Lower Upper

95 100

174 75.4 0.0 150.4

176 70.8 0.0 144.2





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 8.776 min Scan# 2405

Delta R.T. 0.003 min

Lab File: VV039175.D

Acq: 25 Sep 2025 12:55

Instrument :

MSVOA\_V

ClientSampleId :

MW4DL

Tgt Ion:117 Resp: 98744

Ion Ratio Lower Upper





117 100

82 56.6 45.4 68.2

119 32.8 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025

#67

Ethyl Benzene

Concen: 26.369 ug/l

RT: 8.937 min Scan# 2456

Delta R.T. 0.003 min

Lab File: VV039175.D




Acq: 25 Sep 2025 12:55


Tgt Ion: 91 Resp: 1097863

Ion Ratio Lower Upper

91 100


106 30.9 24.6 37.0





#68  
m/p-Xylenes  
Concen: 80.850 ug/l  
RT: 9.063 min Scan# 2495  
Delta R.T. 0.004 min  
Lab File: VV039175.D  
Acq: 25 Sep 2025 12:55




Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4DL



Tgt Ion:106 Resp: 1299190  
Ion Ratio Lower Upper  
106 100  
91 202.8 162.5 243.7




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025



#69  
o-Xylene  
Concen: 5.809 ug/l  
RT: 9.471 min Scan# 2622  
Delta R.T. 0.003 min  
Lab File: VV039175.D  
Acq: 25 Sep 2025 12:55


Tgt Ion:106 Resp: 82647  
Ion Ratio Lower Upper  
106 100  
91 216.7 109.1 327.3





#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 11.172 min Scan# 3151  
Delta R.T. -0.000 min  
Lab File: VV039175.D  
Acq: 25 Sep 2025 12:55




Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4DL




Tgt Ion:152 Resp: 548334  
Ion Ratio Lower Upper  
152 100  
115 60.5 44.8 134.4  
150 156.2 0.0 353.8

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025




#73  
Isopropylbenzene  
Concen: 2.582 ug/l  
RT: 9.860 min Scan# 2743  
Delta R.T. 0.006 min  
Lab File: VV039175.D  
Acq: 25 Sep 2025 12:55



Tgt Ion:105 Resp: 102681  
Ion Ratio Lower Upper  
105 100  
120 25.5 13.1 39.1









#78  
n-propylbenzene  
Concen: 6.642 ug/l  
RT: 10.281 min Scan# 2874  
Delta R.T. 0.006 min  
Lab File: VV039175.D  
Acq: 25 Sep 2025 12:55

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4DL

Tgt Ion: 91 Resp: 32166  
Ion Ratio Lower Upper  
91 100  
120 21.8 11.1 33.3




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025



#80  
1,3,5-Trimethylbenzene  
Concen: 11.711 ug/l  
RT: 10.464 min Scan# 2931  
Delta R.T. -0.000 min  
Lab File: VV039175.D  
Acq: 25 Sep 2025 12:55


Tgt Ion:105 Resp: 387119  
Ion Ratio Lower Upper  
105 100  
120 49.0 24.3 72.8





#84  
1,2,4-Trimethylbenzene  
Concen: 37.710 ug/l  
RT: 10.841 min Scan# 3047  
Delta R.T. 0.003 min  
Lab File: VV039175.D  
Acq: 25 Sep 2025 12:55




Instrument :  
MSVOA\_V  
ClientSampleId :  
MW4DL



Tgt Ion:105 Resp: 1210229  
Ion Ratio Lower Upper  
105 100  
120 45.3 22.4 67.2




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025



#85  
sec-Butylbenzene  
Concen: 0.590 ug/l m  
RT: 11.017 min Scan# 3103  
Delta R.T. 0.006 min  
Lab File: VV039175.D  
Acq: 25 Sep 2025 12:55

Tgt Ion:105 Resp: 25782  
Ion Ratio Lower Upper  
105 100  
134 0.0 9.6 28.8#





#86

p-Isopropyltoluene

Concen: 0.406 ug/l m

RT: 11.172 min Scan# 3149

Delta R.T. 0.006 min

Lab File: VV039175.D

Acq: 25 Sep 2025 12:55

Instrument :

MSVOA\_V

ClientSampleId :

MW4DL

Tgt Ion:119 Resp: 14760

Ion Ratio Lower Upper





119 100

134 39.6 12.7 38.0

91 28.2 12.0 36.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025  
Supervised By :Semsettin Yesilyurt 09/29/2025

#89

n-Butylbenzene

Concen: 1.121 ug/l m

RT: 11.574 min Scan# 3276

Delta R.T. -0.003 min

Lab File: VV039175.D

Acq: 25 Sep 2025 12:55




Tgt Ion: 91 Resp: 38955


Ion Ratio Lower Upper

91 100

92 47.7 26.6 79.8

134 149.8 12.1 36.3#





#95

Naphthalene

Concen: 6.885 ug/l

RT: 13.429 min Scan# 3853

Delta R.T. 0.006 min

Lab File: VV039175.D


Acq: 25 Sep 2025 12:55

Instrument :

MSVOA\_V

ClientSampleId :

MW4DL



Tgt Ion:128 Resp: 248885

Ion Ratio Lower Upper

128 100

127 12.9 10.3 15.5



129 10.8 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025

Supervised By :Semsettin Yesilyurt 09/29/2025





# CALIBRATION SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: GENV01  
 Lab Code: ACE SDG No.: Q3175  
 Instrument ID: MSVOA\_V Calibration Date(s): 09/22/2025 09/22/2025  
 Heated Purge: (Y/N) N Calibration Time(s): 12:26 16:35  
 GC Column: DB-624UI ID: 0.18 (mm)

| LAB FILE ID:                   |        | RRF010 = VV039137.D |        | RRF020 = VV039138.D |        | RRF050 = VV039139.D |       |       |  |
|--------------------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|-------|--|
|                                |        | RRF100 = VV039140.D |        | RRF005 = VV039142.D |        | RRF001 = VV039143.D |       |       |  |
| COMPOUND                       | RRF010 | RRF020              | RRF050 | RRF100              | RRF005 | RRF001              | RRF   | % RSD |  |
| Dichlorodifluoromethane        | 0.848  | 0.757               | 0.740  | 0.699               | 0.772  | 0.934               | 0.792 | 10.8  |  |
| Chloromethane                  | 0.949  | 0.814               | 0.770  | 0.729               | 0.889  | 0.946               | 0.850 | 10.9  |  |
| Vinyl Chloride                 | 1.007  | 0.856               | 0.836  | 0.801               | 0.932  | 1.026               | 0.910 | 10.3  |  |
| Bromomethane                   | 0.631  | 0.548               | 0.508  | 0.524               | 0.655  |                     | 0.573 | 11.5  |  |
| Chloroethane                   | 0.662  | 0.568               | 0.533  | 0.500               | 0.606  | 0.915               | 0.631 | 23.8  |  |
| Trichlorofluoromethane         | 1.394  | 1.253               | 1.193  | 1.151               | 1.365  | 1.546               | 1.317 | 11.2  |  |
| 1,1,2-Trichlorotrifluoroethane | 0.839  | 0.732               | 0.702  | 0.686               | 0.804  | 0.939               | 0.784 | 12.3  |  |
| Tert butyl alcohol             | 0.169  | 0.157               | 0.137  | 0.136               | 0.166  |                     | 0.153 | 10.3  |  |
| 1,1-Dichloroethene             | 0.804  | 0.721               | 0.674  | 0.663               | 0.789  | 0.873               | 0.754 | 10.9  |  |
| Acetone                        | 0.537  | 0.458               | 0.449  | 0.414               | 0.569  | 0.712               | 0.523 | 20.8  |  |
| Carbon Disulfide               | 2.453  | 2.181               | 2.081  | 1.986               | 2.462  | 2.670               | 2.306 | 11.4  |  |
| Methyl tert-butyl Ether        | 2.547  | 2.285               | 2.233  | 2.202               | 2.556  | 2.591               | 2.402 | 7.5   |  |
| Methyl Acetate                 | 1.146  | 1.036               | 0.983  | 0.973               | 1.097  | 1.198               | 1.072 | 8.5   |  |
| Methylene Chloride             | 0.930  | 0.848               | 0.779  | 0.742               | 1.113  | 1.987               | 1.066 | 44    |  |
| trans-1,2-Dichloroethene       | 0.846  | 0.755               | 0.724  | 0.696               | 0.835  | 0.959               | 0.803 | 12.1  |  |
| 1,1-Dichloroethane             | 1.669  | 1.466               | 1.388  | 1.329               | 1.659  | 1.942               | 1.576 | 14.4  |  |
| Cyclohexane                    | 1.325  | 1.222               | 1.245  | 1.254               | 1.407  |                     | 1.291 | 5.9   |  |
| 2-Butanone                     | 0.604  | 0.584               | 0.586  | 0.580               | 0.570  | 0.554               | 0.579 | 2.9   |  |
| Carbon Tetrachloride           | 0.742  | 0.710               | 0.703  | 0.675               | 0.736  | 0.890               | 0.743 | 10.2  |  |
| cis-1,2-Dichloroethene         | 0.908  | 0.842               | 0.868  | 0.860               | 0.832  | 1.001               | 0.885 | 7.1   |  |
| Bromochloromethane             | 0.836  | 0.753               | 0.679  | 0.637               | 0.525  | 0.743               | 0.696 | 15.5  |  |
| Chloroform                     | 1.752  | 1.570               | 1.490  | 1.433               | 1.746  | 1.880               | 1.645 | 10.6  |  |
| 1,1,1-Trichloroethane          | 1.562  | 1.321               | 1.298  | 1.263               | 1.513  | 1.590               | 1.424 | 10.3  |  |
| Methylcyclohexane              | 0.701  | 0.720               | 0.823  | 0.857               | 0.680  | 0.622               | 0.734 | 12.2  |  |
| Benzene                        | 2.077  | 1.978               | 1.990  | 1.931               | 1.963  | 1.936               | 1.979 | 2.7   |  |
| 1,2-Dichloroethane             | 0.725  | 0.694               | 0.666  | 0.643               | 0.693  | 0.740               | 0.693 | 5.2   |  |
| Trichloroethene                | 0.471  | 0.456               | 0.463  | 0.454               | 0.460  | 0.435               | 0.457 | 2.6   |  |
| 1,2-Dichloropropane            | 0.509  | 0.519               | 0.502  | 0.486               | 0.471  | 0.465               | 0.492 | 4.4   |  |
| Bromodichloromethane           | 0.832  | 0.764               | 0.753  | 0.725               | 0.767  | 0.829               | 0.778 | 5.5   |  |
| 4-Methyl-2-Pentanone           | 0.715  | 0.736               | 0.732  | 0.714               | 0.613  | 0.459               | 0.661 | 16.5  |  |

\* Compounds with required minimum RRF and maximum %RSD values.  
 All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: GENV01  
 Lab Code: ACE SDG No.: Q3175  
 Instrument ID: MSVOA\_V Calibration Date(s): 09/22/2025 09/22/2025  
 Heated Purge: (Y/N) N Calibration Time(s): 12:26 16:35  
 GC Column: DB-624UI ID: 0.18 (mm)

| LAB FILE ID:                |        | RRF010 = VV039137.D |        | RRF020 = VV039138.D |        | RRF050 = VV039139.D |       |       |  |
|-----------------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|-------|--|
|                             |        | RRF100 = VV039140.D |        | RRF005 = VV039142.D |        | RRF001 = VV039143.D |       |       |  |
| COMPOUND                    | RRF010 | RRF020              | RRF050 | RRF100              | RRF005 | RRF001              | RRF   | % RSD |  |
| Toluene                     | 1.211  | 1.199               | 1.229  | 1.214               | 1.130  | 0.901               | 1.147 | 10.9  |  |
| t-1,3-Dichloropropene       | 0.701  | 0.696               | 0.728  | 0.753               | 0.645  | 0.548               | 0.678 | 10.8  |  |
| cis-1,3-Dichloropropene     | 0.803  | 0.777               | 0.811  | 0.803               | 0.704  | 0.595               | 0.749 | 11.4  |  |
| 1,1,2-Trichloroethane       | 0.534  | 0.502               | 0.494  | 0.480               | 0.524  | 0.553               | 0.514 | 5.3   |  |
| 2-Hexanone                  | 0.505  | 0.534               | 0.546  | 0.533               | 0.429  | 0.231               | 0.463 | 26.2  |  |
| Dibromochloromethane        | 0.605  | 0.570               | 0.567  | 0.558               | 0.598  | 0.620               | 0.586 | 4.2   |  |
| 1,2-Dibromoethane           | 0.556  | 0.538               | 0.529  | 0.514               | 0.535  | 0.484               | 0.526 | 4.7   |  |
| Tetrachloroethene           | 0.435  | 0.412               | 0.412  | 0.407               | 0.431  | 0.415               | 0.419 | 2.7   |  |
| Chlorobenzene               | 1.403  | 1.368               | 1.383  | 1.358               | 1.417  | 1.428               | 1.393 | 2     |  |
| Ethyl Benzene               | 2.028  | 2.143               | 2.382  | 2.418               | 1.953  | 1.725               | 2.108 | 12.5  |  |
| m/p-Xylenes                 | 0.820  | 0.882               | 0.949  | 0.938               | 0.713  | 0.580               | 0.814 | 17.7  |  |
| o-Xylene                    | 0.703  | 0.742               | 0.846  | 0.880               | 0.635  | 0.518               | 0.720 | 18.7  |  |
| Styrene                     | 1.239  | 1.381               | 1.553  | 1.537               | 1.038  | 0.828               | 1.263 | 22.8  |  |
| Bromoform                   | 0.396  | 0.384               | 0.396  | 0.400               | 0.385  | 0.397               | 0.393 | 1.7   |  |
| Isopropylbenzene            | 3.564  | 3.638               | 3.939  | 4.096               | 3.296  | 3.226               | 3.627 | 9.5   |  |
| 1,1,2,2-Tetrachloroethane   | 1.600  | 1.461               | 1.425  | 1.389               | 1.684  | 1.975               | 1.589 | 13.8  |  |
| 1,3-Dichlorobenzene         | 1.931  | 1.836               | 1.857  | 1.882               | 1.943  | 1.975               | 1.904 | 2.8   |  |
| 1,4-Dichlorobenzene         | 2.070  | 1.951               | 1.934  | 1.919               | 2.139  | 2.522               | 2.089 | 11    |  |
| 1,2-Dichlorobenzene         | 1.751  | 1.652               | 1.738  | 1.802               | 1.797  | 1.935               | 1.779 | 5.3   |  |
| 1,2-Dibromo-3-Chloropropane | 0.299  | 0.283               | 0.285  | 0.302               | 0.333  | 0.476               | 0.330 | 22.4  |  |
| 1,2,4-Trichlorobenzene      | 0.928  | 0.963               | 1.076  | 1.189               | 0.935  | 1.028               | 1.020 | 9.9   |  |
| 1,2,3-Trichlorobenzene      | 0.974  | 0.973               | 1.107  | 1.191               | 0.901  | 1.113               | 1.043 | 10.6  |  |
| 1,2-Dichloroethane-d4       | 0.773  | 0.663               | 0.646  | 0.618               | 0.919  |                     | 0.724 | 17.1  |  |
| Dibromofluoromethane        | 0.435  | 0.382               | 0.377  | 0.360               | 0.454  |                     | 0.402 | 10.1  |  |
| Toluene-d8                  | 1.166  | 1.092               | 1.136  | 1.097               | 1.411  |                     | 1.181 | 11.2  |  |
| 4-Bromofluorobenzene        | 0.481  | 0.464               | 0.507  | 0.501               | 0.477  |                     | 0.486 | 3.6   |  |

\* Compounds with required minimum RRF and maximum %RSD values.  
 All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Method Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\  
 Method File : 82V092225W.M  
 Title : SW846 8260  
 Last Update : Wed Sep 24 08:08:08 2025  
 Response Via : Initial Calibration

## Calibration Files

1 =VV039143.D 5 =VV039142.D 10 =VV039137.D 20 =VV039138.D 50 =VV039139.D 100 =VV039140.D

|        | Compound            | 1              | 5     | 10    | 20    | 50    | 100   | Avg   | %RSD   |
|--------|---------------------|----------------|-------|-------|-------|-------|-------|-------|--------|
| 1) I   | Pentafluorobenzene  | -----ISTD----- |       |       |       |       |       |       |        |
| 2) T   | Dichlorodifluo...   | 0.934          | 0.772 | 0.848 | 0.757 | 0.740 | 0.699 | 0.791 | 10.78  |
| 3) P   | Chloromethane       | 0.946          | 0.889 | 0.949 | 0.814 | 0.770 | 0.729 | 0.850 | 10.90  |
| 4) C   | Vinyl Chloride      | 1.026          | 0.932 | 1.007 | 0.856 | 0.836 | 0.801 | 0.910 | 10.27# |
| 5) T   | Bromomethane        |                | 0.655 | 0.631 | 0.548 | 0.508 | 0.524 | 0.573 | 11.50  |
| 6) T   | Chloroethane        | 0.915          | 0.606 | 0.662 | 0.568 | 0.533 | 0.500 | 0.631 | 23.83  |
| 7) T   | Trichlorofluor...   | 1.546          | 1.365 | 1.394 | 1.253 | 1.193 | 1.151 | 1.317 | 11.15  |
| 8) T   | Diethyl Ether       | 0.572          | 0.513 | 0.505 | 0.461 | 0.447 | 0.435 | 0.489 | 10.50  |
| 9) T   | 1,1,2-Trichlor...   | 0.939          | 0.804 | 0.839 | 0.732 | 0.702 | 0.686 | 0.784 | 12.28  |
| 10) T  | Methyl Iodide       |                | 0.638 | 0.772 | 0.742 | 0.786 | 0.793 | 0.746 | 8.50   |
| 11) T  | Tert butyl alc...   |                | 0.166 | 0.169 | 0.157 | 0.137 | 0.136 | 0.153 | 10.26  |
| 12) CM | 1,1-Dichloroet...   | 0.873          | 0.789 | 0.804 | 0.721 | 0.674 | 0.663 | 0.754 | 10.87# |
| 13) T  | Acrolein            |                | 0.025 | 0.027 | 0.027 | 0.026 | 0.027 | 0.027 | 4.01   |
| 14) T  | Allyl chloride      | 1.523          | 1.434 | 1.388 | 1.283 | 1.243 | 1.208 | 1.347 | 9.06   |
| 15) T  | Acrylonitrile       | 0.587          | 0.522 | 0.507 | 0.462 | 0.436 | 0.427 | 0.490 | 12.38  |
| 16) T  | Acetone             | 0.712          | 0.569 | 0.537 | 0.458 | 0.449 | 0.414 | 0.523 | 20.84  |
| 17) T  | Carbon Disulfide    | 2.670          | 2.462 | 2.453 | 2.181 | 2.081 | 1.986 | 2.306 | 11.43  |
| 18) T  | Methyl Acetate      | 1.198          | 1.097 | 1.146 | 1.036 | 0.983 | 0.973 | 1.072 | 8.46   |
| 19) T  | Methyl tert-bu...   | 2.591          | 2.556 | 2.547 | 2.285 | 2.233 | 2.202 | 2.402 | 7.50   |
| 20) T  | Methylene Chlo...   | 1.987          | 1.113 | 0.930 | 0.848 | 0.779 | 0.742 | 1.066 | 44.05  |
| 21) T  | trans-1,2-Dich...   | 0.959          | 0.835 | 0.846 | 0.755 | 0.724 | 0.696 | 0.803 | 12.11  |
| 22) T  | Diisopropyl ether   | 2.659          | 2.560 | 2.757 | 2.530 | 2.520 | 2.433 | 2.576 | 4.44   |
| 23) T  | Vinyl Acetate       | 2.243          | 2.313 | 2.393 | 2.241 | 2.204 | 2.128 | 2.254 | 4.04   |
| 24) P  | 1,1-Dichloroet...   | 1.942          | 1.659 | 1.669 | 1.466 | 1.388 | 1.329 | 1.576 | 14.42  |
| 25) T  | 2-Butanone          | 0.554          | 0.570 | 0.604 | 0.584 | 0.586 | 0.580 | 0.579 | 2.90   |
| 26) T  | 2,2-Dichloropr...   | 1.906          | 1.393 | 1.364 | 1.275 | 1.239 | 1.204 | 1.397 | 18.57  |
| 27) T  | cis-1,2-Dichlo...   | 1.001          | 0.832 | 0.908 | 0.842 | 0.868 | 0.860 | 0.885 | 7.06   |
| 28) T  | Bromochloromet...   | 0.743          | 0.525 | 0.836 | 0.753 | 0.679 | 0.637 | 0.696 | 15.49  |
| 29) T  | Tetrahydrofuran     | 0.335          | 0.350 | 0.376 | 0.372 | 0.369 | 0.374 | 0.363 | 4.57   |
| 30) C  | Chloroform          | 1.880          | 1.746 | 1.752 | 1.570 | 1.490 | 1.433 | 1.645 | 10.60# |
| 31) T  | Cyclohexane         |                | 1.407 | 1.325 | 1.222 | 1.245 | 1.254 | 1.291 | 5.85   |
| 32) T  | 1,1,1-Trichlor...   | 1.590          | 1.513 | 1.562 | 1.321 | 1.298 | 1.263 | 1.424 | 10.26  |
| 33) S  | 1,2-Dichloroet...   |                | 0.919 | 0.773 | 0.663 | 0.646 | 0.618 | 0.724 | 17.13  |
| 34) I  | 1,4-Difluorobenzene | -----ISTD----- |       |       |       |       |       |       |        |
| 35) S  | Dibromofluorom...   |                | 0.454 | 0.435 | 0.382 | 0.377 | 0.360 | 0.402 | 10.10  |
| 36) T  | 1,1-Dichloropr...   | 0.650          | 0.536 | 0.624 | 0.604 | 0.620 | 0.612 | 0.608 | 6.33   |
| 37) T  | Ethyl Acetate       | 0.825          | 0.722 | 0.715 | 0.740 | 0.714 | 0.710 | 0.737 | 5.96   |
| 38) T  | Carbon Tetrach...   | 0.890          | 0.736 | 0.742 | 0.710 | 0.703 | 0.675 | 0.743 | 10.25  |
| 39) T  | Methylcyclohexane   | 0.622          | 0.680 | 0.701 | 0.720 | 0.823 | 0.857 | 0.734 | 12.15  |
| 40) TM | Benzene             | 1.936          | 1.963 | 2.077 | 1.978 | 1.990 | 1.931 | 1.979 | 2.69   |
| 41) T  | Methacrylonitrile   | 0.185          | 0.215 | 0.273 | 0.292 | 0.330 | 0.342 | 0.273 | 22.92  |
| 42) TM | 1,2-Dichloroet...   | 0.740          | 0.693 | 0.725 | 0.694 | 0.666 | 0.643 | 0.693 | 5.20   |
| 43) T  | Isopropyl Acetate   | 0.869          | 0.933 | 1.032 | 1.029 | 1.069 | 1.100 | 1.005 | 8.71   |
| 44) TM | Trichloroethene     | 0.435          | 0.460 | 0.471 | 0.456 | 0.463 | 0.454 | 0.457 | 2.64   |
| 45) C  | 1,2-Dichloropr...   | 0.465          | 0.471 | 0.509 | 0.519 | 0.502 | 0.486 | 0.492 | 4.40#  |
| 46) T  | Dibromomethane      | 0.396          | 0.381 | 0.392 | 0.367 | 0.368 | 0.354 | 0.376 | 4.40   |
| 47) T  | Bromodichlorom...   | 0.829          | 0.767 | 0.832 | 0.764 | 0.753 | 0.725 | 0.778 | 5.54   |
| 48) T  | Methyl methacr...   | 0.491          | 0.356 | 0.409 | 0.449 | 0.504 | 0.526 | 0.456 | 14.15  |
| 49) T  | 1,4-Dioxane         | 0.009          | 0.005 | 0.007 | 0.007 | 0.007 | 0.008 | 0.007 | 17.72  |
| 50) S  | Toluene-d8          |                | 1.411 | 1.166 | 1.092 | 1.136 | 1.097 | 1.181 | 11.21  |
| 51) T  | 4-Methyl-2-Pen...   | 0.459          | 0.613 | 0.715 | 0.736 | 0.732 | 0.714 | 0.661 | 16.48  |
| 52) CM | Toluene             | 0.901          | 1.130 | 1.211 | 1.199 | 1.229 | 1.214 | 1.147 | 10.93# |
| 53) T  | t-1,3-Dichloro...   | 0.548          | 0.645 | 0.701 | 0.696 | 0.728 | 0.753 | 0.678 | 10.83  |
| 54) T  | cis-1,3-Dichlo...   | 0.595          | 0.704 | 0.803 | 0.777 | 0.811 | 0.803 | 0.749 | 11.37  |
| 55) T  | 1,1,2-Trichlor...   | 0.553          | 0.524 | 0.534 | 0.502 | 0.494 | 0.480 | 0.514 | 5.31   |
| 56) T  | Ethyl methacry...   | 0.507          | 0.510 | 0.569 | 0.603 | 0.716 | 0.761 | 0.611 | 17.43  |


Method Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\  
 Method File : 82V092225W.M

|     |    |                       |                |       |       |       |       |       |       |        |
|-----|----|-----------------------|----------------|-------|-------|-------|-------|-------|-------|--------|
| 57) | T  | 1,3-Dichloropr...     | 0.797          | 0.764 | 0.861 | 0.827 | 0.820 | 0.802 | 0.812 | 4.02   |
| 58) | T  | 2-Chloroethyl ...     | 0.136          | 0.196 | 0.264 | 0.304 | 0.353 | 0.351 | 0.267 | 32.62  |
| 59) | T  | 2-Hexanone            | 0.231          | 0.429 | 0.505 | 0.534 | 0.546 | 0.533 | 0.463 | 26.18  |
| 60) | T  | Dibromochlorom...     | 0.620          | 0.598 | 0.605 | 0.570 | 0.567 | 0.558 | 0.586 | 4.20   |
| 61) | T  | 1,2-Dibromoethane     | 0.484          | 0.535 | 0.556 | 0.538 | 0.529 | 0.514 | 0.526 | 4.70   |
| 62) | S  | 4-Bromofluorob...     | 0.477          | 0.481 | 0.464 | 0.507 | 0.501 | 0.486 |       | 3.62   |
| 63) | I  | Chlorobenzene-d5      | -----ISTD----- |       |       |       |       |       |       |        |
| 64) | T  | Tetrachloroethene     | 0.415          | 0.431 | 0.435 | 0.412 | 0.412 | 0.407 | 0.419 | 2.74   |
| 65) | PM | Chlorobenzene         | 1.428          | 1.417 | 1.403 | 1.368 | 1.383 | 1.358 | 1.393 | 1.99   |
| 66) | T  | 1,1,1,2-Tetrac...     | 0.543          | 0.504 | 0.520 | 0.488 | 0.488 | 0.477 | 0.503 | 4.84   |
| 67) | C  | Ethyl Benzene         | 1.725          | 1.953 | 2.028 | 2.143 | 2.382 | 2.418 | 2.108 | 12.53# |
| 68) | T  | m/p-Xylenes           | 0.580          | 0.713 | 0.820 | 0.882 | 0.949 | 0.938 | 0.814 | 17.68  |
| 69) | T  | o-Xylene              | 0.518          | 0.635 | 0.703 | 0.742 | 0.846 | 0.880 | 0.720 | 18.67  |
| 70) | T  | Styrene               | 0.828          | 1.038 | 1.239 | 1.381 | 1.553 | 1.537 | 1.263 | 22.78  |
| 71) | P  | Bromoform             | 0.397          | 0.385 | 0.396 | 0.384 | 0.396 | 0.400 | 0.393 | 1.68   |
| 72) | I  | 1,4-Dichlorobenzen... | -----ISTD----- |       |       |       |       |       |       |        |
| 73) | T  | Isopropylbenzene      | 3.226          | 3.296 | 3.564 | 3.638 | 3.939 | 4.096 | 3.627 | 9.49   |
| 74) | T  | N-amyl acetate        | 1.031          | 1.279 | 1.425 | 1.506 | 1.673 | 1.764 | 1.447 | 18.47  |
| 75) | P  | 1,1,2,2-Tetrac...     | 1.975          | 1.684 | 1.600 | 1.461 | 1.425 | 1.389 | 1.589 | 13.83  |
| 76) | T  | 1,2,3-Trichlor...     | 1.375          | 1.201 | 1.136 | 1.048 | 1.049 | 1.023 | 1.138 | 11.73  |
| 77) | T  | Bromobenzene          | 0.978          | 0.927 | 0.948 | 0.919 | 0.925 | 0.949 | 0.941 | 2.34   |
| 78) | T  | n-propylbenzene       | 3.873          | 3.967 | 4.330 | 4.442 | 4.899 | 4.986 | 4.416 | 10.44  |
| 79) | T  | 2-Chlorotoluene       | 2.114          | 2.585 | 2.752 | 2.688 | 2.833 | 2.883 | 2.643 | 10.59  |
| 80) | T  | 1,3,5-Trimethy...     | 2.462          | 2.697 | 2.924 | 3.141 | 3.387 | 3.475 | 3.014 | 13.12  |
| 81) | T  | trans-1,4-Dich...     | 0.427          | 0.478 | 0.451 | 0.495 | 0.521 | 0.474 |       | 7.77   |
| 82) | T  | 4-Chlorotoluene       | 2.381          | 2.877 | 3.160 | 3.216 | 3.338 | 3.413 | 3.064 | 12.48  |
| 83) | T  | tert-Butylbenzene     | 2.875          | 2.757 | 2.886 | 2.881 | 3.222 | 3.413 | 3.006 | 8.43   |
| 84) | T  | 1,2,4-Trimethy...     | 2.332          | 2.542 | 2.767 | 3.100 | 3.378 | 3.440 | 2.926 | 15.47  |
| 85) | T  | sec-Butylbenzene      | 3.605          | 3.608 | 3.823 | 3.969 | 4.364 | 4.550 | 3.987 | 9.88   |
| 86) | T  | p-Isopropyltol...     | 2.874          | 2.994 | 3.237 | 3.351 | 3.682 | 3.772 | 3.318 | 10.85  |
| 87) | T  | 1,3-Dichlorobe...     | 1.975          | 1.943 | 1.931 | 1.836 | 1.857 | 1.882 | 1.904 | 2.83   |
| 88) | T  | 1,4-Dichlorobe...     | 2.522          | 2.139 | 2.070 | 1.951 | 1.934 | 1.919 | 2.089 | 10.97  |
| 89) | T  | n-Butylbenzene        | 2.646          | 2.816 | 3.034 | 3.141 | 3.613 | 3.763 | 3.169 | 13.88  |
| 90) | T  | Hexachloroethane      | 1.044          | 0.841 | 0.808 | 0.740 | 0.728 | 0.742 | 0.817 | 14.63  |
| 91) | T  | 1,2-Dichlorobe...     | 1.935          | 1.797 | 1.751 | 1.652 | 1.738 | 1.802 | 1.779 | 5.26   |
| 92) | T  | 1,2-Dibromo-3-...     | 0.476          | 0.333 | 0.299 | 0.283 | 0.285 | 0.302 | 0.330 | 22.39  |
| 93) | T  | 1,2,4-Trichlor...     | 1.028          | 0.935 | 0.928 | 0.963 | 1.076 | 1.189 | 1.020 | 9.87   |
| 94) | T  | Hexachlorobuta...     | 0.683          | 0.530 | 0.512 | 0.474 | 0.480 | 0.501 | 0.530 | 14.65  |
| 95) | T  | Naphthalene           | 3.054          | 2.665 | 2.888 | 3.114 | 3.814 | 4.244 | 3.296 | 18.32  |
| 96) | T  | 1,2,3-Trichlor...     | 1.113          | 0.901 | 0.974 | 0.973 | 1.107 | 1.191 | 1.043 | 10.58  |

(#) = Out of Range

## 2-Hexanone

Response Ratio



$$\text{Response} = 5.390\text{e-}001 * \text{Amt} - 2.507\text{e-}002$$


Coef of Det ( $r^2$ ) = 0.999719 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA V\Method\82V092225W.M

Calibration Table Last Updated: Wed Sep 24 08:08:08 2025

# 2-Chloroethyl Vinyl ether

Response Ratio




$$\text{Response} = 3.588\text{e-}001 * \text{Amt} - 6.905\text{e-}002$$

Coef of Det (r^2) = 0.999314 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA V\Method\82V092225W.M


Calibration Table Last Updated: Wed Sep 24 08:08:08 2025

## Methacrylonitrile



Response =  $3.467 \times 10^{-1} \times \text{Amt} - 1.325 \times 10^{-2}$   
Coef of Det ( $r^2$ ) = 0.999425    Curve Fit: Linear  
Method Name: Z:\voasrv\HPCHEM1\MSVOA V\Method\82V092225W.M  
Calibration Table Last Updated: Wed Sep 24 08:08:08 2025


# 1,2-Dibromo-3-Chloropropane



$R = 1.931e-002 A^2 + 2.606e-001 A + 5.917e-003$   
 Coef of Det ( $r^2$ ) = 0.999981    Curve Fit: Quadratic  
 Method Name:    Z:\voasrv\HPCHEM1\MSVOA V\Method\82V092225W.M  
 Calibration Table Last Updated: Wed Sep 24 08:08:08 2025

# Styrene


Response Ratio



$R = -1.124e-002 A^2 + 1.587e+000 A - 4.955e-002$   
 Coef of Det ( $r^2$ ) = 0.999540    Curve Fit: Quadratic  
 Method Name:    Z:\voasrv\HPCHEM1\MSVOA V\Method\82V092225W.M  
 Calibration Table Last Updated: Wed Sep 24 08:08:08 2025

# Methylene Chloride


Response Ratio



$R = -2.570e-002 A^2 + 7.777e-001 A + 2.991e-002$   
 Coef of Det ( $r^2$ ) = 0.999960    Curve Fit: Quadratic  
 Method Name: Z:\voasrv\HPCHEM1\MSVOA V\Method\82V092225W.M  
 Calibration Table Last Updated: Wed Sep 24 08:08:08 2025

# Acetone


Response Ratio



$R = -5.738e-003 A^2 + 4.676e-001 A + 4.104e-002$   
 Coef of Det ( $r^2$ ) = 0.999745    Curve Fit: Quadratic  
 Method Name:    Z:\voasrv\HPCHEM1\MSVOA V\Method\82V092225W.M  
 Calibration Table Last Updated: Wed Sep 24 08:08:08 2025

# Chloroethane

Response Ratio



$R = -3.006e-002 A^2 + 5.547e-001 A + 1.084e-002$   
 Coef of Det ( $r^2$ ) = 0.999734    Curve Fit: Quadratic  
 Method Name:    Z:\voasrv\HPCHEM1\MSVOA V\Method\82V092225W.M  
 Calibration Table Last Updated: Wed Sep 24 08:08:08 2025

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
 Data File : VV039137.D  
 Acq On : 22 Sep 2025 12:26  
 Operator : SY/MD  
 Sample : VSTDICC010  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VSTDICC010

Quant Time: Sep 24 07:21:49 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 07:14:23 2025  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 4.597  | 168  | 501722   | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 5.529  | 114  | 837272   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 8.773  | 117  | 824583   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 11.172 | 152  | 463438   | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |          |
|-----------------------------|--------|-------|----------|----------|------|----------|
| System Monitoring Compounds |        |       |          |          |      |          |
| 33) 1,2-Dichloroethane-d4   | 4.944  | 65    | 77610    | 10.688   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 81 - 118 | Recovery | =    | 21.380%# |
| 35) Dibromofluoromethane    | 4.503  | 113   | 72914    | 10.833   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 80 - 119 | Recovery | =    | 21.660%# |
| 50) Toluene-d8              | 7.240  | 98    | 195300   | 9.879    | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 89 - 112 | Recovery | =    | 19.760%# |
| 62) 4-Bromofluorobenzene    | 10.005 | 95    | 80587    | 9.902    | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 85 - 114 | Recovery | =    | 19.800%# |

|                              |       |     |         |        |        |     |
|------------------------------|-------|-----|---------|--------|--------|-----|
| Target Compounds             |       |     |         | Qvalue |        |     |
| 2) Dichlorodifluoromethane   | 1.121 | 85  | 85113   | 10.716 | ug/l   | 99  |
| 3) Chloromethane             | 1.230 | 50  | 95215   | 11.167 | ug/l   | 95  |
| 4) Vinyl Chloride            | 1.298 | 62  | 101023  | 11.068 | ug/l   | 99  |
| 5) Bromomethane              | 1.500 | 94  | 63359   | 11.011 | ug/l   | 98  |
| 6) Chloroethane              | 1.561 | 64  | 66412   | 11.087 | ug/l   | 98  |
| 7) Trichlorofluoromethane    | 1.729 | 101 | 139927  | 10.588 | ug/l   | 98  |
| 8) Diethyl Ether             | 1.918 | 74  | 50669   | 10.328 | ug/l   | 99  |
| 9) 1,1,2-Trichlorotrifluo... | 2.082 | 101 | 84221   | 10.711 | ug/l   | 99  |
| 10) Methyl Iodide            | 2.198 | 142 | 77448   | 10.343 | ug/l   | 98  |
| 11) Tert butyl alcohol       | 2.568 | 59  | 84609   | 55.130 | ug/l   | 99  |
| 12) 1,1-Dichloroethene       | 2.082 | 96  | 80704   | 10.669 | ug/l   | 97  |
| 13) Acrolein                 | 2.015 | 56  | 13557   | 50.908 | ug/l # | 80  |
| 14) Allyl chloride           | 2.359 | 41  | 139249  | 10.305 | ug/l   | 98  |
| 15) Acrylonitrile            | 2.674 | 53  | 254235  | 51.676 | ug/l   | 99  |
| 16) Acetone                  | 2.121 | 43  | 269573  | 53.773 | ug/l   | 99  |
| 17) Carbon Disulfide         | 2.253 | 76  | 246125  | 10.639 | ug/l   | 99  |
| 18) Methyl Acetate           | 2.381 | 43  | 115005  | 10.689 | ug/l   | 97  |
| 19) Methyl tert-butyl Ether  | 2.712 | 73  | 255535  | 10.601 | ug/l   | 99  |
| 20) Methylene Chloride       | 2.455 | 84  | 93350   | 10.106 | ug/l   | 99  |
| 21) trans-1,2-Dichloroethene | 2.703 | 96  | 84891   | 10.541 | ug/l   | 95  |
| 22) Diisopropyl ether        | 3.214 | 45  | 276604  | 10.699 | ug/l # | 68  |
| 23) Vinyl Acetate            | 3.191 | 43  | 1200618 | 53.087 | ug/l   | 100 |
| 24) 1,1-Dichloroethane       | 3.118 | 63  | 167496  | 10.594 | ug/l   | 100 |
| 25) 2-Butanone               | 3.873 | 43  | 303016  | 53.015 | ug/l   | 100 |
| 26) 2,2-Dichloropropane      | 3.809 | 77  | 136919  | 9.806  | ug/l   | 100 |
| 27) cis-1,2-Dichloroethene   | 3.822 | 96  | 91069   | 10.254 | ug/l   | 99  |
| 28) Bromochloromethane       | 4.146 | 49  | 83889   | 12.019 | ug/l   | 98  |
| 29) Tetrahydrofuran          | 4.240 | 42  | 188818  | 51.879 | ug/l   | 96  |
| 30) Chloroform               | 4.278 | 83  | 175847  | 10.652 | ug/l   | 97  |
| 31) Cyclohexane              | 4.584 | 56  | 132996  | 10.269 | ug/l   | 96  |
| 32) 1,1,1-Trichloroethane    | 4.513 | 97  | 156733  | 10.966 | ug/l   | 96  |
| 36) 1,1-Dichloropropene      | 4.744 | 75  | 104420  | 10.187 | ug/l   | 99  |
| 37) Ethyl Acetate            | 3.982 | 43  | 119710  | 9.520  | ug/l # | 91  |
| 38) Carbon Tetrachloride     | 4.732 | 117 | 124321  | 10.069 | ug/l   | 98  |
| 39) Methylcyclohexane        | 6.043 | 83  | 117354  | 9.549  | ug/l   | 94  |
| 40) Benzene                  | 5.008 | 78  | 347728  | 10.493 | ug/l   | 100 |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
 Data File : VV039137.D  
 Acq On : 22 Sep 2025 12:26  
 Operator : SY/MD  
 Sample : VSTDIC010  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VSTDIC010

Quant Time: Sep 24 07:21:49 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 07:14:23 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 4.169  | 41   | 45674    | 9.764   | ug/l   | 90       |
| 42) 1,2-Dichloroethane        | 5.043  | 62   | 121437   | 10.458  | ug/l   | 100      |
| 43) Isopropyl Acetate         | 5.195  | 43   | 172880   | 9.980   | ug/l   | 97       |
| 44) Trichloroethene           | 5.834  | 130  | 78943    | 10.324  | ug/l   | 93       |
| 45) 1,2-Dichloropropane       | 6.092  | 63   | 85295    | 10.355  | ug/l   | 99       |
| 46) Dibromomethane            | 6.230  | 93   | 65717    | 10.424  | ug/l   | 98       |
| 47) Bromodichloromethane      | 6.426  | 83   | 139261   | 10.685  | ug/l   | 97       |
| 48) Methyl methacrylate       | 6.304  | 41   | 68465    | 8.639   | ug/l   | 97       |
| 49) 1,4-Dioxane               | 6.297  | 88   | 23699    | 190.260 | ug/l   | 99       |
| 51) 4-Methyl-2-Pentanone      | 7.150  | 43   | 598702   | 54.051  | ug/l   | 97       |
| 52) Toluene                   | 7.310  | 92   | 202741   | 10.552  | ug/l   | 100      |
| 53) t-1,3-Dichloropropene     | 7.580  | 75   | 117314   | 10.329  | ug/l   | 97       |
| 54) cis-1,3-Dichloropropene   | 6.950  | 75   | 134529   | 10.729  | ug/l   | 98       |
| 55) 1,1,2-Trichloroethane     | 7.764  | 97   | 89390    | 10.376  | ug/l   | 98       |
| 56) Ethyl methacrylate        | 7.719  | 69   | 95279    | 9.313   | ug/l   | 99       |
| 57) 1,3-Dichloropropane       | 7.934  | 76   | 144218   | 10.607  | ug/l   | 98       |
| 58) 2-Chloroethyl Vinyl ether | 6.815  | 63   | 220693   | 46.148  | ug/l   | 99       |
| 59) 2-Hexanone                | 8.066  | 43   | 422703   | 49.020  | ug/l   | 93       |
| 60) Dibromochloromethane      | 8.169  | 129  | 101236   | 10.313  | ug/l   | 97       |
| 61) 1,2-Dibromoethane         | 8.275  | 107  | 93182    | 10.576  | ug/l   | 100      |
| 64) Tetrachloroethene         | 7.899  | 164  | 71811    | 10.397  | ug/l   | 98       |
| 65) Chlorobenzene             | 8.805  | 112  | 231329   | 10.072  | ug/l   | 99       |
| 66) 1,1,1,2-Tetrachloroethane | 8.899  | 131  | 85678    | 10.324  | ug/l   | 99       |
| 67) Ethyl Benzene             | 8.937  | 91   | 334394   | 9.618   | ug/l   | 99       |
| 68) m/p-Xylenes               | 9.063  | 106  | 270452   | 20.155  | ug/l   | 99       |
| 69) o-Xylene                  | 9.468  | 106  | 115922   | 9.756   | ug/l   | 96       |
| 70) Styrene                   | 9.487  | 104  | 204368   | 9.382   | ug/l   | 100      |
| 71) Bromoform                 | 9.657  | 173  | 65302    | 10.074  | ug/l # | 98       |
| 73) Isopropylbenzene          | 9.857  | 105  | 330305   | 9.827   | ug/l   | 99       |
| 74) N-amyl acetate            | 9.725  | 43   | 132108   | 9.853   | ug/l   | 96       |
| 75) 1,1,2,2-Tetrachloroethane | 10.165 | 83   | 148267   | 10.067  | ug/l   | 99       |
| 76) 1,2,3-Trichloropropane    | 10.198 | 75   | 105275   | 9.977   | ug/l   | 99       |
| 77) Bromobenzene              | 10.143 | 156  | 87828    | 10.071  | ug/l   | 96       |
| 78) n-propylbenzene           | 10.278 | 91   | 401316   | 9.804   | ug/l   | 100      |
| 79) 2-Chlorotoluene           | 10.349 | 91   | 255053   | 10.413  | ug/l   | 99       |
| 80) 1,3,5-Trimethylbenzene    | 10.464 | 105  | 270977   | 9.699   | ug/l   | 98       |
| 81) trans-1,4-Dichloro-2-b... | 9.931  | 75   | 44314    | 10.082  | ug/l   | 97       |
| 82) 4-Chlorotoluene           | 10.464 | 91   | 292863   | 10.312  | ug/l   | 100      |
| 83) tert-Butylbenzene         | 10.789 | 119  | 267528   | 9.603   | ug/l   | 98       |
| 84) 1,2,4-Trimethylbenzene    | 10.841 | 105  | 256505   | 9.457   | ug/l   | 99       |
| 85) sec-Butylbenzene          | 11.011 | 105  | 354312   | 9.588   | ug/l   | 100      |
| 86) p-Isopropyltoluene        | 11.169 | 119  | 300032   | 9.755   | ug/l   | 99       |
| 87) 1,3-Dichlorobenzene       | 11.108 | 146  | 178990   | 10.141  | ug/l   | 98       |
| 88) 1,4-Dichlorobenzene       | 11.198 | 146  | 191843   | 9.835   | ug/l   | 97       |
| 89) n-Butylbenzene            | 11.580 | 91   | 281258   | 9.575   | ug/l   | 97       |
| 90) Hexachloroethane          | 11.821 | 117  | 74910    | 9.891   | ug/l   | 96       |
| 91) 1,2-Dichlorobenzene       | 11.567 | 146  | 162271   | 9.841   | ug/l   | 98       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.355 | 75   | 27674    | 10.167  | ug/l   | 98       |
| 93) 1,2,4-Trichlorobenzene    | 13.188 | 180  | 85982    | 9.096   | ug/l   | 98       |
| 94) Hexachlorobutadiene       | 13.368 | 225  | 47413    | 9.653   | ug/l   | 97       |
| 95) Naphthalene               | 13.426 | 128  | 267645   | 8.803   | ug/l   | 99       |
| 96) 1,2,3-Trichlorobenzene    | 13.667 | 180  | 90238    | 9.334   | ug/l   | 100      |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
Data File : VV039137.D  
Acq On : 22 Sep 2025 12:26  
Operator : SY/MD  
Sample : VSTDICC010  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VSTDICC010

Quant Time: Sep 24 07:21:49 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 07:14:23 2025  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

**Instrument :**  
MSVOA\_V  
**ClientSampleId :**  
VSTDICC010

[illegible]

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
 Data File : VV039138.D  
 Acq On : 22 Sep 2025 12:48  
 Operator : SY/MD  
 Sample : VSTDICC020  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VSTDICC020

Quant Time: Sep 24 07:22:45 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 07:14:23 2025  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc     | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|----------|--------|----------|
| -----                        |                |      |            |          |        |          |
| Internal Standards           |                |      |            |          |        |          |
| 1) Pentafluorobenzene        | 4.600          | 168  | 545775     | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 5.532          | 114  | 880963     | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 8.773          | 117  | 860340     | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 11.172         | 152  | 509460     | 50.000   | ug/l   | 0.00     |
|                              |                |      |            |          |        |          |
| System Monitoring Compounds  |                |      |            |          |        |          |
| 33) 1,2-Dichloroethane-d4    | 4.947          | 65   | 144689     | 18.317   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 81 - 118 |      | Recovery = | 36.640%# |        |          |
| 35) Dibromofluoromethane     | 4.503          | 113  | 134662     | 19.014   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 80 - 119 |      | Recovery = | 38.020%# |        |          |
| 50) Toluene-d8               | 7.236          | 98   | 384827     | 18.501   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 89 - 112 |      | Recovery = | 37.000%# |        |          |
| 62) 4-Bromofluorobenzene     | 10.002         | 95   | 163597     | 19.105   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 85 - 114 |      | Recovery = | 38.220%# |        |          |
|                              |                |      |            |          |        |          |
| Target Compounds             |                |      |            |          | Qvalue |          |
| 2) Dichlorodifluoromethane   | 1.121          | 85   | 165175     | 19.118   | ug/l   | 96       |
| 3) Chloromethane             | 1.230          | 50   | 177694     | 19.159   | ug/l   | 99       |
| 4) Vinyl Chloride            | 1.298          | 62   | 186887     | 18.822   | ug/l   | 99       |
| 5) Bromomethane              | 1.500          | 94   | 119578     | 19.104   | ug/l   | 98       |
| 6) Chloroethane              | 1.561          | 64   | 124021     | 19.935   | ug/l   | 99       |
| 7) Trichlorofluoromethane    | 1.729          | 101  | 273532     | 19.027   | ug/l   | 100      |
| 8) Diethyl Ether             | 1.921          | 74   | 100661     | 18.862   | ug/l   | 99       |
| 9) 1,1,2-Trichlorotrifluo... | 2.085          | 101  | 159719     | 18.673   | ug/l   | 99       |
| 10) Methyl Iodide            | 2.198          | 142  | 161932     | 19.880   | ug/l   | 99       |
| 11) Tert butyl alcohol       | 2.568          | 59   | 171137     | 102.509  | ug/l   | 98       |
| 12) 1,1-Dichloroethene       | 2.085          | 96   | 157427     | 19.131   | ug/l   | 97       |
| 13) Acrolein                 | 2.011          | 56   | 29609      | 102.210  | ug/l   | 97       |
| 14) Allyl chloride           | 2.359          | 41   | 280200     | 19.062   | ug/l   | 99       |
| 15) Acrylonitrile            | 2.674          | 53   | 504022     | 94.179   | ug/l   | 99       |
| 16) Acetone                  | 2.121          | 43   | 499669     | 95.756   | ug/l   | 99       |
| 17) Carbon Disulfide         | 2.253          | 76   | 476229     | 18.923   | ug/l   | 100      |
| 18) Methyl Acetate           | 2.381          | 43   | 226208     | 19.328   | ug/l   | 98       |
| 19) Methyl tert-butyl Ether  | 2.712          | 73   | 498918     | 19.027   | ug/l   | 99       |
| 20) Methylene Chloride       | 2.458          | 84   | 185050     | 20.143   | ug/l   | 97       |
| 21) trans-1,2-Dichloroethene | 2.703          | 96   | 164898     | 18.823   | ug/l   | 97       |
| 22) Diisopropyl ether        | 3.217          | 45   | 552331     | 19.640   | ug/l   | 83       |
| 23) Vinyl Acetate            | 3.191          | 43   | 2445787    | 99.414   | ug/l   | 100      |
| 24) 1,1-Dichloroethane       | 3.118          | 63   | 320146     | 18.615   | ug/l   | 99       |
| 25) 2-Butanone               | 3.870          | 43   | 637413     | 102.519  | ug/l   | 100      |
| 26) 2,2-Dichloropropane      | 3.812          | 77   | 278362     | 18.326   | ug/l   | 98       |
| 27) cis-1,2-Dichloroethene   | 3.822          | 96   | 183867     | 19.031   | ug/l   | 98       |
| 28) Bromochloromethane       | 4.146          | 49   | 164406     | 21.654   | ug/l   | 99       |
| 29) Tetrahydrofuran          | 4.240          | 42   | 405777     | 102.491  | ug/l   | 99       |
| 30) Chloroform               | 4.281          | 83   | 342700     | 19.083   | ug/l   | 99       |
| 31) Cyclohexane              | 4.587          | 56   | 266753     | 18.934   | ug/l   | 98       |
| 32) 1,1,1-Trichloroethane    | 4.516          | 97   | 288341     | 18.546   | ug/l   | 99       |
| 36) 1,1-Dichloropropene      | 4.748          | 75   | 212835     | 19.734   | ug/l   | 99       |
| 37) Ethyl Acetate            | 3.979          | 43   | 260591     | 19.696   | ug/l   | 97       |
| 38) Carbon Tetrachloride     | 4.735          | 117  | 250103     | 19.252   | ug/l   | 98       |
| 39) Methylcyclohexane        | 6.047          | 83   | 253693     | 19.619   | ug/l   | 95       |
| 40) Benzene                  | 5.011          | 78   | 697141     | 19.994   | ug/l   | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VW092225\  
 Data File : VW039138.D  
 Acq On : 22 Sep 2025 12:48  
 Operator : SY/MD  
 Sample : VSTDIC020  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VSTDIC020

Quant Time: Sep 24 07:22:45 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 07:14:23 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 4.169  | 41   | 102755   | 18.721  | ug/l   | 95       |
| 42) 1,2-Dichloroethane        | 5.043  | 62   | 244487   | 20.012  | ug/l   | 99       |
| 43) Isopropyl Acetate         | 5.195  | 43   | 362512   | 19.889  | ug/l   | 98       |
| 44) Trichloroethene           | 5.831  | 130  | 160692   | 19.972  | ug/l   | 96       |
| 45) 1,2-Dichloropropane       | 6.092  | 63   | 182757   | 21.087  | ug/l   | 96       |
| 46) Dibromomethane            | 6.227  | 93   | 129319   | 19.495  | ug/l   | 98       |
| 47) Bromodichloromethane      | 6.429  | 83   | 269262   | 19.635  | ug/l   | 97       |
| 48) Methyl methacrylate       | 6.301  | 41   | 158251   | 18.977  | ug/l   | 99       |
| 49) 1,4-Dioxane               | 6.294  | 88   | 47860    | 365.172 | ug/l # | 85       |
| 51) 4-Methyl-2-Pentanone      | 7.150  | 43   | 1296796  | 111.268 | ug/l   | 98       |
| 52) Toluene                   | 7.310  | 92   | 422367   | 20.892  | ug/l   | 100      |
| 53) t-1,3-Dichloropropene     | 7.577  | 75   | 245181   | 20.516  | ug/l   | 98       |
| 54) cis-1,3-Dichloropropene   | 6.950  | 75   | 273906   | 20.761  | ug/l   | 100      |
| 55) 1,1,2-Trichloroethane     | 7.760  | 97   | 176812   | 19.505  | ug/l   | 98       |
| 56) Ethyl methacrylate        | 7.719  | 69   | 212320   | 19.723  | ug/l   | 97       |
| 57) 1,3-Dichloropropane       | 7.934  | 76   | 291304   | 20.363  | ug/l   | 100      |
| 58) 2-Chloroethyl Vinyl ether | 6.815  | 63   | 536254   | 94.274  | ug/l   | 99       |
| 59) 2-Hexanone                | 8.066  | 43   | 940265   | 101.225 | ug/l   | 98       |
| 60) Dibromochloromethane      | 8.169  | 129  | 200793   | 19.440  | ug/l   | 99       |
| 61) 1,2-Dibromoethane         | 8.275  | 107  | 189555   | 20.447  | ug/l   | 99       |
| 64) Tetrachloroethene         | 7.899  | 164  | 141876   | 19.688  | ug/l   | 99       |
| 65) Chlorobenzene             | 8.805  | 112  | 470841   | 19.648  | ug/l   | 97       |
| 66) 1,1,1,2-Tetrachloroethane | 8.899  | 131  | 167936   | 19.395  | ug/l   | 99       |
| 67) Ethyl Benzene             | 8.937  | 91   | 737391   | 20.328  | ug/l   | 99       |
| 68) m/p-Xylenes               | 9.063  | 106  | 607362   | 43.381  | ug/l   | 99       |
| 69) o-Xylene                  | 9.468  | 106  | 255212   | 20.587  | ug/l   | 100      |
| 70) Styrene                   | 9.487  | 104  | 475382   | 19.022  | ug/l   | 99       |
| 71) Bromoform                 | 9.654  | 173  | 132242   | 19.552  | ug/l   | 98       |
| 73) Isopropylbenzene          | 9.857  | 105  | 741305   | 20.062  | ug/l   | 99       |
| 74) N-amyl acetate            | 9.725  | 43   | 306963   | 20.826  | ug/l   | 99       |
| 75) 1,1,2,2-Tetrachloroethane | 10.165 | 83   | 297741   | 18.390  | ug/l   | 100      |
| 76) 1,2,3-Trichloropropane    | 10.198 | 75   | 213532   | 18.409  | ug/l   | 99       |
| 77) Bromobenzene              | 10.140 | 156  | 187188   | 19.525  | ug/l   | 96       |
| 78) n-propylbenzene           | 10.278 | 91   | 905281   | 20.118  | ug/l   | 99       |
| 79) 2-Chlorotoluene           | 10.349 | 91   | 547836   | 20.346  | ug/l   | 99       |
| 80) 1,3,5-Trimethylbenzene    | 10.464 | 105  | 640175   | 20.843  | ug/l   | 99       |
| 81) trans-1,4-Dichloro-2-b... | 9.928  | 75   | 91848    | 19.009  | ug/l   | 96       |
| 82) 4-Chlorotoluene           | 10.461 | 91   | 655427   | 20.993  | ug/l   | 99       |
| 83) tert-Butylbenzene         | 10.789 | 119  | 587151   | 19.172  | ug/l   | 99       |
| 84) 1,2,4-Trimethylbenzene    | 10.841 | 105  | 631679   | 21.185  | ug/l   | 100      |
| 85) sec-Butylbenzene          | 11.011 | 105  | 808869   | 19.912  | ug/l   | 99       |
| 86) p-Isopropyltoluene        | 11.165 | 119  | 682814   | 20.196  | ug/l   | 100      |
| 87) 1,3-Dichlorobenzene       | 11.104 | 146  | 374208   | 19.287  | ug/l   | 99       |
| 88) 1,4-Dichlorobenzene       | 11.198 | 146  | 397526   | 18.539  | ug/l   | 100      |
| 89) n-Butylbenzene            | 11.580 | 91   | 640044   | 19.822  | ug/l   | 99       |
| 90) Hexachloroethane          | 11.818 | 117  | 150843   | 18.118  | ug/l   | 99       |
| 91) 1,2-Dichlorobenzene       | 11.567 | 146  | 336566   | 18.567  | ug/l   | 98       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.352 | 75   | 57671    | 19.989  | ug/l   | 98       |
| 93) 1,2,4-Trichlorobenzene    | 13.185 | 180  | 196334   | 18.894  | ug/l   | 99       |
| 94) Hexachlorobutadiene       | 13.368 | 225  | 96610    | 17.893  | ug/l   | 98       |
| 95) Naphthalene               | 13.426 | 128  | 634557   | 18.986  | ug/l   | 99       |
| 96) 1,2,3-Trichlorobenzene    | 13.667 | 180  | 198308   | 18.659  | ug/l   | 97       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
Data File : VV039138.D  
Acq On : 22 Sep 2025 12:48  
Operator : SY/MD  
Sample : VSTDICC020  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VSTDICC020

Quant Time: Sep 24 07:22:45 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 07:14:23 2025  
Response via : Initial Calibration


| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
Data File : VV039138.D  
Acq On : 22 Sep 2025 12:48  
Operator : SY/MD  
Sample : VSTDIC020  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VSTDIC020

Quant Time: Sep 24 07:22:45 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 07:14:23 2025  
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
 Data File : VV039139.D  
 Acq On : 22 Sep 2025 13:26  
 Operator : SY/MD  
 Sample : VSTDICCC050  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VSTDICCC050

Quant Time: Sep 24 07:23:42 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 07:14:23 2025  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 4.597  | 168  | 563410   | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 5.529  | 114  | 894173   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 8.773  | 117  | 863906   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 11.172 | 152  | 516167   | 50.000 | ug/l  | 0.00     |

## System Monitoring Compounds

|                           |        |       |          |          |      |          |
|---------------------------|--------|-------|----------|----------|------|----------|
| 33) 1,2-Dichloroethane-d4 | 4.937  | 65    | 363821   | 44.618   | ug/l | 0.00     |
| Spiked Amount             | 50.000 | Range | 81 - 118 | Recovery | =    | 89.240%  |
| 35) Dibromofluoromethane  | 4.497  | 113   | 337386   | 46.935   | ug/l | 0.00     |
| Spiked Amount             | 50.000 | Range | 80 - 119 | Recovery | =    | 93.860%  |
| 50) Toluene-d8            | 7.233  | 98    | 1016135  | 48.130   | ug/l | 0.00     |
| Spiked Amount             | 50.000 | Range | 89 - 112 | Recovery | =    | 96.260%  |
| 62) 4-Bromofluorobenzene  | 9.998  | 95    | 453081   | 52.130   | ug/l | 0.00     |
| Spiked Amount             | 50.000 | Range | 85 - 114 | Recovery | =    | 104.260% |

## Target Compounds

|                              |       |     |         |         |      | Qvalue |
|------------------------------|-------|-----|---------|---------|------|--------|
| 2) Dichlorodifluoromethane   | 1.121 | 85  | 416798  | 46.733  | ug/l | 100    |
| 3) Chloromethane             | 1.230 | 50  | 433999  | 45.329  | ug/l | 100    |
| 4) Vinyl Chloride            | 1.298 | 62  | 470936  | 45.944  | ug/l | 100    |
| 5) Bromomethane              | 1.494 | 94  | 286332  | 44.312  | ug/l | 100    |
| 6) Chloroethane              | 1.558 | 64  | 300232  | 49.734  | ug/l | 100    |
| 7) Trichlorofluoromethane    | 1.725 | 101 | 672325  | 45.304  | ug/l | 100    |
| 8) Diethyl Ether             | 1.918 | 74  | 251813  | 45.707  | ug/l | 100    |
| 9) 1,1,2-Trichlorotrifluo... | 2.082 | 101 | 395615  | 44.803  | ug/l | 100    |
| 10) Methyl Iodide            | 2.195 | 142 | 442822  | 52.662  | ug/l | 100    |
| 11) Tert butyl alcohol       | 2.564 | 59  | 385746  | 223.826 | ug/l | 100    |
| 12) 1,1-Dichloroethene       | 2.082 | 96  | 379710  | 44.700  | ug/l | 100    |
| 13) Acrolein                 | 2.008 | 56  | 74486   | 249.076 | ug/l | 100    |
| 14) Allyl chloride           | 2.355 | 41  | 700342  | 46.154  | ug/l | 100    |
| 15) Acrylonitrile            | 2.671 | 53  | 1228907 | 222.439 | ug/l | 100    |
| 16) Acetone                  | 2.118 | 43  | 1265424 | 251.263 | ug/l | 100    |
| 17) Carbon Disulfide         | 2.253 | 76  | 1172348 | 45.126  | ug/l | 100    |
| 18) Methyl Acetate           | 2.378 | 43  | 553676  | 45.827  | ug/l | 100    |
| 19) Methyl tert-butyl Ether  | 2.709 | 73  | 1258060 | 46.476  | ug/l | 100    |
| 20) Methylene Chloride       | 2.455 | 84  | 439142  | 49.828  | ug/l | 100    |
| 21) trans-1,2-Dichloroethene | 2.699 | 96  | 407938  | 45.109  | ug/l | 100    |
| 22) Diisopropyl ether        | 3.214 | 45  | 1419527 | 48.895  | ug/l | 100    |
| 23) Vinyl Acetate            | 3.185 | 43  | 6209687 | 244.506 | ug/l | 100    |
| 24) 1,1-Dichloroethane       | 3.114 | 63  | 781832  | 44.037  | ug/l | 100    |
| 25) 2-Butanone               | 3.860 | 43  | 1650213 | 257.105 | ug/l | 100    |
| 26) 2,2-Dichloropropane      | 3.806 | 77  | 698283  | 44.534  | ug/l | 100    |
| 27) cis-1,2-Dichloroethene   | 3.818 | 96  | 488972  | 49.026  | ug/l | 100    |
| 28) Bromochloromethane       | 4.143 | 49  | 382640  | 48.820  | ug/l | 100    |
| 29) Tetrahydrofuran          | 4.230 | 42  | 1039169 | 254.258 | ug/l | 100    |
| 30) Chloroform               | 4.275 | 83  | 839217  | 45.268  | ug/l | 100    |
| 31) Cyclohexane              | 4.580 | 56  | 701482  | 48.232  | ug/l | 100    |
| 32) 1,1,1-Trichloroethane    | 4.510 | 97  | 731030  | 45.547  | ug/l | 100    |
| 36) 1,1-Dichloropropene      | 4.741 | 75  | 554798  | 50.680  | ug/l | 100    |
| 37) Ethyl Acetate            | 3.970 | 43  | 638131  | 47.519  | ug/l | 100    |
| 38) Carbon Tetrachloride     | 4.732 | 117 | 628516  | 47.666  | ug/l | 100    |
| 39) Methylcyclohexane        | 6.043 | 83  | 736231  | 56.095  | ug/l | 100    |
| 40) Benzene                  | 5.005 | 78  | 1779301 | 50.276  | ug/l | 100    |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VW092225\  
 Data File : VW039139.D  
 Acq On : 22 Sep 2025 13:26  
 Operator : SY/MD  
 Sample : VSTDICCC050  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VSTDICCC050

Quant Time: Sep 24 07:23:42 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 07:14:23 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc     | Units | Dev(Min) |
|-------------------------------|--------|------|----------|----------|-------|----------|
| 41) Methacrylonitrile         | 4.153  | 41   | 295316   | 49.541   | ug/l  | 100      |
| 42) 1,2-Dichloroethane        | 5.037  | 62   | 595216   | 48.000   | ug/l  | 100      |
| 43) Isopropyl Acetate         | 5.188  | 43   | 956288   | 51.691   | ug/l  | 100      |
| 44) Trichloroethene           | 5.828  | 130  | 414171   | 50.715   | ug/l  | 100      |
| 45) 1,2-Dichloropropane       | 6.085  | 63   | 448750   | 51.014   | ug/l  | 100      |
| 46) Dibromomethane            | 6.220  | 93   | 329073   | 48.876   | ug/l  | 100      |
| 47) Bromodichloromethane      | 6.423  | 83   | 673646   | 48.398   | ug/l  | 100      |
| 48) Methyl methacrylate       | 6.291  | 41   | 450230   | 53.193   | ug/l  | 100      |
| 49) 1,4-Dioxane               | 6.288  | 88   | 133834   | 1006.069 | ug/l  | 100      |
| 51) 4-Methyl-2-Pentanone      | 7.146  | 43   | 3273840  | 276.753  | ug/l  | 100      |
| 52) Toluene                   | 7.304  | 92   | 1099325  | 53.574   | ug/l  | 100      |
| 53) t-1,3-Dichloropropene     | 7.571  | 75   | 650976   | 53.666   | ug/l  | 100      |
| 54) cis-1,3-Dichloropropene   | 6.944  | 75   | 725015   | 54.142   | ug/l  | 100      |
| 55) 1,1,2-Trichloroethane     | 7.757  | 97   | 441822   | 48.020   | ug/l  | 100      |
| 56) Ethyl methacrylate        | 7.712  | 69   | 640636   | 58.632   | ug/l  | 100      |
| 57) 1,3-Dichloropropane       | 7.931  | 76   | 733341   | 50.505   | ug/l  | 100      |
| 58) 2-Chloroethyl Vinyl ether | 6.809  | 63   | 1579407  | 255.700  | ug/l  | 100      |
| 59) 2-Hexanone                | 8.059  | 43   | 2441880  | 255.630  | ug/l  | 100      |
| 60) Dibromochloromethane      | 8.165  | 129  | 507043   | 48.365   | ug/l  | 100      |
| 61) 1,2-Dibromoethane         | 8.272  | 107  | 473037   | 50.273   | ug/l  | 100      |
| 64) Tetrachloroethene         | 7.895  | 164  | 356209   | 49.227   | ug/l  | 100      |
| 65) Chlorobenzene             | 8.802  | 112  | 1194769  | 49.651   | ug/l  | 100      |
| 66) 1,1,1,2-Tetrachloroethane | 8.895  | 131  | 421762   | 48.507   | ug/l  | 100      |
| 67) Ethyl Benzene             | 8.934  | 91   | 2057558  | 56.487   | ug/l  | 100      |
| 68) m/p-Xylenes               | 9.059  | 106  | 1639130  | 116.591  | ug/l  | 100      |
| 69) o-Xylene                  | 9.468  | 106  | 730558   | 58.688   | ug/l  | 100      |
| 70) Styrene                   | 9.484  | 104  | 1341298  | 50.845   | ug/l  | 100      |
| 71) Bromoform                 | 9.651  | 173  | 341871   | 50.337   | ug/l  | 100      |
| 73) Isopropylbenzene          | 9.854  | 105  | 2033373  | 54.313   | ug/l  | 100      |
| 74) N-amyl acetate            | 9.722  | 43   | 863632   | 57.833   | ug/l  | 100      |
| 75) 1,1,2,2-Tetrachloroethane | 10.165 | 83   | 735590   | 44.844   | ug/l  | 100      |
| 76) 1,2,3-Trichloropropane    | 10.194 | 75   | 541552   | 46.080   | ug/l  | 100      |
| 77) Bromobenzene              | 10.140 | 156  | 477608   | 49.171   | ug/l  | 100      |
| 78) n-propylbenzene           | 10.275 | 91   | 2528458  | 55.461   | ug/l  | 100      |
| 79) 2-Chlorotoluene           | 10.345 | 91   | 1462530  | 53.612   | ug/l  | 100      |
| 80) 1,3,5-Trimethylbenzene    | 10.464 | 105  | 1748435  | 56.188   | ug/l  | 100      |
| 81) trans-1,4-Dichloro-2-b... | 9.928  | 75   | 255259   | 52.143   | ug/l  | 100      |
| 82) 4-Chlorotoluene           | 10.461 | 91   | 1722899  | 54.466   | ug/l  | 100      |
| 83) tert-Butylbenzene         | 10.789 | 119  | 1663245  | 53.602   | ug/l  | 100      |
| 84) 1,2,4-Trimethylbenzene    | 10.837 | 105  | 1743425  | 57.709   | ug/l  | 100      |
| 85) sec-Butylbenzene          | 11.011 | 105  | 2252781  | 54.737   | ug/l  | 100      |
| 86) p-Isopropyltoluene        | 11.165 | 119  | 1900412  | 55.479   | ug/l  | 100      |
| 87) 1,3-Dichlorobenzene       | 11.104 | 146  | 958701   | 48.769   | ug/l  | 100      |
| 88) 1,4-Dichlorobenzene       | 11.194 | 146  | 998426   | 45.957   | ug/l  | 100      |
| 89) n-Butylbenzene            | 11.577 | 91   | 1864844  | 57.003   | ug/l  | 100      |
| 90) Hexachloroethane          | 11.818 | 117  | 375818   | 44.553   | ug/l  | 100      |
| 91) 1,2-Dichlorobenzene       | 11.567 | 146  | 897135   | 48.848   | ug/l  | 100      |
| 92) 1,2-Dibromo-3-Chloropr... | 12.349 | 75   | 147329   | 49.928   | ug/l  | 100      |
| 93) 1,2,4-Trichlorobenzene    | 13.185 | 180  | 555323   | 52.747   | ug/l  | 100      |
| 94) Hexachlorobutadiene       | 13.368 | 225  | 247695   | 45.279   | ug/l  | 100      |
| 95) Naphthalene               | 13.423 | 128  | 1968567  | 58.135   | ug/l  | 100      |
| 96) 1,2,3-Trichlorobenzene    | 13.664 | 180  | 571563   | 53.081   | ug/l  | 100      |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
Data File : VV039139.D  
Acq On : 22 Sep 2025 13:26  
Operator : SY/MD  
Sample : VSTDICCC050  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 6 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VSTDICCC050

Quant Time: Sep 24 07:23:42 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 07:14:23 2025  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

**Instrument :**  
MSVOA\_V  
**ClientSampleId :**  
VSTDICCC050

[illegible]

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
 Data File : VV039140.D  
 Acq On : 22 Sep 2025 13:49  
 Operator : SY/MD  
 Sample : VSTDICC100  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VSTDICC100

Quant Time: Sep 24 07:24:39 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 07:14:23 2025  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 4.600  | 168  | 587274   | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 5.532  | 114  | 935392   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 8.776  | 117  | 904902   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 11.172 | 152  | 523452   | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |             |      |
|-----------------------------|--------|-------|----------|----------|-------------|------|
| System Monitoring Compounds |        |       |          |          |             |      |
| 33) 1,2-Dichloroethane-d4   | 4.941  | 65    | 725543   | 85.362   | ug/l        | 0.00 |
| Spiked Amount               | 50.000 | Range | 81 - 118 | Recovery | = 170.720%# |      |
| 35) Dibromofluoromethane    | 4.500  | 113   | 674408   | 89.684   | ug/l        | 0.00 |
| Spiked Amount               | 50.000 | Range | 80 - 119 | Recovery | = 179.360%# |      |
| 50) Toluene-d8              | 7.236  | 98    | 2052477  | 92.932   | ug/l        | 0.00 |
| Spiked Amount               | 50.000 | Range | 89 - 112 | Recovery | = 185.860%# |      |
| 62) 4-Bromofluorobenzene    | 9.998  | 95    | 937505   | 103.112  | ug/l        | 0.00 |
| Spiked Amount               | 50.000 | Range | 85 - 114 | Recovery | = 206.220%# |      |

|                              |       |     |          |         |        |     |
|------------------------------|-------|-----|----------|---------|--------|-----|
| Target Compounds             |       |     |          |         | Qvalue |     |
| 2) Dichlorodifluoromethane   | 1.121 | 85  | 820676   | 88.278  | ug/l   | 99  |
| 3) Chloromethane             | 1.230 | 50  | 856347   | 85.806  | ug/l   | 98  |
| 4) Vinyl Chloride            | 1.298 | 62  | 940694   | 88.044  | ug/l   | 99  |
| 5) Bromomethane              | 1.487 | 94  | 615924   | 91.446  | ug/l   | 99  |
| 6) Chloroethane              | 1.555 | 64  | 587651   | 100.067 | ug/l   | 99  |
| 7) Trichlorofluoromethane    | 1.722 | 101 | 1351410  | 87.363  | ug/l   | 99  |
| 8) Diethyl Ether             | 1.918 | 74  | 511033   | 88.990  | ug/l   | 100 |
| 9) 1,1,2-Trichlorotrifluo... | 2.079 | 101 | 806216   | 87.593  | ug/l   | 98  |
| 10) Methyl Iodide            | 2.195 | 142 | 931666   | 106.294 | ug/l   | 100 |
| 11) Tert butyl alcohol       | 2.574 | 59  | 798702   | 444.608 | ug/l   | 99  |
| 12) 1,1-Dichloroethene       | 2.079 | 96  | 778436   | 87.914  | ug/l   | 98  |
| 13) Acrolein                 | 2.008 | 56  | 160749   | 515.691 | ug/l   | 99  |
| 14) Allyl chloride           | 2.355 | 41  | 1418889  | 89.708  | ug/l   | 99  |
| 15) Acrylonitrile            | 2.674 | 53  | 2509475  | 435.771 | ug/l   | 100 |
| 16) Acetone                  | 2.118 | 43  | 2432444  | 499.795 | ug/l   | 100 |
| 17) Carbon Disulfide         | 2.253 | 76  | 2332239  | 86.125  | ug/l   | 99  |
| 18) Methyl Acetate           | 2.378 | 43  | 1142900  | 90.753  | ug/l   | 100 |
| 19) Methyl tert-butyl Ether  | 2.712 | 73  | 2586428  | 91.667  | ug/l   | 99  |
| 20) Methylene Chloride       | 2.455 | 84  | 870955   | 100.038 | ug/l   | 99  |
| 21) trans-1,2-Dichloroethene | 2.699 | 96  | 817275   | 86.701  | ug/l   | 98  |
| 22) Diisopropyl ether        | 3.217 | 45  | 2857568  | 94.428  | ug/l   | 94  |
| 23) Vinyl Acetate            | 3.188 | 43  | 12498986 | 472.148 | ug/l   | 99  |
| 24) 1,1-Dichloroethane       | 3.117 | 63  | 1560774  | 84.338  | ug/l   | 99  |
| 25) 2-Butanone               | 3.863 | 43  | 3404031  | 508.801 | ug/l   | 99  |
| 26) 2,2-Dichloropropane      | 3.812 | 77  | 1414724  | 86.559  | ug/l   | 100 |
| 27) cis-1,2-Dichloroethene   | 3.818 | 96  | 1010617  | 97.210  | ug/l   | 99  |
| 28) Bromochloromethane       | 4.146 | 49  | 748549   | 91.624  | ug/l   | 98  |
| 29) Tetrahydrofuran          | 4.233 | 42  | 2198745  | 516.116 | ug/l   | 99  |
| 30) Chloroform               | 4.278 | 83  | 1682895  | 87.087  | ug/l   | 98  |
| 31) Cyclohexane              | 4.580 | 56  | 1473260  | 97.182  | ug/l   | 97  |
| 32) 1,1,1-Trichloroethane    | 4.513 | 97  | 1483938  | 88.700  | ug/l   | 99  |
| 36) 1,1-Dichloropropene      | 4.741 | 75  | 1144520  | 99.944  | ug/l   | 100 |
| 37) Ethyl Acetate            | 3.973 | 43  | 1328655  | 94.580  | ug/l   | 98  |
| 38) Carbon Tetrachloride     | 4.735 | 117 | 1263028  | 91.566  | ug/l   | 100 |
| 39) Methylcyclohexane        | 6.047 | 83  | 1603169  | 116.766 | ug/l   | 98  |
| 40) Benzene                  | 5.008 | 78  | 3611886  | 97.560  | ug/l   | 98  |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VW092225\  
 Data File : VW039140.D  
 Acq On : 22 Sep 2025 13:49  
 Operator : SY/MD  
 Sample : VSTDICC100  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VSTDICC100

Quant Time: Sep 24 07:24:39 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 07:14:23 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc     | Units | Dev(Min) |
|-------------------------------|--------|------|----------|----------|-------|----------|
| 41) Methacrylonitrile         | 4.156  | 41   | 639301   | 100.494  | ug/l  | 99       |
| 42) 1,2-Dichloroethane        | 5.037  | 62   | 1202991  | 92.737   | ug/l  | 100      |
| 43) Isopropyl Acetate         | 5.191  | 43   | 2058622  | 106.372  | ug/l  | 100      |
| 44) Trichloroethene           | 5.828  | 130  | 849813   | 99.474   | ug/l  | 97       |
| 45) 1,2-Dichloropropane       | 6.088  | 63   | 909288   | 98.813   | ug/l  | 95       |
| 46) Dibromomethane            | 6.220  | 93   | 661336   | 93.898   | ug/l  | 98       |
| 47) Bromodichloromethane      | 6.426  | 83   | 1355845  | 93.118   | ug/l  | 99       |
| 48) Methyl methacrylate       | 6.294  | 41   | 984431   | 111.182  | ug/l  | 99       |
| 49) 1,4-Dioxane               | 6.288  | 88   | 304943   | 2191.331 | ug/l  | 95       |
| 51) 4-Methyl-2-Pentanone      | 7.149  | 43   | 6674594  | 539.372  | ug/l  | 99       |
| 52) Toluene                   | 7.307  | 92   | 2271340  | 105.813  | ug/l  | 99       |
| 53) t-1,3-Dichloropropene     | 7.571  | 75   | 1408125  | 110.970  | ug/l  | 97       |
| 54) cis-1,3-Dichloropropene   | 6.947  | 75   | 1501380  | 107.178  | ug/l  | 100      |
| 55) 1,1,2-Trichloroethane     | 7.757  | 97   | 897804   | 93.280   | ug/l  | 99       |
| 56) Ethyl methacrylate        | 7.712  | 69   | 1424432  | 124.621  | ug/l  | 99       |
| 57) 1,3-Dichloropropane       | 7.931  | 76   | 1501225  | 98.832   | ug/l  | 100      |
| 58) 2-Chloroethyl Vinyl ether | 6.812  | 63   | 3282267  | 498.705  | ug/l  | 100      |
| 59) 2-Hexanone                | 8.063  | 43   | 4988835  | 497.186  | ug/l  | 99       |
| 60) Dibromochloromethane      | 8.165  | 129  | 1044126  | 95.206   | ug/l  | 99       |
| 61) 1,2-Dibromoethane         | 8.272  | 107  | 960914   | 97.622   | ug/l  | 98       |
| 64) Tetrachloroethene         | 7.895  | 164  | 736462   | 97.167   | ug/l  | 97       |
| 65) Chlorobenzene             | 8.802  | 112  | 2457021  | 97.480   | ug/l  | 97       |
| 66) 1,1,1,2-Tetrachloroethane | 8.895  | 131  | 863523   | 94.816   | ug/l  | 100      |
| 67) Ethyl Benzene             | 8.934  | 91   | 4376266  | 114.700  | ug/l  | 100      |
| 68) m/p-Xylenes               | 9.059  | 106  | 3396653  | 230.657  | ug/l  | 100      |
| 69) o-Xylene                  | 9.468  | 106  | 1592963  | 122.170  | ug/l  | 98       |
| 70) Styrene                   | 9.484  | 104  | 2781887  | 99.832   | ug/l  | 100      |
| 71) Bromoform                 | 9.651  | 173  | 724467   | 101.838  | ug/l  | 99       |
| 73) Isopropylbenzene          | 9.854  | 105  | 4288139  | 112.946  | ug/l  | 100      |
| 74) N-amyl acetate            | 9.722  | 43   | 1846779  | 121.948  | ug/l  | 98       |
| 75) 1,1,2,2-Tetrachloroethane | 10.165 | 83   | 1453989  | 87.406   | ug/l  | 100      |
| 76) 1,2,3-Trichloropropane    | 10.198 | 75   | 1070580  | 89.827   | ug/l  | 95       |
| 77) Bromobenzene              | 10.140 | 156  | 993863   | 100.897  | ug/l  | 99       |
| 78) n-propylbenzene           | 10.278 | 91   | 5220220  | 112.910  | ug/l  | 99       |
| 79) 2-Chlorotoluene           | 10.345 | 91   | 3018051  | 109.093  | ug/l  | 99       |
| 80) 1,3,5-Trimethylbenzene    | 10.464 | 105  | 3638030  | 115.284  | ug/l  | 100      |
| 81) trans-1,4-Dichloro-2-b... | 9.927  | 75   | 545452   | 109.871  | ug/l  | 97       |
| 82) 4-Chlorotoluene           | 10.461 | 91   | 3573507  | 111.396  | ug/l  | 100      |
| 83) tert-Butylbenzene         | 10.789 | 119  | 3572813  | 113.541  | ug/l  | 99       |
| 84) 1,2,4-Trimethylbenzene    | 10.837 | 105  | 3601355  | 117.550  | ug/l  | 100      |
| 85) sec-Butylbenzene          | 11.011 | 105  | 4763689  | 114.135  | ug/l  | 100      |
| 86) p-Isopropyltoluene        | 11.165 | 119  | 3948556  | 113.666  | ug/l  | 99       |
| 87) 1,3-Dichlorobenzene       | 11.104 | 146  | 1970594  | 98.849   | ug/l  | 99       |
| 88) 1,4-Dichlorobenzene       | 11.194 | 146  | 2008676  | 91.172   | ug/l  | 99       |
| 89) n-Butylbenzene            | 11.577 | 91   | 3939852  | 118.754  | ug/l  | 99       |
| 90) Hexachloroethane          | 11.818 | 117  | 776688   | 90.794   | ug/l  | 99       |
| 91) 1,2-Dichlorobenzene       | 11.564 | 146  | 1886394  | 101.282  | ug/l  | 100      |
| 92) 1,2-Dibromo-3-Chloropr... | 12.349 | 75   | 316446   | 100.015  | ug/l  | 96       |
| 93) 1,2,4-Trichlorobenzene    | 13.185 | 180  | 1244510  | 116.564  | ug/l  | 100      |
| 94) Hexachlorobutadiene       | 13.368 | 225  | 524789   | 94.597   | ug/l  | 99       |
| 95) Naphthalene               | 13.423 | 128  | 4442612  | 129.371  | ug/l  | 100      |
| 96) 1,2,3-Trichlorobenzene    | 13.667 | 180  | 1246406  | 114.143  | ug/l  | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
Data File : VV039140.D  
Acq On : 22 Sep 2025 13:49  
Operator : SY/MD  
Sample : VSTDICC100  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VSTDICC100

Quant Time: Sep 24 07:24:39 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 07:14:23 2025  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

**Instrument :**  
MSVOA\_V  
**ClientSampleId :**  
VSTDICC100

[illegible]

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
 Data File : VV039142.D  
 Acq On : 22 Sep 2025 15:37  
 Operator : SY/MD  
 Sample : VSTDICC005  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VSTDICC005

Manual Integrations  
 APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025  
 Supervised By :Semsettin Yesilyurt 09/26/2025

Quant Time: Sep 24 07:25:48 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 07:14:23 2025  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc     | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|----------|--------|----------|
| -----                        |                |      |            |          |        |          |
| Internal Standards           |                |      |            |          |        |          |
| 1) Pentafluorobenzene        | 4.596          | 168  | 492799     | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 5.529          | 114  | 835334     | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 8.773          | 117  | 807740     | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 11.172         | 152  | 425480     | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds  |                |      |            |          |        |          |
| 33) 1,2-Dichloroethane-d4    | 4.947          | 65   | 45267      | 6.347    | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 81 - 118 |      | Recovery = | 12.700%# |        |          |
| 35) Dibromofluoromethane     | 4.503          | 113  | 37959      | 5.653    | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 80 - 119 |      | Recovery = | 11.300%# |        |          |
| 50) Toluene-d8               | 7.239          | 98   | 117860     | 5.976    | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 89 - 112 |      | Recovery = | 11.960%# |        |          |
| 62) 4-Bromofluorobenzene     | 10.008         | 95   | 39819      | 4.904    | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 85 - 114 |      | Recovery = | 9.800%#  |        |          |
| Target Compounds             |                |      |            |          |        |          |
|                              |                |      |            |          | Qvalue |          |
| 2) Dichlorodifluoromethane   | 1.121          | 85   | 38031      | 4.875    | ug/l   | 93       |
| 3) Chloromethane             | 1.230          | 50   | 43830      | 5.234    | ug/l   | 97       |
| 4) Vinyl Chloride            | 1.298          | 62   | 45941      | 5.124    | ug/l   | 99       |
| 5) Bromomethane              | 1.494          | 94   | 32301      | 5.715    | ug/l   | 100      |
| 6) Chloroethane              | 1.558          | 64   | 29857      | 4.506    | ug/l   | 93       |
| 7) Trichlorofluoromethane    | 1.725          | 101  | 67248      | 5.181    | ug/l   | 100      |
| 8) Diethyl Ether             | 1.918          | 74   | 25304      | 5.251    | ug/l   | 95       |
| 9) 1,1,2-Trichlorotrifluo... | 2.082          | 101  | 39608      | 5.128    | ug/l   | 99       |
| 10) Methyl Iodide            | 2.195          | 142  | 31463      | 4.278    | ug/l   | 95       |
| 11) Tert butyl alcohol       | 2.571          | 59   | 40994      | 27.195   | ug/l   | 97       |
| 12) 1,1-Dichloroethene       | 2.082          | 96   | 38859      | 5.230    | ug/l   | 96       |
| 13) Acrolein                 | 2.015          | 56   | 6095       | 23.302   | ug/l   | 98       |
| 14) Allyl chloride           | 2.355          | 41   | 70683      | 5.326    | ug/l   | 96       |
| 15) Acrylonitrile            | 2.677          | 53   | 128729     | 26.639   | ug/l   | 99       |
| 16) Acetone                  | 2.121          | 43   | 140190     | 26.199   | ug/l   | 98       |
| 17) Carbon Disulfide         | 2.253          | 76   | 121350     | 5.340    | ug/l   | 99       |
| 18) Methyl Acetate           | 2.381          | 43   | 54053      | 5.115    | ug/l   | 97       |
| 19) Methyl tert-butyl Ether  | 2.712          | 73   | 125944     | 5.319    | ug/l   | 96       |
| 20) Methylene Chloride       | 2.455          | 84   | 54859      | 5.252    | ug/l   | 96       |
| 21) trans-1,2-Dichloroethene | 2.703          | 96   | 41151      | 5.202    | ug/l   | 98       |
| 22) Diisopropyl ether        | 3.217          | 45   | 126181     | 4.969    | ug/l # | 64       |
| 23) Vinyl Acetate            | 3.191          | 43   | 570043     | 25.661   | ug/l   | 99       |
| 24) 1,1-Dichloroethane       | 3.117          | 63   | 81764      | 5.265    | ug/l   | 96       |
| 25) 2-Butanone               | 3.879          | 43   | 140360     | 25.002   | ug/l   | 99       |
| 26) 2,2-Dichloropropane      | 3.812          | 77   | 68656      | 5.006    | ug/l   | 99       |
| 27) cis-1,2-Dichloroethene   | 3.825          | 96   | 40984      | 4.698    | ug/l   | 88       |
| 28) Bromochloromethane       | 4.153          | 49   | 25874      | 3.774    | ug/l   | 97       |
| 29) Tetrahydrofuran          | 4.243          | 42   | 86190      | 24.110   | ug/l   | 94       |
| 30) Chloroform               | 4.281          | 83   | 86066      | 5.308    | ug/l   | 99       |
| 31) Cyclohexane              | 4.580          | 56   | 69326      | 5.450    | ug/l # | 94       |
| 32) 1,1,1-Trichloroethane    | 4.513          | 97   | 74539      | 5.310    | ug/l   | 95       |
| 36) 1,1-Dichloropropene      | 4.744          | 75   | 44786      | 4.379    | ug/l   | 99       |
| 37) Ethyl Acetate            | 3.995          | 43   | 60285      | 4.805    | ug/l # | 73       |
| 38) Carbon Tetrachloride     | 4.732          | 117  | 61498      | 4.992    | ug/l   | 96       |
| 39) Methylcyclohexane        | 6.047          | 83   | 56805      | 4.633    | ug/l   | 92       |
| 40) Benzene                  | 5.014          | 78   | 163948     | 4.959    | ug/l   | 100      |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
 Data File : VV039142.D  
 Acq On : 22 Sep 2025 15:37  
 Operator : SY/MD  
 Sample : VSTDICC005  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VSTDICC005

Manual Integrations  
 APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025  
 Supervised By :Semsettin Yesilyurt 09/26/2025

Quant Time: Sep 24 07:25:48 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 07:14:23 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|--------|--------|----------|
| 41) Methacrylonitrile         | 4.188  | 41   | 17959    | 4.995  | ug/l # | 86       |
| 42) 1,2-Dichloroethane        | 5.047  | 62   | 57857    | 4.994  | ug/l   | 96       |
| 43) Isopropyl Acetate         | 5.204  | 43   | 77935    | 4.509  | ug/l # | 93       |
| 44) Trichloroethene           | 5.838  | 130  | 38395    | 5.033  | ug/l   | 89       |
| 45) 1,2-Dichloropropane       | 6.098  | 63   | 39327    | 4.786  | ug/l   | 100      |
| 46) Dibromomethane            | 6.233  | 93   | 31863    | 5.066  | ug/l   | 98       |
| 47) Bromodichloromethane      | 6.429  | 83   | 64039    | 4.925  | ug/l   | 95       |
| 48) Methyl methacrylate       | 6.313  | 41   | 29705    | 3.757  | ug/l   | 93       |
| 49) 1,4-Dioxane               | 6.307  | 88   | 8999     | 72.413 | ug/l # | 79       |
| 51) 4-Methyl-2-Pentanone      | 7.153  | 43   | 255889   | 23.155 | ug/l   | 99       |
| 52) Toluene                   | 7.313  | 92   | 94414    | 4.925  | ug/l   | 99       |
| 53) t-1,3-Dichloropropene     | 7.583  | 75   | 53878    | 4.755  | ug/l   | 94       |
| 54) cis-1,3-Dichloropropene   | 6.957  | 75   | 58800    | 4.700  | ug/l   | 98       |
| 55) 1,1,2-Trichloroethane     | 7.764  | 97   | 43788    | 5.094  | ug/l   | 99       |
| 56) Ethyl methacrylate        | 7.722  | 69   | 42598    | 4.173  | ug/l   | 97       |
| 57) 1,3-Dichloropropane       | 7.937  | 76   | 63818    | 4.705  | ug/l   | 99       |
| 58) 2-Chloroethyl Vinyl ether | 6.818  | 63   | 81984    | 23.077 | ug/l   | 100      |
| 59) 2-Hexanone                | 8.069  | 43   | 179185   | 22.071 | ug/l   | 88       |
| 60) Dibromochloromethane      | 8.169  | 129  | 49953    | 5.100  | ug/l   | 96       |
| 61) 1,2-Dibromoethane         | 8.278  | 107  | 44729    | 5.088  | ug/l   | 98       |
| 64) Tetrachloroethene         | 7.899  | 164  | 34795    | 5.143  | ug/l   | 98       |
| 65) Chlorobenzene             | 8.805  | 112  | 114424   | 5.086  | ug/l   | 98       |
| 66) 1,1,1,2-Tetrachloroethane | 8.899  | 131  | 40690    | 5.005  | ug/l   | 95       |
| 67) Ethyl Benzene             | 8.940  | 91   | 157790   | 4.633  | ug/l   | 97       |
| 68) m/p-Xylenes               | 9.066  | 106  | 115177   | 8.762  | ug/l   | 98       |
| 69) o-Xylene                  | 9.471  | 106  | 51281    | 4.406  | ug/l   | 100      |
| 70) Styrene                   | 9.490  | 104  | 83873    | 4.836  | ug/l   | 96       |
| 71) Bromoform                 | 9.657  | 173  | 31133    | 4.903  | ug/l # | 97       |
| 73) Isopropylbenzene          | 9.857  | 105  | 140251   | 4.545  | ug/l   | 99       |
| 74) N-amyl acetate            | 9.731  | 43   | 54436    | 4.422  | ug/l   | 98       |
| 75) 1,1,2,2-Tetrachloroethane | 10.165 | 83   | 71656    | 5.299  | ug/l   | 99       |
| 76) 1,2,3-Trichloropropane    | 10.201 | 75   | 51079m   | 5.273  | ug/l   |          |
| 77) Bromobenzene              | 10.146 | 156  | 39428    | 4.924  | ug/l   | 95       |
| 78) n-propylbenzene           | 10.281 | 91   | 168794   | 4.492  | ug/l   | 99       |
| 79) 2-Chlorotoluene           | 10.349 | 91   | 110002   | 4.892  | ug/l   | 99       |
| 80) 1,3,5-Trimethylbenzene    | 10.468 | 105  | 114749   | 4.474  | ug/l   | 98       |
| 81) trans-1,4-Dichloro-2-b... | 9.934  | 75   | 18154    | 4.499  | ug/l   | 90       |
| 82) 4-Chlorotoluene           | 10.468 | 91   | 122429   | 4.695  | ug/l   | 99       |
| 83) tert-Butylbenzene         | 10.789 | 119  | 117310   | 4.586  | ug/l   | 99       |
| 84) 1,2,4-Trimethylbenzene    | 10.841 | 105  | 108157   | 4.343  | ug/l   | 96       |
| 85) sec-Butylbenzene          | 11.014 | 105  | 153526   | 4.525  | ug/l   | 99       |
| 86) p-Isopropyltoluene        | 11.169 | 119  | 127377   | 4.511  | ug/l   | 98       |
| 87) 1,3-Dichlorobenzene       | 11.111 | 146  | 82673    | 5.102  | ug/l   | 98       |
| 88) 1,4-Dichlorobenzene       | 11.198 | 146  | 91001    | 5.082  | ug/l   | 94       |
| 89) n-Butylbenzene            | 11.583 | 91   | 119822   | 4.443  | ug/l   | 96       |
| 90) Hexachloroethane          | 11.818 | 117  | 35762    | 5.143  | ug/l   | 95       |
| 91) 1,2-Dichlorobenzene       | 11.570 | 146  | 76477    | 5.052  | ug/l   | 98       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.358 | 75   | 14163    | 5.211  | ug/l   | 92       |
| 93) 1,2,4-Trichlorobenzene    | 13.191 | 180  | 39775    | 4.583  | ug/l   | 100      |
| 94) Hexachlorobutadiene       | 13.368 | 225  | 22549    | 5.001  | ug/l   | 98       |
| 95) Naphthalene               | 13.429 | 128  | 113395   | 4.062  | ug/l   | 99       |
| 96) 1,2,3-Trichlorobenzene    | 13.670 | 180  | 38315    | 4.317  | ug/l   | 96       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
Data File : VV039142.D  
Acq On : 22 Sep 2025 15:37  
Operator : SY/MD  
Sample : VSTDICC005  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VSTDICC005

Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025  
Supervised By :Semsettin Yesilyurt 09/26/2025

Quant Time: Sep 24 07:25:48 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 07:14:23 2025  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|


-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

**Instrument :**  
MSVOA\_V  
**ClientSampleId :**  
VSTDICC005

## Manual Integrations

Reviewed By :Mahesh Dadoda 09/26/2025  
Supervised By :Semsettin Yesilyurt 09/26/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
 Data File : VV039143.D  
 Acq On : 22 Sep 2025 16:35  
 Operator : SY/MD  
 Sample : VSTDICC001  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VSTDICC001

Manual Integrations  
 APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025  
 Supervised By :Semsettin Yesilyurt 09/26/2025

Quant Time: Sep 24 07:26:45 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 07:14:23 2025  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc    | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|---------|--------|----------|
| -----                        |                |      |            |         |        |          |
| Internal Standards           |                |      |            |         |        |          |
| 1) Pentafluorobenzene        | 4.603          | 168  | 406498     | 50.000  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 5.535          | 114  | 735855     | 50.000  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 8.776          | 117  | 700418     | 50.000  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 11.175         | 152  | 322677     | 50.000  | ug/l   | 0.00     |
|                              |                |      |            |         |        |          |
| System Monitoring Compounds  |                |      |            |         |        |          |
| 33) 1,2-Dichloroethane-d4    | 0.000          | 65   | 0d         | 0.000   | ug/l   |          |
| Spiked Amount 50.000         | Range 81 - 118 |      | Recovery = | 0.000%# |        |          |
| 35) Dibromofluoromethane     | 0.000          | 113  | 0d         | 0.000   | ug/l   |          |
| Spiked Amount 50.000         | Range 80 - 119 |      | Recovery = | 0.000%# |        |          |
| 50) Toluene-d8               | 0.000          | 98   | 0d         | 0.000   | ug/l   |          |
| Spiked Amount 50.000         | Range 89 - 112 |      | Recovery = | 0.000%# |        |          |
| 62) 4-Bromofluorobenzene     | 0.000          | 95   | 0d         | 0.000   | ug/l   |          |
| Spiked Amount 50.000         | Range 85 - 114 |      | Recovery = | 0.000%# |        |          |
|                              |                |      |            |         |        |          |
| Target Compounds             |                |      |            |         | Qvalue |          |
| 2) Dichlorodifluoromethane   | 1.121          | 85   | 7593       | 1.180   | ug/l   | 98       |
| 3) Chloromethane             | 1.230          | 50   | 7695       | 1.114   | ug/l   | 99       |
| 4) Vinyl Chloride            | 1.297          | 62   | 8342       | 1.128   | ug/l   | 98       |
| 6) Chloroethane              | 1.564          | 64   | 7440       | 0.673   | ug/l   | 96       |
| 7) Trichlorofluoromethane    | 1.725          | 101  | 12570      | 1.174   | ug/l   | 85       |
| 8) Diethyl Ether             | 1.921          | 74   | 4650       | 1.170   | ug/l   | 92       |
| 9) 1,1,2-Trichlorotrifluo... | 2.082          | 101  | 7630       | 1.198   | ug/l   | 96       |
| 12) 1,1-Dichloroethene       | 2.085          | 96   | 7094       | 1.157   | ug/l # | 66       |
| 14) Allyl chloride           | 2.362          | 41   | 12383      | 1.131   | ug/l   | 97       |
| 15) Acrylonitrile            | 2.683          | 53   | 23873      | 5.989   | ug/l   | 96       |
| 16) Acetone                  | 2.127          | 43   | 28927      | 3.224   | ug/l   | 96       |
| 17) Carbon Disulfide         | 2.256          | 76   | 21708      | 1.158   | ug/l   | 98       |
| 18) Methyl Acetate           | 2.391          | 43   | 9742       | 1.118   | ug/l   | 98       |
| 19) Methyl tert-butyl Ether  | 2.715          | 73   | 21063      | 1.078   | ug/l # | 89       |
| 20) Methylene Chloride       | 2.458          | 84   | 16152      | 0.632   | ug/l   | 90       |
| 21) trans-1,2-Dichloroethene | 2.709          | 96   | 7797       | 1.195   | ug/l   | 92       |
| 22) Diisopropyl ether        | 3.220          | 45   | 21620      | 1.032   | ug/l # | 34       |
| 23) Vinyl Acetate            | 3.204          | 43   | 91192      | 4.977   | ug/l   | 99       |
| 24) 1,1-Dichloroethane       | 3.121          | 63   | 15790      | 1.233   | ug/l # | 94       |
| 25) 2-Butanone               | 3.921          | 43   | 22515m     | 4.862   | ug/l   |          |
| 26) 2,2-Dichloropropane      | 3.815          | 77   | 15492m     | 1.369   | ug/l   |          |
| 27) cis-1,2-Dichloroethene   | 3.834          | 96   | 8138       | 1.131   | ug/l   | 91       |
| 28) Bromochloromethane       | 4.162          | 49   | 6039       | 1.068   | ug/l # | 53       |
| 29) Tetrahydrofuran          | 4.262          | 42   | 13621      | 4.619   | ug/l # | 65       |
| 30) Chloroform               | 4.288          | 83   | 15288      | 1.143   | ug/l   | 96       |
| 32) 1,1,1-Trichloroethane    | 4.519          | 97   | 12926      | 1.116   | ug/l   | 93       |
| 36) 1,1-Dichloropropene      | 4.764          | 75   | 9572m      | 1.063   | ug/l   |          |
| 37) Ethyl Acetate            | 4.030          | 43   | 12136m     | 1.098   | ug/l   |          |
| 38) Carbon Tetrachloride     | 4.735          | 117  | 13098m     | 1.207   | ug/l   |          |
| 39) Methylcyclohexane        | 6.050          | 83   | 9159       | 0.848   | ug/l   | 91       |
| 40) Benzene                  | 5.027          | 78   | 28487      | 0.978   | ug/l   | 98       |
| 41) Methacrylonitrile        | 4.223          | 41   | 2717m      | 2.427   | ug/l   |          |
| 42) 1,2-Dichloroethane       | 5.059          | 62   | 10892      | 1.067   | ug/l   | 95       |
| 43) Isopropyl Acetate        | 5.236          | 43   | 12786m     | 0.840   | ug/l   |          |
| 44) Trichloroethene          | 5.850          | 130  | 6408       | 0.953   | ug/l   | 84       |
| 45) 1,2-Dichloropropane      | 6.104          | 63   | 6838       | 0.945   | ug/l   | 93       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
 Data File : VV039143.D  
 Acq On : 22 Sep 2025 16:35  
 Operator : SY/MD  
 Sample : VSTDICC001  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VSTDICC001

Manual Integrations  
 APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025  
 Supervised By :Semsettin Yesilyurt 09/26/2025

Quant Time: Sep 24 07:26:45 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 07:14:23 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|--------|--------|----------|
| 46) Dibromomethane            | 6.249  | 93   | 5835     | 1.053  | ug/l   | 89       |
| 47) Bromodichloromethane      | 6.439  | 83   | 12206    | 1.066  | ug/l # | 90       |
| 48) Methyl methacrylate       | 6.371  | 41   | 7232m    | 1.038  | ug/l   |          |
| 49) 1,4-Dioxane               | 6.333  | 88   | 2721m    | 24.855 | ug/l   |          |
| 51) 4-Methyl-2-Pentanone      | 7.165  | 43   | 33797    | 3.472  | ug/l   | 96       |
| 52) Toluene                   | 7.326  | 92   | 13265    | 0.786  | ug/l   | 87       |
| 53) t-1,3-Dichloropropene     | 7.606  | 75   | 8060     | 0.807  | ug/l   | 90       |
| 54) cis-1,3-Dichloropropene   | 6.973  | 75   | 8754     | 0.794  | ug/l # | 93       |
| 55) 1,1,2-Trichloroethane     | 7.776  | 97   | 8140     | 1.075  | ug/l   | 91       |
| 56) Ethyl methacrylate        | 7.747  | 69   | 7455m    | 0.829  | ug/l   |          |
| 57) 1,3-Dichloropropane       | 7.956  | 76   | 11732    | 0.982  | ug/l   | 94       |
| 58) 2-Chloroethyl Vinyl ether | 6.841  | 63   | 10011    | 11.289 | ug/l   | 98       |
| 59) 2-Hexanone                | 8.091  | 43   | 17015    | 4.307  | ug/l   | 98       |
| 60) Dibromochloromethane      | 8.172  | 129  | 9122     | 1.057  | ug/l   | 87       |
| 61) 1,2-Dibromoethane         | 8.288  | 107  | 7129     | 0.921  | ug/l   | 90       |
| 64) Tetrachloroethene         | 7.899  | 164  | 5814     | 0.991  | ug/l   | 90       |
| 65) Chlorobenzene             | 8.812  | 112  | 20007    | 1.025  | ug/l # | 88       |
| 66) 1,1,1,2-Tetrachloroethane | 8.902  | 131  | 7603     | 1.079  | ug/l # | 61       |
| 67) Ethyl Benzene             | 8.947  | 91   | 24171    | 0.818  | ug/l   | 94       |
| 68) m/p-Xylenes               | 9.072  | 106  | 16239    | 1.425  | ug/l   | 98       |
| 69) o-Xylene                  | 9.477  | 106  | 7250     | 0.718  | ug/l   | 79       |
| 70) Styrene                   | 9.503  | 104  | 11594    | 2.083  | ug/l # | 80       |
| 71) Bromoform                 | 9.667  | 173  | 5558     | 1.009  | ug/l # | 94       |
| 73) Isopropylbenzene          | 9.860  | 105  | 20820    | 0.890  | ug/l   | 98       |
| 74) N-amyl acetate            | 9.747  | 43   | 6654     | 0.713  | ug/l # | 91       |
| 75) 1,1,2,2-Tetrachloroethane | 10.169 | 83   | 12746    | 1.243  | ug/l   | 95       |
| 76) 1,2,3-Trichloropropane    | 10.207 | 75   | 8871     | 1.207  | ug/l   | 100      |
| 77) Bromobenzene              | 10.159 | 156  | 6311     | 1.039  | ug/l   | 93       |
| 78) n-propylbenzene           | 10.284 | 91   | 24996    | 0.877  | ug/l   | 96       |
| 79) 2-Chlorotoluene           | 10.355 | 91   | 13640    | 0.800  | ug/l   | 89       |
| 80) 1,3,5-Trimethylbenzene    | 10.471 | 105  | 15886    | 0.817  | ug/l   | 96       |
| 82) 4-Chlorotoluene           | 10.474 | 91   | 15363    | 0.777  | ug/l   | 91       |
| 83) tert-Butylbenzene         | 10.789 | 119  | 18552    | 0.956  | ug/l   | 99       |
| 84) 1,2,4-Trimethylbenzene    | 10.847 | 105  | 15048    | 0.797  | ug/l   | 95       |
| 85) sec-Butylbenzene          | 11.014 | 105  | 23268    | 0.904  | ug/l   | 98       |
| 86) p-Isopropyltoluene        | 11.168 | 119  | 18549    | 0.866  | ug/l   | 86       |
| 87) 1,3-Dichlorobenzene       | 11.117 | 146  | 12747    | 1.037  | ug/l   | 97       |
| 88) 1,4-Dichlorobenzene       | 11.201 | 146  | 16279m   | 1.199  | ug/l   |          |
| 89) n-Butylbenzene            | 11.586 | 91   | 17079    | 0.835  | ug/l   | 94       |
| 90) Hexachloroethane          | 11.818 | 117  | 6736     | 1.277  | ug/l   | 94       |
| 91) 1,2-Dichlorobenzene       | 11.577 | 146  | 12486    | 1.088  | ug/l   | 98       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.361 | 75   | 3071     | 0.690  | ug/l   | 75       |
| 93) 1,2,4-Trichlorobenzene    | 13.197 | 180  | 6637     | 1.008  | ug/l   | 99       |
| 94) Hexachlorobutadiene       | 13.368 | 225  | 4406     | 1.288  | ug/l   | 95       |
| 95) Naphthalene               | 13.438 | 128  | 19706m   | 0.931  | ug/l   |          |
| 96) 1,2,3-Trichlorobenzene    | 13.676 | 180  | 7184     | 1.067  | ug/l   | 97       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

**Instrument :**  
MSVOA\_V  
**ClientSampleId :**  
VSTDICC001

## Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025  
Supervised By :Semsettin Yesilyurt 09/26/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
 Data File : VV039144.D  
 Acq On : 22 Sep 2025 16:58  
 Operator : SY/MD  
 Sample : VSTDICV050  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 ICVV092225

Quant Time: Sep 24 08:10:27 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

### Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025  
 Supervised By :Semsettin Yesilyurt 09/26/2025

| Compound                     | R.T.           | QIon | Response   | Conc     | Units | Dev(Min) |
|------------------------------|----------------|------|------------|----------|-------|----------|
| -----                        |                |      |            |          |       |          |
| Internal Standards           |                |      |            |          |       |          |
| 1) Pentafluorobenzene        | 4.600          | 168  | 530687     | 50.000   | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene      | 5.535          | 114  | 821122     | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5         | 8.776          | 117  | 811351     | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 11.172         | 152  | 500088     | 50.000   | ug/l  | 0.00     |
| System Monitoring Compounds  |                |      |            |          |       |          |
| 33) 1,2-Dichloroethane-d4    | 4.940          | 65   | 368982     | 48.041   | ug/l  | 0.00     |
| Spiked Amount 50.000         | Range 81 - 118 |      | Recovery = | 96.080%  |       |          |
| 35) Dibromofluoromethane     | 4.500          | 113  | 335549     | 50.832   | ug/l  | 0.00     |
| Spiked Amount 50.000         | Range 80 - 119 |      | Recovery = | 101.660% |       |          |
| 50) Toluene-d8               | 7.236          | 98   | 1012009    | 52.199   | ug/l  | 0.00     |
| Spiked Amount 50.000         | Range 89 - 112 |      | Recovery = | 104.400% |       |          |
| 62) 4-Bromofluorobenzene     | 10.001         | 95   | 453523     | 56.823   | ug/l  | 0.00     |
| Spiked Amount 50.000         | Range 85 - 114 |      | Recovery = | 113.640% |       |          |
| Target Compounds             |                |      |            |          |       |          |
|                              |                |      |            |          |       | Qvalue   |
| 2) Dichlorodifluoromethane   | 1.121          | 85   | 380034     | 45.238   | ug/l  | 98       |
| 3) Chloromethane             | 1.230          | 50   | 419560     | 46.523   | ug/l  | 99       |
| 4) Vinyl Chloride            | 1.298          | 62   | 438971     | 45.467   | ug/l  | 100      |
| 5) Bromomethane              | 1.494          | 94   | 274408     | 45.085   | ug/l  | 100      |
| 6) Chloroethane              | 1.558          | 64   | 288768     | 50.873   | ug/l  | 97       |
| 7) Trichlorofluoromethane    | 1.725          | 101  | 649327     | 46.452   | ug/l  | 98       |
| 8) Diethyl Ether             | 1.918          | 74   | 239410     | 46.136   | ug/l  | 100      |
| 9) 1,1,2-Trichlorotrifluo... | 2.082          | 101  | 386460     | 46.465   | ug/l  | 98       |
| 10) Methyl Iodide            | 2.198          | 142  | 394299     | 49.783   | ug/l  | 100      |
| 11) Tert butyl alcohol       | 2.571          | 59   | 409358     | 252.173  | ug/l  | 99       |
| 12) 1,1-Dichloroethene       | 2.082          | 96   | 367440     | 45.922   | ug/l  | 99       |
| 13) Acrolein                 | 2.011          | 56   | 76596      | 271.925  | ug/l  | 95       |
| 14) Allyl chloride           | 2.359          | 41   | 673070     | 47.092   | ug/l  | 100      |
| 15) Acrylonitrile            | 2.674          | 53   | 1247029    | 239.637  | ug/l  | 100      |
| 16) Acetone                  | 2.117          | 43   | 1131047    | 237.325  | ug/l  | 99       |
| 17) Carbon Disulfide         | 2.252          | 76   | 1114937    | 45.563   | ug/l  | 99       |
| 18) Methyl Acetate           | 2.381          | 43   | 564211     | 49.579   | ug/l  | 99       |
| 19) Methyl tert-butyl Ether  | 2.712          | 73   | 1204021    | 47.223   | ug/l  | 100      |
| 20) Methylene Chloride       | 2.458          | 84   | 426790     | 51.536   | ug/l  | 99       |
| 21) trans-1,2-Dichloroethene | 2.703          | 96   | 391103     | 45.915   | ug/l  | 100      |
| 22) Diisopropyl ether        | 3.217          | 45   | 1355158    | 49.556   | ug/l  | 95       |
| 23) Vinyl Acetate            | 3.188          | 43   | 5946893    | 248.597  | ug/l  | 100      |
| 24) 1,1-Dichloroethane       | 3.117          | 63   | 761495     | 45.536   | ug/l  | 100      |
| 25) 2-Butanone               | 3.867          | 43   | 1611713    | 262.050  | ug/l  | 99       |
| 26) 2,2-Dichloropropane      | 3.812          | 77   | 648277     | 43.721   | ug/l  | 99       |
| 27) cis-1,2-Dichloroethene   | 3.818          | 96   | 459958     | 48.960   | ug/l  | 99       |
| 28) Bromochloromethane       | 4.146          | 49   | 384995     | 52.149   | ug/l  | 100      |
| 29) Tetrahydrofuran          | 4.236          | 42   | 1064545    | 276.528  | ug/l  | 99       |
| 30) Chloroform               | 4.278          | 83   | 820367     | 46.980   | ug/l  | 99       |
| 31) Cyclohexane              | 4.584          | 56   | 654806     | 47.799   | ug/l  | 99       |
| 32) 1,1,1-Trichloroethane    | 4.513          | 97   | 709351     | 46.922   | ug/l  | 99       |
| 36) 1,1-Dichloropropene      | 4.744          | 75   | 537027     | 53.809   | ug/l  | 99       |
| 37) Ethyl Acetate            | 3.976          | 43   | 630342     | 52.050   | ug/l  | 99       |
| 38) Carbon Tetrachloride     | 4.735          | 117  | 603187     | 49.452   | ug/l  | 99       |
| 39) Methylcyclohexane        | 6.047          | 83   | 692131     | 57.426   | ug/l  | 98       |
| 40) Benzene                  | 5.008          | 78   | 1699505    | 52.293   | ug/l  | 98       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
 Data File : VV039144.D  
 Acq On : 22 Sep 2025 16:58  
 Operator : SY/MD  
 Sample : VSTDICV050  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 ICVV092225

Quant Time: Sep 24 08:10:27 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

### Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025  
 Supervised By :Semsettin Yesilyurt 09/26/2025

| Compound                      | R.T.   | QIon | Response | Conc     | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|----------|--------|----------|
| 41) Methacrylonitrile         | 4.159  | 41   | 282895   | 51.602   | ug/l   | 99       |
| 42) 1,2-Dichloroethane        | 5.040  | 62   | 583550   | 51.245   | ug/l   | 99       |
| 43) Isopropyl Acetate         | 5.194  | 43   | 924654   | 55.998   | ug/l   | 99       |
| 44) Trichloroethene           | 5.831  | 130  | 384092   | 51.216   | ug/l   | 95       |
| 45) 1,2-Dichloropropane       | 6.088  | 63   | 440211   | 54.495   | ug/l   | 97       |
| 46) Dibromomethane            | 6.223  | 93   | 316032   | 51.116   | ug/l   | 97       |
| 47) Bromodichloromethane      | 6.426  | 83   | 654836   | 51.232   | ug/l   | 99       |
| 48) Methyl methacrylate       | 6.297  | 41   | 442395   | 59.104   | ug/l   | 98       |
| 49) 1,4-Dioxane               | 6.291  | 88   | 136899   | 1133.346 | ug/l   | 99       |
| 51) 4-Methyl-2-Pentanone      | 7.149  | 43   | 3292794  | 303.119  | ug/l   | 100      |
| 52) Toluene                   | 7.307  | 92   | 1064972  | 56.517   | ug/l   | 99       |
| 53) t-1,3-Dichloropropene     | 7.574  | 75   | 621814   | 55.823   | ug/l   | 99       |
| 54) cis-1,3-Dichloropropene   | 6.947  | 75   | 690532   | 56.154   | ug/l   | 98       |
| 55) 1,1,2-Trichloroethane     | 7.760  | 97   | 432856   | 51.231   | ug/l   | 99       |
| 56) Ethyl methacrylate        | 7.715  | 69   | 610159   | 60.810   | ug/l   | 99       |
| 57) 1,3-Dichloropropane       | 7.934  | 76   | 716955   | 53.769   | ug/l   | 100      |
| 58) 2-Chloroethyl Vinyl ether | 6.812  | 63   | 1474938  | 259.936  | ug/l   | 99       |
| 59) 2-Hexanone                | 8.063  | 43   | 2462493  | 280.545  | ug/l   | 100      |
| 60) Dibromochloromethane      | 8.169  | 129  | 499908   | 51.926   | ug/l   | 99       |
| 61) 1,2-Dibromoethane         | 8.272  | 107  | 460510   | 53.295   | ug/l   | 100      |
| 64) Tetrachloroethene         | 7.895  | 164  | 346271   | 50.954   | ug/l   | 99       |
| 65) Chlorobenzene             | 8.802  | 112  | 1153503  | 51.041   | ug/l   | 99       |
| 66) 1,1,1,2-Tetrachloroethane | 8.899  | 131  | 409401   | 50.136   | ug/l   | 100      |
| 67) Ethyl Benzene             | 8.937  | 91   | 1964943  | 57.438   | ug/l   | 100      |
| 68) m/p-Xylenes               | 9.062  | 106  | 1592118  | 120.582  | ug/l   | 100      |
| 69) o-Xylene                  | 9.468  | 106  | 706301   | 60.414   | ug/l   | 98       |
| 70) Styrene                   | 9.484  | 104  | 1306032  | 52.671   | ug/l   | 99       |
| 71) Bromoform                 | 9.654  | 173  | 335410   | 52.585   | ug/l # | 98       |
| 73) Isopropylbenzene          | 9.857  | 105  | 1966280  | 54.210   | ug/l   | 100      |
| 74) N-amyl acetate            | 9.725  | 43   | 841099   | 58.135   | ug/l   | 99       |
| 75) 1,1,2,2-Tetrachloroethane | 10.165 | 83   | 724860   | 45.610   | ug/l   | 100      |
| 76) 1,2,3-Trichloropropane    | 10.197 | 75   | 563785m  | 49.515   | ug/l   |          |
| 77) Bromobenzene              | 10.140 | 156  | 462931   | 49.193   | ug/l   | 99       |
| 78) n-propylbenzene           | 10.278 | 91   | 2444658  | 55.347   | ug/l   | 99       |
| 79) 2-Chlorotoluene           | 10.349 | 91   | 1412046  | 53.426   | ug/l   | 99       |
| 80) 1,3,5-Trimethylbenzene    | 10.464 | 105  | 1709353  | 56.698   | ug/l   | 99       |
| 81) trans-1,4-Dichloro-2-b... | 9.927  | 75   | 247871   | 52.262   | ug/l   | 99       |
| 82) 4-Chlorotoluene           | 10.461 | 91   | 1680201  | 54.824   | ug/l   | 100      |
| 83) tert-Butylbenzene         | 10.789 | 119  | 1610870  | 53.584   | ug/l   | 100      |
| 84) 1,2,4-Trimethylbenzene    | 10.837 | 105  | 1695994  | 57.944   | ug/l   | 100      |
| 85) sec-Butylbenzene          | 11.011 | 105  | 2189540  | 54.911   | ug/l   | 100      |
| 86) p-Isopropyltoluene        | 11.165 | 119  | 1841747  | 55.495   | ug/l   | 100      |
| 87) 1,3-Dichlorobenzene       | 11.104 | 146  | 939483   | 49.328   | ug/l   | 99       |
| 88) 1,4-Dichlorobenzene       | 11.194 | 146  | 971665   | 46.502   | ug/l   | 99       |
| 89) n-Butylbenzene            | 11.580 | 91   | 1798412  | 56.740   | ug/l   | 100      |
| 90) Hexachloroethane          | 11.818 | 117  | 370188   | 45.296   | ug/l   | 98       |
| 91) 1,2-Dichlorobenzene       | 11.567 | 146  | 878347   | 49.362   | ug/l   | 99       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.349 | 75   | 150399   | 52.479   | ug/l   | 98       |
| 93) 1,2,4-Trichlorobenzene    | 13.185 | 180  | 533045   | 52.259   | ug/l   | 99       |
| 94) Hexachlorobutadiene       | 13.368 | 225  | 243496   | 45.942   | ug/l   | 99       |
| 95) Naphthalene               | 13.422 | 128  | 1919772  | 58.231   | ug/l   | 100      |
| 96) 1,2,3-Trichlorobenzene    | 13.667 | 180  | 549079   | 52.633   | ug/l   | 98       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
Data File : VV039144.D  
Acq On : 22 Sep 2025 16:58  
Operator : SY/MD  
Sample : VSTDICV050  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 12 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
ICVVV092225

Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025  
Supervised By :Semsettin Yesilyurt 09/26/2025

Quant Time: Sep 24 08:10:27 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

-----  
(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VW092225\  
Data File : VW039144.D  
Acq On    : 22 Sep 2025  16:58  
Operator  : SY/MD  
Sample    : VSTDICV050  
Misc      : 5.0mL/MSVOA_V/WATER  
ALS Vial  : 12   Sample Multiplier: 1
```

**Instrument :**  
MSVOA\_V  
**ClientSampleId :**  
ICVV092225

## Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 09/26/2025  
Supervised By :Semsettin Yesilyurt 09/26/2025

Quant Time: Sep 24 08:10:27 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
 Data File : VV039144.D  
 Acq On : 22 Sep 2025 16:58  
 Operator : SY/MD  
 Sample : VSTDICV050  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 ICSVV092225

Quant Time: Sep 24 08:10:27 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 20% Max. Rel. Area : 150%

|       | Compound                    | AvgRF | CCRF  | %Dev   | Area% | Dev(min) |
|-------|-----------------------------|-------|-------|--------|-------|----------|
| 1 I   | Pentafluorobenzene          | 1.000 | 1.000 | 0.0    | 94    | 0.00     |
| 2 T   | Dichlorodifluoromethane     | 0.791 | 0.716 | 9.5    | 91    | 0.00     |
| 3 P   | Chloromethane               | 0.850 | 0.791 | 6.9    | 97    | 0.00     |
| 4 C   | Vinyl Chloride              | 0.910 | 0.827 | 9.1#   | 93    | 0.00     |
| 5 T   | Bromomethane                | 0.573 | 0.517 | 9.8    | 96    | 0.00     |
| 6 T   | Chloroethane                | 0.631 | 0.544 | 13.8   | 96    | 0.00     |
| 7 T   | Trichlorofluoromethane      | 1.317 | 1.224 | 7.1    | 97    | 0.00     |
| 8 T   | Diethyl Ether               | 0.489 | 0.451 | 7.8    | 95    | 0.00     |
| 9 T   | 1,1,2-Trichlorotrifluoroeth | 0.784 | 0.728 | 7.1    | 98    | 0.00     |
| 10 T  | Methyl Iodide               | 0.746 | 0.743 | 0.4    | 89    | 0.00     |
| 11 T  | Tert butyl alcohol          | 0.153 | 0.154 | -0.7   | 106   | 0.00     |
| 12 CM | 1,1-Dichloroethene          | 0.754 | 0.692 | 8.2#   | 97    | 0.00     |
| 13 T  | Acrolein                    | 0.027 | 0.029 | -7.4   | 103   | 0.00     |
| 14 T  | Allyl chloride              | 1.347 | 1.268 | 5.9    | 96    | 0.00     |
| 15 T  | Acrylonitrile               | 0.490 | 0.470 | 4.1    | 101   | 0.00     |
| 16 T  | Acetone                     | 0.523 | 0.426 | 18.5   | 89    | 0.00     |
| 17 T  | Carbon Disulfide            | 2.306 | 2.101 | 8.9    | 95    | 0.00     |
| 18 T  | Methyl Acetate              | 1.072 | 1.063 | 0.8    | 102   | 0.00     |
| 19 T  | Methyl tert-butyl Ether     | 2.402 | 2.269 | 5.5    | 96    | 0.00     |
| 20 T  | Methylene Chloride          | 1.066 | 0.804 | 24.6#  | 97    | 0.00     |
| 21 T  | trans-1,2-Dichloroethene    | 0.803 | 0.737 | 8.2    | 96    | 0.00     |
| 22 T  | Diisopropyl ether           | 2.576 | 2.554 | 0.9    | 95    | 0.00     |
| 23 T  | Vinyl Acetate               | 2.254 | 2.241 | 0.6    | 96    | 0.00     |
| 24 P  | 1,1-Dichloroethane          | 1.576 | 1.435 | 8.9    | 97    | 0.00     |
| 25 T  | 2-Butanone                  | 0.579 | 0.607 | -4.8   | 98    | 0.00     |
| 26 T  | 2,2-Dichloropropane         | 1.397 | 1.222 | 12.5   | 93    | 0.00     |
| 27 T  | cis-1,2-Dichloroethene      | 0.885 | 0.867 | 2.0    | 94    | 0.00     |
| 28 T  | Bromochloromethane          | 0.696 | 0.725 | -4.2   | 101   | 0.00     |
| 29 T  | Tetrahydrofuran             | 0.363 | 0.401 | -10.5  | 102   | 0.00     |
| 30 C  | Chloroform                  | 1.645 | 1.546 | 6.0#   | 98    | 0.00     |
| 31 T  | Cyclohexane                 | 1.291 | 1.234 | 4.4    | 93    | 0.00     |
| 32 T  | 1,1,1-Trichloroethane       | 1.424 | 1.337 | 6.1    | 97    | 0.00     |
| 33 S  | 1,2-Dichloroethane-d4       | 0.724 | 0.695 | 4.0    | 101   | 0.00     |
| 34 I  | 1,4-Difluorobenzene         | 1.000 | 1.000 | 0.0    | 92    | 0.00     |
| 35 S  | Dibromofluoromethane        | 0.402 | 0.409 | -1.7   | 99    | 0.00     |
| 36 T  | 1,1-Dichloropropene         | 0.608 | 0.654 | -7.6   | 97    | 0.00     |
| 37 T  | Ethyl Acetate               | 0.737 | 0.768 | -4.2   | 99    | 0.00     |
| 38 T  | Carbon Tetrachloride        | 0.743 | 0.735 | 1.1    | 96    | 0.00     |
| 39 T  | Methylcyclohexane           | 0.734 | 0.843 | -14.9  | 94    | 0.00     |
| 40 TM | Benzene                     | 1.979 | 2.070 | -4.6   | 96    | 0.00     |
| 41 T  | Methacrylonitrile           | 0.273 | 0.345 | -26.4# | 96    | 0.00     |
| 42 TM | 1,2-Dichloroethane          | 0.693 | 0.711 | -2.6   | 98    | 0.00     |
| 43 T  | Isopropyl Acetate           | 1.005 | 1.126 | -12.0  | 97    | 0.00     |
| 44 TM | Trichloroethene             | 0.457 | 0.468 | -2.4   | 93    | 0.00     |
| 45 C  | 1,2-Dichloropropane         | 0.492 | 0.536 | -8.9#  | 98    | 0.00     |
| 46 T  | Dibromomethane              | 0.376 | 0.385 | -2.4   | 96    | 0.00     |
| 47 T  | Bromodichloromethane        | 0.778 | 0.797 | -2.4   | 97    | 0.00     |
| 48 T  | Methyl methacrylate         | 0.456 | 0.539 | -18.2  | 98    | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
 Data File : VV039144.D  
 Acq On : 22 Sep 2025 16:58  
 Operator : SY/MD  
 Sample : VSTDICV050  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 ICVV092225

Quant Time: Sep 24 08:10:27 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 20% Max. Rel. Area : 150%

|       | Compound                    | AvgRF | CCRF  | %Dev   | Area% | Dev(min) |
|-------|-----------------------------|-------|-------|--------|-------|----------|
| 49 T  | 1,4-Dioxane                 | 0.007 | 0.008 | -14.3  | 102   | 0.00     |
| 50 S  | Toluene-d8                  | 1.181 | 1.232 | -4.3   | 100   | 0.00     |
| 51 T  | 4-Methyl-2-Pentanone        | 0.661 | 0.802 | -21.3# | 101   | 0.00     |
| 52 CM | Toluene                     | 1.147 | 1.297 | -13.1# | 97    | 0.00     |
| 53 T  | t-1,3-Dichloropropene       | 0.678 | 0.757 | -11.7  | 96    | 0.00     |
| 54 T  | cis-1,3-Dichloropropene     | 0.749 | 0.841 | -12.3  | 95    | 0.00     |
| 55 T  | 1,1,2-Trichloroethane       | 0.514 | 0.527 | -2.5   | 98    | 0.00     |
| 56 T  | Ethyl methacrylate          | 0.611 | 0.743 | -21.6# | 95    | 0.00     |
| 57 T  | 1,3-Dichloropropane         | 0.812 | 0.873 | -7.5   | 98    | 0.00     |
| 58 T  | 2-Chloroethyl Vinyl ether   | 0.267 | 0.359 | -34.5# | 93    | 0.00     |
| 59 T  | 2-Hexanone                  | 0.463 | 0.600 | -29.6# | 101   | 0.00     |
| 60 T  | Dibromochloromethane        | 0.586 | 0.609 | -3.9   | 99    | 0.00     |
| 61 T  | 1,2-Dibromoethane           | 0.526 | 0.561 | -6.7   | 97    | 0.00     |
| 62 S  | 4-Bromofluorobenzene        | 0.486 | 0.552 | -13.6  | 100   | 0.00     |
| 63 I  | Chlorobenzene-d5            | 1.000 | 1.000 | 0.0    | 94    | 0.00     |
| 64 T  | Tetrachloroethene           | 0.419 | 0.427 | -1.9   | 97    | 0.00     |
| 65 PM | Chlorobenzene               | 1.393 | 1.422 | -2.1   | 97    | 0.00     |
| 66 T  | 1,1,1,2-Tetrachloroethane   | 0.503 | 0.505 | -0.4   | 97    | 0.00     |
| 67 C  | Ethyl Benzene               | 2.108 | 2.422 | -14.9# | 95    | 0.00     |
| 68 T  | m/p-Xylenes                 | 0.814 | 0.981 | -20.5# | 97    | 0.00     |
| 69 T  | o-Xylene                    | 0.720 | 0.871 | -21.0# | 97    | 0.00     |
| 70 T  | Styrene                     | 1.263 | 1.610 | -27.5# | 97    | 0.00     |
| 71 P  | Bromoform                   | 0.393 | 0.413 | -5.1   | 98    | 0.00     |
| 72 I  | 1,4-Dichlorobenzene-d4      | 1.000 | 1.000 | 0.0    | 97    | 0.00     |
| 73 T  | Isopropylbenzene            | 3.627 | 3.932 | -8.4   | 97    | 0.00     |
| 74 T  | N-amyl acetate              | 1.447 | 1.682 | -16.2  | 97    | 0.00     |
| 75 P  | 1,1,2,2-Tetrachloroethane   | 1.589 | 1.449 | 8.8    | 99    | 0.00     |
| 76 T  | 1,2,3-Trichloropropane      | 1.138 | 1.127 | 1.0    | 104   | 0.00     |
| 77 T  | Bromobenzene                | 0.941 | 0.926 | 1.6    | 97    | 0.00     |
| 78 T  | n-propylbenzene             | 4.416 | 4.888 | -10.7  | 97    | 0.00     |
| 79 T  | 2-Chlorotoluene             | 2.643 | 2.824 | -6.8   | 97    | 0.00     |
| 80 T  | 1,3,5-Trimethylbenzene      | 3.014 | 3.418 | -13.4  | 98    | 0.00     |
| 81 T  | trans-1,4-Dichloro-2-butene | 0.474 | 0.496 | -4.6   | 97    | 0.00     |
| 82 T  | 4-Chlorotoluene             | 3.064 | 3.360 | -9.7   | 98    | 0.00     |
| 83 T  | tert-Butylbenzene           | 3.006 | 3.221 | -7.2   | 97    | 0.00     |
| 84 T  | 1,2,4-Trimethylbenzene      | 2.926 | 3.391 | -15.9  | 97    | 0.00     |
| 85 T  | sec-Butylbenzene            | 3.987 | 4.378 | -9.8   | 97    | 0.00     |
| 86 T  | p-Isopropyltoluene          | 3.318 | 3.683 | -11.0  | 97    | 0.00     |
| 87 T  | 1,3-Dichlorobenzene         | 1.904 | 1.879 | 1.3    | 98    | 0.00     |
| 88 T  | 1,4-Dichlorobenzene         | 2.089 | 1.943 | 7.0    | 97    | 0.00     |
| 89 T  | n-Butylbenzene              | 3.169 | 3.596 | -13.5  | 96    | 0.00     |
| 90 T  | Hexachloroethane            | 0.817 | 0.740 | 9.4    | 99    | 0.00     |
| 91 T  | 1,2-Dichlorobenzene         | 1.779 | 1.756 | 1.3    | 98    | 0.00     |
| 92 T  | 1,2-Dibromo-3-Chloropropane | 0.330 | 0.301 | 8.8    | 102   | 0.00     |
| 93 T  | 1,2,4-Trichlorobenzene      | 1.020 | 1.066 | -4.5   | 96    | 0.00     |
| 94 T  | Hexachlorobutadiene         | 0.530 | 0.487 | 8.1    | 98    | 0.00     |
| 95 T  | Naphthalene                 | 3.296 | 3.839 | -16.5  | 98    | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
Data File : VV039144.D  
Acq On : 22 Sep 2025 16:58  
Operator : SY/MD  
Sample : VSTDICV050  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 12 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
ICVVV092225

Quant Time: Sep 24 08:10:27 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 20% Max. Rel. Area : 150%

|      | Compound               | AvgRF | CCRF  | %Dev | Area% | Dev(min) |
|------|------------------------|-------|-------|------|-------|----------|
| 96 T | 1,2,3-Trichlorobenzene | 1.043 | 1.098 | -5.3 | 96    | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
 Data File : VV039144.D  
 Acq On : 22 Sep 2025 16:58  
 Operator : SY/MD  
 Sample : VSTDICV050  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 ICVV092225

Quant Time: Sep 24 08:10:27 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 20% Max. Rel. Area : 150%

|       | Compound                    | Amount  | Calc.   | %Dev  | Area% | Dev(min) |
|-------|-----------------------------|---------|---------|-------|-------|----------|
| 1 I   | Pentafluorobenzene          | 50.000  | 50.000  | 0.0   | 94    | 0.00     |
| 2 T   | Dichlorodifluoromethane     | 50.000  | 45.238  | 9.5   | 91    | 0.00     |
| 3 P   | Chloromethane               | 50.000  | 46.523  | 7.0   | 97    | 0.00     |
| 4 C   | Vinyl Chloride              | 50.000  | 45.467  | 9.1#  | 93    | 0.00     |
| 5 T   | Bromomethane                | 50.000  | 45.085  | 9.8   | 96    | 0.00     |
| 6 T   | Chloroethane                | 50.000  | 50.873  | -1.7  | 96    | 0.00     |
| 7 T   | Trichlorofluoromethane      | 50.000  | 46.452  | 7.1   | 97    | 0.00     |
| 8 T   | Diethyl Ether               | 50.000  | 46.136  | 7.7   | 95    | 0.00     |
| 9 T   | 1,1,2-Trichlorotrifluoroeth | 50.000  | 46.465  | 7.1   | 98    | 0.00     |
| 10 T  | Methyl Iodide               | 50.000  | 49.783  | 0.4   | 89    | 0.00     |
| 11 T  | Tert butyl alcohol          | 250.000 | 252.173 | -0.9  | 106   | 0.00     |
| 12 CM | 1,1-Dichloroethene          | 50.000  | 45.922  | 8.2#  | 97    | 0.00     |
| 13 T  | Acrolein                    | 250.000 | 271.925 | -8.8  | 103   | 0.00     |
| 14 T  | Allyl chloride              | 50.000  | 47.092  | 5.8   | 96    | 0.00     |
| 15 T  | Acrylonitrile               | 250.000 | 239.637 | 4.1   | 101   | 0.00     |
| 16 T  | Acetone                     | 250.000 | 237.325 | 5.1   | 89    | 0.00     |
| 17 T  | Carbon Disulfide            | 50.000  | 45.563  | 8.9   | 95    | 0.00     |
| 18 T  | Methyl Acetate              | 50.000  | 49.579  | 0.8   | 102   | 0.00     |
| 19 T  | Methyl tert-butyl Ether     | 50.000  | 47.223  | 5.6   | 96    | 0.00     |
| 20 T  | Methylene Chloride          | 50.000  | 51.536  | -3.1  | 97    | 0.00     |
| 21 T  | trans-1,2-Dichloroethene    | 50.000  | 45.915  | 8.2   | 96    | 0.00     |
| 22 T  | Diisopropyl ether           | 50.000  | 49.556  | 0.9   | 95    | 0.00     |
| 23 T  | Vinyl Acetate               | 250.000 | 248.597 | 0.6   | 96    | 0.00     |
| 24 P  | 1,1-Dichloroethane          | 50.000  | 45.536  | 8.9   | 97    | 0.00     |
| 25 T  | 2-Butanone                  | 250.000 | 262.050 | -4.8  | 98    | 0.00     |
| 26 T  | 2,2-Dichloropropane         | 50.000  | 43.721  | 12.6  | 93    | 0.00     |
| 27 T  | cis-1,2-Dichloroethene      | 50.000  | 48.960  | 2.1   | 94    | 0.00     |
| 28 T  | Bromochloromethane          | 50.000  | 52.149  | -4.3  | 101   | 0.00     |
| 29 T  | Tetrahydrofuran             | 250.000 | 276.528 | -10.6 | 102   | 0.00     |
| 30 C  | Chloroform                  | 50.000  | 46.980  | 6.0#  | 98    | 0.00     |
| 31 T  | Cyclohexane                 | 50.000  | 47.799  | 4.4   | 93    | 0.00     |
| 32 T  | 1,1,1-Trichloroethane       | 50.000  | 46.922  | 6.2   | 97    | 0.00     |
| 33 S  | 1,2-Dichloroethane-d4       | 50.000  | 48.041  | 3.9   | 101   | 0.00     |
| 34 I  | 1,4-Difluorobenzene         | 50.000  | 50.000  | 0.0   | 92    | 0.00     |
| 35 S  | Dibromofluoromethane        | 50.000  | 50.832  | -1.7  | 99    | 0.00     |
| 36 T  | 1,1-Dichloropropene         | 50.000  | 53.809  | -7.6  | 97    | 0.00     |
| 37 T  | Ethyl Acetate               | 50.000  | 52.050  | -4.1  | 99    | 0.00     |
| 38 T  | Carbon Tetrachloride        | 50.000  | 49.452  | 1.1   | 96    | 0.00     |
| 39 T  | Methylcyclohexane           | 50.000  | 57.426  | -14.9 | 94    | 0.00     |
| 40 TM | Benzene                     | 50.000  | 52.293  | -4.6  | 96    | 0.00     |
| 41 T  | Methacrylonitrile           | 50.000  | 51.602  | -3.2  | 96    | 0.00     |
| 42 TM | 1,2-Dichloroethane          | 50.000  | 51.245  | -2.5  | 98    | 0.00     |
| 43 T  | Isopropyl Acetate           | 50.000  | 55.998  | -12.0 | 97    | 0.00     |
| 44 TM | Trichloroethene             | 50.000  | 51.216  | -2.4  | 93    | 0.00     |
| 45 C  | 1,2-Dichloropropane         | 50.000  | 54.495  | -9.0# | 98    | 0.00     |
| 46 T  | Dibromomethane              | 50.000  | 51.116  | -2.2  | 96    | 0.00     |
| 47 T  | Bromodichloromethane        | 50.000  | 51.232  | -2.5  | 97    | 0.00     |
| 48 T  | Methyl methacrylate         | 50.000  | 59.104  | -18.2 | 98    | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
 Data File : VV039144.D  
 Acq On : 22 Sep 2025 16:58  
 Operator : SY/MD  
 Sample : VSTDICV050  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 ICVV092225

Quant Time: Sep 24 08:10:27 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 20% Max. Rel. Area : 150%

|       | Compound                    | Amount   | Calc.    | %Dev   | Area% | Dev(min) |
|-------|-----------------------------|----------|----------|--------|-------|----------|
| 49 T  | 1,4-Dioxane                 | 1000.000 | 1133.346 | -13.3  | 102   | 0.00     |
| 50 S  | Toluene-d8                  | 50.000   | 52.199   | -4.4   | 100   | 0.00     |
| 51 T  | 4-Methyl-2-Pentanone        | 250.000  | 303.119  | -21.2# | 101   | 0.00     |
| 52 CM | Toluene                     | 50.000   | 56.517   | -13.0# | 97    | 0.00     |
| 53 T  | t-1,3-Dichloropropene       | 50.000   | 55.823   | -11.6  | 96    | 0.00     |
| 54 T  | cis-1,3-Dichloropropene     | 50.000   | 56.154   | -12.3  | 95    | 0.00     |
| 55 T  | 1,1,2-Trichloroethane       | 50.000   | 51.231   | -2.5   | 98    | 0.00     |
| 56 T  | Ethyl methacrylate          | 50.000   | 60.810   | -21.6# | 95    | 0.00     |
| 57 T  | 1,3-Dichloropropane         | 50.000   | 53.769   | -7.5   | 98    | 0.00     |
| 58 T  | 2-Chloroethyl Vinyl ether   | 250.000  | 259.936  | -4.0   | 93    | 0.00     |
| 59 T  | 2-Hexanone                  | 250.000  | 280.545  | -12.2  | 101   | 0.00     |
| 60 T  | Dibromochloromethane        | 50.000   | 51.926   | -3.9   | 99    | 0.00     |
| 61 T  | 1,2-Dibromoethane           | 50.000   | 53.295   | -6.6   | 97    | 0.00     |
| 62 S  | 4-Bromofluorobenzene        | 50.000   | 56.823   | -13.6  | 100   | 0.00     |
| 63 I  | Chlorobenzene-d5            | 50.000   | 50.000   | 0.0    | 94    | 0.00     |
| 64 T  | Tetrachloroethene           | 50.000   | 50.954   | -1.9   | 97    | 0.00     |
| 65 PM | Chlorobenzene               | 50.000   | 51.041   | -2.1   | 97    | 0.00     |
| 66 T  | 1,1,1,2-Tetrachloroethane   | 50.000   | 50.136   | -0.3   | 97    | 0.00     |
| 67 C  | Ethyl Benzene               | 50.000   | 57.438   | -14.9# | 95    | 0.00     |
| 68 T  | m/p-Xylenes                 | 100.000  | 120.582  | -20.6# | 97    | 0.00     |
| 69 T  | o-Xylene                    | 50.000   | 60.414   | -20.8# | 97    | 0.00     |
| 70 T  | Styrene                     | 50.000   | 52.671   | -5.3   | 97    | 0.00     |
| 71 P  | Bromoform                   | 50.000   | 52.585   | -5.2   | 98    | 0.00     |
| 72 I  | 1,4-Dichlorobenzene-d4      | 50.000   | 50.000   | 0.0    | 97    | 0.00     |
| 73 T  | Isopropylbenzene            | 50.000   | 54.210   | -8.4   | 97    | 0.00     |
| 74 T  | N-amyl acetate              | 50.000   | 58.135   | -16.3  | 97    | 0.00     |
| 75 P  | 1,1,2,2-Tetrachloroethane   | 50.000   | 45.610   | 8.8    | 99    | 0.00     |
| 76 T  | 1,2,3-Trichloropropane      | 50.000   | 49.515   | 1.0    | 104   | 0.00     |
| 77 T  | Bromobenzene                | 50.000   | 49.193   | 1.6    | 97    | 0.00     |
| 78 T  | n-propylbenzene             | 50.000   | 55.347   | -10.7  | 97    | 0.00     |
| 79 T  | 2-Chlorotoluene             | 50.000   | 53.426   | -6.9   | 97    | 0.00     |
| 80 T  | 1,3,5-Trimethylbenzene      | 50.000   | 56.698   | -13.4  | 98    | 0.00     |
| 81 T  | trans-1,4-Dichloro-2-butene | 50.000   | 52.262   | -4.5   | 97    | 0.00     |
| 82 T  | 4-Chlorotoluene             | 50.000   | 54.824   | -9.6   | 98    | 0.00     |
| 83 T  | tert-Butylbenzene           | 50.000   | 53.584   | -7.2   | 97    | 0.00     |
| 84 T  | 1,2,4-Trimethylbenzene      | 50.000   | 57.944   | -15.9  | 97    | 0.00     |
| 85 T  | sec-Butylbenzene            | 50.000   | 54.911   | -9.8   | 97    | 0.00     |
| 86 T  | p-Isopropyltoluene          | 50.000   | 55.495   | -11.0  | 97    | 0.00     |
| 87 T  | 1,3-Dichlorobenzene         | 50.000   | 49.328   | 1.3    | 98    | 0.00     |
| 88 T  | 1,4-Dichlorobenzene         | 50.000   | 46.502   | 7.0    | 97    | 0.00     |
| 89 T  | n-Butylbenzene              | 50.000   | 56.740   | -13.5  | 96    | 0.00     |
| 90 T  | Hexachloroethane            | 50.000   | 45.296   | 9.4    | 99    | 0.00     |
| 91 T  | 1,2-Dichlorobenzene         | 50.000   | 49.362   | 1.3    | 98    | 0.00     |
| 92 T  | 1,2-Dibromo-3-Chloropropane | 50.000   | 52.479   | -5.0   | 102   | 0.00     |
| 93 T  | 1,2,4-Trichlorobenzene      | 50.000   | 52.259   | -4.5   | 96    | 0.00     |
| 94 T  | Hexachlorobutadiene         | 50.000   | 45.942   | 8.1    | 98    | 0.00     |
| 95 T  | Naphthalene                 | 50.000   | 58.231   | -16.5  | 98    | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
Data File : VV039144.D  
Acq On : 22 Sep 2025 16:58  
Operator : SY/MD  
Sample : VSTDICV050  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 12 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
ICVVV092225

Quant Time: Sep 24 08:10:27 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 20% Max. Rel. Area : 150%

|      | Compound               | Amount | Calc.  | %Dev | Area% | Dev(min) |
|------|------------------------|--------|--------|------|-------|----------|
| 96 T | 1,2,3-Trichlorobenzene | 50.000 | 52.633 | -5.3 | 96    | 0.00     |

(#) = Out of Range                      SPCC's out = 0    CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## VOLATILE CONTINUING CALIBRATION CHECK

|                     |                   |                        |                              |
|---------------------|-------------------|------------------------|------------------------------|
| Lab Name:           | <u>Alliance</u>   | Contract:              | <u>GENV01</u>                |
| Lab Code:           | <u>ACE</u>        | SDG No.:               | <u>Q3175</u>                 |
| Instrument ID:      | <u>MSVOA_V</u>    | Calibration Date/Time: | <u>09/24/2025 10:49</u>      |
| Lab File ID:        | <u>VV039146.D</u> | Init. Calib. Date(s):  | <u>09/22/2025 09/22/2025</u> |
| Heated Purge: (Y/N) | <u>N</u>          | Init. Calib. Time(s):  | <u>12:26 16:35</u>           |
| GC Column:          | <u>DB-624UI</u>   | ID:                    | <u>0.18</u> (mm)             |

| COMPOUND                       | RRF   | RRF050 | MIN RRF | %D     | MAX%D |
|--------------------------------|-------|--------|---------|--------|-------|
| Dichlorodifluoromethane        | 0.792 | 0.627  |         | -20.83 | 20    |
| Chloromethane                  | 0.850 | 0.730  | 0.1     | -14.12 | 20    |
| Vinyl Chloride                 | 0.910 | 0.742  |         | -18.46 | 20    |
| Bromomethane                   | 0.573 | 0.504  |         | -12.04 | 20    |
| Chloroethane                   | 0.631 | 0.488  |         | -22.66 | 20    |
| Trichlorofluoromethane         | 1.317 | 1.061  |         | -19.44 | 20    |
| 1,1,2-Trichlorotrifluoroethane | 0.784 | 0.628  |         | -19.9  | 20    |
| Tert butyl alcohol             | 0.153 | 0.157  |         | 2.61   | 20    |
| 1,1-Dichloroethene             | 0.754 | 0.615  |         | -18.43 | 20    |
| Acetone                        | 0.523 | 0.448  |         | -14.34 | 20    |
| Carbon Disulfide               | 2.306 | 1.849  |         | -19.82 | 20    |
| Methyl tert-butyl Ether        | 2.402 | 2.294  |         | -4.5   | 20    |
| Methyl Acetate                 | 1.072 | 1.090  |         | 1.68   | 20    |
| Methylene Chloride             | 1.066 | 0.777  |         | -27.11 | 20    |
| trans-1,2-Dichloroethene       | 0.803 | 0.683  |         | -14.94 | 20    |
| 1,1-Dichloroethane             | 1.576 | 1.323  | 0.1     | -16.05 | 20    |
| Cyclohexane                    | 1.291 | 1.047  |         | -18.9  | 20    |
| 2-Butanone                     | 0.579 | 0.608  |         | 5.01   | 20    |
| Carbon Tetrachloride           | 0.743 | 0.609  |         | -18.03 | 20    |
| cis-1,2-Dichloroethene         | 0.885 | 0.808  |         | -8.7   | 20    |
| Bromochloromethane             | 0.696 | 0.683  |         | -1.87  | 20    |
| Chloroform                     | 1.645 | 1.449  |         | -11.91 | 20    |
| 1,1,1-Trichloroethane          | 1.424 | 1.169  |         | -17.91 | 20    |
| Methylcyclohexane              | 0.734 | 0.667  |         | -9.13  | 20    |
| Benzene                        | 1.979 | 1.817  |         | -8.19  | 20    |
| 1,2-Dichloroethane             | 0.693 | 0.655  |         | -5.48  | 20    |
| Trichloroethene                | 0.457 | 0.405  |         | -11.38 | 20    |
| 1,2-Dichloropropane            | 0.492 | 0.476  |         | -3.25  | 20    |
| Bromodichloromethane           | 0.778 | 0.727  |         | -6.55  | 20    |
| 4-Methyl-2-Pentanone           | 0.661 | 0.751  |         | 13.62  | 20    |
| Toluene                        | 1.147 | 1.120  |         | -2.35  | 20    |
| t-1,3-Dichloropropene          | 0.678 | 0.704  |         | 3.84   | 20    |
| cis-1,3-Dichloropropene        | 0.749 | 0.769  |         | 2.67   | 20    |
| 1,1,2-Trichloroethane          | 0.514 | 0.500  |         | -2.72  | 20    |
| 2-Hexanone                     | 0.463 | 0.557  |         | 20.3   | 20    |
| Dibromochloromethane           | 0.586 | 0.559  |         | -4.61  | 20    |
| 1,2-Dibromoethane              | 0.526 | 0.530  |         | 0.76   | 20    |
| Tetrachloroethene              | 0.419 | 0.347  |         | -17.18 | 20    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## VOLATILE CONTINUING CALIBRATION CHECK

|                     |                   |                        |                              |
|---------------------|-------------------|------------------------|------------------------------|
| Lab Name:           | <u>Alliance</u>   | Contract:              | <u>GENV01</u>                |
| Lab Code:           | <u>ACE</u>        | SDG No.:               | <u>Q3175</u>                 |
| Instrument ID:      | <u>MSVOA_V</u>    | Calibration Date/Time: | <u>09/24/2025 10:49</u>      |
| Lab File ID:        | <u>VV039146.D</u> | Init. Calib. Date(s):  | <u>09/22/2025 09/22/2025</u> |
| Heated Purge: (Y/N) | <u>N</u>          | Init. Calib. Time(s):  | <u>12:26 16:35</u>           |
| GC Column:          | <u>DB-624UI</u>   | ID:                    | <u>0.18</u> (mm)             |

| COMPOUND                    | RRF   | RRF050 | MIN<br>RRF | %D     | MAX%D |
|-----------------------------|-------|--------|------------|--------|-------|
| Chlorobenzene               | 1.393 | 1.221  | 0.3        | -12.35 | 20    |
| Ethyl Benzene               | 2.108 | 1.997  |            | -5.27  | 20    |
| m/p-Xylenes                 | 0.814 | 0.814  |            | 0      | 20    |
| o-Xylene                    | 0.720 | 0.732  |            | 1.67   | 20    |
| Styrene                     | 1.263 | 1.382  |            | 9.42   | 20    |
| Bromoform                   | 0.393 | 0.378  | 0.1        | -3.82  | 20    |
| Isopropylbenzene            | 3.627 | 3.308  |            | -8.8   | 20    |
| 1,1,2,2-Tetrachloroethane   | 1.589 | 1.354  | 0.3        | -14.79 | 20    |
| 1,3-Dichlorobenzene         | 1.904 | 1.649  |            | -13.39 | 20    |
| 1,4-Dichlorobenzene         | 2.089 | 1.730  |            | -17.18 | 20    |
| 1,2-Dichlorobenzene         | 1.779 | 1.564  |            | -12.09 | 20    |
| 1,2-Dibromo-3-Chloropropane | 0.330 | 0.288  |            | -12.73 | 20    |
| 1,2,4-Trichlorobenzene      | 1.020 | 0.948  |            | -7.06  | 20    |
| 1,2,3-Trichlorobenzene      | 1.043 | 1.000  |            | -4.12  | 20    |
| 1,2-Dichloroethane-d4       | 0.724 | 0.661  |            | -8.7   | 20    |
| Dibromofluoromethane        | 0.402 | 0.366  |            | -8.95  | 20    |
| Toluene-d8                  | 1.181 | 1.053  |            | -10.84 | 20    |
| 4-Bromofluorobenzene        | 0.486 | 0.489  |            | 0.62   | 20    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
 Data File : VV039146.D  
 Acq On : 24 Sep 2025 10:49  
 Operator : SY/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VSTDCCC050

Quant Time: Sep 25 04:42:18 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 4.600  | 168  | 558154   | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 5.532  | 114  | 907128   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 8.777  | 117  | 923608   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 11.172 | 152  | 551620   | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |          |
|-----------------------------|--------|-------|----------|----------|------|----------|
| System Monitoring Compounds |        |       |          |          |      |          |
| 33) 1,2-Dichloroethane-d4   | 4.941  | 65    | 369186   | 45.702   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 81 - 118 | Recovery | =    | 91.400%  |
| 35) Dibromofluoromethane    | 4.500  | 113   | 332304   | 45.567   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 80 - 119 | Recovery | =    | 91.140%  |
| 50) Toluene-d8              | 7.236  | 98    | 954978   | 44.587   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 89 - 112 | Recovery | =    | 89.180%  |
| 62) 4-Bromofluorobenzene    | 10.002 | 95    | 443437   | 50.291   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 85 - 114 | Recovery | =    | 100.580% |

|                              |       |     |         |         |        |     |
|------------------------------|-------|-----|---------|---------|--------|-----|
| Target Compounds             |       |     |         |         | Qvalue |     |
| 2) Dichlorodifluoromethane   | 1.121 | 85  | 349754  | 39.585  | ug/l   | 96  |
| 3) Chloromethane             | 1.227 | 50  | 407305  | 42.941  | ug/l   | 100 |
| 4) Vinyl Chloride            | 1.295 | 62  | 414391  | 40.808  | ug/l   | 100 |
| 5) Bromomethane              | 1.494 | 94  | 281451  | 43.967  | ug/l   | 100 |
| 6) Chloroethane              | 1.558 | 64  | 272410  | 45.230  | ug/l   | 97  |
| 7) Trichlorofluoromethane    | 1.722 | 101 | 592362  | 40.292  | ug/l   | 99  |
| 8) Diethyl Ether             | 1.918 | 74  | 252704  | 46.301  | ug/l   | 99  |
| 9) 1,1,2-Trichlorotrifluo... | 2.079 | 101 | 350274  | 40.042  | ug/l   | 98  |
| 10) Methyl Iodide            | 2.195 | 142 | 408452  | 49.032  | ug/l   | 100 |
| 11) Tert butyl alcohol       | 2.571 | 59  | 439522  | 257.431 | ug/l   | 98  |
| 12) 1,1-Dichloroethene       | 2.079 | 96  | 343170  | 40.778  | ug/l   | 97  |
| 13) Acrolein                 | 2.008 | 56  | 82132   | 277.230 | ug/l   | 99  |
| 14) Allyl chloride           | 2.356 | 41  | 678394  | 45.129  | ug/l   | 97  |
| 15) Acrylonitrile            | 2.674 | 53  | 1317414 | 240.705 | ug/l   | 100 |
| 16) Acetone                  | 2.118 | 43  | 1249190 | 250.295 | ug/l   | 99  |
| 17) Carbon Disulfide         | 2.249 | 76  | 1032299 | 40.110  | ug/l   | 99  |
| 18) Methyl Acetate           | 2.378 | 43  | 608422  | 50.833  | ug/l   | 100 |
| 19) Methyl tert-butyl Ether  | 2.712 | 73  | 1280541 | 47.752  | ug/l   | 100 |
| 20) Methylene Chloride       | 2.455 | 84  | 433679  | 49.660  | ug/l   | 100 |
| 21) trans-1,2-Dichloroethene | 2.700 | 96  | 381303  | 42.561  | ug/l   | 100 |
| 22) Diisopropyl ether        | 3.214 | 45  | 1373474 | 47.754  | ug/l   | 92  |
| 23) Vinyl Acetate            | 3.188 | 43  | 6174734 | 245.419 | ug/l   | 99  |
| 24) 1,1-Dichloroethane       | 3.114 | 63  | 738164  | 41.969  | ug/l   | 100 |
| 25) 2-Butanone               | 3.864 | 43  | 1696999 | 262.339 | ug/l   | 98  |
| 26) 2,2-Dichloropropane      | 3.809 | 77  | 608343  | 39.009  | ug/l   | 100 |
| 27) cis-1,2-Dichloroethene   | 3.815 | 96  | 451121  | 45.657  | ug/l   | 99  |
| 28) Bromochloromethane       | 4.143 | 49  | 381004  | 49.069  | ug/l   | 97  |
| 29) Tetrahydrofuran          | 4.233 | 42  | 1116689 | 275.798 | ug/l   | 100 |
| 30) Chloroform               | 4.275 | 83  | 808825  | 44.039  | ug/l   | 100 |
| 31) Cyclohexane              | 4.581 | 56  | 584423  | 40.562  | ug/l   | 99  |
| 32) 1,1,1-Trichloroethane    | 4.513 | 97  | 652596  | 41.043  | ug/l   | 99  |
| 36) 1,1-Dichloropropene      | 4.741 | 75  | 487252  | 44.192  | ug/l   | 98  |
| 37) Ethyl Acetate            | 3.973 | 43  | 663808  | 49.616  | ug/l   | 100 |
| 38) Carbon Tetrachloride     | 4.732 | 117 | 552633  | 41.012  | ug/l   | 99  |
| 39) Methylcyclohexane        | 6.047 | 83  | 605143  | 45.449  | ug/l   | 99  |
| 40) Benzene                  | 5.005 | 78  | 1647897 | 45.898  | ug/l   | 100 |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
 Data File : VV039146.D  
 Acq On : 24 Sep 2025 10:49  
 Operator : SY/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VSTDCCC050

Quant Time: Sep 25 04:42:18 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc     | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|----------|--------|----------|
| 41) Methacrylonitrile         | 4.156  | 41   | 299616   | 49.549   | ug/l   | 99       |
| 42) 1,2-Dichloroethane        | 5.037  | 62   | 594558   | 47.262   | ug/l   | 100      |
| 43) Isopropyl Acetate         | 5.191  | 43   | 967617   | 53.044   | ug/l   | 100      |
| 44) Trichloroethene           | 5.828  | 130  | 367209   | 44.323   | ug/l   | 96       |
| 45) 1,2-Dichloropropane       | 6.085  | 63   | 431931   | 48.401   | ug/l   | 98       |
| 46) Dibromomethane            | 6.224  | 93   | 328651   | 48.117   | ug/l   | 99       |
| 47) Bromodichloromethane      | 6.426  | 83   | 659260   | 46.688   | ug/l   | 99       |
| 48) Methyl methacrylate       | 6.294  | 41   | 457609   | 55.340   | ug/l   | 99       |
| 49) 1,4-Dioxane               | 6.291  | 88   | 155709   | 1166.851 | ug/l   | 97       |
| 51) 4-Methyl-2-Pentanone      | 7.146  | 43   | 3404796  | 283.713  | ug/l   | 100      |
| 52) Toluene                   | 7.307  | 92   | 1016283  | 48.820   | ug/l   | 99       |
| 53) t-1,3-Dichloropropene     | 7.571  | 75   | 638267   | 51.867   | ug/l   | 98       |
| 54) cis-1,3-Dichloropropene   | 6.947  | 75   | 697782   | 51.364   | ug/l   | 100      |
| 55) 1,1,2-Trichloroethane     | 7.760  | 97   | 453142   | 48.547   | ug/l   | 98       |
| 56) Ethyl methacrylate        | 7.715  | 69   | 630396   | 56.871   | ug/l   | 100      |
| 57) 1,3-Dichloropropane       | 7.931  | 76   | 727051   | 49.356   | ug/l   | 100      |
| 58) 2-Chloroethyl Vinyl ether | 6.809  | 63   | 1515308  | 242.405  | ug/l   | 100      |
| 59) 2-Hexanone                | 8.063  | 43   | 2528312  | 260.898  | ug/l   | 99       |
| 60) Dibromochloromethane      | 8.166  | 129  | 506709   | 47.642   | ug/l   | 99       |
| 61) 1,2-Dibromoethane         | 8.272  | 107  | 480605   | 50.348   | ug/l   | 100      |
| 64) Tetrachloroethene         | 7.896  | 164  | 320555   | 41.437   | ug/l   | 99       |
| 65) Chlorobenzene             | 8.802  | 112  | 1127747  | 43.836   | ug/l   | 98       |
| 66) 1,1,1,2-Tetrachloroethane | 8.895  | 131  | 407431   | 43.830   | ug/l   | 100      |
| 67) Ethyl Benzene             | 8.934  | 91   | 1844871  | 47.374   | ug/l   | 100      |
| 68) m/p-Xylenes               | 9.063  | 106  | 1502908  | 99.991   | ug/l   | 100      |
| 69) o-Xylene                  | 9.468  | 106  | 676401   | 50.825   | ug/l   | 99       |
| 70) Styrene                   | 9.484  | 104  | 1276348  | 45.393   | ug/l   | 99       |
| 71) Bromoform                 | 9.651  | 173  | 348806   | 48.038   | ug/l # | 100      |
| 73) Isopropylbenzene          | 9.857  | 105  | 1824633  | 45.605   | ug/l   | 100      |
| 74) N-amyl acetate            | 9.725  | 43   | 854806   | 53.563   | ug/l   | 98       |
| 75) 1,1,2,2-Tetrachloroethane | 10.166 | 83   | 747109   | 42.619   | ug/l   | 99       |
| 76) 1,2,3-Trichloropropane    | 10.198 | 75   | 575337   | 45.809   | ug/l   | 98       |
| 77) Bromobenzene              | 10.140 | 156  | 460599   | 44.372   | ug/l   | 99       |
| 78) n-propylbenzene           | 10.278 | 91   | 2250074  | 46.182   | ug/l   | 99       |
| 79) 2-Chlorotoluene           | 10.349 | 91   | 1358304  | 46.591   | ug/l   | 99       |
| 80) 1,3,5-Trimethylbenzene    | 10.465 | 105  | 1607013  | 48.324   | ug/l   | 99       |
| 81) trans-1,4-Dichloro-2-b... | 9.928  | 75   | 256388   | 49.007   | ug/l   | 99       |
| 82) 4-Chlorotoluene           | 10.461 | 91   | 1605527  | 47.493   | ug/l   | 99       |
| 83) tert-Butylbenzene         | 10.789 | 119  | 1497786  | 45.168   | ug/l   | 99       |
| 84) 1,2,4-Trimethylbenzene    | 10.838 | 105  | 1608644  | 49.826   | ug/l   | 100      |
| 85) sec-Butylbenzene          | 11.011 | 105  | 1967674  | 44.737   | ug/l   | 99       |
| 86) p-Isopropyltoluene        | 11.165 | 119  | 1677532  | 45.825   | ug/l   | 100      |
| 87) 1,3-Dichlorobenzene       | 11.104 | 146  | 909637   | 43.299   | ug/l   | 99       |
| 88) 1,4-Dichlorobenzene       | 11.194 | 146  | 954565   | 41.416   | ug/l   | 100      |
| 89) n-Butylbenzene            | 11.580 | 91   | 1645246  | 47.058   | ug/l   | 100      |
| 90) Hexachloroethane          | 11.821 | 117  | 343097   | 38.060   | ug/l   | 99       |
| 91) 1,2-Dichlorobenzene       | 11.567 | 146  | 862856   | 43.962   | ug/l   | 99       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.349 | 75   | 158828   | 50.345   | ug/l   | 97       |
| 93) 1,2,4-Trichlorobenzene    | 13.185 | 180  | 522821   | 46.468   | ug/l   | 100      |
| 94) Hexachlorobutadiene       | 13.368 | 225  | 224298   | 38.367   | ug/l   | 100      |
| 95) Naphthalene               | 13.426 | 128  | 1960806  | 53.919   | ug/l   | 99       |
| 96) 1,2,3-Trichlorobenzene    | 13.667 | 180  | 551841   | 47.956   | ug/l   | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039146.D  
Acq On : 24 Sep 2025 10:49  
Operator : SY/MD  
Sample : VSTDCCC050  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VSTDCCC050

Quant Time: Sep 25 04:42:18 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration


| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092425\
Data File : VV039146.D
Acq On    : 24 Sep 2025  10:49
Operator  : SY/MD
Sample    : VSTDCCC050
Misc      : 5.0mL/MSVOA_V/WATER
ALS Vial  : 2      Sample Multiplier: 1
```

**Instrument :**  
MSVOA\_V  
**ClientSampleId :**  
VSTDCCC050

Quant Time: Sep 25 04:42:18 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
 Data File : VV039146.D  
 Acq On : 24 Sep 2025 10:49  
 Operator : SY/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 LabSampleId :  
 VSTDCCC050

Quant Time: Sep 25 04:42:18 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 20% Max. Rel. Area : 150%

|       | Compound                    | AvgRF | CCRF  | %Dev   | Area% | Dev(min) |
|-------|-----------------------------|-------|-------|--------|-------|----------|
| 1 I   | Pentafluorobenzene          | 1.000 | 1.000 | 0.0    | 99    | 0.00     |
| 2 T   | Dichlorodifluoromethane     | 0.791 | 0.627 | 20.7#  | 84    | 0.00     |
| 3 P   | Chloromethane               | 0.850 | 0.730 | 14.1   | 94    | 0.00     |
| 4 C   | Vinyl Chloride              | 0.910 | 0.742 | 18.5#  | 88    | 0.00     |
| 5 T   | Bromomethane                | 0.573 | 0.504 | 12.0   | 98    | 0.00     |
| 6 T   | Chloroethane                | 0.631 | 0.488 | 22.7#  | 91    | 0.00     |
| 7 T   | Trichlorofluoromethane      | 1.317 | 1.061 | 19.4   | 88    | 0.00     |
| 8 T   | Diethyl Ether               | 0.489 | 0.453 | 7.4    | 100   | 0.00     |
| 9 T   | 1,1,2-Trichlorotrifluoroeth | 0.784 | 0.628 | 19.9   | 89    | 0.00     |
| 10 T  | Methyl Iodide               | 0.746 | 0.732 | 1.9    | 92    | 0.00     |
| 11 T  | Tert butyl alcohol          | 0.153 | 0.157 | -2.6   | 114   | 0.00     |
| 12 CM | 1,1-Dichloroethene          | 0.754 | 0.615 | 18.4#  | 90    | 0.00     |
| 13 T  | Acrolein                    | 0.027 | 0.029 | -7.4   | 110   | 0.00     |
| 14 T  | Allyl chloride              | 1.347 | 1.215 | 9.8    | 97    | 0.00     |
| 15 T  | Acrylonitrile               | 0.490 | 0.472 | 3.7    | 107   | 0.00     |
| 16 T  | Acetone                     | 0.523 | 0.448 | 14.3   | 99    | 0.00     |
| 17 T  | Carbon Disulfide            | 2.306 | 1.849 | 19.8   | 88    | 0.00     |
| 18 T  | Methyl Acetate              | 1.072 | 1.090 | -1.7   | 110   | 0.00     |
| 19 T  | Methyl tert-butyl Ether     | 2.402 | 2.294 | 4.5    | 102   | 0.00     |
| 20 T  | Methylene Chloride          | 1.066 | 0.777 | 27.1#  | 99    | 0.00     |
| 21 T  | trans-1,2-Dichloroethene    | 0.803 | 0.683 | 14.9   | 93    | 0.00     |
| 22 T  | Diisopropyl ether           | 2.576 | 2.461 | 4.5    | 97    | 0.00     |
| 23 T  | Vinyl Acetate               | 2.254 | 2.213 | 1.8    | 99    | 0.00     |
| 24 P  | 1,1-Dichloroethane          | 1.576 | 1.323 | 16.1   | 94    | 0.00     |
| 25 T  | 2-Butanone                  | 0.579 | 0.608 | -5.0   | 103   | 0.00     |
| 26 T  | 2,2-Dichloropropane         | 1.397 | 1.090 | 22.0#  | 87    | 0.00     |
| 27 T  | cis-1,2-Dichloroethene      | 0.885 | 0.808 | 8.7    | 92    | 0.00     |
| 28 T  | Bromochloromethane          | 0.696 | 0.683 | 1.9    | 100   | 0.00     |
| 29 T  | Tetrahydrofuran             | 0.363 | 0.400 | -10.2  | 107   | 0.00     |
| 30 C  | Chloroform                  | 1.645 | 1.449 | 11.9#  | 96    | 0.00     |
| 31 T  | Cyclohexane                 | 1.291 | 1.047 | 18.9   | 83    | 0.00     |
| 32 T  | 1,1,1-Trichloroethane       | 1.424 | 1.169 | 17.9   | 89    | 0.00     |
| 33 S  | 1,2-Dichloroethane-d4       | 0.724 | 0.661 | 8.7    | 101   | 0.00     |
| 34 I  | 1,4-Difluorobenzene         | 1.000 | 1.000 | 0.0    | 101   | 0.00     |
| 35 S  | Dibromofluoromethane        | 0.402 | 0.366 | 9.0    | 98    | 0.00     |
| 36 T  | 1,1-Dichloropropene         | 0.608 | 0.537 | 11.7   | 88    | 0.00     |
| 37 T  | Ethyl Acetate               | 0.737 | 0.732 | 0.7    | 104   | 0.00     |
| 38 T  | Carbon Tetrachloride        | 0.743 | 0.609 | 18.0   | 88    | 0.00     |
| 39 T  | Methylcyclohexane           | 0.734 | 0.667 | 9.1    | 82    | 0.00     |
| 40 TM | Benzene                     | 1.979 | 1.817 | 8.2    | 93    | 0.00     |
| 41 T  | Methacrylonitrile           | 0.273 | 0.330 | -20.9# | 101   | 0.00     |
| 42 TM | 1,2-Dichloroethane          | 0.693 | 0.655 | 5.5    | 100   | 0.00     |
| 43 T  | Isopropyl Acetate           | 1.005 | 1.067 | -6.2   | 101   | 0.00     |
| 44 TM | Trichloroethene             | 0.457 | 0.405 | 11.4   | 89    | 0.00     |
| 45 C  | 1,2-Dichloropropane         | 0.492 | 0.476 | 3.3#   | 96    | 0.00     |
| 46 T  | Dibromomethane              | 0.376 | 0.362 | 3.7    | 100   | 0.00     |
| 47 T  | Bromodichloromethane        | 0.778 | 0.727 | 6.6    | 98    | 0.00     |
| 48 T  | Methyl methacrylate         | 0.456 | 0.504 | -10.5  | 102   | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
 Data File : VV039146.D  
 Acq On : 24 Sep 2025 10:49  
 Operator : SY/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 LabSampleId :  
 VSTDCCC050

Quant Time: Sep 25 04:42:18 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 20% Max. Rel. Area : 150%

|       | Compound                    | AvgRF | CCRF  | %Dev   | Area% | Dev(min) |
|-------|-----------------------------|-------|-------|--------|-------|----------|
| 49 T  | 1,4-Dioxane                 | 0.007 | 0.009 | -28.6# | 116   | 0.00     |
| 50 S  | Toluene-d8                  | 1.181 | 1.053 | 10.8   | 94    | 0.00     |
| 51 T  | 4-Methyl-2-Pentanone        | 0.661 | 0.751 | -13.6  | 104   | 0.00     |
| 52 CM | Toluene                     | 1.147 | 1.120 | 2.4#   | 92    | 0.00     |
| 53 T  | t-1,3-Dichloropropene       | 0.678 | 0.704 | -3.8   | 98    | 0.00     |
| 54 T  | cis-1,3-Dichloropropene     | 0.749 | 0.769 | -2.7   | 96    | 0.00     |
| 55 T  | 1,1,2-Trichloroethane       | 0.514 | 0.500 | 2.7    | 103   | 0.00     |
| 56 T  | Ethyl methacrylate          | 0.611 | 0.695 | -13.7  | 98    | 0.00     |
| 57 T  | 1,3-Dichloropropane         | 0.812 | 0.801 | 1.4    | 99    | 0.00     |
| 58 T  | 2-Chloroethyl Vinyl ether   | 0.267 | 0.334 | -25.1# | 96    | 0.00     |
| 59 T  | 2-Hexanone                  | 0.463 | 0.557 | -20.3# | 104   | 0.00     |
| 60 T  | Dibromochloromethane        | 0.586 | 0.559 | 4.6    | 100   | 0.00     |
| 61 T  | 1,2-Dibromoethane           | 0.526 | 0.530 | -0.8   | 102   | 0.00     |
| 62 S  | 4-Bromofluorobenzene        | 0.486 | 0.489 | -0.6   | 98    | 0.00     |
| 63 I  | Chlorobenzene-d5            | 1.000 | 1.000 | 0.0    | 107   | 0.00     |
| 64 T  | Tetrachloroethene           | 0.419 | 0.347 | 17.2   | 90    | 0.00     |
| 65 PM | Chlorobenzene               | 1.393 | 1.221 | 12.3   | 94    | 0.00     |
| 66 T  | 1,1,1,2-Tetrachloroethane   | 0.503 | 0.441 | 12.3   | 97    | 0.00     |
| 67 C  | Ethyl Benzene               | 2.108 | 1.997 | 5.3#   | 90    | 0.00     |
| 68 T  | m/p-Xylenes                 | 0.814 | 0.814 | 0.0    | 92    | 0.00     |
| 69 T  | o-Xylene                    | 0.720 | 0.732 | -1.7   | 93    | 0.00     |
| 70 T  | Styrene                     | 1.263 | 1.382 | -9.4   | 95    | 0.00     |
| 71 P  | Bromoform                   | 0.393 | 0.378 | 3.8    | 102   | 0.00     |
| 72 I  | 1,4-Dichlorobenzene-d4      | 1.000 | 1.000 | 0.0    | 107   | 0.00     |
| 73 T  | Isopropylbenzene            | 3.627 | 3.308 | 8.8    | 90    | 0.00     |
| 74 T  | N-amyl acetate              | 1.447 | 1.550 | -7.1   | 99    | 0.00     |
| 75 P  | 1,1,2,2-Tetrachloroethane   | 1.589 | 1.354 | 14.8   | 102   | 0.00     |
| 76 T  | 1,2,3-Trichloropropane      | 1.138 | 1.043 | 8.3    | 106   | 0.00     |
| 77 T  | Bromobenzene                | 0.941 | 0.835 | 11.3   | 96    | 0.00     |
| 78 T  | n-propylbenzene             | 4.416 | 4.079 | 7.6    | 89    | 0.00     |
| 79 T  | 2-Chlorotoluene             | 2.643 | 2.462 | 6.8    | 93    | 0.00     |
| 80 T  | 1,3,5-Trimethylbenzene      | 3.014 | 2.913 | 3.4    | 92    | 0.00     |
| 81 T  | trans-1,4-Dichloro-2-butene | 0.474 | 0.465 | 1.9    | 100   | 0.00     |
| 82 T  | 4-Chlorotoluene             | 3.064 | 2.911 | 5.0    | 93    | 0.00     |
| 83 T  | tert-Butylbenzene           | 3.006 | 2.715 | 9.7    | 90    | 0.00     |
| 84 T  | 1,2,4-Trimethylbenzene      | 2.926 | 2.916 | 0.3    | 92    | 0.00     |
| 85 T  | sec-Butylbenzene            | 3.987 | 3.567 | 10.5   | 87    | 0.00     |
| 86 T  | p-Isopropyltoluene          | 3.318 | 3.041 | 8.3    | 88    | 0.00     |
| 87 T  | 1,3-Dichlorobenzene         | 1.904 | 1.649 | 13.4   | 95    | 0.00     |
| 88 T  | 1,4-Dichlorobenzene         | 2.089 | 1.730 | 17.2   | 96    | 0.00     |
| 89 T  | n-Butylbenzene              | 3.169 | 2.983 | 5.9    | 88    | 0.00     |
| 90 T  | Hexachloroethane            | 0.817 | 0.622 | 23.9#  | 91    | 0.00     |
| 91 T  | 1,2-Dichlorobenzene         | 1.779 | 1.564 | 12.1   | 96    | 0.00     |
| 92 T  | 1,2-Dibromo-3-Chloropropane | 0.330 | 0.288 | 12.7   | 108   | 0.00     |
| 93 T  | 1,2,4-Trichlorobenzene      | 1.020 | 0.948 | 7.1    | 94    | 0.00     |
| 94 T  | Hexachlorobutadiene         | 0.530 | 0.407 | 23.2#  | 91    | 0.00     |
| 95 T  | Naphthalene                 | 3.296 | 3.555 | -7.9   | 100   | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039146.D  
Acq On : 24 Sep 2025 10:49  
Operator : SY/MD  
Sample : VSTDCCC050  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
LabSampleId :  
VSTDCCC050

Quant Time: Sep 25 04:42:18 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 20% Max. Rel. Area : 150%

|      | Compound               | AvgRF | CCRF  | %Dev | Area% | Dev(min) |
|------|------------------------|-------|-------|------|-------|----------|
| 96 T | 1,2,3-Trichlorobenzene | 1.043 | 1.000 | 4.1  | 97    | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
 Data File : VV039146.D  
 Acq On : 24 Sep 2025 10:49  
 Operator : SY/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 LabSampleId :  
 VSTDCCC050

Quant Time: Sep 25 04:42:18 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 20% Max. Rel. Area : 150%

|       | Compound                    | Amount  | Calc.   | %Dev  | Area% | Dev(min) |
|-------|-----------------------------|---------|---------|-------|-------|----------|
| 1 I   | Pentafluorobenzene          | 50.000  | 50.000  | 0.0   | 99    | 0.00     |
| 2 T   | Dichlorodifluoromethane     | 50.000  | 39.585  | 20.8# | 84    | 0.00     |
| 3 P   | Chloromethane               | 50.000  | 42.941  | 14.1  | 94    | 0.00     |
| 4 C   | Vinyl Chloride              | 50.000  | 40.808  | 18.4# | 88    | 0.00     |
| 5 T   | Bromomethane                | 50.000  | 43.967  | 12.1  | 98    | 0.00     |
| 6 T   | Chloroethane                | 50.000  | 45.230  | 9.5   | 91    | 0.00     |
| 7 T   | Trichlorofluoromethane      | 50.000  | 40.292  | 19.4  | 88    | 0.00     |
| 8 T   | Diethyl Ether               | 50.000  | 46.301  | 7.4   | 100   | 0.00     |
| 9 T   | 1,1,2-Trichlorotrifluoroeth | 50.000  | 40.042  | 19.9  | 89    | 0.00     |
| 10 T  | Methyl Iodide               | 50.000  | 49.032  | 1.9   | 92    | 0.00     |
| 11 T  | Tert butyl alcohol          | 250.000 | 257.431 | -3.0  | 114   | 0.00     |
| 12 CM | 1,1-Dichloroethene          | 50.000  | 40.778  | 18.4# | 90    | 0.00     |
| 13 T  | Acrolein                    | 250.000 | 277.230 | -10.9 | 110   | 0.00     |
| 14 T  | Allyl chloride              | 50.000  | 45.129  | 9.7   | 97    | 0.00     |
| 15 T  | Acrylonitrile               | 250.000 | 240.705 | 3.7   | 107   | 0.00     |
| 16 T  | Acetone                     | 250.000 | 250.295 | -0.1  | 99    | 0.00     |
| 17 T  | Carbon Disulfide            | 50.000  | 40.110  | 19.8  | 88    | 0.00     |
| 18 T  | Methyl Acetate              | 50.000  | 50.833  | -1.7  | 110   | 0.00     |
| 19 T  | Methyl tert-butyl Ether     | 50.000  | 47.752  | 4.5   | 102   | 0.00     |
| 20 T  | Methylene Chloride          | 50.000  | 49.660  | 0.7   | 99    | 0.00     |
| 21 T  | trans-1,2-Dichloroethene    | 50.000  | 42.561  | 14.9  | 93    | 0.00     |
| 22 T  | Diisopropyl ether           | 50.000  | 47.754  | 4.5   | 97    | 0.00     |
| 23 T  | Vinyl Acetate               | 250.000 | 245.419 | 1.8   | 99    | 0.00     |
| 24 P  | 1,1-Dichloroethane          | 50.000  | 41.969  | 16.1  | 94    | 0.00     |
| 25 T  | 2-Butanone                  | 250.000 | 262.339 | -4.9  | 103   | 0.00     |
| 26 T  | 2,2-Dichloropropane         | 50.000  | 39.009  | 22.0# | 87    | 0.00     |
| 27 T  | cis-1,2-Dichloroethene      | 50.000  | 45.657  | 8.7   | 92    | 0.00     |
| 28 T  | Bromochloromethane          | 50.000  | 49.069  | 1.9   | 100   | 0.00     |
| 29 T  | Tetrahydrofuran             | 250.000 | 275.798 | -10.3 | 107   | 0.00     |
| 30 C  | Chloroform                  | 50.000  | 44.039  | 11.9# | 96    | 0.00     |
| 31 T  | Cyclohexane                 | 50.000  | 40.562  | 18.9  | 83    | 0.00     |
| 32 T  | 1,1,1-Trichloroethane       | 50.000  | 41.043  | 17.9  | 89    | 0.00     |
| 33 S  | 1,2-Dichloroethane-d4       | 50.000  | 45.702  | 8.6   | 101   | 0.00     |
| 34 I  | 1,4-Difluorobenzene         | 50.000  | 50.000  | 0.0   | 101   | 0.00     |
| 35 S  | Dibromofluoromethane        | 50.000  | 45.567  | 8.9   | 98    | 0.00     |
| 36 T  | 1,1-Dichloropropene         | 50.000  | 44.192  | 11.6  | 88    | 0.00     |
| 37 T  | Ethyl Acetate               | 50.000  | 49.616  | 0.8   | 104   | 0.00     |
| 38 T  | Carbon Tetrachloride        | 50.000  | 41.012  | 18.0  | 88    | 0.00     |
| 39 T  | Methylcyclohexane           | 50.000  | 45.449  | 9.1   | 82    | 0.00     |
| 40 TM | Benzene                     | 50.000  | 45.898  | 8.2   | 93    | 0.00     |
| 41 T  | Methacrylonitrile           | 50.000  | 49.549  | 0.9   | 101   | 0.00     |
| 42 TM | 1,2-Dichloroethane          | 50.000  | 47.262  | 5.5   | 100   | 0.00     |
| 43 T  | Isopropyl Acetate           | 50.000  | 53.044  | -6.1  | 101   | 0.00     |
| 44 TM | Trichloroethene             | 50.000  | 44.323  | 11.4  | 89    | 0.00     |
| 45 C  | 1,2-Dichloropropane         | 50.000  | 48.401  | 3.2#  | 96    | 0.00     |
| 46 T  | Dibromomethane              | 50.000  | 48.117  | 3.8   | 100   | 0.00     |
| 47 T  | Bromodichloromethane        | 50.000  | 46.688  | 6.6   | 98    | 0.00     |
| 48 T  | Methyl methacrylate         | 50.000  | 55.340  | -10.7 | 102   | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
 Data File : VV039146.D  
 Acq On : 24 Sep 2025 10:49  
 Operator : SY/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 LabSampleId :  
 VSTDCCC050

Quant Time: Sep 25 04:42:18 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 20% Max. Rel. Area : 150%

|       | Compound                    | Amount   | Calc.    | %Dev  | Area% | Dev(min) |
|-------|-----------------------------|----------|----------|-------|-------|----------|
| 49 T  | 1,4-Dioxane                 | 1000.000 | 1166.851 | -16.7 | 116   | 0.00     |
| 50 S  | Toluene-d8                  | 50.000   | 44.587   | 10.8  | 94    | 0.00     |
| 51 T  | 4-Methyl-2-Pentanone        | 250.000  | 283.713  | -13.5 | 104   | 0.00     |
| 52 CM | Toluene                     | 50.000   | 48.820   | 2.4#  | 92    | 0.00     |
| 53 T  | t-1,3-Dichloropropene       | 50.000   | 51.867   | -3.7  | 98    | 0.00     |
| 54 T  | cis-1,3-Dichloropropene     | 50.000   | 51.364   | -2.7  | 96    | 0.00     |
| 55 T  | 1,1,2-Trichloroethane       | 50.000   | 48.547   | 2.9   | 103   | 0.00     |
| 56 T  | Ethyl methacrylate          | 50.000   | 56.871   | -13.7 | 98    | 0.00     |
| 57 T  | 1,3-Dichloropropane         | 50.000   | 49.356   | 1.3   | 99    | 0.00     |
| 58 T  | 2-Chloroethyl Vinyl ether   | 250.000  | 242.405  | 3.0   | 96    | 0.00     |
| 59 T  | 2-Hexanone                  | 250.000  | 260.898  | -4.4  | 104   | 0.00     |
| 60 T  | Dibromochloromethane        | 50.000   | 47.642   | 4.7   | 100   | 0.00     |
| 61 T  | 1,2-Dibromoethane           | 50.000   | 50.348   | -0.7  | 102   | 0.00     |
| 62 S  | 4-Bromofluorobenzene        | 50.000   | 50.291   | -0.6  | 98    | 0.00     |
| 63 I  | Chlorobenzene-d5            | 50.000   | 50.000   | 0.0   | 107   | 0.00     |
| 64 T  | Tetrachloroethene           | 50.000   | 41.437   | 17.1  | 90    | 0.00     |
| 65 PM | Chlorobenzene               | 50.000   | 43.836   | 12.3  | 94    | 0.00     |
| 66 T  | 1,1,1,2-Tetrachloroethane   | 50.000   | 43.830   | 12.3  | 97    | 0.00     |
| 67 C  | Ethyl Benzene               | 50.000   | 47.374   | 5.3#  | 90    | 0.00     |
| 68 T  | m/p-Xylenes                 | 100.000  | 99.991   | 0.0   | 92    | 0.00     |
| 69 T  | o-Xylene                    | 50.000   | 50.825   | -1.7  | 93    | 0.00     |
| 70 T  | Styrene                     | 50.000   | 45.393   | 9.2   | 95    | 0.00     |
| 71 P  | Bromoform                   | 50.000   | 48.038   | 3.9   | 102   | 0.00     |
| 72 I  | 1,4-Dichlorobenzene-d4      | 50.000   | 50.000   | 0.0   | 107   | 0.00     |
| 73 T  | Isopropylbenzene            | 50.000   | 45.605   | 8.8   | 90    | 0.00     |
| 74 T  | N-amyl acetate              | 50.000   | 53.563   | -7.1  | 99    | 0.00     |
| 75 P  | 1,1,2,2-Tetrachloroethane   | 50.000   | 42.619   | 14.8  | 102   | 0.00     |
| 76 T  | 1,2,3-Trichloropropane      | 50.000   | 45.809   | 8.4   | 106   | 0.00     |
| 77 T  | Bromobenzene                | 50.000   | 44.372   | 11.3  | 96    | 0.00     |
| 78 T  | n-propylbenzene             | 50.000   | 46.182   | 7.6   | 89    | 0.00     |
| 79 T  | 2-Chlorotoluene             | 50.000   | 46.591   | 6.8   | 93    | 0.00     |
| 80 T  | 1,3,5-Trimethylbenzene      | 50.000   | 48.324   | 3.4   | 92    | 0.00     |
| 81 T  | trans-1,4-Dichloro-2-butene | 50.000   | 49.007   | 2.0   | 100   | 0.00     |
| 82 T  | 4-Chlorotoluene             | 50.000   | 47.493   | 5.0   | 93    | 0.00     |
| 83 T  | tert-Butylbenzene           | 50.000   | 45.168   | 9.7   | 90    | 0.00     |
| 84 T  | 1,2,4-Trimethylbenzene      | 50.000   | 49.826   | 0.3   | 92    | 0.00     |
| 85 T  | sec-Butylbenzene            | 50.000   | 44.737   | 10.5  | 87    | 0.00     |
| 86 T  | p-Isopropyltoluene          | 50.000   | 45.825   | 8.3   | 88    | 0.00     |
| 87 T  | 1,3-Dichlorobenzene         | 50.000   | 43.299   | 13.4  | 95    | 0.00     |
| 88 T  | 1,4-Dichlorobenzene         | 50.000   | 41.416   | 17.2  | 96    | 0.00     |
| 89 T  | n-Butylbenzene              | 50.000   | 47.058   | 5.9   | 88    | 0.00     |
| 90 T  | Hexachloroethane            | 50.000   | 38.060   | 23.9# | 91    | 0.00     |
| 91 T  | 1,2-Dichlorobenzene         | 50.000   | 43.962   | 12.1  | 96    | 0.00     |
| 92 T  | 1,2-Dibromo-3-Chloropropane | 50.000   | 50.345   | -0.7  | 108   | 0.00     |
| 93 T  | 1,2,4-Trichlorobenzene      | 50.000   | 46.468   | 7.1   | 94    | 0.00     |
| 94 T  | Hexachlorobutadiene         | 50.000   | 38.367   | 23.3# | 91    | 0.00     |
| 95 T  | Naphthalene                 | 50.000   | 53.919   | -7.8  | 100   | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039146.D  
Acq On : 24 Sep 2025 10:49  
Operator : SY/MD  
Sample : VSTDCCC050  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
LabSampleId :  
VSTDCCC050

Quant Time: Sep 25 04:42:18 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 20% Max. Rel. Area : 150%

|      | Compound               | Amount | Calc.  | %Dev | Area% | Dev(min) |
|------|------------------------|--------|--------|------|-------|----------|
| 96 T | 1,2,3-Trichlorobenzene | 50.000 | 47.956 | 4.1  | 97    | 0.00     |

(#) = Out of Range                      SPCC's out = 0    CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## VOLATILE CONTINUING CALIBRATION CHECK

|                     |                   |                        |                              |
|---------------------|-------------------|------------------------|------------------------------|
| Lab Name:           | <u>Alliance</u>   | Contract:              | <u>GENV01</u>                |
| Lab Code:           | <u>ACE</u>        | SDG No.:               | <u>Q3175</u>                 |
| Instrument ID:      | <u>MSVOA_V</u>    | Calibration Date/Time: | <u>09/25/2025 09:59</u>      |
| Lab File ID:        | <u>VV039168.D</u> | Init. Calib. Date(s):  | <u>09/22/2025 09/22/2025</u> |
| Heated Purge: (Y/N) | <u>N</u>          | Init. Calib. Time(s):  | <u>12:26 16:35</u>           |
| GC Column:          | <u>DB-624UI</u>   | ID:                    | <u>0.18</u> (mm)             |

| COMPOUND                       | RRF   | RRF050 | MIN RRF | %D     | MAX%D |
|--------------------------------|-------|--------|---------|--------|-------|
| Dichlorodifluoromethane        | 0.792 | 0.682  |         | -13.89 | 20    |
| Chloromethane                  | 0.850 | 0.729  | 0.1     | -14.23 | 20    |
| Vinyl Chloride                 | 0.910 | 0.819  |         | -10    | 20    |
| Bromomethane                   | 0.573 | 0.458  |         | -20.07 | 20    |
| Chloroethane                   | 0.631 | 0.507  |         | -19.65 | 20    |
| Trichlorofluoromethane         | 1.317 | 1.199  |         | -8.96  | 20    |
| 1,1,2-Trichlorotrifluoroethane | 0.784 | 0.710  |         | -9.44  | 20    |
| Tert butyl alcohol             | 0.153 | 0.153  |         | 0      | 20    |
| 1,1-Dichloroethene             | 0.754 | 0.698  |         | -7.43  | 20    |
| Acetone                        | 0.523 | 0.406  |         | -22.37 | 20    |
| Carbon Disulfide               | 2.306 | 2.004  |         | -13.1  | 20    |
| Methyl tert-butyl Ether        | 2.402 | 2.247  |         | -6.45  | 20    |
| Methyl Acetate                 | 1.072 | 0.995  |         | -7.18  | 20    |
| Methylene Chloride             | 1.066 | 0.788  |         | -26.08 | 20    |
| trans-1,2-Dichloroethene       | 0.803 | 0.724  |         | -9.84  | 20    |
| 1,1-Dichloroethane             | 1.576 | 1.367  | 0.1     | -13.26 | 20    |
| Cyclohexane                    | 1.291 | 1.147  |         | -11.15 | 20    |
| 2-Butanone                     | 0.579 | 0.544  |         | -6.05  | 20    |
| Carbon Tetrachloride           | 0.743 | 0.689  |         | -7.27  | 20    |
| cis-1,2-Dichloroethene         | 0.885 | 0.825  |         | -6.78  | 20    |
| Bromochloromethane             | 0.696 | 0.612  |         | -12.07 | 20    |
| Chloroform                     | 1.645 | 1.426  |         | -13.31 | 20    |
| 1,1,1-Trichloroethane          | 1.424 | 1.255  |         | -11.87 | 20    |
| Methylcyclohexane              | 0.734 | 0.797  |         | 8.58   | 20    |
| Benzene                        | 1.979 | 1.926  |         | -2.68  | 20    |
| 1,2-Dichloroethane             | 0.693 | 0.639  |         | -7.79  | 20    |
| Trichloroethene                | 0.457 | 0.463  |         | 1.31   | 20    |
| 1,2-Dichloropropane            | 0.492 | 0.493  |         | 0.2    | 20    |
| Bromodichloromethane           | 0.778 | 0.729  |         | -6.3   | 20    |
| 4-Methyl-2-Pentanone           | 0.661 | 0.702  |         | 6.2    | 20    |
| Toluene                        | 1.147 | 1.204  |         | 4.97   | 20    |
| t-1,3-Dichloropropene          | 0.678 | 0.709  |         | 4.57   | 20    |
| cis-1,3-Dichloropropene        | 0.749 | 0.783  |         | 4.54   | 20    |
| 1,1,2-Trichloroethane          | 0.514 | 0.498  |         | -3.11  | 20    |
| 2-Hexanone                     | 0.463 | 0.528  |         | 14.04  | 20    |
| Dibromochloromethane           | 0.586 | 0.574  |         | -2.05  | 20    |
| 1,2-Dibromoethane              | 0.526 | 0.534  |         | 1.52   | 20    |
| Tetrachloroethene              | 0.419 | 0.418  |         | -0.24  | 20    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## VOLATILE CONTINUING CALIBRATION CHECK

|                     |                   |                        |                              |
|---------------------|-------------------|------------------------|------------------------------|
| Lab Name:           | <u>Alliance</u>   | Contract:              | <u>GENV01</u>                |
| Lab Code:           | <u>ACE</u>        | SDG No.:               | <u>Q3175</u>                 |
| Instrument ID:      | <u>MSVOA_V</u>    | Calibration Date/Time: | <u>09/25/2025 09:59</u>      |
| Lab File ID:        | <u>VV039168.D</u> | Init. Calib. Date(s):  | <u>09/22/2025 09/22/2025</u> |
| Heated Purge: (Y/N) | <u>N</u>          | Init. Calib. Time(s):  | <u>12:26 16:35</u>           |
| GC Column:          | <u>DB-624UI</u>   | ID:                    | <u>0.18</u> (mm)             |

| COMPOUND                    | RRF   | RRF050 | MIN<br>RRF | %D     | MAX%D |
|-----------------------------|-------|--------|------------|--------|-------|
| Chlorobenzene               | 1.393 | 1.377  | 0.3        | -1.15  | 20    |
| Ethyl Benzene               | 2.108 | 2.320  |            | 10.06  | 20    |
| m/p-Xylenes                 | 0.814 | 0.936  |            | 14.99  | 20    |
| o-Xylene                    | 0.720 | 0.841  |            | 16.81  | 20    |
| Styrene                     | 1.263 | 1.530  |            | 21.14  | 20    |
| Bromoform                   | 0.393 | 0.402  | 0.1        | 2.29   | 20    |
| Isopropylbenzene            | 3.627 | 3.946  |            | 8.8    | 20    |
| 1,1,2,2-Tetrachloroethane   | 1.589 | 1.396  | 0.3        | -12.15 | 20    |
| 1,3-Dichlorobenzene         | 1.904 | 1.857  |            | -2.47  | 20    |
| 1,4-Dichlorobenzene         | 2.089 | 1.903  |            | -8.9   | 20    |
| 1,2-Dichlorobenzene         | 1.779 | 1.736  |            | -2.42  | 20    |
| 1,2-Dibromo-3-Chloropropane | 0.330 | 0.292  |            | -11.52 | 20    |
| 1,2,4-Trichlorobenzene      | 1.020 | 1.062  |            | 4.12   | 20    |
| 1,2,3-Trichlorobenzene      | 1.043 | 1.082  |            | 3.74   | 20    |
| 1,2-Dichloroethane-d4       | 0.724 | 0.609  |            | -15.88 | 20    |
| Dibromofluoromethane        | 0.402 | 0.384  |            | -4.48  | 20    |
| Toluene-d8                  | 1.181 | 1.121  |            | -5.08  | 20    |
| 4-Bromofluorobenzene        | 0.486 | 0.508  |            | 4.53   | 20    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
 Data File : VV039168.D  
 Acq On : 25 Sep 2025 09:59  
 Operator : SY/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VSTDCCC050

Quant Time: Sep 26 01:18:56 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc     | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|----------|--------|----------|
| -----                        |                |      |            |          |        |          |
| Internal Standards           |                |      |            |          |        |          |
| 1) Pentafluorobenzene        | 4.596          | 168  | 605297     | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 5.529          | 114  | 933338     | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 8.773          | 117  | 901987     | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 11.172         | 152  | 535710     | 50.000   | ug/l   | 0.00     |
|                              |                |      |            |          |        |          |
| System Monitoring Compounds  |                |      |            |          |        |          |
| 33) 1,2-Dichloroethane-d4    | 4.937          | 65   | 368829     | 42.102   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 81 - 118 |      | Recovery = | 84.200%  |        |          |
| 35) Dibromofluoromethane     | 4.497          | 113  | 358851     | 47.826   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 80 - 119 |      | Recovery = | 95.660%  |        |          |
| 50) Toluene-d8               | 7.233          | 98   | 1046447    | 47.485   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 89 - 112 |      | Recovery = | 94.980%  |        |          |
| 62) 4-Bromofluorobenzene     | 9.998          | 95   | 473731     | 52.218   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 85 - 114 |      | Recovery = | 104.440% |        |          |
|                              |                |      |            |          |        |          |
| Target Compounds             |                |      |            |          | Qvalue |          |
| 2) Dichlorodifluoromethane   | 1.121          | 85   | 412859     | 43.088   | ug/l   | 96       |
| 3) Chloromethane             | 1.227          | 50   | 441526     | 42.924   | ug/l   | 99       |
| 4) Vinyl Chloride            | 1.294          | 62   | 495985     | 45.040   | ug/l   | 99       |
| 5) Bromomethane              | 1.494          | 94   | 277073     | 39.912   | ug/l   | 99       |
| 6) Chloroethane              | 1.558          | 64   | 307095     | 47.163   | ug/l   | 96       |
| 7) Trichlorofluoromethane    | 1.722          | 101  | 725756     | 45.520   | ug/l   | 99       |
| 8) Diethyl Ether             | 1.915          | 74   | 273800     | 46.259   | ug/l   | 95       |
| 9) 1,1,2-Trichlorotrifluo... | 2.079          | 101  | 429502     | 45.275   | ug/l   | 98       |
| 10) Methyl Iodide            | 2.195          | 142  | 394240     | 43.640   | ug/l   | 99       |
| 11) Tert butyl alcohol       | 2.571          | 59   | 462139     | 249.596  | ug/l   | 98       |
| 12) 1,1-Dichloroethene       | 2.079          | 96   | 422387     | 46.283   | ug/l   | 93       |
| 13) Acrolein                 | 2.008          | 56   | 73826      | 229.786  | ug/l   | 96       |
| 14) Allyl chloride           | 2.355          | 41   | 725818     | 44.523   | ug/l   | 97       |
| 15) Acrylonitrile            | 2.670          | 53   | 1328524    | 223.829  | ug/l   | 99       |
| 16) Acetone                  | 2.117          | 43   | 1227421    | 224.844  | ug/l   | 98       |
| 17) Carbon Disulfide         | 2.249          | 76   | 1213128    | 43.464   | ug/l   | 99       |
| 18) Methyl Acetate           | 2.378          | 43   | 602375     | 46.408   | ug/l   | 97       |
| 19) Methyl tert-butyl Ether  | 2.709          | 73   | 1360066    | 46.768   | ug/l   | 99       |
| 20) Methylene Chloride       | 2.455          | 84   | 477153     | 50.438   | ug/l   | 94       |
| 21) trans-1,2-Dichloroethene | 2.699          | 96   | 438480     | 45.131   | ug/l   | 98       |
| 22) Diisopropyl ether        | 3.214          | 45   | 1401924    | 44.947   | ug/l   | 97       |
| 23) Vinyl Acetate            | 3.185          | 43   | 6223966    | 228.109  | ug/l   | 98       |
| 24) 1,1-Dichloroethane       | 3.114          | 63   | 827604     | 43.389   | ug/l   | 100      |
| 25) 2-Butanone               | 3.863          | 43   | 1647490    | 234.849  | ug/l   | 96       |
| 26) 2,2-Dichloropropane      | 3.805          | 77   | 690029     | 40.800   | ug/l   | 100      |
| 27) cis-1,2-Dichloroethene   | 3.815          | 96   | 499109     | 46.579   | ug/l   | 98       |
| 28) Bromochloromethane       | 4.143          | 49   | 370658     | 44.019   | ug/l   | 91       |
| 29) Tetrahydrofuran          | 4.233          | 42   | 1075144    | 244.856  | ug/l   | 97       |
| 30) Chloroform               | 4.275          | 83   | 862902     | 43.324   | ug/l   | 98       |
| 31) Cyclohexane              | 4.580          | 56   | 694256     | 44.432   | ug/l   | 97       |
| 32) 1,1,1-Trichloroethane    | 4.510          | 97   | 759916     | 44.071   | ug/l   | 98       |
| 36) 1,1-Dichloropropene      | 4.741          | 75   | 562517     | 49.586   | ug/l   | 99       |
| 37) Ethyl Acetate            | 3.973          | 43   | 650966     | 47.290   | ug/l   | 98       |
| 38) Carbon Tetrachloride     | 4.731          | 117  | 642615     | 46.350   | ug/l   | 99       |
| 39) Methylcyclohexane        | 6.043          | 83   | 743575     | 54.277   | ug/l   | 99       |
| 40) Benzene                  | 5.005          | 78   | 1797939    | 48.671   | ug/l   | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
 Data File : VV039168.D  
 Acq On : 25 Sep 2025 09:59  
 Operator : SY/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VSTDCCC050

Quant Time: Sep 26 01:18:56 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc     | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|----------|--------|----------|
| 41) Methacrylonitrile         | 4.156  | 41   | 291884   | 47.017   | ug/l   | 98       |
| 42) 1,2-Dichloroethane        | 5.037  | 62   | 596048   | 46.050   | ug/l   | 98       |
| 43) Isopropyl Acetate         | 5.191  | 43   | 954350   | 50.848   | ug/l   | 98       |
| 44) Trichloroethene           | 5.828  | 130  | 432014   | 50.680   | ug/l   | 99       |
| 45) 1,2-Dichloropropane       | 6.085  | 63   | 459902   | 50.088   | ug/l   | 99       |
| 46) Dibromomethane            | 6.220  | 93   | 337738   | 48.059   | ug/l   | 97       |
| 47) Bromodichloromethane      | 6.423  | 83   | 680557   | 46.843   | ug/l   | 98       |
| 48) Methyl methacrylate       | 6.294  | 41   | 449431   | 52.825   | ug/l   | 98       |
| 49) 1,4-Dioxane               | 6.291  | 88   | 155938   | 1135.751 | ug/l   | 97       |
| 51) 4-Methyl-2-Pentanone      | 7.146  | 43   | 3274291  | 265.177  | ug/l   | 98       |
| 52) Toluene                   | 7.307  | 92   | 1123906  | 52.474   | ug/l   | 99       |
| 53) t-1,3-Dichloropropene     | 7.571  | 75   | 661472   | 52.243   | ug/l   | 99       |
| 54) cis-1,3-Dichloropropene   | 6.944  | 75   | 731190   | 52.312   | ug/l   | 97       |
| 55) 1,1,2-Trichloroethane     | 7.757  | 97   | 465236   | 48.443   | ug/l   | 98       |
| 56) Ethyl methacrylate        | 7.712  | 69   | 658147   | 57.707   | ug/l   | 95       |
| 57) 1,3-Dichloropropane       | 7.931  | 76   | 748312   | 49.373   | ug/l   | 100      |
| 58) 2-Chloroethyl Vinyl ether | 6.808  | 63   | 1657102  | 257.039  | ug/l   | 99       |
| 59) 2-Hexanone                | 8.062  | 43   | 2465950  | 247.438  | ug/l   | 97       |
| 60) Dibromochloromethane      | 8.165  | 129  | 535532   | 48.938   | ug/l   | 99       |
| 61) 1,2-Dibromoethane         | 8.271  | 107  | 498142   | 50.719   | ug/l   | 100      |
| 64) Tetrachloroethene         | 7.895  | 164  | 376666   | 49.857   | ug/l   | 97       |
| 65) Chlorobenzene             | 8.802  | 112  | 1242390  | 49.450   | ug/l   | 98       |
| 66) 1,1,1,2-Tetrachloroethane | 8.895  | 131  | 435516   | 47.975   | ug/l   | 98       |
| 67) Ethyl Benzene             | 8.934  | 91   | 2092639  | 55.024   | ug/l   | 99       |
| 68) m/p-Xylenes               | 9.059  | 106  | 1688294  | 115.018  | ug/l   | 100      |
| 69) o-Xylene                  | 9.468  | 106  | 758328   | 58.347   | ug/l   | 98       |
| 70) Styrene                   | 9.484  | 104  | 1379790  | 50.114   | ug/l   | 99       |
| 71) Bromoform                 | 9.651  | 173  | 362481   | 51.118   | ug/l # | 99       |
| 73) Isopropylbenzene          | 9.853  | 105  | 2113988  | 54.407   | ug/l   | 99       |
| 74) N-amyl acetate            | 9.725  | 43   | 853417   | 55.064   | ug/l   | 98       |
| 75) 1,1,2,2-Tetrachloroethane | 10.165 | 83   | 747856   | 43.928   | ug/l   | 99       |
| 76) 1,2,3-Trichloropropane    | 10.197 | 75   | 573502   | 47.019   | ug/l   | 98       |
| 77) Bromobenzene              | 10.140 | 156  | 497564   | 49.357   | ug/l   | 98       |
| 78) n-propylbenzene           | 10.278 | 91   | 2554851  | 53.995   | ug/l   | 99       |
| 79) 2-Chlorotoluene           | 10.345 | 91   | 1498849  | 52.939   | ug/l   | 100      |
| 80) 1,3,5-Trimethylbenzene    | 10.464 | 105  | 1809221  | 56.020   | ug/l   | 100      |
| 81) trans-1,4-Dichloro-2-b... | 9.927  | 75   | 258743   | 50.926   | ug/l   | 97       |
| 82) 4-Chlorotoluene           | 10.461 | 91   | 1769699  | 53.904   | ug/l   | 100      |
| 83) tert-Butylbenzene         | 10.789 | 119  | 1725454  | 53.579   | ug/l   | 100      |
| 84) 1,2,4-Trimethylbenzene    | 10.837 | 105  | 1777572  | 56.693   | ug/l   | 98       |
| 85) sec-Butylbenzene          | 11.011 | 105  | 2294341  | 53.713   | ug/l   | 99       |
| 86) p-Isopropyltoluene        | 11.165 | 119  | 1933306  | 54.380   | ug/l   | 99       |
| 87) 1,3-Dichlorobenzene       | 11.104 | 146  | 995077   | 48.773   | ug/l   | 100      |
| 88) 1,4-Dichlorobenzene       | 11.194 | 146  | 1019216  | 45.535   | ug/l   | 99       |
| 89) n-Butylbenzene            | 11.580 | 91   | 1854102  | 54.607   | ug/l   | 100      |
| 90) Hexachloroethane          | 11.821 | 117  | 387482   | 44.260   | ug/l   | 99       |
| 91) 1,2-Dichlorobenzene       | 11.567 | 146  | 930166   | 48.799   | ug/l   | 99       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.348 | 75   | 156226   | 50.961   | ug/l   | 97       |
| 93) 1,2,4-Trichlorobenzene    | 13.184 | 180  | 568781   | 52.055   | ug/l   | 99       |
| 94) Hexachlorobutadiene       | 13.368 | 225  | 261028   | 45.975   | ug/l   | 99       |
| 95) Naphthalene               | 13.422 | 128  | 2039508  | 57.749   | ug/l   | 100      |
| 96) 1,2,3-Trichlorobenzene    | 13.667 | 180  | 579422   | 51.848   | ug/l   | 97       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
Data File : VV039168.D  
Acq On : 25 Sep 2025 09:59  
Operator : SY/MD  
Sample : VSTDCCC050  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VSTDCCC050

Quant Time: Sep 26 01:18:56 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration


| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VW092525\  
Data File : VW039168.D  
Acq On    : 25 Sep 2025   09:59  
Operator  : SY/MD  
Sample    : VSTDCCC050  
Misc      : 5.0mL/MSVOA_V/WATER  
ALS Vial  : 2      Sample Multiplier: 1
```

**Instrument :**  
MSVOA\_V  
**ClientSampleId :**  
VSTDCCC050

Quant Time: Sep 26 01:18:56 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
 Data File : VV039168.D  
 Acq On : 25 Sep 2025 09:59  
 Operator : SY/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 LabSampleId :  
 VSTDCCC050

Quant Time: Sep 26 01:18:56 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 20% Max. Rel. Area : 150%

|       | Compound                    | AvgRF | CCRF  | %Dev  | Area% | Dev(min) |
|-------|-----------------------------|-------|-------|-------|-------|----------|
| 1 I   | Pentafluorobenzene          | 1.000 | 1.000 | 0.0   | 107   | 0.00     |
| 2 T   | Dichlorodifluoromethane     | 0.791 | 0.682 | 13.8  | 99    | 0.00     |
| 3 P   | Chloromethane               | 0.850 | 0.729 | 14.2  | 102   | 0.00     |
| 4 C   | Vinyl Chloride              | 0.910 | 0.819 | 10.0# | 105   | 0.00     |
| 5 T   | Bromomethane                | 0.573 | 0.458 | 20.1# | 97    | 0.00     |
| 6 T   | Chloroethane                | 0.631 | 0.507 | 19.7  | 102   | 0.00     |
| 7 T   | Trichlorofluoromethane      | 1.317 | 1.199 | 9.0   | 108   | 0.00     |
| 8 T   | Diethyl Ether               | 0.489 | 0.452 | 7.6   | 109   | 0.00     |
| 9 T   | 1,1,2-Trichlorotrifluoroeth | 0.784 | 0.710 | 9.4   | 109   | 0.00     |
| 10 T  | Methyl Iodide               | 0.746 | 0.651 | 12.7  | 89    | 0.00     |
| 11 T  | Tert butyl alcohol          | 0.153 | 0.153 | 0.0   | 120   | 0.00     |
| 12 CM | 1,1-Dichloroethene          | 0.754 | 0.698 | 7.4#  | 111   | 0.00     |
| 13 T  | Acrolein                    | 0.027 | 0.024 | 11.1  | 99    | 0.00     |
| 14 T  | Allyl chloride              | 1.347 | 1.199 | 11.0  | 104   | 0.00     |
| 15 T  | Acrylonitrile               | 0.490 | 0.439 | 10.4  | 108   | 0.00     |
| 16 T  | Acetone                     | 0.523 | 0.406 | 22.4# | 97    | 0.00     |
| 17 T  | Carbon Disulfide            | 2.306 | 2.004 | 13.1  | 103   | 0.00     |
| 18 T  | Methyl Acetate              | 1.072 | 0.995 | 7.2   | 109   | 0.00     |
| 19 T  | Methyl tert-butyl Ether     | 2.402 | 2.247 | 6.5   | 108   | 0.00     |
| 20 T  | Methylene Chloride          | 1.066 | 0.788 | 26.1# | 109   | 0.00     |
| 21 T  | trans-1,2-Dichloroethene    | 0.803 | 0.724 | 9.8   | 107   | 0.00     |
| 22 T  | Diisopropyl ether           | 2.576 | 2.316 | 10.1  | 99    | 0.00     |
| 23 T  | Vinyl Acetate               | 2.254 | 2.056 | 8.8   | 100   | 0.00     |
| 24 P  | 1,1-Dichloroethane          | 1.576 | 1.367 | 13.3  | 106   | 0.00     |
| 25 T  | 2-Butanone                  | 0.579 | 0.544 | 6.0   | 100   | 0.00     |
| 26 T  | 2,2-Dichloropropane         | 1.397 | 1.140 | 18.4  | 99    | 0.00     |
| 27 T  | cis-1,2-Dichloroethene      | 0.885 | 0.825 | 6.8   | 102   | 0.00     |
| 28 T  | Bromochloromethane          | 0.696 | 0.612 | 12.1  | 97    | 0.00     |
| 29 T  | Tetrahydrofuran             | 0.363 | 0.355 | 2.2   | 103   | 0.00     |
| 30 C  | Chloroform                  | 1.645 | 1.426 | 13.3# | 103   | 0.00     |
| 31 T  | Cyclohexane                 | 1.291 | 1.147 | 11.2  | 99    | 0.00     |
| 32 T  | 1,1,1-Trichloroethane       | 1.424 | 1.255 | 11.9  | 104   | 0.00     |
| 33 S  | 1,2-Dichloroethane-d4       | 0.724 | 0.609 | 15.9  | 101   | 0.00     |
| 34 I  | 1,4-Difluorobenzene         | 1.000 | 1.000 | 0.0   | 104   | 0.00     |
| 35 S  | Dibromofluoromethane        | 0.402 | 0.384 | 4.5   | 106   | 0.00     |
| 36 T  | 1,1-Dichloropropene         | 0.608 | 0.603 | 0.8   | 101   | 0.00     |
| 37 T  | Ethyl Acetate               | 0.737 | 0.697 | 5.4   | 102   | 0.00     |
| 38 T  | Carbon Tetrachloride        | 0.743 | 0.689 | 7.3   | 102   | 0.00     |
| 39 T  | Methylcyclohexane           | 0.734 | 0.797 | -8.6  | 101   | 0.00     |
| 40 TM | Benzene                     | 1.979 | 1.926 | 2.7   | 101   | 0.00     |
| 41 T  | Methacrylonitrile           | 0.273 | 0.313 | -14.7 | 99    | 0.00     |
| 42 TM | 1,2-Dichloroethane          | 0.693 | 0.639 | 7.8   | 100   | 0.00     |
| 43 T  | Isopropyl Acetate           | 1.005 | 1.023 | -1.8  | 100   | 0.00     |
| 44 TM | Trichloroethene             | 0.457 | 0.463 | -1.3  | 104   | 0.00     |
| 45 C  | 1,2-Dichloropropane         | 0.492 | 0.493 | -0.2# | 102   | 0.00     |
| 46 T  | Dibromomethane              | 0.376 | 0.362 | 3.7   | 103   | 0.00     |
| 47 T  | Bromodichloromethane        | 0.778 | 0.729 | 6.3   | 101   | 0.00     |
| 48 T  | Methyl methacrylate         | 0.456 | 0.482 | -5.7  | 100   | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
 Data File : VV039168.D  
 Acq On : 25 Sep 2025 09:59  
 Operator : SY/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 LabSampleId :  
 VSTDCCC050

Quant Time: Sep 26 01:18:56 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 20% Max. Rel. Area : 150%

|       | Compound                    | AvgRF | CCRF  | %Dev   | Area% | Dev(min) |
|-------|-----------------------------|-------|-------|--------|-------|----------|
| 49 T  | 1,4-Dioxane                 | 0.007 | 0.008 | -14.3  | 117   | 0.00     |
| 50 S  | Toluene-d8                  | 1.181 | 1.121 | 5.1    | 103   | 0.00     |
| 51 T  | 4-Methyl-2-Pentanone        | 0.661 | 0.702 | -6.2   | 100   | 0.00     |
| 52 CM | Toluene                     | 1.147 | 1.204 | -5.0#  | 102   | 0.00     |
| 53 T  | t-1,3-Dichloropropene       | 0.678 | 0.709 | -4.6   | 102   | 0.00     |
| 54 T  | cis-1,3-Dichloropropene     | 0.749 | 0.783 | -4.5   | 101   | 0.00     |
| 55 T  | 1,1,2-Trichloroethane       | 0.514 | 0.498 | 3.1    | 105   | 0.00     |
| 56 T  | Ethyl methacrylate          | 0.611 | 0.705 | -15.4  | 103   | 0.00     |
| 57 T  | 1,3-Dichloropropane         | 0.812 | 0.802 | 1.2    | 102   | 0.00     |
| 58 T  | 2-Chloroethyl Vinyl ether   | 0.267 | 0.355 | -33.0# | 105   | 0.00     |
| 59 T  | 2-Hexanone                  | 0.463 | 0.528 | -14.0  | 101   | 0.00     |
| 60 T  | Dibromochloromethane        | 0.586 | 0.574 | 2.0    | 106   | 0.00     |
| 61 T  | 1,2-Dibromoethane           | 0.526 | 0.534 | -1.5   | 105   | 0.00     |
| 62 S  | 4-Bromofluorobenzene        | 0.486 | 0.508 | -4.5   | 105   | 0.00     |
| 63 I  | Chlorobenzene-d5            | 1.000 | 1.000 | 0.0    | 104   | 0.00     |
| 64 T  | Tetrachloroethene           | 0.419 | 0.418 | 0.2    | 106   | 0.00     |
| 65 PM | Chlorobenzene               | 1.393 | 1.377 | 1.1    | 104   | 0.00     |
| 66 T  | 1,1,1,2-Tetrachloroethane   | 0.503 | 0.483 | 4.0    | 103   | 0.00     |
| 67 C  | Ethyl Benzene               | 2.108 | 2.320 | -10.1# | 102   | 0.00     |
| 68 T  | m/p-Xylenes                 | 0.814 | 0.936 | -15.0  | 103   | 0.00     |
| 69 T  | o-Xylene                    | 0.720 | 0.841 | -16.8  | 104   | 0.00     |
| 70 T  | Styrene                     | 1.263 | 1.530 | -21.1# | 103   | 0.00     |
| 71 P  | Bromoform                   | 0.393 | 0.402 | -2.3   | 106   | 0.00     |
| 72 I  | 1,4-Dichlorobenzene-d4      | 1.000 | 1.000 | 0.0    | 104   | 0.00     |
| 73 T  | Isopropylbenzene            | 3.627 | 3.946 | -8.8   | 104   | 0.00     |
| 74 T  | N-amyl acetate              | 1.447 | 1.593 | -10.1  | 99    | 0.00     |
| 75 P  | 1,1,2,2-Tetrachloroethane   | 1.589 | 1.396 | 12.1   | 102   | 0.00     |
| 76 T  | 1,2,3-Trichloropropane      | 1.138 | 1.071 | 5.9    | 106   | 0.00     |
| 77 T  | Bromobenzene                | 0.941 | 0.929 | 1.3    | 104   | 0.00     |
| 78 T  | n-propylbenzene             | 4.416 | 4.769 | -8.0   | 101   | 0.00     |
| 79 T  | 2-Chlorotoluene             | 2.643 | 2.798 | -5.9   | 102   | 0.00     |
| 80 T  | 1,3,5-Trimethylbenzene      | 3.014 | 3.377 | -12.0  | 103   | 0.00     |
| 81 T  | trans-1,4-Dichloro-2-butene | 0.474 | 0.483 | -1.9   | 101   | 0.00     |
| 82 T  | 4-Chlorotoluene             | 3.064 | 3.303 | -7.8   | 103   | 0.00     |
| 83 T  | tert-Butylbenzene           | 3.006 | 3.221 | -7.2   | 104   | 0.00     |
| 84 T  | 1,2,4-Trimethylbenzene      | 2.926 | 3.318 | -13.4  | 102   | 0.00     |
| 85 T  | sec-Butylbenzene            | 3.987 | 4.283 | -7.4   | 102   | 0.00     |
| 86 T  | p-Isopropyltoluene          | 3.318 | 3.609 | -8.8   | 102   | 0.00     |
| 87 T  | 1,3-Dichlorobenzene         | 1.904 | 1.857 | 2.5    | 104   | 0.00     |
| 88 T  | 1,4-Dichlorobenzene         | 2.089 | 1.903 | 8.9    | 102   | 0.00     |
| 89 T  | n-Butylbenzene              | 3.169 | 3.461 | -9.2   | 99    | 0.00     |
| 90 T  | Hexachloroethane            | 0.817 | 0.723 | 11.5   | 103   | 0.00     |
| 91 T  | 1,2-Dichlorobenzene         | 1.779 | 1.736 | 2.4    | 104   | 0.00     |
| 92 T  | 1,2-Dibromo-3-Chloropropane | 0.330 | 0.292 | 11.5   | 106   | 0.00     |
| 93 T  | 1,2,4-Trichlorobenzene      | 1.020 | 1.062 | -4.1   | 102   | 0.00     |
| 94 T  | Hexachlorobutadiene         | 0.530 | 0.487 | 8.1    | 105   | 0.00     |
| 95 T  | Naphthalene                 | 3.296 | 3.807 | -15.5  | 104   | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
Data File : VV039168.D  
Acq On : 25 Sep 2025 09:59  
Operator : SY/MD  
Sample : VSTDCCC050  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
LabSampleId :  
VSTDCCC050

Quant Time: Sep 26 01:18:56 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 20% Max. Rel. Area : 150%

|      | Compound               | AvgRF | CCRF  | %Dev | Area% | Dev(min) |
|------|------------------------|-------|-------|------|-------|----------|
| 96 T | 1,2,3-Trichlorobenzene | 1.043 | 1.082 | -3.7 | 101   | 0.00     |

(#) = Out of Range                      SPCC's out = 0    CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
 Data File : VV039168.D  
 Acq On : 25 Sep 2025 09:59  
 Operator : SY/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 LabSampleId :  
 VSTDCCC050

Quant Time: Sep 26 01:18:56 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 20% Max. Rel. Area : 150%

|       | Compound                    | Amount  | Calc.   | %Dev  | Area% | Dev(min) |
|-------|-----------------------------|---------|---------|-------|-------|----------|
| 1 I   | Pentafluorobenzene          | 50.000  | 50.000  | 0.0   | 107   | 0.00     |
| 2 T   | Dichlorodifluoromethane     | 50.000  | 43.088  | 13.8  | 99    | 0.00     |
| 3 P   | Chloromethane               | 50.000  | 42.924  | 14.2  | 102   | 0.00     |
| 4 C   | Vinyl Chloride              | 50.000  | 45.040  | 9.9#  | 105   | 0.00     |
| 5 T   | Bromomethane                | 50.000  | 39.912  | 20.2# | 97    | 0.00     |
| 6 T   | Chloroethane                | 50.000  | 47.163  | 5.7   | 102   | 0.00     |
| 7 T   | Trichlorofluoromethane      | 50.000  | 45.520  | 9.0   | 108   | 0.00     |
| 8 T   | Diethyl Ether               | 50.000  | 46.259  | 7.5   | 109   | 0.00     |
| 9 T   | 1,1,2-Trichlorotrifluoroeth | 50.000  | 45.275  | 9.5   | 109   | 0.00     |
| 10 T  | Methyl Iodide               | 50.000  | 43.640  | 12.7  | 89    | 0.00     |
| 11 T  | Tert butyl alcohol          | 250.000 | 249.596 | 0.2   | 120   | 0.00     |
| 12 CM | 1,1-Dichloroethene          | 50.000  | 46.283  | 7.4#  | 111   | 0.00     |
| 13 T  | Acrolein                    | 250.000 | 229.786 | 8.1   | 99    | 0.00     |
| 14 T  | Allyl chloride              | 50.000  | 44.523  | 11.0  | 104   | 0.00     |
| 15 T  | Acrylonitrile               | 250.000 | 223.829 | 10.5  | 108   | 0.00     |
| 16 T  | Acetone                     | 250.000 | 224.844 | 10.1  | 97    | 0.00     |
| 17 T  | Carbon Disulfide            | 50.000  | 43.464  | 13.1  | 103   | 0.00     |
| 18 T  | Methyl Acetate              | 50.000  | 46.408  | 7.2   | 109   | 0.00     |
| 19 T  | Methyl tert-butyl Ether     | 50.000  | 46.768  | 6.5   | 108   | 0.00     |
| 20 T  | Methylene Chloride          | 50.000  | 50.438  | -0.9  | 109   | 0.00     |
| 21 T  | trans-1,2-Dichloroethene    | 50.000  | 45.131  | 9.7   | 107   | 0.00     |
| 22 T  | Diisopropyl ether           | 50.000  | 44.947  | 10.1  | 99    | 0.00     |
| 23 T  | Vinyl Acetate               | 250.000 | 228.109 | 8.8   | 100   | 0.00     |
| 24 P  | 1,1-Dichloroethane          | 50.000  | 43.389  | 13.2  | 106   | 0.00     |
| 25 T  | 2-Butanone                  | 250.000 | 234.849 | 6.1   | 100   | 0.00     |
| 26 T  | 2,2-Dichloropropane         | 50.000  | 40.800  | 18.4  | 99    | 0.00     |
| 27 T  | cis-1,2-Dichloroethene      | 50.000  | 46.579  | 6.8   | 102   | 0.00     |
| 28 T  | Bromochloromethane          | 50.000  | 44.019  | 12.0  | 97    | 0.00     |
| 29 T  | Tetrahydrofuran             | 250.000 | 244.856 | 2.1   | 103   | 0.00     |
| 30 C  | Chloroform                  | 50.000  | 43.324  | 13.4# | 103   | 0.00     |
| 31 T  | Cyclohexane                 | 50.000  | 44.432  | 11.1  | 99    | 0.00     |
| 32 T  | 1,1,1-Trichloroethane       | 50.000  | 44.071  | 11.9  | 104   | 0.00     |
| 33 S  | 1,2-Dichloroethane-d4       | 50.000  | 42.102  | 15.8  | 101   | 0.00     |
| 34 I  | 1,4-Difluorobenzene         | 50.000  | 50.000  | 0.0   | 104   | 0.00     |
| 35 S  | Dibromofluoromethane        | 50.000  | 47.826  | 4.3   | 106   | 0.00     |
| 36 T  | 1,1-Dichloropropene         | 50.000  | 49.586  | 0.8   | 101   | 0.00     |
| 37 T  | Ethyl Acetate               | 50.000  | 47.290  | 5.4   | 102   | 0.00     |
| 38 T  | Carbon Tetrachloride        | 50.000  | 46.350  | 7.3   | 102   | 0.00     |
| 39 T  | Methylcyclohexane           | 50.000  | 54.277  | -8.6  | 101   | 0.00     |
| 40 TM | Benzene                     | 50.000  | 48.671  | 2.7   | 101   | 0.00     |
| 41 T  | Methacrylonitrile           | 50.000  | 47.017  | 6.0   | 99    | 0.00     |
| 42 TM | 1,2-Dichloroethane          | 50.000  | 46.050  | 7.9   | 100   | 0.00     |
| 43 T  | Isopropyl Acetate           | 50.000  | 50.848  | -1.7  | 100   | 0.00     |
| 44 TM | Trichloroethene             | 50.000  | 50.680  | -1.4  | 104   | 0.00     |
| 45 C  | 1,2-Dichloropropane         | 50.000  | 50.088  | -0.2# | 102   | 0.00     |
| 46 T  | Dibromomethane              | 50.000  | 48.059  | 3.9   | 103   | 0.00     |
| 47 T  | Bromodichloromethane        | 50.000  | 46.843  | 6.3   | 101   | 0.00     |
| 48 T  | Methyl methacrylate         | 50.000  | 52.825  | -5.7  | 100   | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
 Data File : VV039168.D  
 Acq On : 25 Sep 2025 09:59  
 Operator : SY/MD  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 LabSampleId :  
 VSTDCCC050

Quant Time: Sep 26 01:18:56 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 20% Max. Rel. Area : 150%

|       | Compound                    | Amount   | Calc.    | %Dev   | Area% | Dev(min) |
|-------|-----------------------------|----------|----------|--------|-------|----------|
| 49 T  | 1,4-Dioxane                 | 1000.000 | 1135.751 | -13.6  | 117   | 0.00     |
| 50 S  | Toluene-d8                  | 50.000   | 47.485   | 5.0    | 103   | 0.00     |
| 51 T  | 4-Methyl-2-Pentanone        | 250.000  | 265.177  | -6.1   | 100   | 0.00     |
| 52 CM | Toluene                     | 50.000   | 52.474   | -4.9#  | 102   | 0.00     |
| 53 T  | t-1,3-Dichloropropene       | 50.000   | 52.243   | -4.5   | 102   | 0.00     |
| 54 T  | cis-1,3-Dichloropropene     | 50.000   | 52.312   | -4.6   | 101   | 0.00     |
| 55 T  | 1,1,2-Trichloroethane       | 50.000   | 48.443   | 3.1    | 105   | 0.00     |
| 56 T  | Ethyl methacrylate          | 50.000   | 57.707   | -15.4  | 103   | 0.00     |
| 57 T  | 1,3-Dichloropropane         | 50.000   | 49.373   | 1.3    | 102   | 0.00     |
| 58 T  | 2-Chloroethyl Vinyl ether   | 250.000  | 257.039  | -2.8   | 105   | 0.00     |
| 59 T  | 2-Hexanone                  | 250.000  | 247.438  | 1.0    | 101   | 0.00     |
| 60 T  | Dibromochloromethane        | 50.000   | 48.938   | 2.1    | 106   | 0.00     |
| 61 T  | 1,2-Dibromoethane           | 50.000   | 50.719   | -1.4   | 105   | 0.00     |
| 62 S  | 4-Bromofluorobenzene        | 50.000   | 52.218   | -4.4   | 105   | 0.00     |
| 63 I  | Chlorobenzene-d5            | 50.000   | 50.000   | 0.0    | 104   | 0.00     |
| 64 T  | Tetrachloroethene           | 50.000   | 49.857   | 0.3    | 106   | 0.00     |
| 65 PM | Chlorobenzene               | 50.000   | 49.450   | 1.1    | 104   | 0.00     |
| 66 T  | 1,1,1,2-Tetrachloroethane   | 50.000   | 47.975   | 4.0    | 103   | 0.00     |
| 67 C  | Ethyl Benzene               | 50.000   | 55.024   | -10.0# | 102   | 0.00     |
| 68 T  | m/p-Xylenes                 | 100.000  | 115.018  | -15.0  | 103   | 0.00     |
| 69 T  | o-Xylene                    | 50.000   | 58.347   | -16.7  | 104   | 0.00     |
| 70 T  | Styrene                     | 50.000   | 50.114   | -0.2   | 103   | 0.00     |
| 71 P  | Bromoform                   | 50.000   | 51.118   | -2.2   | 106   | 0.00     |
| 72 I  | 1,4-Dichlorobenzene-d4      | 50.000   | 50.000   | 0.0    | 104   | 0.00     |
| 73 T  | Isopropylbenzene            | 50.000   | 54.407   | -8.8   | 104   | 0.00     |
| 74 T  | N-amyl acetate              | 50.000   | 55.064   | -10.1  | 99    | 0.00     |
| 75 P  | 1,1,2,2-Tetrachloroethane   | 50.000   | 43.928   | 12.1   | 102   | 0.00     |
| 76 T  | 1,2,3-Trichloropropane      | 50.000   | 47.019   | 6.0    | 106   | 0.00     |
| 77 T  | Bromobenzene                | 50.000   | 49.357   | 1.3    | 104   | 0.00     |
| 78 T  | n-propylbenzene             | 50.000   | 53.995   | -8.0   | 101   | 0.00     |
| 79 T  | 2-Chlorotoluene             | 50.000   | 52.939   | -5.9   | 102   | 0.00     |
| 80 T  | 1,3,5-Trimethylbenzene      | 50.000   | 56.020   | -12.0  | 103   | 0.00     |
| 81 T  | trans-1,4-Dichloro-2-butene | 50.000   | 50.926   | -1.9   | 101   | 0.00     |
| 82 T  | 4-Chlorotoluene             | 50.000   | 53.904   | -7.8   | 103   | 0.00     |
| 83 T  | tert-Butylbenzene           | 50.000   | 53.579   | -7.2   | 104   | 0.00     |
| 84 T  | 1,2,4-Trimethylbenzene      | 50.000   | 56.693   | -13.4  | 102   | 0.00     |
| 85 T  | sec-Butylbenzene            | 50.000   | 53.713   | -7.4   | 102   | 0.00     |
| 86 T  | p-Isopropyltoluene          | 50.000   | 54.380   | -8.8   | 102   | 0.00     |
| 87 T  | 1,3-Dichlorobenzene         | 50.000   | 48.773   | 2.5    | 104   | 0.00     |
| 88 T  | 1,4-Dichlorobenzene         | 50.000   | 45.535   | 8.9    | 102   | 0.00     |
| 89 T  | n-Butylbenzene              | 50.000   | 54.607   | -9.2   | 99    | 0.00     |
| 90 T  | Hexachloroethane            | 50.000   | 44.260   | 11.5   | 103   | 0.00     |
| 91 T  | 1,2-Dichlorobenzene         | 50.000   | 48.799   | 2.4    | 104   | 0.00     |
| 92 T  | 1,2-Dibromo-3-Chloropropane | 50.000   | 50.961   | -1.9   | 106   | 0.00     |
| 93 T  | 1,2,4-Trichlorobenzene      | 50.000   | 52.055   | -4.1   | 102   | 0.00     |
| 94 T  | Hexachlorobutadiene         | 50.000   | 45.975   | 8.0    | 105   | 0.00     |
| 95 T  | Naphthalene                 | 50.000   | 57.749   | -15.5  | 104   | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
Data File : VV039168.D  
Acq On : 25 Sep 2025 09:59  
Operator : SY/MD  
Sample : VSTDCCC050  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
LabSampleId :  
VSTDCCC050

Quant Time: Sep 26 01:18:56 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 20% Max. Rel. Area : 150%

|      | Compound               | Amount | Calc.  | %Dev | Area% | Dev(min) |
|------|------------------------|--------|--------|------|-------|----------|
| 96 T | 1,2,3-Trichlorobenzene | 50.000 | 51.848 | -3.7 | 101   | 0.00     |

(#) = Out of Range                      SPCC's out = 0    CCC's out = 6





# QC SAMPLE DATA

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092225\  
Data File : VV039134.D  
Acq On : 22 Sep 2025 10:15  
Operator : SY/MD  
Sample : BFB  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Title : SW846 8260  
Last Update : Wed Sep 24 08:08:08 2025



AutoFind: Scans 2785, 2786, 2787; Background Corrected with Scan 2774



| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 50             | 95              | 15              | 40              | 19.6         | 22014      | PASS                |
| 75             | 95              | 30              | 60              | 51.9         | 58147      | PASS                |
| 95             | 95              | 100             | 100             | 100.0        | 112096     | PASS                |
| 96             | 95              | 5               | 9               | 7.1          | 7955       | PASS                |
| 173            | 174             | 0.00            | 2               | 1.6          | 1275       | PASS                |
| 174            | 95              | 50              | 100             | 73.3         | 82200      | PASS                |
| 175            | 174             | 5               | 9               | 7.5          | 6152       | PASS                |
| 176            | 174             | 95              | 101             | 96.0         | 78917      | PASS                |
| 177            | 176             | 5               | 9               | 6.3          | 4933       | PASS                |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039145.D  
Acq On : 24 Sep 2025 09:52  
Operator : SY/MD  
Sample : BFB  
Misc : 25.0mL/MSVOA\_V/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Title : SW846 8260  
Last Update : Wed Sep 24 08:08:08 2025



AutoFind: Scans 2786, 2787, 2788; Background Corrected with Scan 2776



| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 50             | 95              | 15              | 40              | 19.6         | 13613      | PASS                |
| 75             | 95              | 30              | 60              | 51.7         | 35965      | PASS                |
| 95             | 95              | 100             | 100             | 100.0        | 69512      | PASS                |
| 96             | 95              | 5               | 9               | 6.7          | 4635       | PASS                |
| 173            | 174             | 0.00            | 2               | 1.9          | 995        | PASS                |
| 174            | 95              | 50              | 100             | 77.1         | 53579      | PASS                |
| 175            | 174             | 5               | 9               | 6.7          | 3610       | PASS                |
| 176            | 174             | 95              | 101             | 95.3         | 51064      | PASS                |
| 177            | 176             | 5               | 9               | 6.6          | 3395       | PASS                |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
Data File : VV039167.D  
Acq On : 25 Sep 2025 09:14  
Operator : SY/MD  
Sample : BFB  
Misc : 25.0mL/MSVOA\_V/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Title : SW846 8260  
Last Update : Wed Sep 24 08:08:08 2025



AutoFind: Scans 2787, 2788, 2789; Background Corrected with Scan 2776

| Target | Rel. to | Lower  | Upper  | Rel.  | Raw   | Result    |
|--------|---------|--------|--------|-------|-------|-----------|
| Mass   | Mass    | Limit% | Limit% | Abn%  | Abn   | Pass/Fail |
| 50     | 95      | 15     | 40     | 16.2  | 12139 | PASS      |
| 75     | 95      | 30     | 60     | 49.9  | 37445 | PASS      |
| 95     | 95      | 100    | 100    | 100.0 | 75093 | PASS      |
| 96     | 95      | 5      | 9      | 6.3   | 4756  | PASS      |
| 173    | 174     | 0.00   | 2      | 1.7   | 944   | PASS      |
| 174    | 95      | 50     | 100    | 72.3  | 54320 | PASS      |
| 175    | 174     | 5      | 9      | 7.9   | 4279  | PASS      |
| 176    | 174     | 95     | 101    | 96.6  | 52461 | PASS      |
| 177    | 176     | 5      | 9      | 6.3   | 3317  | PASS      |

## Report of Analysis

|                    |                 |        |      |                 |              |
|--------------------|-----------------|--------|------|-----------------|--------------|
| Client:            | G Environmental |        |      | Date Collected: |              |
| Project:           | Summer          |        |      | Date Received:  |              |
| Client Sample ID:  | VV0924WBL01     |        |      | SDG No.:        | Q3175        |
| Lab Sample ID:     | VV0924WBL01     |        |      | Matrix:         | Water        |
| Analytical Method: | 8260D           |        |      | % Solid:        | 0            |
| Sample Wt/Vol:     | 5               | Units: | mL   | Final Vol:      | 5000 uL      |
| Soil Aliquot Vol:  |                 |        | uL   | Test:           | VOCMS Group1 |
| GC Column:         | DB-624UI        | ID :   | 0.18 | Level :         | LOW          |
| Prep Method :      |                 |        |      |                 |              |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039148.D        | 1         | 09/24/25 13:09 | VV092425      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL   | LOQ / CRQL | Units |
|----------------|--------------------------------|-------|-----------|-------|------------|-------|
| <b>TARGETS</b> |                                |       |           |       |            |       |
| 75-71-8        | Dichlorodifluoromethane        | 0.22  | U         | 0.22  | 1.00       | ug/L  |
| 74-87-3        | Chloromethane                  | 0.32  | U         | 0.32  | 1.00       | ug/L  |
| 75-01-4        | Vinyl Chloride                 | 0.26  | U         | 0.26  | 1.00       | ug/L  |
| 74-83-9        | Bromomethane                   | 1.40  | U         | 1.40  | 5.00       | ug/L  |
| 75-00-3        | Chloroethane                   | 0.47  | U         | 0.47  | 1.00       | ug/L  |
| 75-69-4        | Trichlorofluoromethane         | 0.33  | U         | 0.33  | 1.00       | ug/L  |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 0.25  | U         | 0.25  | 1.00       | ug/L  |
| 75-65-0        | Tert butyl alcohol             | 5.50  | U         | 5.50  | 25.0       | ug/L  |
| 75-35-4        | 1,1-Dichloroethene             | 0.23  | U         | 0.23  | 1.00       | ug/L  |
| 67-64-1        | Acetone                        | 1.50  | U         | 1.50  | 5.00       | ug/L  |
| 75-15-0        | Carbon Disulfide               | 0.21  | U         | 0.21  | 1.00       | ug/L  |
| 1634-04-4      | Methyl tert-butyl Ether        | 0.16  | U         | 0.16  | 1.00       | ug/L  |
| 79-20-9        | Methyl Acetate                 | 0.27  | U         | 0.27  | 1.00       | ug/L  |
| 75-09-2        | Methylene Chloride             | 0.28  | U         | 0.28  | 1.00       | ug/L  |
| 156-60-5       | trans-1,2-Dichloroethene       | 0.23  | U         | 0.23  | 1.00       | ug/L  |
| 75-34-3        | 1,1-Dichloroethane             | 0.23  | U         | 0.23  | 1.00       | ug/L  |
| 110-82-7       | Cyclohexane                    | 1.50  | U         | 1.50  | 5.00       | ug/L  |
| 78-93-3        | 2-Butanone                     | 0.98  | U         | 0.98  | 5.00       | ug/L  |
| 56-23-5        | Carbon Tetrachloride           | 0.25  | U         | 0.25  | 1.00       | ug/L  |
| 156-59-2       | cis-1,2-Dichloroethene         | 0.19  | U         | 0.19  | 1.00       | ug/L  |
| 74-97-5        | Bromochloromethane             | 0.22  | U         | 0.22  | 1.00       | ug/L  |
| 67-66-3        | Chloroform                     | 0.25  | U         | 0.25  | 1.00       | ug/L  |
| 71-55-6        | 1,1,1-Trichloroethane          | 0.20  | U         | 0.20  | 1.00       | ug/L  |
| 108-87-2       | Methylcyclohexane              | 0.16  | U         | 0.16  | 1.00       | ug/L  |
| 71-43-2        | Benzene                        | 0.15  | U         | 0.15  | 1.00       | ug/L  |
| 107-06-2       | 1,2-Dichloroethane             | 0.22  | U         | 0.22  | 1.00       | ug/L  |
| 79-01-6        | Trichloroethene                | 0.090 | U         | 0.090 | 1.00       | ug/L  |
| 78-87-5        | 1,2-Dichloropropane            | 0.20  | U         | 0.20  | 1.00       | ug/L  |
| 75-27-4        | Bromodichloromethane           | 0.22  | U         | 0.22  | 1.00       | ug/L  |
| 108-10-1       | 4-Methyl-2-Pentanone           | 0.68  | U         | 0.68  | 5.00       | ug/L  |

### Report of Analysis

|                    |                 |           |                 |              |
|--------------------|-----------------|-----------|-----------------|--------------|
| Client:            | G Environmental |           | Date Collected: |              |
| Project:           | Summer          |           | Date Received:  |              |
| Client Sample ID:  | VV0924WBL01     |           | SDG No.:        | Q3175        |
| Lab Sample ID:     | VV0924WBL01     |           | Matrix:         | Water        |
| Analytical Method: | 8260D           |           | % Solid:        | 0            |
| Sample Wt/Vol:     | 5               | Units: mL | Final Vol:      | 5000 uL      |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOCMS Group1 |
| GC Column:         | DB-624UI        | ID : 0.18 | Level :         | LOW          |
| Prep Method :      |                 |           |                 |              |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039148.D        | 1         | 09/24/25 13:09 | VV092425      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units   |
|---------------------------|-----------------------------|--------|-----------|----------|------------|---------|
| 108-88-3                  | Toluene                     | 0.14   | U         | 0.14     | 1.00       | ug/L    |
| 10061-02-6                | t-1,3-Dichloropropene       | 0.17   | U         | 0.17     | 1.00       | ug/L    |
| 10061-01-5                | cis-1,3-Dichloropropene     | 0.16   | U         | 0.16     | 1.00       | ug/L    |
| 79-00-5                   | 1,1,2-Trichloroethane       | 0.21   | U         | 0.21     | 1.00       | ug/L    |
| 591-78-6                  | 2-Hexanone                  | 0.89   | U         | 0.89     | 5.00       | ug/L    |
| 124-48-1                  | Dibromochloromethane        | 0.18   | U         | 0.18     | 1.00       | ug/L    |
| 106-93-4                  | 1,2-Dibromoethane           | 0.15   | U         | 0.15     | 1.00       | ug/L    |
| 127-18-4                  | Tetrachloroethene           | 0.23   | U         | 0.23     | 1.00       | ug/L    |
| 108-90-7                  | Chlorobenzene               | 0.12   | U         | 0.12     | 1.00       | ug/L    |
| 100-41-4                  | Ethyl Benzene               | 0.13   | U         | 0.13     | 1.00       | ug/L    |
| 179601-23-1               | m/p-Xylenes                 | 0.24   | U         | 0.24     | 2.00       | ug/L    |
| 95-47-6                   | o-Xylene                    | 0.12   | U         | 0.12     | 1.00       | ug/L    |
| 100-42-5                  | Styrene                     | 0.15   | U         | 0.15     | 1.00       | ug/L    |
| 75-25-2                   | Bromoform                   | 0.19   | U         | 0.19     | 1.00       | ug/L    |
| 98-82-8                   | Isopropylbenzene            | 0.12   | U         | 0.12     | 1.00       | ug/L    |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 0.26   | U         | 0.26     | 1.00       | ug/L    |
| 541-73-1                  | 1,3-Dichlorobenzene         | 0.16   | U         | 0.16     | 1.00       | ug/L    |
| 106-46-7                  | 1,4-Dichlorobenzene         | 0.19   | U         | 0.19     | 1.00       | ug/L    |
| 95-50-1                   | 1,2-Dichlorobenzene         | 0.16   | U         | 0.16     | 1.00       | ug/L    |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 0.53   | U         | 0.53     | 1.00       | ug/L    |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 0.20   | U         | 0.20     | 1.00       | ug/L    |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 0.20   | U         | 0.20     | 1.00       | ug/L    |
| <b>SURROGATES</b>         |                             |        |           |          |            |         |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 62.2   |           | 74 - 125 | 124%       | SPK: 50 |
| 1868-53-7                 | Dibromofluoromethane        | 54.6   |           | 75 - 124 | 109%       | SPK: 50 |
| 2037-26-5                 | Toluene-d8                  | 44.1   |           | 86 - 113 | 88%        | SPK: 50 |
| 460-00-4                  | 4-Bromofluorobenzene        | 48.0   |           | 77 - 121 | 96%        | SPK: 50 |
| <b>INTERNAL STANDARDS</b> |                             |        |           |          |            |         |
| 363-72-4                  | Pentafluorobenzene          | 464000 | 4.596     |          |            |         |
| 540-36-3                  | 1,4-Difluorobenzene         | 859000 | 5.532     |          |            |         |
| 3114-55-4                 | Chlorobenzene-d5            | 832000 | 8.773     |          |            |         |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 379000 | 11.175    |          |            |         |

## Report of Analysis

|                    |                 |        |      |                 |              |
|--------------------|-----------------|--------|------|-----------------|--------------|
| Client:            | G Environmental |        |      | Date Collected: |              |
| Project:           | Summer          |        |      | Date Received:  |              |
| Client Sample ID:  | VV0924WBL01     |        |      | SDG No.:        | Q3175        |
| Lab Sample ID:     | VV0924WBL01     |        |      | Matrix:         | Water        |
| Analytical Method: | 8260D           |        |      | % Solid:        | 0            |
| Sample Wt/Vol:     | 5               | Units: | mL   | Final Vol:      | 5000 uL      |
| Soil Aliquot Vol:  |                 |        | uL   | Test:           | VOCMS Group1 |
| GC Column:         | DB-624UI        | ID :   | 0.18 | Level :         | LOW          |
| Prep Method :      |                 |        |      |                 |              |

| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
|-------------------|-----------|----------------|---------------|
| VV039148.D        | 1         | 09/24/25 13:09 | VV092425      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected  
LOQ = Limit of Quantitation  
MDL = Method Detection Limit  
LOD = Limit of Detection  
E = Value Exceeds Calibration Range  
Q = indicates LCS control criteria did not meet requirements  
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value  
B = Analyte Found in Associated Method Blank  
N = Presumptive Evidence of a Compound  
\* = Values outside of QC limits  
D = Dilution  
( ) = Laboratory InHouse Limit  
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039148.D  
Acq On : 24 Sep 2025 13:09  
Operator : SY/MD  
Sample : VV0924WBL01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VV0924WBL01

Quant Time: Sep 25 04:44:19 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 4.596  | 168  | 463770   | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 5.532  | 114  | 858892   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 8.773  | 117  | 831675   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 11.175 | 152  | 379013   | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |           |
|-----------------------------|--------|-------|----------|----------|------|-----------|
| System Monitoring Compounds |        |       |          |          |      |           |
| 33) 1,2-Dichloroethane-d4   | 4.937  | 65    | 417696   | 62.230   | ug/l | 0.00      |
| Spiked Amount               | 50.000 | Range | 81 - 118 | Recovery | =    | 124.460%# |
| 35) Dibromofluoromethane    | 4.497  | 113   | 376781   | 54.568   | ug/l | 0.00      |
| Spiked Amount               | 50.000 | Range | 80 - 119 | Recovery | =    | 109.140%  |
| 50) Toluene-d8              | 7.236  | 98    | 894968   | 44.132   | ug/l | 0.00      |
| Spiked Amount               | 50.000 | Range | 89 - 112 | Recovery | =    | 88.260%#  |
| 62) 4-Bromofluorobenzene    | 10.001 | 95    | 400947   | 48.026   | ug/l | 0.00      |
| Spiked Amount               | 50.000 | Range | 85 - 114 | Recovery | =    | 96.060%   |


| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039148.D  
Acq On : 24 Sep 2025 13:09  
Operator : SY/MD  
Sample : VV0924WBL01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VV0924WBL01

Quant Time: Sep 25 04:44:19 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration









#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 4.596 min Scan# 1106  
Delta R.T. -0.001 min  
Lab File: VV039148.D  
Acq: 24 Sep 2025 13:09




Instrument :  
MSVOA\_V  
ClientSampleId :  
VV0924WBL01


Tgt Ion:168 Resp: 463770  
Ion Ratio Lower Upper  
168 100  
99 66.7 49.6 74.4



#33  
1,2-Dichloroethane-d4  
Concen: 62.230 ug/l  
RT: 4.937 min Scan# 1212  
Delta R.T. -0.000 min  
Lab File: VV039148.D  
Acq: 24 Sep 2025 13:09

Tgt Ion: 65 Resp: 417696  
Ion Ratio Lower Upper  
65 100  
67 53.6 0.0 107.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 5.532 min Scan# 1396

Delta R.T. 0.003 min

Lab File: VV039148.D

Acq: 24 Sep 2025 13:09

Instrument :

MSVOA\_V

ClientSampleId :

VV0924WBL01





Tgt Ion:114 Resp: 858892

Ion Ratio Lower Upper

114 100

63 20.4 0.0 42.6

88 15.4 0.0 30.8



#35

Dibromofluoromethane

Concen: 54.568 ug/l

RT: 4.497 min Scan# 1075

Delta R.T. -0.000 min

Lab File: VV039148.D

Acq: 24 Sep 2025 13:09




Tgt Ion:113 Resp: 376781


Ion Ratio Lower Upper

113 100

111 102.7 82.4 123.6

192 16.6 13.8 20.8









#50  
Toluene-d8  
Concen: 44.132 ug/l  
RT: 7.236 min Scan# 1926  
Delta R.T. 0.003 min  
Lab File: VV039148.D  
Acq: 24 Sep 2025 13:09




Instrument :  
MSVOA\_V  
ClientSampleId :  
VV0924WBL01


Tgt Ion: 98 Resp: 894968  
Ion Ratio Lower Upper  
98 100  
100 63.4 52.2 78.2



#62  
4-Bromofluorobenzene  
Concen: 48.026 ug/l  
RT: 10.001 min Scan# 2787  
Delta R.T. 0.003 min  
Lab File: VV039148.D  
Acq: 24 Sep 2025 13:09

Tgt Ion: 95 Resp: 400947  
Ion Ratio Lower Upper  
95 100  
174 74.8 0.0 150.4  
176 70.4 0.0 144.2





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 8.773 min Scan# 2405

Delta R.T. 0.000 min

Lab File: VV039148.D

Acq: 24 Sep 2025 13:09

Instrument :

MSVOA\_V

ClientSampleId :

VV0924WBL01





Tgt Ion:117 Resp: 831675

Ion Ratio Lower Upper

117 100

82 56.5 45.4 68.2

119 31.7 25.6 38.4



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 11.175 min Scan# 3152

Delta R.T. 0.003 min

Lab File: VV039148.D

Acq: 24 Sep 2025 13:09




Tgt Ion:152 Resp: 379013

Ion Ratio Lower Upper

152 100

115 61.8 44.8 134.4

150 157.4 0.0 353.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039148.D  
Acq On : 24 Sep 2025 13:09  
Operator : SY/MD  
Sample : VV0924WBL01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VV0924WBL01

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M

Title : SW846 8260

Signal : TIC: VV039148.D\data.ms

| peak<br># | R.T.<br>min | first<br>scan | max<br>scan | last<br>scan | PK<br>TY | peak<br>height | corr.<br>area | corr.<br>% max. | % of<br>total |
|-----------|-------------|---------------|-------------|--------------|----------|----------------|---------------|-----------------|---------------|
| 1         | 1.426       | 111           | 120         | 130          | rBV5     | 16630          | 35090         | 1.38%           | 0.230%        |
| 2         | 2.455       | 433           | 440         | 451          | rVB4     | 18087          | 31588         | 1.24%           | 0.207%        |
| 3         | 4.497       | 1057          | 1075        | 1094         | rBV2     | 450386         | 1192254       | 46.76%          | 7.800%        |
| 4         | 4.596       | 1094          | 1106        | 1145         | rVB2     | 557832         | 1545205       | 60.61%          | 10.110%       |
| 5         | 4.937       | 1195          | 1212        | 1245         | rBV      | 467561         | 1171324       | 45.94%          | 7.664%        |
| 6         | 5.532       | 1383          | 1397        | 1433         | rBV      | 882266         | 2082399       | 81.67%          | 13.624%       |
| 7         | 7.236       | 1915          | 1927        | 1952         | rBV      | 1260224        | 2372338       | 93.05%          | 15.521%       |
| 8         | 8.773       | 2395          | 2405        | 2443         | rBV      | 1482069        | 2549623       | 100.00%         | 16.681%       |
| 9         | 10.001      | 2776          | 2787        | 2809         | rBV      | 1194118        | 1970915       | 77.30%          | 12.895%       |
| 10        | 11.172      | 3140          | 3151        | 3175         | rBV      | 1430252        | 2333631       | 91.53%          | 15.268%       |




Sum of corrected areas: 15284367

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039148.D  
Acq On : 24 Sep 2025 13:09  
Operator : SY/MD  
Sample : VV0924WBL01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VV0924WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039148.D  
Acq On : 24 Sep 2025 13:09  
Operator : SY/MD  
Sample : VV0924WBL01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VV0924WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

\*\*\*\*\*

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039148.D  
Acq On : 24 Sep 2025 13:09  
Operator : SY/MD  
Sample : VV0924WBL01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VV0924WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST0.L  
TIC Integration Parameters: LSCINT.P

| TIC Top Hit name | RT | EstConc | Units | Response | --Internal Standard-- |    |      |      |
|------------------|----|---------|-------|----------|-----------------------|----|------|------|
|                  |    |         |       |          | #                     | RT | Resp | Conc |

-----



## Report of Analysis

|                    |                                            |                 |                              |
|--------------------|--------------------------------------------|-----------------|------------------------------|
| Client:            | G Environmental                            | Date Collected: |                              |
| Project:           | Summer                                     | Date Received:  |                              |
| Client Sample ID:  | VV0925WBL01                                | SDG No.:        | Q3175                        |
| Lab Sample ID:     | VV0925WBL01                                | Matrix:         | Water                        |
| Analytical Method: | 8260D                                      | % Solid:        | 0                            |
| Sample Wt/Vol:     | 5                      Units:    mL        | Final Vol:      | 5000                      uL |
| Soil Aliquot Vol:  | uL                                         | Test:           | VOCMS Group1                 |
| GC Column:         | DB-624UI                      ID :    0.18 | Level :         | LOW                          |
| Prep Method :      |                                            |                 |                              |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039170.D        | 1         | 09/25/25 11:01 | VV092525      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units   |
|---------------------------|-----------------------------|--------|-----------|----------|------------|---------|
| 108-88-3                  | Toluene                     | 0.14   | U         | 0.14     | 1.00       | ug/L    |
| 10061-02-6                | t-1,3-Dichloropropene       | 0.17   | U         | 0.17     | 1.00       | ug/L    |
| 10061-01-5                | cis-1,3-Dichloropropene     | 0.16   | U         | 0.16     | 1.00       | ug/L    |
| 79-00-5                   | 1,1,2-Trichloroethane       | 0.21   | U         | 0.21     | 1.00       | ug/L    |
| 591-78-6                  | 2-Hexanone                  | 0.89   | U         | 0.89     | 5.00       | ug/L    |
| 124-48-1                  | Dibromochloromethane        | 0.18   | U         | 0.18     | 1.00       | ug/L    |
| 106-93-4                  | 1,2-Dibromoethane           | 0.15   | U         | 0.15     | 1.00       | ug/L    |
| 127-18-4                  | Tetrachloroethene           | 0.23   | U         | 0.23     | 1.00       | ug/L    |
| 108-90-7                  | Chlorobenzene               | 0.12   | U         | 0.12     | 1.00       | ug/L    |
| 100-41-4                  | Ethyl Benzene               | 0.13   | U         | 0.13     | 1.00       | ug/L    |
| 179601-23-1               | m/p-Xylenes                 | 0.24   | U         | 0.24     | 2.00       | ug/L    |
| 95-47-6                   | o-Xylene                    | 0.12   | U         | 0.12     | 1.00       | ug/L    |
| 100-42-5                  | Styrene                     | 0.15   | U         | 0.15     | 1.00       | ug/L    |
| 75-25-2                   | Bromoform                   | 0.19   | U         | 0.19     | 1.00       | ug/L    |
| 98-82-8                   | Isopropylbenzene            | 0.12   | U         | 0.12     | 1.00       | ug/L    |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 0.26   | U         | 0.26     | 1.00       | ug/L    |
| 541-73-1                  | 1,3-Dichlorobenzene         | 0.16   | U         | 0.16     | 1.00       | ug/L    |
| 106-46-7                  | 1,4-Dichlorobenzene         | 0.19   | U         | 0.19     | 1.00       | ug/L    |
| 95-50-1                   | 1,2-Dichlorobenzene         | 0.16   | U         | 0.16     | 1.00       | ug/L    |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 0.53   | U         | 0.53     | 1.00       | ug/L    |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 0.20   | U         | 0.20     | 1.00       | ug/L    |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 0.20   | U         | 0.20     | 1.00       | ug/L    |
| <b>SURROGATES</b>         |                             |        |           |          |            |         |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 60.1   |           | 74 - 125 | 120%       | SPK: 50 |
| 1868-53-7                 | Dibromofluoromethane        | 53.3   |           | 75 - 124 | 107%       | SPK: 50 |
| 2037-26-5                 | Toluene-d8                  | 47.3   |           | 86 - 113 | 95%        | SPK: 50 |
| 460-00-4                  | 4-Bromofluorobenzene        | 44.1   |           | 77 - 121 | 88%        | SPK: 50 |
| <b>INTERNAL STANDARDS</b> |                             |        |           |          |            |         |
| 363-72-4                  | Pentafluorobenzene          | 480000 | 4.597     |          |            |         |
| 540-36-3                  | 1,4-Difluorobenzene         | 865000 | 5.532     |          |            |         |
| 3114-55-4                 | Chlorobenzene-d5            | 871000 | 8.776     |          |            |         |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 414000 | 11.175    |          |            |         |

## Report of Analysis

|                    |                 |        |      |                 |              |    |  |
|--------------------|-----------------|--------|------|-----------------|--------------|----|--|
| Client:            | G Environmental |        |      | Date Collected: |              |    |  |
| Project:           | Summer          |        |      | Date Received:  |              |    |  |
| Client Sample ID:  | VV0925WBL01     |        |      | SDG No.:        | Q3175        |    |  |
| Lab Sample ID:     | VV0925WBL01     |        |      | Matrix:         | Water        |    |  |
| Analytical Method: | 8260D           |        |      | % Solid:        | 0            |    |  |
| Sample Wt/Vol:     | 5               | Units: | mL   | Final Vol:      | 5000         | uL |  |
| Soil Aliquot Vol:  | uL              |        |      | Test:           | VOCMS Group1 |    |  |
| GC Column:         | DB-624UI        | ID :   | 0.18 | Level :         | LOW          |    |  |
| Prep Method :      |                 |        |      |                 |              |    |  |

| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
|-------------------|-----------|----------------|---------------|
| VV039170.D        | 1         | 09/25/25 11:01 | VV092525      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

( ) = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
 Data File : VV039170.D  
 Acq On : 25 Sep 2025 11:01  
 Operator : SY/MD  
 Sample : VV0925WBL01  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VV0925WBL01

Quant Time: Sep 26 01:20:45 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units       | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------------|----------|
| -----                       |        |       |          |          |             |          |
| Internal Standards          |        |       |          |          |             |          |
| 1) Pentafluorobenzene       | 4.597  | 168   | 480047   | 50.000   | ug/l        | 0.00     |
| 34) 1,4-Difluorobenzene     | 5.532  | 114   | 864912   | 50.000   | ug/l        | 0.00     |
| 63) Chlorobenzene-d5        | 8.776  | 117   | 870930   | 50.000   | ug/l        | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 11.175 | 152   | 413592   | 50.000   | ug/l        | 0.00     |
| System Monitoring Compounds |        |       |          |          |             |          |
| 33) 1,2-Dichloroethane-d4   | 4.941  | 65    | 417535   | 60.097   | ug/l        | 0.00     |
| Spiked Amount               | 50.000 | Range | 81 - 118 | Recovery | = 120.200%# |          |
| 35) Dibromofluoromethane    | 4.500  | 113   | 370419   | 53.273   | ug/l        | 0.00     |
| Spiked Amount               | 50.000 | Range | 80 - 119 | Recovery | = 106.540%  |          |
| 50) Toluene-d8              | 7.236  | 98    | 965988   | 47.302   | ug/l        | 0.00     |
| Spiked Amount               | 50.000 | Range | 89 - 112 | Recovery | = 94.600%   |          |
| 62) 4-Bromofluorobenzene    | 10.001 | 95    | 371030   | 44.133   | ug/l        | 0.00     |
| Spiked Amount               | 50.000 | Range | 85 - 114 | Recovery | = 88.260%   |          |


| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
Data File : VV039170.D  
Acq On : 25 Sep 2025 11:01  
Operator : SY/MD  
Sample : VV0925WBL01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VV0925WBL01

Quant Time: Sep 26 01:20:45 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration









#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 4.597 min Scan# 1106  
Delta R.T. -0.000 min  
Lab File: VV039170.D  
Acq: 25 Sep 2025 11:01




Instrument :  
MSVOA\_V  
ClientSampleId :  
VV0925WBL01


Tgt Ion:168 Resp: 480047  
Ion Ratio Lower Upper  
168 100  
99 63.9 49.6 74.4



#33  
1,2-Dichloroethane-d4  
Concen: 60.097 ug/l  
RT: 4.941 min Scan# 1213  
Delta R.T. 0.003 min  
Lab File: VV039170.D  
Acq: 25 Sep 2025 11:01

Tgt Ion: 65 Resp: 417535  
Ion Ratio Lower Upper  
65 100  
67 54.2 0.0 107.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 5.532 min Scan# 11

Delta R.T. 0.003 min

Lab File: VV039170.D

Acq: 25 Sep 2025 11:01

Instrument :

MSVOA\_V

ClientSampleId :

VV0925WBL01





Tgt Ion:114 Resp: 864912

Ion Ratio Lower Upper

114 100

63 20.3 0.0 42.6

88 15.1 0.0 30.8



#35

Dibromofluoromethane

Concen: 53.273 ug/l

RT: 4.500 min Scan# 1076

Delta R.T. 0.003 min

Lab File: VV039170.D

Acq: 25 Sep 2025 11:01




Tgt Ion:113 Resp: 370419


Ion Ratio Lower Upper

113 100

111 104.0 82.4 123.6

192 17.8 13.8 20.8









#50  
Toluene-d8  
Concen: 47.302 ug/l  
RT: 7.236 min Scan# 1926  
Delta R.T. 0.003 min  
Lab File: VV039170.D  
Acq: 25 Sep 2025 11:01




Instrument :  
MSVOA\_V  
ClientSampleId :  
VV0925WBL01


Tgt Ion: 98 Resp: 965988  
Ion Ratio Lower Upper  
98 100  
100 64.0 52.2 78.2



#62  
4-Bromofluorobenzene  
Concen: 44.133 ug/l  
RT: 10.001 min Scan# 2787  
Delta R.T. 0.003 min  
Lab File: VV039170.D  
Acq: 25 Sep 2025 11:01

Tgt Ion: 95 Resp: 371030  
Ion Ratio Lower Upper  
95 100  
174 77.2 0.0 150.4  
176 71.4 0.0 144.2





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 8.776 min Scan# 2405

Delta R.T. 0.003 min

Lab File: VV039170.D

Acq: 25 Sep 2025 11:01

Instrument :

MSVOA\_V

ClientSampleId :

VV0925WBL01





Tgt Ion:117 Resp: 870930

Ion Ratio Lower Upper

117 100

82 56.4 45.4 68.2

119 32.3 25.6 38.4



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 11.175 min Scan# 3152

Delta R.T. 0.003 min

Lab File: VV039170.D

Acq: 25 Sep 2025 11:01




Tgt Ion:152 Resp: 413592

Ion Ratio Lower Upper

152 100

115 62.1 44.8 134.4

150 158.0 0.0 353.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
Data File : VV039170.D  
Acq On : 25 Sep 2025 11:01  
Operator : SY/MD  
Sample : VV0925WBL01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VV0925WBL01

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M

Title : SW846 8260

Signal : TIC: VV039170.D\data.ms

| peak<br># | R.T.<br>min | first<br>scan | max<br>scan | last<br>scan | PK<br>TY | peak<br>height | corr.<br>area | corr.<br>% max. | % of<br>total |
|-----------|-------------|---------------|-------------|--------------|----------|----------------|---------------|-----------------|---------------|
| 1         | 4.500       | 1060          | 1076        | 1095         | rBV      | 443480         | 1175595       | 44.03%          | 7.554%        |
| 2         | 4.597       | 1095          | 1106        | 1148         | rVB      | 568175         | 1566147       | 58.66%          | 10.064%       |
| 3         | 4.941       | 1200          | 1213        | 1243         | rBV      | 475406         | 1163011       | 43.56%          | 7.473%        |
| 4         | 5.532       | 1383          | 1397        | 1436         | rBV      | 887588         | 2082143       | 77.99%          | 13.379%       |
| 5         | 7.236       | 1916          | 1927        | 1970         | rBV      | 1332205        | 2555376       | 95.72%          | 16.420%       |
| 6         | 8.773       | 2395          | 2405        | 2437         | rBV      | 1535667        | 2669706       | 100.00%         | 17.155%       |
| 7         | 10.001      | 2776          | 2787        | 2816         | rBV      | 1114083        | 1822624       | 68.27%          | 11.712%       |
| 8         | 11.172      | 3140          | 3151        | 3183         | rBV      | 1569081        | 2527734       | 94.68%          | 16.243%       |


Sum of corrected areas: 15562336

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
Data File : VV039170.D  
Acq On : 25 Sep 2025 11:01  
Operator : SY/MD  
Sample : VV0925WBL01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VV0925WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
Data File : VV039170.D  
Acq On : 25 Sep 2025 11:01  
Operator : SY/MD  
Sample : VV0925WBL01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VV0925WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

\*\*\*\*\*

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
Data File : VV039170.D  
Acq On : 25 Sep 2025 11:01  
Operator : SY/MD  
Sample : VV0925WBL01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VV0925WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L  
TIC Integration Parameters: LSCINT.P

| TIC Top Hit name | RT | EstConc | Units | Response | # | RT | Resp | Conc |
|------------------|----|---------|-------|----------|---|----|------|------|
|------------------|----|---------|-------|----------|---|----|------|------|

### Report of Analysis

|                    |                 |           |                 |              |
|--------------------|-----------------|-----------|-----------------|--------------|
| Client:            | G Environmental |           | Date Collected: |              |
| Project:           | Summer          |           | Date Received:  |              |
| Client Sample ID:  | VV0924WBS02     |           | SDG No.:        | Q3175        |
| Lab Sample ID:     | VV0924WBS02     |           | Matrix:         | Water        |
| Analytical Method: | 8260D           |           | % Solid:        | 0            |
| Sample Wt/Vol:     | 5               | Units: mL | Final Vol:      | 5000 uL      |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOCMS Group1 |
| GC Column:         | DB-624UI        | ID : 0.18 | Level :         | LOW          |
| Prep Method :      |                 |           |                 |              |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039150.D        | 1         | 09/24/25 14:01 | VV092425      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL   | LOQ / CRQL | Units |
|----------------|--------------------------------|-------|-----------|-------|------------|-------|
| <b>TARGETS</b> |                                |       |           |       |            |       |
| 75-71-8        | Dichlorodifluoromethane        | 19.1  |           | 0.22  | 1.00       | ug/L  |
| 74-87-3        | Chloromethane                  | 19.1  |           | 0.32  | 1.00       | ug/L  |
| 75-01-4        | Vinyl Chloride                 | 19.4  |           | 0.26  | 1.00       | ug/L  |
| 74-83-9        | Bromomethane                   | 20.9  |           | 1.40  | 5.00       | ug/L  |
| 75-00-3        | Chloroethane                   | 21.1  |           | 0.47  | 1.00       | ug/L  |
| 75-69-4        | Trichlorofluoromethane         | 20.0  |           | 0.33  | 1.00       | ug/L  |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 19.7  |           | 0.25  | 1.00       | ug/L  |
| 75-65-0        | Tert butyl alcohol             | 110   |           | 5.50  | 25.0       | ug/L  |
| 75-35-4        | 1,1-Dichloroethene             | 20.0  |           | 0.23  | 1.00       | ug/L  |
| 67-64-1        | Acetone                        | 110   |           | 1.50  | 5.00       | ug/L  |
| 75-15-0        | Carbon Disulfide               | 19.6  |           | 0.21  | 1.00       | ug/L  |
| 1634-04-4      | Methyl tert-butyl Ether        | 20.9  |           | 0.16  | 1.00       | ug/L  |
| 79-20-9        | Methyl Acetate                 | 22.4  |           | 0.27  | 1.00       | ug/L  |
| 75-09-2        | Methylene Chloride             | 22.7  |           | 0.28  | 1.00       | ug/L  |
| 156-60-5       | trans-1,2-Dichloroethene       | 19.7  |           | 0.23  | 1.00       | ug/L  |
| 75-34-3        | 1,1-Dichloroethane             | 19.6  |           | 0.23  | 1.00       | ug/L  |
| 110-82-7       | Cyclohexane                    | 18.7  |           | 1.50  | 5.00       | ug/L  |
| 78-93-3        | 2-Butanone                     | 110   |           | 0.98  | 5.00       | ug/L  |
| 56-23-5        | Carbon Tetrachloride           | 19.7  |           | 0.25  | 1.00       | ug/L  |
| 156-59-2       | cis-1,2-Dichloroethene         | 19.5  |           | 0.19  | 1.00       | ug/L  |
| 74-97-5        | Bromochloromethane             | 20.3  |           | 0.22  | 1.00       | ug/L  |
| 67-66-3        | Chloroform                     | 19.9  |           | 0.25  | 1.00       | ug/L  |
| 71-55-6        | 1,1,1-Trichloroethane          | 20.0  |           | 0.20  | 1.00       | ug/L  |
| 108-87-2       | Methylcyclohexane              | 19.6  |           | 0.16  | 1.00       | ug/L  |
| 71-43-2        | Benzene                        | 20.1  |           | 0.15  | 1.00       | ug/L  |
| 107-06-2       | 1,2-Dichloroethane             | 20.7  |           | 0.22  | 1.00       | ug/L  |
| 79-01-6        | Trichloroethene                | 20.3  |           | 0.090 | 1.00       | ug/L  |
| 78-87-5        | 1,2-Dichloropropane            | 20.9  |           | 0.20  | 1.00       | ug/L  |
| 75-27-4        | Bromodichloromethane           | 20.6  |           | 0.22  | 1.00       | ug/L  |
| 108-10-1       | 4-Methyl-2-Pentanone           | 120   |           | 0.68  | 5.00       | ug/L  |

## Report of Analysis

|                    |                 |        |      |                 |              |
|--------------------|-----------------|--------|------|-----------------|--------------|
| Client:            | G Environmental |        |      | Date Collected: |              |
| Project:           | Summer          |        |      | Date Received:  |              |
| Client Sample ID:  | VV0924WBS02     |        |      | SDG No.:        | Q3175        |
| Lab Sample ID:     | VV0924WBS02     |        |      | Matrix:         | Water        |
| Analytical Method: | 8260D           |        |      | % Solid:        | 0            |
| Sample Wt/Vol:     | 5               | Units: | mL   | Final Vol:      | 5000 uL      |
| Soil Aliquot Vol:  |                 |        | uL   | Test:           | VOCMS Group1 |
| GC Column:         | DB-624UI        | ID :   | 0.18 | Level :         | LOW          |
| Prep Method :      |                 |        |      |                 |              |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039150.D        | 1         | 09/24/25 14:01 | VV092425      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units   |
|---------------------------|-----------------------------|--------|-----------|----------|------------|---------|
| 108-88-3                  | Toluene                     | 21.8   |           | 0.14     | 1.00       | ug/L    |
| 10061-02-6                | t-1,3-Dichloropropene       | 21.6   |           | 0.17     | 1.00       | ug/L    |
| 10061-01-5                | cis-1,3-Dichloropropene     | 21.5   |           | 0.16     | 1.00       | ug/L    |
| 79-00-5                   | 1,1,2-Trichloroethane       | 21.0   |           | 0.21     | 1.00       | ug/L    |
| 591-78-6                  | 2-Hexanone                  | 110    |           | 0.89     | 5.00       | ug/L    |
| 124-48-1                  | Dibromochloromethane        | 20.5   |           | 0.18     | 1.00       | ug/L    |
| 106-93-4                  | 1,2-Dibromoethane           | 21.5   |           | 0.15     | 1.00       | ug/L    |
| 127-18-4                  | Tetrachloroethene           | 20.2   |           | 0.23     | 1.00       | ug/L    |
| 108-90-7                  | Chlorobenzene               | 19.8   |           | 0.12     | 1.00       | ug/L    |
| 100-41-4                  | Ethyl Benzene               | 19.9   |           | 0.13     | 1.00       | ug/L    |
| 179601-23-1               | m/p-Xylenes                 | 43.2   |           | 0.24     | 2.00       | ug/L    |
| 95-47-6                   | o-Xylene                    | 20.4   |           | 0.12     | 1.00       | ug/L    |
| 100-42-5                  | Styrene                     | 18.9   |           | 0.15     | 1.00       | ug/L    |
| 75-25-2                   | Bromoform                   | 21.4   |           | 0.19     | 1.00       | ug/L    |
| 98-82-8                   | Isopropylbenzene            | 19.8   |           | 0.12     | 1.00       | ug/L    |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 19.2   |           | 0.26     | 1.00       | ug/L    |
| 541-73-1                  | 1,3-Dichlorobenzene         | 19.6   |           | 0.16     | 1.00       | ug/L    |
| 106-46-7                  | 1,4-Dichlorobenzene         | 19.2   |           | 0.19     | 1.00       | ug/L    |
| 95-50-1                   | 1,2-Dichlorobenzene         | 19.0   |           | 0.16     | 1.00       | ug/L    |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 21.3   |           | 0.53     | 1.00       | ug/L    |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 18.7   |           | 0.20     | 1.00       | ug/L    |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 19.1   |           | 0.20     | 1.00       | ug/L    |
| <b>SURROGATES</b>         |                             |        |           |          |            |         |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 49.1   |           | 74 - 125 | 98%        | SPK: 50 |
| 1868-53-7                 | Dibromofluoromethane        | 50.0   |           | 75 - 124 | 100%       | SPK: 50 |
| 2037-26-5                 | Toluene-d8                  | 49.5   |           | 86 - 113 | 99%        | SPK: 50 |
| 460-00-4                  | 4-Bromofluorobenzene        | 52.0   |           | 77 - 121 | 104%       | SPK: 50 |
| <b>INTERNAL STANDARDS</b> |                             |        |           |          |            |         |
| 363-72-4                  | Pentafluorobenzene          | 492000 | 4.6       |          |            |         |
| 540-36-3                  | 1,4-Difluorobenzene         | 796000 | 5.532     |          |            |         |
| 3114-55-4                 | Chlorobenzene-d5            | 785000 | 8.776     |          |            |         |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 469000 | 11.172    |          |            |         |

## Report of Analysis

|                    |                                            |                 |                              |
|--------------------|--------------------------------------------|-----------------|------------------------------|
| Client:            | G Environmental                            | Date Collected: |                              |
| Project:           | Summer                                     | Date Received:  |                              |
| Client Sample ID:  | VV0924WBS02                                | SDG No.:        | Q3175                        |
| Lab Sample ID:     | VV0924WBS02                                | Matrix:         | Water                        |
| Analytical Method: | 8260D                                      | % Solid:        | 0                            |
| Sample Wt/Vol:     | 5                      Units:    mL        | Final Vol:      | 5000                      uL |
| Soil Aliquot Vol:  | uL                                         | Test:           | VOCMS Group1                 |
| GC Column:         | DB-624UI                      ID :    0.18 | Level :         | LOW                          |
| Prep Method :      |                                            |                 |                              |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039150.D        | 1         | 09/24/25 14:01 | VV092425      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
 Data File : VV039150.D  
 Acq On : 24 Sep 2025 14:01  
 Operator : SY/MD  
 Sample : VV0924WBS02  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VV0924WBS02

Quant Time: Sep 25 04:45:44 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

| Compound                     | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|------------------------------|--------|-------|----------|----------|--------|----------|
| -----                        |        |       |          |          |        |          |
| Internal Standards           |        |       |          |          |        |          |
| 1) Pentafluorobenzene        | 4.600  | 168   | 492318   | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 5.532  | 114   | 795602   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 8.776  | 117   | 784865   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 11.172 | 152   | 468522   | 50.000   | ug/l   | 0.00     |
|                              |        |       |          |          |        |          |
| System Monitoring Compounds  |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4    | 4.941  | 65    | 349518   | 49.053   | ug/l   | 0.00     |
| Spiked Amount                | 50.000 | Range | 81 - 118 | Recovery | =      | 98.100%  |
| 35) Dibromofluoromethane     | 4.500  | 113   | 319944   | 50.023   | ug/l   | 0.00     |
| Spiked Amount                | 50.000 | Range | 80 - 119 | Recovery | =      | 100.040% |
| 50) Toluene-d8               | 7.236  | 98    | 929285   | 49.469   | ug/l   | 0.00     |
| Spiked Amount                | 50.000 | Range | 89 - 112 | Recovery | =      | 98.940%  |
| 62) 4-Bromofluorobenzene     | 10.002 | 95    | 402187   | 52.007   | ug/l   | 0.00     |
| Spiked Amount                | 50.000 | Range | 85 - 114 | Recovery | =      | 104.020% |
|                              |        |       |          |          |        |          |
| Target Compounds             |        |       |          |          |        | Qvalue   |
| 2) Dichlorodifluoromethane   | 1.121  | 85    | 148756   | 19.087   | ug/l   | 99       |
| 3) Chloromethane             | 1.230  | 50    | 159926   | 19.115   | ug/l   | 98       |
| 4) Vinyl Chloride            | 1.298  | 62    | 173519   | 19.373   | ug/l   | 100      |
| 5) Bromomethane              | 1.497  | 94    | 117993   | 20.897   | ug/l   | 100      |
| 6) Chloroethane              | 1.561  | 64    | 117810   | 21.073   | ug/l   | 99       |
| 7) Trichlorofluoromethane    | 1.725  | 101   | 259123   | 19.982   | ug/l   | 99       |
| 8) Diethyl Ether             | 1.918  | 74    | 101938   | 21.175   | ug/l   | 99       |
| 9) 1,1,2-Trichlorotrifluo... | 2.082  | 101   | 151673   | 19.657   | ug/l   | 99       |
| 10) Methyl Iodide            | 2.198  | 142   | 155103   | 21.109   | ug/l   | 98       |
| 11) Tert butyl alcohol       | 2.568  | 59    | 166515   | 110.571  | ug/l   | 99       |
| 12) 1,1-Dichloroethene       | 2.082  | 96    | 148258   | 19.973   | ug/l   | 96       |
| 13) Acrolein                 | 2.011  | 56    | 30535    | 116.851  | ug/l   | 97       |
| 14) Allyl chloride           | 2.359  | 41    | 270301   | 20.386   | ug/l   | 98       |
| 15) Acrylonitrile            | 2.674  | 53    | 510399   | 105.726  | ug/l   | 100      |
| 16) Acetone                  | 2.118  | 43    | 500511   | 107.135  | ug/l   | 98       |
| 17) Carbon Disulfide         | 2.253  | 76    | 445608   | 19.629   | ug/l   | 99       |
| 18) Methyl Acetate           | 2.381  | 43    | 236287   | 22.381   | ug/l   | 97       |
| 19) Methyl tert-butyl Ether  | 2.712  | 73    | 494520   | 20.907   | ug/l   | 96       |
| 20) Methylene Chloride       | 2.458  | 84    | 185787   | 22.678   | ug/l   | 97       |
| 21) trans-1,2-Dichloroethene | 2.703  | 96    | 155320   | 19.655   | ug/l   | 98       |
| 22) Diisopropyl ether        | 3.217  | 45    | 514292   | 20.273   | ug/l   | 85       |
| 23) Vinyl Acetate            | 3.192  | 43    | 2297382  | 103.522  | ug/l   | 100      |
| 24) 1,1-Dichloroethane       | 3.118  | 63    | 304665   | 19.638   | ug/l   | 99       |
| 25) 2-Butanone               | 3.870  | 43    | 603602   | 105.789  | ug/l   | 98       |
| 26) 2,2-Dichloropropane      | 3.812  | 77    | 246178   | 17.897   | ug/l   | 100      |
| 27) cis-1,2-Dichloroethene   | 3.822  | 96    | 170071   | 19.514   | ug/l   | 100      |
| 28) Bromochloromethane       | 4.146  | 49    | 138833   | 20.271   | ug/l   | 96       |
| 29) Tetrahydrofuran          | 4.240  | 42    | 389964   | 109.192  | ug/l   | 100      |
| 30) Chloroform               | 4.278  | 83    | 322270   | 19.894   | ug/l   | 99       |
| 31) Cyclohexane              | 4.584  | 56    | 238107   | 18.736   | ug/l   | 99       |
| 32) 1,1,1-Trichloroethane    | 4.513  | 97    | 279806   | 19.951   | ug/l   | 97       |
| 36) 1,1-Dichloropropene      | 4.745  | 75    | 187909   | 19.432   | ug/l   | 98       |
| 37) Ethyl Acetate            | 3.982  | 43    | 243790   | 20.776   | ug/l # | 92       |
| 38) Carbon Tetrachloride     | 4.735  | 117   | 232933   | 19.709   | ug/l   | 100      |
| 39) Methylcyclohexane        | 6.047  | 83    | 228816   | 19.594   | ug/l   | 96       |
| 40) Benzene                  | 5.011  | 78    | 633991   | 20.134   | ug/l   | 97       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
 Data File : VV039150.D  
 Acq On : 24 Sep 2025 14:01  
 Operator : SY/MD  
 Sample : VV0924WBS02  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VV0924WBS02

Quant Time: Sep 25 04:45:44 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 4.166  | 41   | 96391    | 19.385  | ug/l   | 93       |
| 42) 1,2-Dichloroethane        | 5.040  | 62   | 228142   | 20.677  | ug/l   | 99       |
| 43) Isopropyl Acetate         | 5.195  | 43   | 337530   | 21.097  | ug/l   | 99       |
| 44) Trichloroethene           | 5.831  | 130  | 147483   | 20.297  | ug/l   | 99       |
| 45) 1,2-Dichloropropane       | 6.092  | 63   | 163901   | 20.941  | ug/l   | 96       |
| 46) Dibromomethane            | 6.227  | 93   | 126157   | 21.059  | ug/l   | 99       |
| 47) Bromodichloromethane      | 6.426  | 83   | 254692   | 20.565  | ug/l   | 99       |
| 48) Methyl methacrylate       | 6.301  | 41   | 145228   | 20.025  | ug/l   | 98       |
| 49) 1,4-Dioxane               | 6.294  | 88   | 46282    | 395.445 | ug/l   | 91       |
| 51) 4-Methyl-2-Pentanone      | 7.150  | 43   | 1236058  | 117.436 | ug/l   | 98       |
| 52) Toluene                   | 7.310  | 92   | 398034   | 21.801  | ug/l   | 99       |
| 53) t-1,3-Dichloropropene     | 7.577  | 75   | 233157   | 21.603  | ug/l   | 96       |
| 54) cis-1,3-Dichloropropene   | 6.950  | 75   | 256176   | 21.501  | ug/l   | 96       |
| 55) 1,1,2-Trichloroethane     | 7.764  | 97   | 171861   | 20.993  | ug/l   | 97       |
| 56) Ethyl methacrylate        | 7.719  | 69   | 198833   | 20.452  | ug/l   | 100      |
| 57) 1,3-Dichloropropane       | 7.934  | 76   | 277597   | 21.486  | ug/l   | 97       |
| 58) 2-Chloroethyl Vinyl ether | 6.815  | 63   | 517964   | 100.347 | ug/l   | 99       |
| 59) 2-Hexanone                | 8.066  | 43   | 894828   | 106.669 | ug/l   | 98       |
| 60) Dibromochloromethane      | 8.166  | 129  | 191297   | 20.508  | ug/l   | 99       |
| 61) 1,2-Dibromoethane         | 8.275  | 107  | 179580   | 21.450  | ug/l   | 99       |
| 64) Tetrachloroethene         | 7.899  | 164  | 132876   | 20.213  | ug/l   | 98       |
| 65) Chlorobenzene             | 8.805  | 112  | 432358   | 19.777  | ug/l   | 97       |
| 66) 1,1,1,2-Tetrachloroethane | 8.899  | 131  | 156609   | 19.826  | ug/l   | 100      |
| 67) Ethyl Benzene             | 8.937  | 91   | 658467   | 19.898  | ug/l   | 100      |
| 68) m/p-Xylenes               | 9.063  | 106  | 551462   | 43.176  | ug/l   | 99       |
| 69) o-Xylene                  | 9.468  | 106  | 230630   | 20.393  | ug/l   | 98       |
| 70) Styrene                   | 9.487  | 104  | 431651   | 18.940  | ug/l   | 98       |
| 71) Bromoform                 | 9.654  | 173  | 131945   | 21.384  | ug/l # | 98       |
| 73) Isopropylbenzene          | 9.857  | 105  | 671517   | 19.761  | ug/l   | 100      |
| 74) N-amyl acetate            | 9.725  | 43   | 283249   | 20.897  | ug/l   | 99       |
| 75) 1,1,2,2-Tetrachloroethane | 10.166 | 83   | 286264   | 19.226  | ug/l   | 100      |
| 76) 1,2,3-Trichloropropane    | 10.198 | 75   | 206903   | 19.396  | ug/l   | 100      |
| 77) Bromobenzene              | 10.140 | 156  | 173461   | 19.674  | ug/l   | 99       |
| 78) n-propylbenzene           | 10.278 | 91   | 836225   | 20.207  | ug/l   | 100      |
| 79) 2-Chlorotoluene           | 10.349 | 91   | 511892   | 20.673  | ug/l   | 98       |
| 80) 1,3,5-Trimethylbenzene    | 10.465 | 105  | 583511   | 20.659  | ug/l   | 99       |
| 81) trans-1,4-Dichloro-2-b... | 9.931  | 75   | 86872    | 19.550  | ug/l   | 99       |
| 82) 4-Chlorotoluene           | 10.465 | 91   | 594451   | 20.703  | ug/l   | 100      |
| 83) tert-Butylbenzene         | 10.789 | 119  | 525337   | 18.652  | ug/l   | 100      |
| 84) 1,2,4-Trimethylbenzene    | 10.841 | 105  | 564955   | 20.602  | ug/l   | 100      |
| 85) sec-Butylbenzene          | 11.014 | 105  | 728021   | 19.488  | ug/l   | 100      |
| 86) p-Isopropyltoluene        | 11.169 | 119  | 612071   | 19.685  | ug/l   | 100      |
| 87) 1,3-Dichlorobenzene       | 11.108 | 146  | 349017   | 19.560  | ug/l   | 100      |
| 88) 1,4-Dichlorobenzene       | 11.198 | 146  | 375846   | 19.199  | ug/l   | 99       |
| 89) n-Butylbenzene            | 11.580 | 91   | 577604   | 19.451  | ug/l   | 98       |
| 90) Hexachloroethane          | 11.821 | 117  | 135944   | 17.755  | ug/l   | 99       |
| 91) 1,2-Dichlorobenzene       | 11.567 | 146  | 316105   | 18.962  | ug/l   | 100      |
| 92) 1,2-Dibromo-3-Chloropr... | 12.352 | 75   | 56358    | 21.271  | ug/l   | 98       |
| 93) 1,2,4-Trichlorobenzene    | 13.185 | 180  | 178967   | 18.728  | ug/l   | 99       |
| 94) Hexachlorobutadiene       | 13.368 | 225  | 89231    | 17.970  | ug/l   | 99       |
| 95) Naphthalene               | 13.426 | 128  | 578144   | 18.718  | ug/l   | 100      |
| 96) 1,2,3-Trichlorobenzene    | 13.667 | 180  | 187044   | 19.137  | ug/l   | 100      |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039150.D  
Acq On : 24 Sep 2025 14:01  
Operator : SY/MD  
Sample : VV0924WBS02  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 6 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VV0924WBS02

Quant Time: Sep 25 04:45:44 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration


| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092425\  
Data File : VV039150.D  
Acq On    : 24 Sep 2025  14:01  
Operator  : SY/MD  
Sample    : VV0924WBS02  
Misc      : 5.0mL/MSVOA_V/WATER  
ALS Vial  : 6      Sample Multiplier: 1
```

**Instrument :**  
MSVOA\_V  
**ClientSampleId :**  
VV0924WBS02

Quant Time: Sep 25 04:45:44 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration



### Report of Analysis

|                    |                 |           |                 |              |
|--------------------|-----------------|-----------|-----------------|--------------|
| Client:            | G Environmental |           | Date Collected: |              |
| Project:           | Summer          |           | Date Received:  |              |
| Client Sample ID:  | VV0925WBS01     |           | SDG No.:        | Q3175        |
| Lab Sample ID:     | VV0925WBS01     |           | Matrix:         | Water        |
| Analytical Method: | 8260D           |           | % Solid:        | 0            |
| Sample Wt/Vol:     | 5               | Units: mL | Final Vol:      | 5000 uL      |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOCMS Group1 |
| GC Column:         | DB-624UI        | ID : 0.18 | Level :         | LOW          |
| Prep Method :      |                 |           |                 |              |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039171.D        | 1         | 09/25/25 11:24 | VV092525      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL   | LOQ / CRQL | Units |
|----------------|--------------------------------|-------|-----------|-------|------------|-------|
| <b>TARGETS</b> |                                |       |           |       |            |       |
| 75-71-8        | Dichlorodifluoromethane        | 18.2  |           | 0.22  | 1.00       | ug/L  |
| 74-87-3        | Chloromethane                  | 19.1  |           | 0.32  | 1.00       | ug/L  |
| 75-01-4        | Vinyl Chloride                 | 19.4  |           | 0.26  | 1.00       | ug/L  |
| 74-83-9        | Bromomethane                   | 18.7  |           | 1.40  | 5.00       | ug/L  |
| 75-00-3        | Chloroethane                   | 19.3  |           | 0.47  | 1.00       | ug/L  |
| 75-69-4        | Trichlorofluoromethane         | 19.4  |           | 0.33  | 1.00       | ug/L  |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 20.0  |           | 0.25  | 1.00       | ug/L  |
| 75-65-0        | Tert butyl alcohol             | 110   |           | 5.50  | 25.0       | ug/L  |
| 75-35-4        | 1,1-Dichloroethene             | 19.9  |           | 0.23  | 1.00       | ug/L  |
| 67-64-1        | Acetone                        | 89.0  |           | 1.50  | 5.00       | ug/L  |
| 75-15-0        | Carbon Disulfide               | 19.0  |           | 0.21  | 1.00       | ug/L  |
| 1634-04-4      | Methyl tert-butyl Ether        | 20.0  |           | 0.16  | 1.00       | ug/L  |
| 79-20-9        | Methyl Acetate                 | 21.2  |           | 0.27  | 1.00       | ug/L  |
| 75-09-2        | Methylene Chloride             | 20.6  |           | 0.28  | 1.00       | ug/L  |
| 156-60-5       | trans-1,2-Dichloroethene       | 19.7  |           | 0.23  | 1.00       | ug/L  |
| 75-34-3        | 1,1-Dichloroethane             | 19.3  |           | 0.23  | 1.00       | ug/L  |
| 110-82-7       | Cyclohexane                    | 18.2  |           | 1.50  | 5.00       | ug/L  |
| 78-93-3        | 2-Butanone                     | 97.1  |           | 0.98  | 5.00       | ug/L  |
| 56-23-5        | Carbon Tetrachloride           | 19.4  |           | 0.25  | 1.00       | ug/L  |
| 156-59-2       | cis-1,2-Dichloroethene         | 19.5  |           | 0.19  | 1.00       | ug/L  |
| 74-97-5        | Bromochloromethane             | 20.5  |           | 0.22  | 1.00       | ug/L  |
| 67-66-3        | Chloroform                     | 19.2  |           | 0.25  | 1.00       | ug/L  |
| 71-55-6        | 1,1,1-Trichloroethane          | 19.0  |           | 0.20  | 1.00       | ug/L  |
| 108-87-2       | Methylcyclohexane              | 19.2  |           | 0.16  | 1.00       | ug/L  |
| 71-43-2        | Benzene                        | 19.9  |           | 0.15  | 1.00       | ug/L  |
| 107-06-2       | 1,2-Dichloroethane             | 19.6  |           | 0.22  | 1.00       | ug/L  |
| 79-01-6        | Trichloroethene                | 20.4  |           | 0.090 | 1.00       | ug/L  |
| 78-87-5        | 1,2-Dichloropropane            | 20.5  |           | 0.20  | 1.00       | ug/L  |
| 75-27-4        | Bromodichloromethane           | 19.3  |           | 0.22  | 1.00       | ug/L  |
| 108-10-1       | 4-Methyl-2-Pentanone           | 110   |           | 0.68  | 5.00       | ug/L  |

### Report of Analysis

|                    |                 |           |                 |              |
|--------------------|-----------------|-----------|-----------------|--------------|
| Client:            | G Environmental |           | Date Collected: |              |
| Project:           | Summer          |           | Date Received:  |              |
| Client Sample ID:  | VV0925WBS01     |           | SDG No.:        | Q3175        |
| Lab Sample ID:     | VV0925WBS01     |           | Matrix:         | Water        |
| Analytical Method: | 8260D           |           | % Solid:        | 0            |
| Sample Wt/Vol:     | 5               | Units: mL | Final Vol:      | 5000 uL      |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOCMS Group1 |
| GC Column:         | DB-624UI        | ID : 0.18 | Level :         | LOW          |
| Prep Method :      |                 |           |                 |              |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039171.D        | 1         | 09/25/25 11:24 | VV092525      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units   |
|---------------------------|-----------------------------|--------|-----------|----------|------------|---------|
| 108-88-3                  | Toluene                     | 21.5   |           | 0.14     | 1.00       | ug/L    |
| 10061-02-6                | t-1,3-Dichloropropene       | 20.1   |           | 0.17     | 1.00       | ug/L    |
| 10061-01-5                | cis-1,3-Dichloropropene     | 20.2   |           | 0.16     | 1.00       | ug/L    |
| 79-00-5                   | 1,1,2-Trichloroethane       | 20.0   |           | 0.21     | 1.00       | ug/L    |
| 591-78-6                  | 2-Hexanone                  | 99.6   |           | 0.89     | 5.00       | ug/L    |
| 124-48-1                  | Dibromochloromethane        | 20.2   |           | 0.18     | 1.00       | ug/L    |
| 106-93-4                  | 1,2-Dibromoethane           | 20.6   |           | 0.15     | 1.00       | ug/L    |
| 127-18-4                  | Tetrachloroethene           | 20.2   |           | 0.23     | 1.00       | ug/L    |
| 108-90-7                  | Chlorobenzene               | 20.1   |           | 0.12     | 1.00       | ug/L    |
| 100-41-4                  | Ethyl Benzene               | 20.0   |           | 0.13     | 1.00       | ug/L    |
| 179601-23-1               | m/p-Xylenes                 | 43.1   |           | 0.24     | 2.00       | ug/L    |
| 95-47-6                   | o-Xylene                    | 21.0   |           | 0.12     | 1.00       | ug/L    |
| 100-42-5                  | Styrene                     | 19.3   |           | 0.15     | 1.00       | ug/L    |
| 75-25-2                   | Bromoform                   | 20.5   |           | 0.19     | 1.00       | ug/L    |
| 98-82-8                   | Isopropylbenzene            | 20.6   |           | 0.12     | 1.00       | ug/L    |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 19.4   |           | 0.26     | 1.00       | ug/L    |
| 541-73-1                  | 1,3-Dichlorobenzene         | 19.6   |           | 0.16     | 1.00       | ug/L    |
| 106-46-7                  | 1,4-Dichlorobenzene         | 18.9   |           | 0.19     | 1.00       | ug/L    |
| 95-50-1                   | 1,2-Dichlorobenzene         | 19.8   |           | 0.16     | 1.00       | ug/L    |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 22.1   |           | 0.53     | 1.00       | ug/L    |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 19.2   |           | 0.20     | 1.00       | ug/L    |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 19.9   |           | 0.20     | 1.00       | ug/L    |
| <b>SURROGATES</b>         |                             |        |           |          |            |         |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 45.1   |           | 74 - 125 | 90%        | SPK: 50 |
| 1868-53-7                 | Dibromofluoromethane        | 47.8   |           | 75 - 124 | 96%        | SPK: 50 |
| 2037-26-5                 | Toluene-d8                  | 47.9   |           | 86 - 113 | 96%        | SPK: 50 |
| 460-00-4                  | 4-Bromofluorobenzene        | 51.2   |           | 77 - 121 | 102%       | SPK: 50 |
| <b>INTERNAL STANDARDS</b> |                             |        |           |          |            |         |
| 363-72-4                  | Pentafluorobenzene          | 557000 | 4.6       |          |            |         |
| 540-36-3                  | 1,4-Difluorobenzene         | 899000 | 5.535     |          |            |         |
| 3114-55-4                 | Chlorobenzene-d5            | 872000 | 8.776     |          |            |         |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 509000 | 11.172    |          |            |         |

## Report of Analysis

|                    |                 |        |      |                 |              |
|--------------------|-----------------|--------|------|-----------------|--------------|
| Client:            | G Environmental |        |      | Date Collected: |              |
| Project:           | Summer          |        |      | Date Received:  |              |
| Client Sample ID:  | VV0925WBS01     |        |      | SDG No.:        | Q3175        |
| Lab Sample ID:     | VV0925WBS01     |        |      | Matrix:         | Water        |
| Analytical Method: | 8260D           |        |      | % Solid:        | 0            |
| Sample Wt/Vol:     | 5               | Units: | mL   | Final Vol:      | 5000 uL      |
| Soil Aliquot Vol:  |                 |        | uL   | Test:           | VOCMS Group1 |
| GC Column:         | DB-624UI        | ID :   | 0.18 | Level :         | LOW          |
| Prep Method :      |                 |        |      |                 |              |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039171.D        | 1         | 09/25/25 11:24 | VV092525      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
 Data File : VV039171.D  
 Acq On : 25 Sep 2025 11:24  
 Operator : SY/MD  
 Sample : VV0925WBS01  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VV0925WBS01

Quant Time: Sep 26 01:21:22 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 4.600  | 168  | 557017   | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 5.535  | 114  | 898902   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 8.776  | 117  | 871912   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 11.172 | 152  | 508947   | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |          |
|-----------------------------|--------|-------|----------|----------|------|----------|
| System Monitoring Compounds |        |       |          |          |      |          |
| 33) 1,2-Dichloroethane-d4   | 4.944  | 65    | 363796   | 45.127   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 81 - 118 | Recovery | =    | 90.260%  |
| 35) Dibromofluoromethane    | 4.500  | 113   | 345732   | 47.843   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 80 - 119 | Recovery | =    | 95.680%  |
| 50) Toluene-d8              | 7.236  | 98    | 1017460  | 47.939   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 89 - 112 | Recovery | =    | 95.880%  |
| 62) 4-Bromofluorobenzene    | 10.001 | 95    | 447073   | 51.168   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 85 - 114 | Recovery | =    | 102.340% |

|                              |       |     |         |         |      |     |
|------------------------------|-------|-----|---------|---------|------|-----|
| Target Compounds             |       |     |         | Qvalue  |      |     |
| 2) Dichlorodifluoromethane   | 1.121 | 85  | 160404  | 18.191  | ug/l | 99  |
| 3) Chloromethane             | 1.230 | 50  | 180629  | 19.082  | ug/l | 99  |
| 4) Vinyl Chloride            | 1.298 | 62  | 196152  | 19.356  | ug/l | 99  |
| 5) Bromomethane              | 1.497 | 94  | 119487  | 18.704  | ug/l | 98  |
| 6) Chloroethane              | 1.561 | 64  | 122694  | 19.279  | ug/l | 97  |
| 7) Trichlorofluoromethane    | 1.725 | 101 | 284580  | 19.396  | ug/l | 98  |
| 8) Diethyl Ether             | 1.918 | 74  | 106026  | 19.466  | ug/l | 99  |
| 9) 1,1,2-Trichlorotrifluo... | 2.082 | 101 | 174658  | 20.007  | ug/l | 97  |
| 10) Methyl Iodide            | 2.198 | 142 | 135972  | 16.356  | ug/l | 99  |
| 11) Tert butyl alcohol       | 2.571 | 59  | 187721  | 110.174 | ug/l | 99  |
| 12) 1,1-Dichloroethene       | 2.082 | 96  | 166976  | 19.882  | ug/l | 98  |
| 13) Acrolein                 | 2.011 | 56  | 34025   | 115.083 | ug/l | 98  |
| 14) Allyl chloride           | 2.359 | 41  | 286242  | 19.080  | ug/l | 98  |
| 15) Acrylonitrile            | 2.677 | 53  | 538540  | 98.597  | ug/l | 100 |
| 16) Acetone                  | 2.121 | 43  | 476447  | 89.017  | ug/l | 97  |
| 17) Carbon Disulfide         | 2.253 | 76  | 487416  | 18.977  | ug/l | 99  |
| 18) Methyl Acetate           | 2.381 | 43  | 252836  | 21.167  | ug/l | 98  |
| 19) Methyl tert-butyl Ether  | 2.716 | 73  | 535106  | 19.995  | ug/l | 99  |
| 20) Methylene Chloride       | 2.458 | 84  | 192332  | 20.555  | ug/l | 95  |
| 21) trans-1,2-Dichloroethene | 2.703 | 96  | 176159  | 19.703  | ug/l | 96  |
| 22) Diisopropyl ether        | 3.220 | 45  | 549959  | 19.161  | ug/l | 99  |
| 23) Vinyl Acetate            | 3.191 | 43  | 2415006 | 96.182  | ug/l | 98  |
| 24) 1,1-Dichloroethane       | 3.117 | 63  | 339160  | 19.322  | ug/l | 98  |
| 25) 2-Butanone               | 3.873 | 43  | 626977  | 97.122  | ug/l | 99  |
| 26) 2,2-Dichloropropane      | 3.812 | 77  | 266277  | 17.109  | ug/l | 98  |
| 27) cis-1,2-Dichloroethene   | 3.822 | 96  | 191963  | 19.468  | ug/l | 98  |
| 28) Bromochloromethane       | 4.150 | 49  | 158654  | 20.475  | ug/l | 97  |
| 29) Tetrahydrofuran          | 4.240 | 42  | 407358  | 100.814 | ug/l | 98  |
| 30) Chloroform               | 4.281 | 83  | 352541  | 19.234  | ug/l | 97  |
| 31) Cyclohexane              | 4.587 | 56  | 262018  | 18.223  | ug/l | 94  |
| 32) 1,1,1-Trichloroethane    | 4.513 | 97  | 301922  | 19.027  | ug/l | 98  |
| 36) 1,1-Dichloropropene      | 4.744 | 75  | 210259  | 19.244  | ug/l | 98  |
| 37) Ethyl Acetate            | 3.986 | 43  | 249892  | 18.849  | ug/l | 99  |
| 38) Carbon Tetrachloride     | 4.735 | 117 | 258780  | 19.380  | ug/l | 96  |
| 39) Methylcyclohexane        | 6.047 | 83  | 253234  | 19.193  | ug/l | 98  |
| 40) Benzene                  | 5.011 | 78  | 706824  | 19.867  | ug/l | 98  |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
 Data File : VV039171.D  
 Acq On : 25 Sep 2025 11:24  
 Operator : SY/MD  
 Sample : VV0925WBS01  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VV0925WBS01

Quant Time: Sep 26 01:21:22 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 4.166  | 41   | 101017   | 18.119  | ug/l   | 92       |
| 42) 1,2-Dichloroethane        | 5.043  | 62   | 243861   | 19.562  | ug/l   | 99       |
| 43) Isopropyl Acetate         | 5.198  | 43   | 354785   | 19.627  | ug/l   | 99       |
| 44) Trichloroethene           | 5.834  | 130  | 167082   | 20.352  | ug/l   | 99       |
| 45) 1,2-Dichloropropane       | 6.092  | 63   | 181382   | 20.511  | ug/l   | 96       |
| 46) Dibromomethane            | 6.227  | 93   | 135579   | 20.031  | ug/l   | 98       |
| 47) Bromodichloromethane      | 6.429  | 83   | 269985   | 19.295  | ug/l   | 100      |
| 48) Methyl methacrylate       | 6.301  | 41   | 154615   | 18.869  | ug/l   | 97       |
| 49) 1,4-Dioxane               | 6.294  | 88   | 55817    | 422.109 | ug/l   | 98       |
| 51) 4-Methyl-2-Pentanone      | 7.149  | 43   | 1295291  | 108.921 | ug/l   | 99       |
| 52) Toluene                   | 7.310  | 92   | 442966   | 21.474  | ug/l   | 98       |
| 53) t-1,3-Dichloropropene     | 7.577  | 75   | 244520   | 20.052  | ug/l   | 99       |
| 54) cis-1,3-Dichloropropene   | 6.950  | 75   | 272587   | 20.249  | ug/l   | 98       |
| 55) 1,1,2-Trichloroethane     | 7.764  | 97   | 184964   | 19.997  | ug/l   | 95       |
| 56) Ethyl methacrylate        | 7.719  | 69   | 220741   | 20.096  | ug/l   | 97       |
| 57) 1,3-Dichloropropane       | 7.937  | 76   | 293432   | 20.102  | ug/l   | 99       |
| 58) 2-Chloroethyl Vinyl ether | 6.815  | 63   | 541689   | 93.599  | ug/l   | 99       |
| 59) 2-Hexanone                | 8.066  | 43   | 942254   | 99.572  | ug/l   | 99       |
| 60) Dibromochloromethane      | 8.169  | 129  | 212792   | 20.190  | ug/l   | 100      |
| 61) 1,2-Dibromoethane         | 8.275  | 107  | 195149   | 20.631  | ug/l   | 99       |
| 64) Tetrachloroethene         | 7.899  | 164  | 147205   | 20.157  | ug/l   | 95       |
| 65) Chlorobenzene             | 8.805  | 112  | 488359   | 20.108  | ug/l   | 98       |
| 66) 1,1,1,2-Tetrachloroethane | 8.899  | 131  | 174648   | 19.902  | ug/l   | 100      |
| 67) Ethyl Benzene             | 8.937  | 91   | 736424   | 20.032  | ug/l   | 99       |
| 68) m/p-Xylenes               | 9.063  | 106  | 611502   | 43.097  | ug/l   | 100      |
| 69) o-Xylene                  | 9.468  | 106  | 263271   | 20.955  | ug/l   | 99       |
| 70) Styrene                   | 9.487  | 104  | 489732   | 19.311  | ug/l   | 99       |
| 71) Bromoform                 | 9.654  | 173  | 140287   | 20.466  | ug/l # | 98       |
| 73) Isopropylbenzene          | 9.857  | 105  | 759505   | 20.575  | ug/l   | 100      |
| 74) N-amyl acetate            | 9.725  | 43   | 297254   | 20.188  | ug/l   | 99       |
| 75) 1,1,2,2-Tetrachloroethane | 10.165 | 83   | 313809   | 19.402  | ug/l   | 100      |
| 76) 1,2,3-Trichloropropane    | 10.198 | 75   | 227880   | 19.665  | ug/l   | 98       |
| 77) Bromobenzene              | 10.143 | 156  | 194148   | 20.272  | ug/l   | 97       |
| 78) n-propylbenzene           | 10.278 | 91   | 917928   | 20.420  | ug/l   | 99       |
| 79) 2-Chlorotoluene           | 10.349 | 91   | 569887   | 21.187  | ug/l   | 98       |
| 80) 1,3,5-Trimethylbenzene    | 10.464 | 105  | 649406   | 21.165  | ug/l   | 100      |
| 81) trans-1,4-Dichloro-2-b... | 9.931  | 75   | 93631    | 19.398  | ug/l   | 98       |
| 82) 4-Chlorotoluene           | 10.464 | 91   | 658860   | 21.124  | ug/l   | 100      |
| 83) tert-Butylbenzene         | 10.789 | 119  | 599275   | 19.587  | ug/l   | 99       |
| 84) 1,2,4-Trimethylbenzene    | 10.841 | 105  | 634517   | 21.301  | ug/l   | 100      |
| 85) sec-Butylbenzene          | 11.014 | 105  | 814764   | 20.078  | ug/l   | 99       |
| 86) p-Isopropyltoluene        | 11.169 | 119  | 692579   | 20.505  | ug/l   | 99       |
| 87) 1,3-Dichlorobenzene       | 11.108 | 146  | 380476   | 19.629  | ug/l   | 100      |
| 88) 1,4-Dichlorobenzene       | 11.198 | 146  | 400938   | 18.854  | ug/l   | 99       |
| 89) n-Butylbenzene            | 11.580 | 91   | 629906   | 19.528  | ug/l   | 98       |
| 90) Hexachloroethane          | 11.821 | 117  | 153004   | 18.396  | ug/l   | 99       |
| 91) 1,2-Dichlorobenzene       | 11.567 | 146  | 359391   | 19.846  | ug/l   | 99       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.352 | 75   | 63577    | 22.105  | ug/l   | 98       |
| 93) 1,2,4-Trichlorobenzene    | 13.185 | 180  | 198844   | 19.155  | ug/l   | 98       |
| 94) Hexachlorobutadiene       | 13.368 | 225  | 100713   | 18.672  | ug/l   | 99       |
| 95) Naphthalene               | 13.426 | 128  | 654241   | 19.499  | ug/l   | 99       |
| 96) 1,2,3-Trichlorobenzene    | 13.667 | 180  | 210809   | 19.856  | ug/l   | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092525\  
Data File : VV039171.D  
Acq On : 25 Sep 2025 11:24  
Operator : SY/MD  
Sample : VV0925WBS01  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VV0925WBS01

Quant Time: Sep 26 01:21:22 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

**Instrument :**  
MSVOA\_V  
**ClientSampleId :**  
VV0925WBS01

[illegible]

## Report of Analysis

|                    |                            |                 |              |
|--------------------|----------------------------|-----------------|--------------|
| Client:            | G Environmental            | Date Collected: |              |
| Project:           | Summer                     | Date Received:  |              |
| Client Sample ID:  | VV0924WBSD02               | SDG No.:        | Q3175        |
| Lab Sample ID:     | VV0924WBSD02               | Matrix:         | Water        |
| Analytical Method: | 8260D                      | % Solid:        | 0            |
| Sample Wt/Vol:     | 5      Units:    mL        | Final Vol:      | 5000      uL |
| Soil Aliquot Vol:  | uL                         | Test:           | VOCMS Group1 |
| GC Column:         | DB-624UI      ID :    0.18 | Level :         | LOW          |
| Prep Method :      |                            |                 |              |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039165.D        | 1         | 09/24/25 19:59 | VV092425      |

| CAS Number     | Parameter                      | Conc. | Qualifier | MDL   | LOQ / CRQL | Units |
|----------------|--------------------------------|-------|-----------|-------|------------|-------|
| <b>TARGETS</b> |                                |       |           |       |            |       |
| 75-71-8        | Dichlorodifluoromethane        | 18.1  |           | 0.22  | 1.00       | ug/L  |
| 74-87-3        | Chloromethane                  | 18.2  |           | 0.32  | 1.00       | ug/L  |
| 75-01-4        | Vinyl Chloride                 | 18.1  |           | 0.26  | 1.00       | ug/L  |
| 74-83-9        | Bromomethane                   | 14.7  |           | 1.40  | 5.00       | ug/L  |
| 75-00-3        | Chloroethane                   | 20.0  |           | 0.47  | 1.00       | ug/L  |
| 75-69-4        | Trichlorofluoromethane         | 18.9  |           | 0.33  | 1.00       | ug/L  |
| 76-13-1        | 1,1,2-Trichlorotrifluoroethane | 19.0  |           | 0.25  | 1.00       | ug/L  |
| 75-65-0        | Tert butyl alcohol             | 110   |           | 5.50  | 25.0       | ug/L  |
| 75-35-4        | 1,1-Dichloroethene             | 19.4  |           | 0.23  | 1.00       | ug/L  |
| 67-64-1        | Acetone                        | 79.7  |           | 1.50  | 5.00       | ug/L  |
| 75-15-0        | Carbon Disulfide               | 18.3  |           | 0.21  | 1.00       | ug/L  |
| 1634-04-4      | Methyl tert-butyl Ether        | 19.3  |           | 0.16  | 1.00       | ug/L  |
| 79-20-9        | Methyl Acetate                 | 20.1  |           | 0.27  | 1.00       | ug/L  |
| 75-09-2        | Methylene Chloride             | 20.9  |           | 0.28  | 1.00       | ug/L  |
| 156-60-5       | trans-1,2-Dichloroethene       | 18.6  |           | 0.23  | 1.00       | ug/L  |
| 75-34-3        | 1,1-Dichloroethane             | 18.1  |           | 0.23  | 1.00       | ug/L  |
| 110-82-7       | Cyclohexane                    | 18.4  |           | 1.50  | 5.00       | ug/L  |
| 78-93-3        | 2-Butanone                     | 92.0  |           | 0.98  | 5.00       | ug/L  |
| 56-23-5        | Carbon Tetrachloride           | 18.2  |           | 0.25  | 1.00       | ug/L  |
| 156-59-2       | cis-1,2-Dichloroethene         | 18.8  |           | 0.19  | 1.00       | ug/L  |
| 74-97-5        | Bromochloromethane             | 19.2  |           | 0.22  | 1.00       | ug/L  |
| 67-66-3        | Chloroform                     | 18.4  |           | 0.25  | 1.00       | ug/L  |
| 71-55-6        | 1,1,1-Trichloroethane          | 18.4  |           | 0.20  | 1.00       | ug/L  |
| 108-87-2       | Methylcyclohexane              | 18.2  |           | 0.16  | 1.00       | ug/L  |
| 71-43-2        | Benzene                        | 18.8  |           | 0.15  | 1.00       | ug/L  |
| 107-06-2       | 1,2-Dichloroethane             | 18.5  |           | 0.22  | 1.00       | ug/L  |
| 79-01-6        | Trichloroethene                | 19.2  |           | 0.090 | 1.00       | ug/L  |
| 78-87-5        | 1,2-Dichloropropane            | 19.4  |           | 0.20  | 1.00       | ug/L  |
| 75-27-4        | Bromodichloromethane           | 18.3  |           | 0.22  | 1.00       | ug/L  |
| 108-10-1       | 4-Methyl-2-Pentanone           | 100   |           | 0.68  | 5.00       | ug/L  |

### Report of Analysis

|                    |                 |           |                 |              |
|--------------------|-----------------|-----------|-----------------|--------------|
| Client:            | G Environmental |           | Date Collected: |              |
| Project:           | Summer          |           | Date Received:  |              |
| Client Sample ID:  | VV0924WBSD02    |           | SDG No.:        | Q3175        |
| Lab Sample ID:     | VV0924WBSD02    |           | Matrix:         | Water        |
| Analytical Method: | 8260D           |           | % Solid:        | 0            |
| Sample Wt/Vol:     | 5               | Units: mL | Final Vol:      | 5000 uL      |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOCMS Group1 |
| GC Column:         | DB-624UI        | ID : 0.18 | Level :         | LOW          |
| Prep Method :      |                 |           |                 |              |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039165.D        | 1         | 09/24/25 19:59 | VV092425      |

| CAS Number                | Parameter                   | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units   |
|---------------------------|-----------------------------|--------|-----------|----------|------------|---------|
| 108-88-3                  | Toluene                     | 20.0   |           | 0.14     | 1.00       | ug/L    |
| 10061-02-6                | t-1,3-Dichloropropene       | 18.5   |           | 0.17     | 1.00       | ug/L    |
| 10061-01-5                | cis-1,3-Dichloropropene     | 19.2   |           | 0.16     | 1.00       | ug/L    |
| 79-00-5                   | 1,1,2-Trichloroethane       | 18.5   |           | 0.21     | 1.00       | ug/L    |
| 591-78-6                  | 2-Hexanone                  | 91.4   |           | 0.89     | 5.00       | ug/L    |
| 124-48-1                  | Dibromochloromethane        | 18.5   |           | 0.18     | 1.00       | ug/L    |
| 106-93-4                  | 1,2-Dibromoethane           | 19.3   |           | 0.15     | 1.00       | ug/L    |
| 127-18-4                  | Tetrachloroethene           | 19.7   |           | 0.23     | 1.00       | ug/L    |
| 108-90-7                  | Chlorobenzene               | 19.9   |           | 0.12     | 1.00       | ug/L    |
| 100-41-4                  | Ethyl Benzene               | 19.2   |           | 0.13     | 1.00       | ug/L    |
| 179601-23-1               | m/p-Xylenes                 | 40.4   |           | 0.24     | 2.00       | ug/L    |
| 95-47-6                   | o-Xylene                    | 19.8   |           | 0.12     | 1.00       | ug/L    |
| 100-42-5                  | Styrene                     | 17.9   |           | 0.15     | 1.00       | ug/L    |
| 75-25-2                   | Bromoform                   | 18.8   |           | 0.19     | 1.00       | ug/L    |
| 98-82-8                   | Isopropylbenzene            | 19.6   |           | 0.12     | 1.00       | ug/L    |
| 79-34-5                   | 1,1,2,2-Tetrachloroethane   | 17.9   |           | 0.26     | 1.00       | ug/L    |
| 541-73-1                  | 1,3-Dichlorobenzene         | 18.7   |           | 0.16     | 1.00       | ug/L    |
| 106-46-7                  | 1,4-Dichlorobenzene         | 17.9   |           | 0.19     | 1.00       | ug/L    |
| 95-50-1                   | 1,2-Dichlorobenzene         | 18.5   |           | 0.16     | 1.00       | ug/L    |
| 96-12-8                   | 1,2-Dibromo-3-Chloropropane | 19.7   |           | 0.53     | 1.00       | ug/L    |
| 120-82-1                  | 1,2,4-Trichlorobenzene      | 18.4   |           | 0.20     | 1.00       | ug/L    |
| 87-61-6                   | 1,2,3-Trichlorobenzene      | 18.7   |           | 0.20     | 1.00       | ug/L    |
| <b>SURROGATES</b>         |                             |        |           |          |            |         |
| 17060-07-0                | 1,2-Dichloroethane-d4       | 49.5   |           | 74 - 125 | 99%        | SPK: 50 |
| 1868-53-7                 | Dibromofluoromethane        | 50.1   |           | 75 - 124 | 100%       | SPK: 50 |
| 2037-26-5                 | Toluene-d8                  | 50.7   |           | 86 - 113 | 101%       | SPK: 50 |
| 460-00-4                  | 4-Bromofluorobenzene        | 55.0   |           | 77 - 121 | 110%       | SPK: 50 |
| <b>INTERNAL STANDARDS</b> |                             |        |           |          |            |         |
| 363-72-4                  | Pentafluorobenzene          | 545000 | 4.6       |          |            |         |
| 540-36-3                  | 1,4-Difluorobenzene         | 904000 | 5.532     |          |            |         |
| 3114-55-4                 | Chlorobenzene-d5            | 885000 | 8.776     |          |            |         |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4      | 511000 | 11.172    |          |            |         |

## Report of Analysis

|                    |                 |           |                 |              |
|--------------------|-----------------|-----------|-----------------|--------------|
| Client:            | G Environmental |           | Date Collected: |              |
| Project:           | Summer          |           | Date Received:  |              |
| Client Sample ID:  | VV0924WBSD02    |           | SDG No.:        | Q3175        |
| Lab Sample ID:     | VV0924WBSD02    |           | Matrix:         | Water        |
| Analytical Method: | 8260D           |           | % Solid:        | 0            |
| Sample Wt/Vol:     | 5               | Units: mL | Final Vol:      | 5000 uL      |
| Soil Aliquot Vol:  |                 | uL        | Test:           | VOCMS Group1 |
| GC Column:         | DB-624UI        | ID : 0.18 | Level :         | LOW          |
| Prep Method :      |                 |           |                 |              |

|                   |           |                |               |
|-------------------|-----------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Date Analyzed  | Prep Batch ID |
| VV039165.D        | 1         | 09/24/25 19:59 | VV092425      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
 Data File : VV039165.D  
 Acq On : 24 Sep 2025 19:59  
 Operator : SY/MD  
 Sample : VV0924WBSD02  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VV0924WBSD02

Quant Time: Sep 25 08:24:33 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc     | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|----------|--------|----------|
| -----                        |                |      |            |          |        |          |
| Internal Standards           |                |      |            |          |        |          |
| 1) Pentafluorobenzene        | 4.600          | 168  | 545364     | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 5.532          | 114  | 903558     | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 8.776          | 117  | 885424     | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 11.172         | 152  | 510716     | 50.000   | ug/l   | 0.00     |
|                              |                |      |            |          |        |          |
| System Monitoring Compounds  |                |      |            |          |        |          |
| 33) 1,2-Dichloroethane-d4    | 4.944          | 65   | 390603     | 49.487   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 81 - 118 |      | Recovery = | 98.980%  |        |          |
| 35) Dibromofluoromethane     | 4.503          | 113  | 363843     | 50.089   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 80 - 119 |      | Recovery = | 100.180% |        |          |
| 50) Toluene-d8               | 7.236          | 98   | 1080761    | 50.659   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 89 - 112 |      | Recovery = | 101.320% |        |          |
| 62) 4-Bromofluorobenzene     | 10.001         | 95   | 483167     | 55.014   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 85 - 114 |      | Recovery = | 110.020% |        |          |
|                              |                |      |            |          |        |          |
| Target Compounds             |                |      |            |          | Qvalue |          |
| 2) Dichlorodifluoromethane   | 1.121          | 85   | 156320     | 18.107   | ug/l   | 98       |
| 3) Chloromethane             | 1.230          | 50   | 168337     | 18.164   | ug/l   | 100      |
| 4) Vinyl Chloride            | 1.298          | 62   | 180006     | 18.142   | ug/l   | 99       |
| 5) Bromomethane              | 1.497          | 94   | 91963      | 14.703   | ug/l   | 99       |
| 6) Chloroethane              | 1.561          | 64   | 124416     | 20.020   | ug/l   | 100      |
| 7) Trichlorofluoromethane    | 1.725          | 101  | 271403     | 18.893   | ug/l   | 99       |
| 8) Diethyl Ether             | 1.918          | 74   | 98678      | 18.504   | ug/l   | 98       |
| 9) 1,1,2-Trichlorotrifluo... | 2.082          | 101  | 162784     | 19.045   | ug/l   | 98       |
| 10) Methyl Iodide            | 2.198          | 142  | 97066      | 11.925   | ug/l   | 99       |
| 11) Tert butyl alcohol       | 2.568          | 59   | 177476     | 106.387  | ug/l   | 97       |
| 12) 1,1-Dichloroethene       | 2.082          | 96   | 159305     | 19.374   | ug/l   | 92       |
| 13) Acrolein                 | 2.011          | 56   | 32659      | 112.823  | ug/l   | 95       |
| 14) Allyl chloride           | 2.359          | 41   | 262106     | 17.845   | ug/l   | 98       |
| 15) Acrylonitrile            | 2.677          | 53   | 500749     | 93.637   | ug/l   | 99       |
| 16) Acetone                  | 2.121          | 43   | 420916     | 79.697   | ug/l   | 98       |
| 17) Carbon Disulfide         | 2.253          | 76   | 461436     | 18.349   | ug/l   | 99       |
| 18) Methyl Acetate           | 2.381          | 43   | 235110     | 20.104   | ug/l   | 99       |
| 19) Methyl tert-butyl Ether  | 2.712          | 73   | 506464     | 19.329   | ug/l   | 99       |
| 20) Methylene Chloride       | 2.458          | 84   | 191133     | 20.897   | ug/l   | 98       |
| 21) trans-1,2-Dichloroethene | 2.703          | 96   | 163017     | 18.623   | ug/l   | 99       |
| 22) Diisopropyl ether        | 3.217          | 45   | 524404     | 18.661   | ug/l   | 84       |
| 23) Vinyl Acetate            | 3.191          | 43   | 2227046    | 90.591   | ug/l   | 99       |
| 24) 1,1-Dichloroethane       | 3.117          | 63   | 311443     | 18.122   | ug/l   | 99       |
| 25) 2-Butanone               | 3.873          | 43   | 581279     | 91.967   | ug/l   | 99       |
| 26) 2,2-Dichloropropane      | 3.812          | 77   | 223757     | 14.684   | ug/l   | 99       |
| 27) cis-1,2-Dichloroethene   | 3.822          | 96   | 181034     | 18.752   | ug/l   | 99       |
| 28) Bromochloromethane       | 4.150          | 49   | 145773     | 19.214   | ug/l   | 95       |
| 29) Tetrahydrofuran          | 4.243          | 42   | 385403     | 97.419   | ug/l   | 99       |
| 30) Chloroform               | 4.278          | 83   | 330400     | 18.412   | ug/l   | 97       |
| 31) Cyclohexane              | 4.584          | 56   | 258917     | 18.392   | ug/l   | 96       |
| 32) 1,1,1-Trichloroethane    | 4.513          | 97   | 285397     | 18.370   | ug/l   | 98       |
| 36) 1,1-Dichloropropene      | 4.748          | 75   | 205485     | 18.711   | ug/l   | 98       |
| 37) Ethyl Acetate            | 3.982          | 43   | 238455     | 17.894   | ug/l   | 98       |
| 38) Carbon Tetrachloride     | 4.735          | 117  | 243831     | 18.166   | ug/l   | 98       |
| 39) Methylcyclohexane        | 6.050          | 83   | 241644     | 18.220   | ug/l   | 97       |
| 40) Benzene                  | 5.011          | 78   | 673840     | 18.842   | ug/l   | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
 Data File : VV039165.D  
 Acq On : 24 Sep 2025 19:59  
 Operator : SY/MD  
 Sample : VV0924WBSD02  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VV0924WBSD02

Quant Time: Sep 25 08:24:33 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 4.166  | 41   | 96133    | 17.256  | ug/l   | 93       |
| 42) 1,2-Dichloroethane        | 5.043  | 62   | 232378   | 18.545  | ug/l   | 100      |
| 43) Isopropyl Acetate         | 5.198  | 43   | 339665   | 18.694  | ug/l   | 99       |
| 44) Trichloroethene           | 5.831  | 130  | 158174   | 19.167  | ug/l   | 100      |
| 45) 1,2-Dichloropropane       | 6.092  | 63   | 172229   | 19.376  | ug/l   | 97       |
| 46) Dibromomethane            | 6.230  | 93   | 128940   | 18.952  | ug/l   | 97       |
| 47) Bromodichloromethane      | 6.429  | 83   | 257271   | 18.292  | ug/l   | 97       |
| 48) Methyl methacrylate       | 6.301  | 41   | 148719   | 18.056  | ug/l   | 99       |
| 49) 1,4-Dioxane               | 6.297  | 88   | 52205    | 392.759 | ug/l   | 94       |
| 51) 4-Methyl-2-Pentanone      | 7.149  | 43   | 1220207  | 102.079 | ug/l   | 99       |
| 52) Toluene                   | 7.310  | 92   | 415722   | 20.049  | ug/l   | 100      |
| 53) t-1,3-Dichloropropene     | 7.577  | 75   | 226898   | 18.511  | ug/l   | 99       |
| 54) cis-1,3-Dichloropropene   | 6.950  | 75   | 259523   | 19.179  | ug/l   | 98       |
| 55) 1,1,2-Trichloroethane     | 7.764  | 97   | 171629   | 18.460  | ug/l   | 97       |
| 56) Ethyl methacrylate        | 7.719  | 69   | 204966   | 18.564  | ug/l   | 98       |
| 57) 1,3-Dichloropropane       | 7.934  | 76   | 277934   | 18.942  | ug/l   | 100      |
| 58) 2-Chloroethyl Vinyl ether | 6.815  | 63   | 486078   | 84.589  | ug/l   | 99       |
| 59) 2-Hexanone                | 8.066  | 43   | 867936   | 91.441  | ug/l   | 99       |
| 60) Dibromochloromethane      | 8.169  | 129  | 195919   | 18.494  | ug/l   | 100      |
| 61) 1,2-Dibromoethane         | 8.275  | 107  | 183626   | 19.312  | ug/l   | 99       |
| 64) Tetrachloroethene         | 7.899  | 164  | 146324   | 19.730  | ug/l   | 98       |
| 65) Chlorobenzene             | 8.805  | 112  | 491329   | 19.922  | ug/l   | 99       |
| 66) 1,1,1,2-Tetrachloroethane | 8.899  | 131  | 162674   | 18.255  | ug/l   | 100      |
| 67) Ethyl Benzene             | 8.937  | 91   | 716157   | 19.183  | ug/l   | 99       |
| 68) m/p-Xylenes               | 9.063  | 106  | 582336   | 40.415  | ug/l   | 99       |
| 69) o-Xylene                  | 9.468  | 106  | 252664   | 19.804  | ug/l   | 98       |
| 70) Styrene                   | 9.487  | 104  | 457807   | 17.897  | ug/l   | 99       |
| 71) Bromoform                 | 9.654  | 173  | 131160   | 18.843  | ug/l # | 99       |
| 73) Isopropylbenzene          | 9.857  | 105  | 724475   | 19.558  | ug/l   | 99       |
| 74) N-amyl acetate            | 9.725  | 43   | 279661   | 18.927  | ug/l   | 99       |
| 75) 1,1,2,2-Tetrachloroethane | 10.165 | 83   | 290148   | 17.877  | ug/l   | 100      |
| 76) 1,2,3-Trichloropropane    | 10.198 | 75   | 219335   | 18.862  | ug/l   | 100      |
| 77) Bromobenzene              | 10.140 | 156  | 183095   | 19.051  | ug/l   | 97       |
| 78) n-propylbenzene           | 10.278 | 91   | 881210   | 19.535  | ug/l   | 100      |
| 79) 2-Chlorotoluene           | 10.349 | 91   | 539084   | 19.972  | ug/l   | 99       |
| 80) 1,3,5-Trimethylbenzene    | 10.464 | 105  | 620336   | 20.148  | ug/l   | 99       |
| 81) trans-1,4-Dichloro-2-b... | 9.931  | 75   | 82955    | 17.126  | ug/l   | 98       |
| 82) 4-Chlorotoluene           | 10.464 | 91   | 626744   | 20.025  | ug/l   | 100      |
| 83) tert-Butylbenzene         | 10.789 | 119  | 563652   | 18.359  | ug/l   | 100      |
| 84) 1,2,4-Trimethylbenzene    | 10.841 | 105  | 604069   | 20.209  | ug/l   | 99       |
| 85) sec-Butylbenzene          | 11.014 | 105  | 783211   | 19.233  | ug/l   | 100      |
| 86) p-Isopropyltoluene        | 11.165 | 119  | 661055   | 19.504  | ug/l   | 100      |
| 87) 1,3-Dichlorobenzene       | 11.104 | 146  | 362792   | 18.652  | ug/l   | 99       |
| 88) 1,4-Dichlorobenzene       | 11.198 | 146  | 381649   | 17.885  | ug/l   | 98       |
| 89) n-Butylbenzene            | 11.580 | 91   | 608918   | 18.812  | ug/l   | 98       |
| 90) Hexachloroethane          | 11.818 | 117  | 138165   | 16.554  | ug/l   | 99       |
| 91) 1,2-Dichlorobenzene       | 11.567 | 146  | 335832   | 18.481  | ug/l   | 100      |
| 92) 1,2-Dibromo-3-Chloropr... | 12.352 | 75   | 57082    | 19.730  | ug/l   | 99       |
| 93) 1,2,4-Trichlorobenzene    | 13.185 | 180  | 191728   | 18.406  | ug/l   | 100      |
| 94) Hexachlorobutadiene       | 13.368 | 225  | 91410    | 16.888  | ug/l   | 99       |
| 95) Naphthalene               | 13.426 | 128  | 623030   | 18.505  | ug/l   | 100      |
| 96) 1,2,3-Trichlorobenzene    | 13.667 | 180  | 199614   | 18.736  | ug/l   | 98       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092425\  
Data File : VV039165.D  
Acq On : 24 Sep 2025 19:59  
Operator : SY/MD  
Sample : VV0924WBSD02  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 22 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VV0924WBSD02

Quant Time: Sep 25 08:24:33 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration


| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092425\  
Data File : VV039165.D  
Acq On    : 24 Sep 2025   19:59  
Operator  : SY/MD  
Sample    : VV0924WBSD02  
Misc      : 5.0mL/MSVOA_V/WATER  
ALS Vial  : 22   Sample Multiplier: 1
```

**Instrument :**  
MSVOA\_V  
**ClientSampleId :**  
VV0924WBSD02

Quant Time: Sep 25 08:24:33 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration



## Manual Integration Report

Sequence:

VV092225

Instrument

MSVOA\_v

| Sample ID  | File ID    | Parameter              | Review By    | Review On               | Supervised By | Supervised On           | Reason                      |
|------------|------------|------------------------|--------------|-------------------------|---------------|-------------------------|-----------------------------|
| VSTDICC005 | VV039142.D | 1,2,3-Trichloropropane | MMDadod<br>a | 9/26/2025 1:38:50<br>AM | SAM           | 9/26/2025 1:57:11<br>AM | Peak Integrated by Software |
| VSTDICC001 | VV039143.D | 1,1-Dichloropropene    | MMDadod<br>a | 9/26/2025 1:38:52<br>AM | SAM           | 9/26/2025 1:57:13<br>AM | Peak Integrated by Software |
| VSTDICC001 | VV039143.D | 1,4-Dichlorobenzene    | MMDadod<br>a | 9/26/2025 1:38:52<br>AM | SAM           | 9/26/2025 1:57:13<br>AM | Peak Integrated by Software |
| VSTDICC001 | VV039143.D | 1,4-Dioxane            | MMDadod<br>a | 9/26/2025 1:38:52<br>AM | SAM           | 9/26/2025 1:57:13<br>AM | Peak Integrated by Software |
| VSTDICC001 | VV039143.D | 2,2-Dichloropropane    | MMDadod<br>a | 9/26/2025 1:38:52<br>AM | SAM           | 9/26/2025 1:57:13<br>AM | Peak Integrated by Software |
| VSTDICC001 | VV039143.D | 2-Butanone             | MMDadod<br>a | 9/26/2025 1:38:52<br>AM | SAM           | 9/26/2025 1:57:13<br>AM | Peak Integrated by Software |
| VSTDICC001 | VV039143.D | Carbon Tetrachloride   | MMDadod<br>a | 9/26/2025 1:38:52<br>AM | SAM           | 9/26/2025 1:57:13<br>AM | Peak Integrated by Software |
| VSTDICC001 | VV039143.D | Ethyl Acetate          | MMDadod<br>a | 9/26/2025 1:38:52<br>AM | SAM           | 9/26/2025 1:57:13<br>AM | Peak Integrated by Software |
| VSTDICC001 | VV039143.D | Ethyl methacrylate     | MMDadod<br>a | 9/26/2025 1:38:52<br>AM | SAM           | 9/26/2025 1:57:13<br>AM | Peak Integrated by Software |
| VSTDICC001 | VV039143.D | Isopropyl Acetate      | MMDadod<br>a | 9/26/2025 1:38:52<br>AM | SAM           | 9/26/2025 1:57:13<br>AM | Peak Integrated by Software |
| VSTDICC001 | VV039143.D | Methacrylonitrile      | MMDadod<br>a | 9/26/2025 1:38:52<br>AM | SAM           | 9/26/2025 1:57:13<br>AM | Peak Integrated by Software |
| VSTDICC001 | VV039143.D | Methyl methacrylate    | MMDadod<br>a | 9/26/2025 1:38:52<br>AM | SAM           | 9/26/2025 1:57:13<br>AM | Peak Integrated by Software |
| VSTDICC001 | VV039143.D | Naphthalene            | MMDadod<br>a | 9/26/2025 1:38:52<br>AM | SAM           | 9/26/2025 1:57:13<br>AM | Peak Integrated by Software |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

## Manual Integration Report

|           |          |            |         |
|-----------|----------|------------|---------|
| Sequence: | VV092225 | Instrument | MSVOA_v |
|-----------|----------|------------|---------|

| Sample ID  | File ID    | Parameter              | Review By | Review On            | Supervised By | Supervised On        | Reason                      |
|------------|------------|------------------------|-----------|----------------------|---------------|----------------------|-----------------------------|
| VSTDICV050 | VV039144.D | 1,2,3-Trichloropropane | MMDadoda  | 9/26/2025 1:38:55 AM | SAM           | 9/26/2025 1:57:15 AM | Peak Integrated by Software |

## Manual Integration Report

|           |          |            |         |
|-----------|----------|------------|---------|
| Sequence: | VV092425 | Instrument | MSVOA_v |
|-----------|----------|------------|---------|

| Sample ID | File ID    | Parameter              | Review By    | Review On                | Supervised By | Supervised On           | Reason                      |
|-----------|------------|------------------------|--------------|--------------------------|---------------|-------------------------|-----------------------------|
| Q3175-01  | VV039153.D | 1,2,4-Trimethylbenzene | MMdadod<br>a | 9/29/2025<br>12:51:06 AM | sam           | 9/29/2025 1:03:18<br>AM | Peak Integrated by Software |
| Q3175-01  | VV039153.D | 1,3,5-Trimethylbenzene | MMdadod<br>a | 9/29/2025<br>12:51:06 AM | sam           | 9/29/2025 1:03:18<br>AM | Peak Integrated by Software |
| Q3175-01  | VV039153.D | Ethyl Benzene          | MMdadod<br>a | 9/29/2025<br>12:51:06 AM | sam           | 9/29/2025 1:03:18<br>AM | Peak Integrated by Software |
| Q3175-01  | VV039153.D | m/p-Xylenes            | MMdadod<br>a | 9/29/2025<br>12:51:06 AM | sam           | 9/29/2025 1:03:18<br>AM | Peak Integrated by Software |
| Q3175-01  | VV039153.D | n-Butylbenzene         | MMdadod<br>a | 9/29/2025<br>12:51:06 AM | sam           | 9/29/2025 1:03:18<br>AM | Peak Integrated by Software |
| Q3175-01  | VV039153.D | n-propylbenzene        | MMdadod<br>a | 9/29/2025<br>12:51:06 AM | sam           | 9/29/2025 1:03:18<br>AM | Peak Integrated by Software |
| Q3175-01  | VV039153.D | Naphthalene            | MMdadod<br>a | 9/29/2025<br>12:51:06 AM | sam           | 9/29/2025 1:03:18<br>AM | Peak Integrated by Software |
| Q3175-01  | VV039153.D | p-Isopropyltoluene     | MMdadod<br>a | 9/29/2025<br>12:51:06 AM | sam           | 9/29/2025 1:03:18<br>AM | Peak Integrated by Software |
| Q3175-02  | VV039154.D | 1,2,4-Trimethylbenzene | MMDadod<br>a | 9/26/2025 1:44:44<br>AM  | SAM           | 9/26/2025 2:03:19<br>AM | Peak Integrated by Software |
| Q3175-02  | VV039154.D | Benzene                | MMDadod<br>a | 9/26/2025 1:44:44<br>AM  | SAM           | 9/26/2025 2:03:19<br>AM | Peak Integrated by Software |
| Q3175-02  | VV039154.D | Ethyl Benzene          | MMDadod<br>a | 9/26/2025 1:44:44<br>AM  | SAM           | 9/26/2025 2:03:19<br>AM | Peak Integrated by Software |
| Q3175-02  | VV039154.D | m/p-Xylenes            | MMDadod<br>a | 9/26/2025 1:44:44<br>AM  | SAM           | 9/26/2025 2:03:19<br>AM | Peak Integrated by Software |
| Q3175-02  | VV039154.D | n-Butylbenzene         | MMDadod<br>a | 9/26/2025 1:44:44<br>AM  | SAM           | 9/26/2025 2:03:19<br>AM | Peak Integrated by Software |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

## Manual Integration Report

Sequence:

VV092425

Instrument

MSVOA\_v

Sample ID

File ID

Parameter

Review By

Review On

Supervised  
By

Supervised On

Reason

## Manual Integration Report

Sequence:

VV092525

Instrument

MSVOA\_v

| Sample ID   | File ID    | Parameter          | Review By    | Review On                | Supervised By | Supervised On           | Reason                      |
|-------------|------------|--------------------|--------------|--------------------------|---------------|-------------------------|-----------------------------|
| Q3175-01DL  | VV039174.D | n-Butylbenzene     | MMdadod<br>a | 9/29/2025<br>12:54:28 AM | sam           | 9/29/2025 1:07:56<br>AM | Peak Integrated by Software |
| Q3175-01DL  | VV039174.D | p-Isopropyltoluene | MMdadod<br>a | 9/29/2025<br>12:54:28 AM | sam           | 9/29/2025 1:07:56<br>AM | Peak Integrated by Software |
| Q3175-02DL  | VV039175.D | n-Butylbenzene     | MMdadod<br>a | 9/29/2025<br>12:54:39 AM | sam           | 9/29/2025 1:08:03<br>AM | Peak Integrated by Software |
| Q3175-02DL  | VV039175.D | p-Isopropyltoluene | MMdadod<br>a | 9/29/2025<br>12:54:39 AM | sam           | 9/29/2025 1:08:03<br>AM | Peak Integrated by Software |
| Q3175-01DL2 | VV039179.D | n-Butylbenzene     | MMdadod<br>a | 9/29/2025<br>12:56:03 AM | sam           | 9/29/2025 1:08:19<br>AM | Peak Integrated by Software |
| Q3175-01DL2 | VV039179.D | p-Isopropyltoluene | MMdadod<br>a | 9/29/2025<br>12:56:03 AM | sam           | 9/29/2025 1:08:19<br>AM | Peak Integrated by Software |
| Q3175-01DL2 | VV039179.D | sec-Butylbenzene   | MMdadod<br>a | 9/29/2025<br>12:56:03 AM | sam           | 9/29/2025 1:08:19<br>AM | Peak Integrated by Software |

Instrument ID: **MSVOA\_V**

**Daily Analysis Runlog For Sequence/QC Batch ID # VV092225**

|                          |                     |                                                       |                      |                      |              |
|--------------------------|---------------------|-------------------------------------------------------|----------------------|----------------------|--------------|
| Review By                | Mahesh Dadoda       | Review On                                             | 9/26/2025 1:39:32 AM |                      |              |
| Supervise By             | Semsettin Yesilyurt | Supervise On                                          | 9/26/2025 1:56:56 AM |                      |              |
| SubDirectory             | VV092225            | HP Acquire Method                                     | MSVOA_V              | HP Processing Method | 82v092225w.m |
| STD. NAME                |                     | STD REF.#                                             |                      |                      |              |
| Tune/Reschk              |                     | VP135580                                              |                      |                      |              |
| Initial Calibration Stds |                     | VP135581,VP135582,VP135583,VP135584,VP135585,VP135586 |                      |                      |              |
| CCC                      |                     |                                                       |                      |                      |              |
| Internal Standard/PEM    |                     | VP133935                                              |                      |                      |              |
| ICV/I.BLK                |                     | VP135587                                              |                      |                      |              |
| Surrogate Standard       |                     |                                                       |                      |                      |              |
| MS/MSD Standard          |                     |                                                       |                      |                      |              |
| LCS Standard             |                     |                                                       |                      |                      |              |

| Sr# | SampleId   | Data File Name | Date-Time         | Operator | Status |
|-----|------------|----------------|-------------------|----------|--------|
| 1   | BFB        | VV039134.D     | 22 Sep 2025 10:15 | SY/MD    | Ok     |
| 2   | VSTDIC001  | VV039135.D     | 22 Sep 2025 10:59 | SY/MD    | Not Ok |
| 3   | VSTDIC005  | VV039136.D     | 22 Sep 2025 11:30 | SY/MD    | Not Ok |
| 4   | VSTDIC010  | VV039137.D     | 22 Sep 2025 12:26 | SY/MD    | Ok     |
| 5   | VSTDIC020  | VV039138.D     | 22 Sep 2025 12:48 | SY/MD    | Ok     |
| 6   | VSTDIC050  | VV039139.D     | 22 Sep 2025 13:26 | SY/MD    | Ok     |
| 7   | VSTDIC100  | VV039140.D     | 22 Sep 2025 13:49 | SY/MD    | Ok     |
| 8   | VIBLK      | VV039141.D     | 22 Sep 2025 14:21 | SY/MD    | Ok     |
| 9   | VSTDIC005  | VV039142.D     | 22 Sep 2025 15:37 | SY/MD    | Ok,M   |
| 10  | VSTDIC001  | VV039143.D     | 22 Sep 2025 16:35 | SY/MD    | Ok,M   |
| 11  | VSTDICV050 | VV039144.D     | 22 Sep 2025 16:58 | SY/MD    | Ok,M   |

M : Manual Integration

Instrument ID: **MSVOA\_V**

**Daily Analysis Runlog For Sequence/QC Batch ID # VV092425**

|                                         |                     |                   |                      |                      |
|-----------------------------------------|---------------------|-------------------|----------------------|----------------------|
| Review By                               | Mahesh Dadoda       | Review On         | 9/26/2025 1:45:06 AM |                      |
| Supervise By                            | Semsettin Yesilyurt | Supervise On      | 9/26/2025 2:03:33 AM |                      |
| SubDirectory                            | VV092425            | HP Acquire Method | MSVOA_V              | HP Processing Method |
| STD. NAME                               |                     | STD REF.#         |                      |                      |
| Tune/Reschk<br>Initial Calibration Stds |                     | VP135622          |                      |                      |
| CCC                                     |                     | VP135623,VP135624 |                      |                      |
| Internal Standard/PEM                   |                     | VP133935          |                      |                      |
| ICV/I.BLK                               |                     |                   |                      |                      |
| Surrogate Standard                      |                     |                   |                      |                      |
| MS/MSD Standard                         |                     |                   |                      |                      |
| LCS Standard                            |                     |                   |                      |                      |

| Sr# | SampleId     | Data File Name | Date-Time         | Operator | Status   |
|-----|--------------|----------------|-------------------|----------|----------|
| 1   | BFB          | VV039145.D     | 24 Sep 2025 09:52 | SY/MD    | Ok       |
| 2   | VSTDCCC050   | VV039146.D     | 24 Sep 2025 10:49 | SY/MD    | Ok       |
| 3   | VV0924MBL01  | VV039147.D     | 24 Sep 2025 12:08 | SY/MD    | Ok       |
| 4   | VV0924WBL01  | VV039148.D     | 24 Sep 2025 13:09 | SY/MD    | Ok       |
| 5   | VV0924WBS01  | VV039149.D     | 24 Sep 2025 13:37 | SY/MD    | Not Ok   |
| 6   | VV0924WBS02  | VV039150.D     | 24 Sep 2025 14:01 | SY/MD    | Ok       |
| 7   | Q3173-02     | VV039151.D     | 24 Sep 2025 14:46 | SY/MD    | Ok       |
| 8   | Q3173-01     | VV039152.D     | 24 Sep 2025 15:08 | SY/MD    | Ok       |
| 9   | Q3175-01     | VV039153.D     | 24 Sep 2025 15:31 | SY/MD    | Dilution |
| 10  | Q3175-02     | VV039154.D     | 24 Sep 2025 15:53 | SY/MD    | Dilution |
| 11  | VIBLK        | VV039155.D     | 24 Sep 2025 16:16 | SY/MD    | Ok       |
| 12  | VIBLK        | VV039156.D     | 24 Sep 2025 16:38 | SY/MD    | Ok       |
| 13  | Q3172-01     | VV039157.D     | 24 Sep 2025 17:00 | SY/MD    | Dilution |
| 14  | Q3172-02     | VV039158.D     | 24 Sep 2025 17:22 | SY/MD    | Dilution |
| 15  | Q3172-03     | VV039159.D     | 24 Sep 2025 17:45 | SY/MD    | ReRun    |
| 16  | Q3164-01     | VV039160.D     | 24 Sep 2025 18:07 | SY/MD    | Ok       |
| 17  | Q3164-02     | VV039161.D     | 24 Sep 2025 18:30 | SY/MD    | Not Ok   |
| 18  | Q3164-03     | VV039162.D     | 24 Sep 2025 18:52 | SY/MD    | Ok       |
| 19  | VIBLK        | VV039163.D     | 24 Sep 2025 19:15 | SY/MD    | Ok       |
| 20  | VIBLK        | VV039164.D     | 24 Sep 2025 19:37 | SY/MD    | Ok,M     |
| 21  | VV0924WBSD02 | VV039165.D     | 24 Sep 2025 19:59 | SY/MD    | Ok       |

Instrument ID: **MSVOA\_V**

**Daily Analysis Runlog For Sequence/QCBatch ID # VV092425**

| Review By                               | Mahesh Dadoda       | Review On         | 9/26/2025 1:45:06 AM |                      |
|-----------------------------------------|---------------------|-------------------|----------------------|----------------------|
| Supervise By                            | Semsettin Yesilyurt | Supervise On      | 9/26/2025 2:03:33 AM |                      |
| SubDirectory                            | VV092425            | HP Acquire Method | MSVOA_V              | HP Processing Method |
| STD. NAME                               | STD REF.#           |                   |                      |                      |
| Tune/Reschk<br>Initial Calibration Stds | VP135622            |                   |                      |                      |
| CCC                                     | VP135623,VP135624   |                   |                      |                      |
| Internal Standard/PEM                   | VP133935            |                   |                      |                      |
| ICV/I.BLK                               |                     |                   |                      |                      |
| Surrogate Standard                      |                     |                   |                      |                      |
| MS/MSD Standard                         |                     |                   |                      |                      |
| LCS Standard                            |                     |                   |                      |                      |

|    |            |            |                   |       |    |
|----|------------|------------|-------------------|-------|----|
| 22 | VSTDCCC050 | VV039166.D | 24 Sep 2025 20:22 | SY/MD | Ok |
|----|------------|------------|-------------------|-------|----|

M : Manual Integration

Instrument ID: **MSVOA\_V**

**Daily Analysis Runlog For Sequence/QC Batch ID # VV092525**

|                                                                                                    |                               |                      |                       |
|----------------------------------------------------------------------------------------------------|-------------------------------|----------------------|-----------------------|
| Review By                                                                                          | Mahesh Dadoda                 | Review On            | 9/29/2025 12:56:22 AM |
| Supervise By                                                                                       | Semsettin Yesilyurt           | Supervise On         | 9/29/2025 1:08:36 AM  |
| SubDirectory                                                                                       | VV092525                      | HP Acquire Method    | 8260_V                |
|                                                                                                    |                               | HP Processing Method | 82v092225w.m          |
| <b>STD. NAME</b>                                                                                   | <b>STD REF.#</b>              |                      |                       |
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP135649                      |                      |                       |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP135650,VP135651<br>VP133935 |                      |                       |

| Sr# | SampleId     | Data File Name | Date-Time         | Operator | Status   |
|-----|--------------|----------------|-------------------|----------|----------|
| 1   | BFB          | VV039167.D     | 25 Sep 2025 09:14 | SY/MD    | Ok       |
| 2   | VSTDCCC050   | VV039168.D     | 25 Sep 2025 09:59 | SY/MD    | Ok       |
| 3   | VV0925MBL01  | VV039169.D     | 25 Sep 2025 10:31 | SY/MD    | Ok       |
| 4   | VV0925WBL01  | VV039170.D     | 25 Sep 2025 11:01 | SY/MD    | Ok       |
| 5   | VV0925WBS01  | VV039171.D     | 25 Sep 2025 11:24 | SY/MD    | Ok       |
| 6   | VV0925WBSD01 | VV039172.D     | 25 Sep 2025 11:48 | SY/MD    | Ok       |
| 7   | Q3164-02     | VV039173.D     | 25 Sep 2025 12:10 | SY/MD    | Ok       |
| 8   | Q3175-01DL   | VV039174.D     | 25 Sep 2025 12:33 | SY/MD    | Dilution |
| 9   | Q3175-02DL   | VV039175.D     | 25 Sep 2025 12:55 | SY/MD    | Ok,M     |
| 10  | VIBLK        | VV039176.D     | 25 Sep 2025 13:18 | SY/MD    | Ok,M     |
| 11  | Q3168-05     | VV039177.D     | 25 Sep 2025 13:40 | SY/MD    | ReRun    |
| 12  | Q3168-06     | VV039178.D     | 25 Sep 2025 14:03 | SY/MD    | ReRun    |
| 13  | Q3175-01DL2  | VV039179.D     | 25 Sep 2025 14:25 | SY/MD    | Ok,M     |
| 14  | Q3172-02DL   | VV039180.D     | 25 Sep 2025 14:48 | SY/MD    | Ok       |
| 15  | VIBLK        | VV039181.D     | 25 Sep 2025 15:10 | SY/MD    | Ok       |
| 16  | Q3172-01DL   | VV039182.D     | 25 Sep 2025 15:33 | SY/MD    | Ok       |
| 17  | Q3168-05RE   | VV039183.D     | 25 Sep 2025 15:55 | SY/MD    | Confirms |
| 18  | Q3168-06RE   | VV039184.D     | 25 Sep 2025 16:18 | SY/MD    | Confirms |
| 19  | Q3172-03     | VV039185.D     | 25 Sep 2025 16:40 | SY/MD    | Ok       |
| 20  | Q3195-22     | VV039186.D     | 25 Sep 2025 17:03 | SY/MD    | Ok       |
| 21  | Q3195-23     | VV039187.D     | 25 Sep 2025 17:25 | SY/MD    | Ok       |

Instrument ID: **MSVOA\_V**

**Daily Analysis Runlog For Sequence/QCBatch ID # VV092525**

| Review By                               | Maresh Dadoda       | Review On         | 9/29/2025 12:56:22 AM |                      |              |
|-----------------------------------------|---------------------|-------------------|-----------------------|----------------------|--------------|
| Supervise By                            | Semsettin Yesilyurt | Supervise On      | 9/29/2025 1:08:36 AM  |                      |              |
| SubDirectory                            | VV092525            | HP Acquire Method | 8260_V                | HP Processing Method | 82v092225w.m |
| STD. NAME                               |                     | STD REF.#         |                       |                      |              |
| Tune/Reschk<br>Initial Calibration Stds |                     | VP135649          |                       |                      |              |
| CCC                                     |                     | VP135650,VP135651 |                       |                      |              |
| Internal Standard/PEM                   |                     | VP133935          |                       |                      |              |
| ICV/I.BLK                               |                     |                   |                       |                      |              |
| Surrogate Standard                      |                     |                   |                       |                      |              |
| MS/MSD Standard                         |                     |                   |                       |                      |              |
| LCS Standard                            |                     |                   |                       |                      |              |

|    |            |            |                   |       |       |
|----|------------|------------|-------------------|-------|-------|
| 22 | Q3195-24   | VV039188.D | 25 Sep 2025 17:47 | SY/MD | ReRun |
| 23 | Q3195-25   | VV039189.D | 25 Sep 2025 18:10 | SY/MD | Ok    |
| 24 | Q3195-26   | VV039190.D | 25 Sep 2025 18:32 | SY/MD | Ok    |
| 25 | Q3195-27   | VV039191.D | 25 Sep 2025 18:55 | SY/MD | Ok    |
| 26 | Q3195-02   | VV039192.D | 25 Sep 2025 19:17 | SY/MD | Ok    |
| 27 | Q3195-09   | VV039193.D | 25 Sep 2025 19:40 | SY/MD | Ok    |
| 28 | Q3195-10   | VV039194.D | 25 Sep 2025 20:02 | SY/MD | Ok    |
| 29 | Q3195-11   | VV039195.D | 25 Sep 2025 20:25 | SY/MD | Ok    |
| 30 | VSTDCCC050 | VV039196.D | 25 Sep 2025 20:47 | SY/MD | Ok    |

M : Manual Integration

Instrument ID: MSVOA\_V

**Daily Analysis Runlog For Sequence/QC Batch ID # VV092225**

|              |                     |                   |                      |                      |              |
|--------------|---------------------|-------------------|----------------------|----------------------|--------------|
| Review By    | Mahesh Dadoda       | Review On         | 9/26/2025 1:39:32 AM |                      |              |
| Supervise By | Semsettin Yesilyurt | Supervise On      | 9/26/2025 1:56:56 AM |                      |              |
| SubDirectory | VV092225            | HP Acquire Method | MSVOA_V              | HP Processing Method | 82v092225w.m |

| STD. NAME                | STD REF.#                                             |
|--------------------------|-------------------------------------------------------|
| Tune/Reschk              | VP135580                                              |
| Initial Calibration Stds | VP135581,VP135582,VP135583,VP135584,VP135585,VP135586 |
| CCC                      |                                                       |
| Internal Standard/PEM    | VP133935                                              |
| ICV/I.BLK                | VP135587                                              |
| Surrogate Standard       |                                                       |
| MS/MSD Standard          |                                                       |
| LCS Standard             |                                                       |

| Sr# | Sampleld   | ClientID    | Data File Name | Date-Time         | Comment            | Operator | Status |
|-----|------------|-------------|----------------|-------------------|--------------------|----------|--------|
| 1   | BFB        | BFB         | VV039134.D     | 22 Sep 2025 10:15 |                    | SY/MD    | Ok     |
| 2   | VSTDIC001  | VSTDIC001   | VV039135.D     | 22 Sep 2025 10:59 | Not use            | SY/MD    | Not Ok |
| 3   | VSTDIC005  | VSTDIC005   | VV039136.D     | 22 Sep 2025 11:30 | Not use            | SY/MD    | Not Ok |
| 4   | VSTDIC010  | VSTDIC010   | VV039137.D     | 22 Sep 2025 12:26 |                    | SY/MD    | Ok     |
| 5   | VSTDIC020  | VSTDIC020   | VV039138.D     | 22 Sep 2025 12:48 | QR- 06,16,20,70,92 | SY/MD    | Ok     |
| 6   | VSTDIC050  | VSTDIC050   | VV039139.D     | 22 Sep 2025 13:26 | LR- 41,58,59       | SY/MD    | Ok     |
| 7   | VSTDIC100  | VSTDIC100   | VV039140.D     | 22 Sep 2025 13:49 |                    | SY/MD    | Ok     |
| 8   | VIBLK      | VIBLK       | VV039141.D     | 22 Sep 2025 14:21 |                    | SY/MD    | Ok     |
| 9   | VSTDIC005  | VSTDIC005   | VV039142.D     | 22 Sep 2025 15:37 |                    | SY/MD    | Ok,M   |
| 10  | VSTDIC001  | VSTDIC001   | VV039143.D     | 22 Sep 2025 16:35 |                    | SY/MD    | Ok,M   |
| 11  | VSTDICV050 | ICVVV092225 | VV039144.D     | 22 Sep 2025 16:58 | Goof for Non-DOD   | SY/MD    | Ok,M   |

M : Manual Integration

Instrument ID: MSVOA\_V

**Daily Analysis Runlog For Sequence/QC Batch ID # VV092425**

|              |                     |                   |                              |
|--------------|---------------------|-------------------|------------------------------|
| Review By    | Maresh Dadoda       | Review On         | 9/26/2025 1:45:06 AM         |
| Supervise By | Semsettin Yesilyurt | Supervise On      | 9/26/2025 2:03:33 AM         |
| SubDirectory | VV092425            | HP Acquire Method | MSVOA_V HP Processing Method |

| STD. NAME                                                                                          | STD REF.#                     |
|----------------------------------------------------------------------------------------------------|-------------------------------|
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP135622                      |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP135623,VP135624<br>VP133935 |

| Sr# | SampleID    | ClientID    | Data File Name | Date-Time         | Comment                             | Operator | Status   |
|-----|-------------|-------------|----------------|-------------------|-------------------------------------|----------|----------|
| 1   | BFB         | BFB         | VV039145.D     | 24 Sep 2025 09:52 |                                     | SY/MD    | Ok       |
| 2   | VSTDCCC050  | VSTDCCC050  | VV039146.D     | 24 Sep 2025 10:49 | pH#lot#v14222                       | SY/MD    | Ok       |
| 3   | VV0924MBL01 | VV0924MBL01 | VV039147.D     | 24 Sep 2025 12:08 |                                     | SY/MD    | Ok       |
| 4   | VV0924WBL01 | VV0924WBL01 | VV039148.D     | 24 Sep 2025 13:09 |                                     | SY/MD    | Ok       |
| 5   | VV0924WBS01 | VV0924WBS01 | VV039149.D     | 24 Sep 2025 13:37 | Recovery fail                       | SY/MD    | Not Ok   |
| 6   | VV0924WBS02 | VV0924WBS02 | VV039150.D     | 24 Sep 2025 14:01 |                                     | SY/MD    | Ok       |
| 7   | Q3173-02    | Field Blank | VV039151.D     | 24 Sep 2025 14:46 | pH<2.0 A,FB                         | SY/MD    | Ok       |
| 8   | Q3173-01    | OBS1        | VV039152.D     | 24 Sep 2025 15:08 | pH<2.0 A                            | SY/MD    | Ok       |
| 9   | Q3175-01    | MW2         | VV039153.D     | 24 Sep 2025 15:31 | pH<2.0 A ,Need 20X                  | SY/MD    | Dilution |
| 10  | Q3175-02    | MW4         | VV039154.D     | 24 Sep 2025 15:53 | pH<2.0 A ,Need 20X                  | SY/MD    | Dilution |
| 11  | VIBLK       | VIBLK       | VV039155.D     | 24 Sep 2025 16:16 |                                     | SY/MD    | Ok       |
| 12  | VIBLK       | VIBLK       | VV039156.D     | 24 Sep 2025 16:38 |                                     | SY/MD    | Ok       |
| 13  | Q3172-01    | FW7D        | VV039157.D     | 24 Sep 2025 17:00 | pH<2.0 A ,Surrogate Fail;Need 100X  | SY/MD    | Dilution |
| 14  | Q3172-02    | FW6D        | VV039158.D     | 24 Sep 2025 17:22 | pH<2.0 A ,Need 4X                   | SY/MD    | Dilution |
| 15  | Q3172-03    | GB3W        | VV039159.D     | 24 Sep 2025 17:45 | pH<2.0 A ,E flag of previous sample | SY/MD    | ReRun    |
| 16  | Q3164-01    | DA-1        | VV039160.D     | 24 Sep 2025 18:07 | pH<2.0 A                            | SY/MD    | Ok       |
| 17  | Q3164-02    | DA-2        | VV039161.D     | 24 Sep 2025 18:30 | pH<2.0 A ,Confirm Concentration     | SY/MD    | Not Ok   |

Instrument ID: MSVOA\_V

**Daily Analysis Runlog For Sequence/QC Batch ID # VV092425**

| Review By                               | Mahesh Dadoda       | Review On         | 9/26/2025 1:45:06 AM         |
|-----------------------------------------|---------------------|-------------------|------------------------------|
| Supervise By                            | Semsettin Yesilyurt | Supervise On      | 9/26/2025 2:03:33 AM         |
| SubDirectory                            | VV092425            | HP Acquire Method | MSVOA_V HP Processing Method |
| STD. NAME                               | STD REF.#           |                   |                              |
| Tune/Reschk<br>Initial Calibration Stds | VP135622            |                   |                              |
| CCC                                     | VP135623,VP135624   |                   |                              |
| Internal Standard/PEM                   | VP133935            |                   |                              |
| ICV/I.BLK                               |                     |                   |                              |
| Surrogate Standard                      |                     |                   |                              |
| MS/MSD Standard                         |                     |                   |                              |
| LCS Standard                            |                     |                   |                              |

|    |              |              |            |                   |          |       |      |
|----|--------------|--------------|------------|-------------------|----------|-------|------|
| 18 | Q3164-03     | DA-3         | VV039162.D | 24 Sep 2025 18:52 | pH<2.0 A | SY/MD | Ok   |
| 19 | VIBLK        | VIBLK        | VV039163.D | 24 Sep 2025 19:15 |          | SY/MD | Ok   |
| 20 | VIBLK        | VIBLK        | VV039164.D | 24 Sep 2025 19:37 |          | SY/MD | Ok,M |
| 21 | VV0924WBSD02 | VV0924WBSD02 | VV039165.D | 24 Sep 2025 19:59 |          | SY/MD | Ok   |
| 22 | VSTDCCC050   | VSTDCCC050EC | VV039166.D | 24 Sep 2025 20:22 |          | SY/MD | Ok   |

M : Manual Integration

Instrument ID: MSVOA\_V

**Daily Analysis Runlog For Sequence/QC Batch ID # VV092525**

|              |                     |                   |                       |                      |              |
|--------------|---------------------|-------------------|-----------------------|----------------------|--------------|
| Review By    | Mahesh Dadoda       | Review On         | 9/29/2025 12:56:22 AM |                      |              |
| Supervise By | Semsettin Yesilyurt | Supervise On      | 9/29/2025 1:08:36 AM  |                      |              |
| SubDirectory | VV092525            | HP Acquire Method | 8260_V                | HP Processing Method | 82v092225w.m |

| STD. NAME                                                                                          | STD REF.#                     |
|----------------------------------------------------------------------------------------------------|-------------------------------|
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP135649                      |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP135650,VP135651<br>VP133935 |

| Sr# | SampleID     | ClientID       | Data File Name | Date-Time         | Comment                  | Operator | Status   |
|-----|--------------|----------------|----------------|-------------------|--------------------------|----------|----------|
| 1   | BFB          | BFB            | VV039167.D     | 25 Sep 2025 09:14 |                          | SY/MD    | Ok       |
| 2   | VSTDCCC050   | VSTDCCC050     | VV039168.D     | 25 Sep 2025 09:59 | pH#lot#v14222            | SY/MD    | Ok       |
| 3   | VV0925MBL01  | VV0925MBL01    | VV039169.D     | 25 Sep 2025 10:31 |                          | SY/MD    | Ok       |
| 4   | VV0925WBL01  | VV0925WBL01    | VV039170.D     | 25 Sep 2025 11:01 |                          | SY/MD    | Ok       |
| 5   | VV0925WBS01  | VV0925WBS01    | VV039171.D     | 25 Sep 2025 11:24 |                          | SY/MD    | Ok       |
| 6   | VV0925WBSD01 | VV0925WBSD01   | VV039172.D     | 25 Sep 2025 11:48 |                          | SY/MD    | Ok       |
| 7   | Q3164-02     | DA-2           | VV039173.D     | 25 Sep 2025 12:10 | pH<2.0 B                 | SY/MD    | Ok       |
| 8   | Q3175-01DL   | MW2DL          | VV039174.D     | 25 Sep 2025 12:33 | Need 100X                | SY/MD    | Dilution |
| 9   | Q3175-02DL   | MW4DL          | VV039175.D     | 25 Sep 2025 12:55 |                          | SY/MD    | Ok,M     |
| 10  | VIBLK        | VIBLK          | VV039176.D     | 25 Sep 2025 13:18 |                          | SY/MD    | Ok,M     |
| 11  | Q3168-05     | TOTES COMP#1   | VV039177.D     | 25 Sep 2025 13:40 | pH<2.0 A, Surrogate fail | SY/MD    | ReRun    |
| 12  | Q3168-06     | TOTES COMP#2   | VV039178.D     | 25 Sep 2025 14:03 | pH<2.0 A, Surrogate fail | SY/MD    | ReRun    |
| 13  | Q3175-01DL2  | MW2DL2         | VV039179.D     | 25 Sep 2025 14:25 |                          | SY/MD    | Ok,M     |
| 14  | Q3172-02DL   | FW6DDL         | VV039180.D     | 25 Sep 2025 14:48 |                          | SY/MD    | Ok       |
| 15  | VIBLK        | VIBLK          | VV039181.D     | 25 Sep 2025 15:10 |                          | SY/MD    | Ok       |
| 16  | Q3172-01DL   | FW7DDL         | VV039182.D     | 25 Sep 2025 15:33 | Surrogate fail           | SY/MD    | Ok       |
| 17  | Q3168-05RE   | TOTES COMP#1RE | VV039183.D     | 25 Sep 2025 15:55 | pH<2.0 B, Surrogate fail | SY/MD    | Confirms |
| 18  | Q3168-06RE   | TOTES COMP#2RE | VV039184.D     | 25 Sep 2025 16:18 | pH<2.0 B, Surrogate fail | SY/MD    | Confirms |

Instrument ID: MSVOA\_V

**Daily Analysis Runlog For Sequence/QC Batch ID # VV092525**

|                                         |                     |                   |                       |                      |              |
|-----------------------------------------|---------------------|-------------------|-----------------------|----------------------|--------------|
| Review By                               | Mahesh Dadoda       | Review On         | 9/29/2025 12:56:22 AM |                      |              |
| Supervise By                            | Semsettin Yesilyurt | Supervise On      | 9/29/2025 1:08:36 AM  |                      |              |
| SubDirectory                            | VV092525            | HP Acquire Method | 8260_V                | HP Processing Method | 82v092225w.m |
| STD. NAME                               |                     | STD REF.#         |                       |                      |              |
| Tune/Reschk<br>Initial Calibration Stds |                     | VP135649          |                       |                      |              |
| CCC                                     |                     | VP135650,VP135651 |                       |                      |              |
| Internal Standard/PEM                   |                     | VP133935          |                       |                      |              |
| ICV/I.BLK                               |                     |                   |                       |                      |              |
| Surrogate Standard                      |                     |                   |                       |                      |              |
| MS/MSD Standard                         |                     |                   |                       |                      |              |
| LCS Standard                            |                     |                   |                       |                      |              |

|    |            |              |            |                   |                                  |       |       |
|----|------------|--------------|------------|-------------------|----------------------------------|-------|-------|
| 19 | Q3172-03   | GB3W         | VV039185.D | 25 Sep 2025 16:40 | pH<2.0 A                         | SY/MD | Ok    |
| 20 | Q3195-22   | RW-1-092425  | VV039186.D | 25 Sep 2025 17:03 | pH<2.0 A                         | SY/MD | Ok    |
| 21 | Q3195-23   | EB-01-091925 | VV039187.D | 25 Sep 2025 17:25 | pH<2.0 A ,EB                     | SY/MD | Ok    |
| 22 | Q3195-24   | EB-02-092225 | VV039188.D | 25 Sep 2025 17:47 | pH<2.0 A ,Internal Standard Fail | SY/MD | ReRun |
| 23 | Q3195-25   | FB-01-091925 | VV039189.D | 25 Sep 2025 18:10 | pH<2.0 A ,FB                     | SY/MD | Ok    |
| 24 | Q3195-26   | FB-02-092225 | VV039190.D | 25 Sep 2025 18:32 | pH<2.0 A FB                      | SY/MD | Ok    |
| 25 | Q3195-27   | TB-01-091525 | VV039191.D | 25 Sep 2025 18:55 | pH<2.0 A ,TB                     | SY/MD | Ok    |
| 26 | Q3195-02   | MW-1-091725  | VV039192.D | 25 Sep 2025 19:17 | pH<2.0 A                         | SY/MD | Ok    |
| 27 | Q3195-09   | MW-22-091825 | VV039193.D | 25 Sep 2025 19:40 | pH<2.0 A                         | SY/MD | Ok    |
| 28 | Q3195-10   | MW-23-092325 | VV039194.D | 25 Sep 2025 20:02 | pH<2.0 A                         | SY/MD | Ok    |
| 29 | Q3195-11   | MW-24-091925 | VV039195.D | 25 Sep 2025 20:25 | pH<2.0 A                         | SY/MD | Ok    |
| 30 | VSTDCCC050 | VSTDCCC050EC | VV039196.D | 25 Sep 2025 20:47 |                                  | SY/MD | Ok    |

M : Manual Integration



# SHIPPING DOCUMENTS

CLIENT INFORMATION

COMPANY: G Environmental  
ADDRESS: 8 CARRIDGE  
CITY: Succasunna STATE: NJ ZIP: 07816  
ATTENTION:  
PHONE: FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: Summer  
PROJECT NO.: LOCATION: NJ  
PROJECT MANAGER: GL  
e-mail:  
PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: G Environmental  
ADDRESS: 8 CARRIDGE  
CITY: Succasunna STATE: NJ ZIP:  
ATTENTION: PHONE:  
ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) DAYS\*  
HARDCOPY (DATA PACKAGE): Standard DAYS\*  
EDD: DAYS\*  
\*TO BE APPROVED BY CHEMTECH  
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)  
☐ Level 2 (Results + QC) ☒ NJ Reduced ☐ US EPA CLP  
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B  
+ Raw Data ☐ Other  
☒ EDD FORMAT PDF SRP

PRESERVATIVES

COMMENTS

ALLIANCE  
SAMPLE  
ID

PROJECT  
SAMPLE IDENTIFICATION

SAMPLE  
MATRIX

SAMPLE  
TYPE  
COMP GRAB

SAMPLE  
COLLECTION  
DATE TIME

# OF BOTTLES

1 2 3 4 5 6 7 8 9  
← Specify Preservatives  
A-HCl D-NaOH  
B-HNO3 E-ICE  
C-H2SO4 F-OTHER

1.  
2.  
3.  
4.  
5.  
6.  
7.  
8.  
9.  
10.

MW2  
MW4

GW  
GW

X 9/23/15 1400 2  
X 9/23/15 1500 2

1 2 3 4 5 6 7 8 9

X  
X

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: DATE/TIME: RECEIVED BY:  
1. 9/22/15 15:15 [Signature]  
RELINQUISHED BY SAMPLER: DATE/TIME: RECEIVED BY:  
2. 9/23/15 15:15 [Signature]  
RELINQUISHED BY SAMPLER: DATE/TIME: RECEIVED BY:  
3. [Signature]

Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 2.2 °C  
Comments:

Page \_\_\_\_ of

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete  
☐ YES ☐ NO

## LOGIN REPORT/SAMPLE TRANSFER

|                                       |        |                                                            |                                 |
|---------------------------------------|--------|------------------------------------------------------------|---------------------------------|
| <b>Order ID :</b> Q3175               | GENV01 | <b>Order Date :</b> 9/24/2025 9:56:00 AM                   | <b>Project Mgr :</b>            |
| <b>Client Name :</b> G Environmental  |        | <b>Project Name :</b> Summer                               | <b>Report Type :</b> NJ Reduced |
| <b>Client Contact :</b> Gary Landis   |        | <b>Receive DateTime :</b> <del>9/24</del> /2025 1:15:00 PM | <b>EDD Type :</b> HAZ/EXCEL     |
| <b>Invoice Name :</b> G Environmental |        | <b>Purchase Order :</b>                                    | <b>Hard Copy Date :</b>         |
| <b>Invoice Contact :</b> Gary Landis  |        |                                                            | <b>Date Signoff :</b>           |

| LAB ID   | CLIENT ID | MATRIX | SAMPLE DATE | SAMPLE TIME | TEST         | TEST GROUP | METHOD   | FAX DATE     | DUE DATES |
|----------|-----------|--------|-------------|-------------|--------------|------------|----------|--------------|-----------|
| Q3175-01 | MW2       | Water  | 09/23/2025  | 14:05       |              |            |          |              |           |
|          |           |        |             |             | VOCMS Group1 |            | 8260-Low | 10 Bus. Days |           |
| Q3175-02 | MW4       | Water  | 09/23/2025  | 15:00       |              |            |          |              |           |
|          |           |        |             |             | VOCMS Group1 |            | 8260-Low | 10 Bus. Days |           |

Relinquished By :

Date / Time :



9/24/25 0950

Received By :

Date / Time :



09/24/25

09:50 OG #4

Storage Area : VOA Refridgerator Room

### Laboratory Certification

| Certified By    | License No.      |
|-----------------|------------------|
|                 |                  |
| Connecticut     | PH-0830          |
|                 |                  |
| DOD ELAP (ANAB) | L2219            |
|                 |                  |
| Maine           | 2024021          |
|                 |                  |
| Maryland        | 296              |
|                 |                  |
| New Hampshire   | 255425           |
|                 |                  |
| New Jersey      | 20012            |
|                 |                  |
| New York        | 11376            |
|                 |                  |
| Pennsylvania    | 68-00548         |
|                 |                  |
| Soil Permit     | 525-24-234-08441 |
|                 |                  |
| Texas           | TX-C25-00189     |
|                 |                  |
| Virginia        | 460312           |