

Surrogate Summary

SDG No.: **Q3175**

Client: **G Environmental**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3175-01	MW2	1,2-Dichloroethane-d4	50	44.7	89		74	125
		Dibromofluoromethane	50	38.9	78		75	124
		Toluene-d8	50	47.6	95		86	113
		4-Bromofluorobenzene	50	45.2	90		77	121
Q3175-01DL	MW2DL	1,2-Dichloroethane-d4	50	52.4	105		74	125
		Dibromofluoromethane	50	48.9	98		75	124
		Toluene-d8	50	48.3	97		86	113
		4-Bromofluorobenzene	50	50.2	100		77	121
Q3175-01DL2	MW2DL2	1,2-Dichloroethane-d4	50	55.6	111		74	125
		Dibromofluoromethane	50	49.9	100		75	124
		Toluene-d8	50	45.3	91		86	113
		4-Bromofluorobenzene	50	46.4	93		77	121
Q3175-02	MW4	1,2-Dichloroethane-d4	50	46.6	93		74	125
		Dibromofluoromethane	50	38.8	78		75	124
		Toluene-d8	50	47.7	95		86	113
		4-Bromofluorobenzene	50	48.4	97		77	121
Q3175-02DL	MW4DL	1,2-Dichloroethane-d4	50	54.6	109		74	125
		Dibromofluoromethane	50	48.8	98		75	124
		Toluene-d8	50	47.6	95		86	113
		4-Bromofluorobenzene	50	48.5	97		77	121
VV0924WBL01	VV0924WBL01	1,2-Dichloroethane-d4	50	62.2	124		74	125
		Dibromofluoromethane	50	54.6	109		75	124
		Toluene-d8	50	44.1	88		86	113
		4-Bromofluorobenzene	50	48.0	96		77	121
VV0924WBS02	VV0924WBS02	1,2-Dichloroethane-d4	50	49.0	98		74	125
		Dibromofluoromethane	50	50.0	100		75	124
		Toluene-d8	50	49.5	99		86	113
		4-Bromofluorobenzene	50	52.0	104		77	121
VV0924WBSD0	VV0924WBSD02	1,2-Dichloroethane-d4	50	49.5	99		74	125
		Dibromofluoromethane	50	50.1	100		75	124
		Toluene-d8	50	50.7	101		86	113
		4-Bromofluorobenzene	50	55.0	110		77	121
VV0925WBL01	VV0925WBL01	1,2-Dichloroethane-d4	50	60.1	120		74	125
		Dibromofluoromethane	50	53.3	107		75	124
		Toluene-d8	50	47.3	95		86	113
		4-Bromofluorobenzene	50	44.1	88		77	121
VV0925WBS01	VV0925WBS01	1,2-Dichloroethane-d4	50	45.1	90		74	125
		Dibromofluoromethane	50	47.8	96		75	124
		Toluene-d8	50	47.9	96		86	113
		4-Bromofluorobenzene	50	51.2	102		77	121