

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3175</u>
Lab File ID:	<u>VV039168.D</u>	Date Analyzed:	<u>09/25/2025</u>
Instrument ID:	<u>MSVOA_V</u>	Time Analyzed:	<u>09:59</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	605297	4.60	933338	5.53	901987	8.77
UPPER LIMIT	1210590	5.096	1866680	6.029	1803970	9.273
LOWER LIMIT	302649	4.096	466669	5.029	450994	8.273
EPA SAMPLE NO.						
MW2DL	563178	4.60	979530	5.54	954674	8.78
MW2DL2	457907	4.60	814896	5.54	779302	8.78
MW4DL	560555	4.60	1013068	5.54	987441	8.78
VV0925WBL01	480047	4.60	864912	5.53	870930	8.78
VV0925WBS01	557017	4.60	898902	5.54	871912	8.78

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.