

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3175</u>
Lab File ID:	<u>VV039146.D</u>	Date Analyzed:	<u>09/24/2025</u>
Instrument ID:	<u>MSVOA_V</u>	Time Analyzed:	<u>10:49</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	551620	11.172				
UPPER LIMIT	1103240	11.672				
LOWER LIMIT	275810	10.672				
EPA SAMPLE NO.						
MW2	318916	11.18				
MW4	492612	11.18				
VV0924WBL01	379013	11.18				
VV0924WBS02	468522	11.17				
VV0924WBSD02	510716	11.17				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.