

Initial and Continuing Calibration Verification

Client: Weston Solutions, Inc.

SDG No.: Q3176

Project: RFP 916

RunNo.: LB137340

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.095	0.099	96	90-110	09/26/2025
Sample ID: Cyanide	CCV1	mg/L	0.24	0.25	96	90-110	09/26/2025
Sample ID: Cyanide	CCV2	mg/L	0.24	0.25	96	90-110	09/26/2025
Sample ID: Cyanide	CCV3	mg/L	0.24	0.25	96	90-110	09/26/2025