

LB137305


7471B

INSTRUMENT ID: CV1

Linear

μ Abs. :

4131E



A= 0.0000e+000

B= 2.3789e-004

C= -1.5400e-002

Rho= 0.9994511

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.011	-0.011	19	0.000	19				
0.2	0.200	0.165	-0.035	757	0.0 %	757				
2.5	2.500	2.426	-0.074	10264	0.0 %	10264				
5.0	5.000	5.132	0.132	21639	0.0 %	21639				
7.5	7.500	7.675	0.175	32326	0.0 %	32326				
10.0	10.000	9.813	-0.187	41316	0.0 %	41316				

LB137305 INSTRUMENT ID : CV1

Method: 7471B Operator: Admin Date of Analysis: 25 Sep 2025 12:02:32

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50	-	19	PPB	Std	0.0000	7471B	25 Sep 2025 12:11:14
S	0.2 - 1	50-2	-	757	PPB	Std	0.2000	7471B	25 Sep 2025 12:13:31
S	2.5 - 1	52.5	-	10264	PPB	Std	2.5000	7471B	25 Sep 2025 12:15:48
S	5.0 - 1	55	-	21639	PPB	Std	5.0000	7471B	25 Sep 2025 12:18:05
S	7.5 - 1	57.5	-	32326	PPB	Std	7.5000	7471B	25 Sep 2025 12:20:35
S	10.0 - 1	510	-	41316	PPB	Std	10.0000	7471B	25 Sep 2025 12:23:11
U	ICV46 - 1	ICV46	4.3903	18520	PPB	SMPL	-7471B		25 Sep 2025 12:26:26
U	ICB46 - 1	ICB46	-0.0751	-251	PPB	SMPL	-7471B		25 Sep 2025 12:28:55
U	CCV39 - 1	CCV39	5.1410	21676	PPB	SMPL	-7471B		25 Sep 2025 12:31:32
U	CCB39 - 1	CCB39	-0.0784	-265	PPB	SMPL	-7471B		25 Sep 2025 12:33:47
U	CRA - 1	CRA	0.1611	742	PPB	SMPL	-7471B		25 Sep 2025 12:36:24
U	HighStd - 1	HighStd	9.7872	41207	PPB	SMPL	-7471B		25 Sep 2025 12:38:39
U	ChkStd - 1	ChkStd	7.6182	32089	PPB	SMPL	-7471B		25 Sep 2025 12:41:02
U	PB169842BL - 1	PB169842BL	-0.0682	-222	PPB	SMPL	-7471B		25 Sep 2025 12:43:47
U	PB169842BS - 1	PB169842BS	4.7526	20043	PPB	SMPL	-7471B		25 Sep 2025 12:46:31
U	Q3167-01 - 1	RT4784	6.3747	26862	PPB	SMPL	-7471B		25 Sep 2025 12:48:47
U	Q3176-01 - 1	EME-TS12-01	0.6864	2950	PPB	SMPL	-7471B		25 Sep 2025 12:51:25
U	Q3176-02 - 1	EME-TS13-01	0.8391	3592	PPB	SMPL	-7471B		25 Sep 2025 12:54:08
U	Q3176-03 - 1	EME-TS14-02	0.4670	2028	PPB	SMPL	-7471B		25 Sep 2025 12:56:27
U	Q3178-01 - 1	EME-TS14-01	0.0981	477	PPB	SMPL	-7471B		25 Sep 2025 12:58:45
U	CCV40 - 1	CCV40	5.2495	22132	PPB	SMPL	-7471B		25 Sep 2025 13:01:01
U	CCB40 - 1	CCB40	-0.0727	-241	PPB	SMPL	-7471B		25 Sep 2025 13:03:17
U	Q3178-01DUP - 1	EME-TS14-01DUP	0.0553	297	PPB	SMPL	-7471B		25 Sep 2025 13:08:28
U	Q3178-02 - 1	EME-TS14-01MS	4.5373	19138	PPB	SMPL	-7471B		25 Sep 2025 13:10:44
U	Q3178-03 - 1	EME-TS14-01MSD	4.4174	18634	PPB	SMPL	-7471B		25 Sep 2025 13:12:59
U	Q3183-01 - 1	WC-1	4.4588	18808	PPB	SMPL	-7471B		25 Sep 2025 13:15:36
U	Q3183-04 - 1	WC-4	1.4428	6130	PPB	SMPL	-7471B		25 Sep 2025 13:18:07
U	Q3183-07 - 1	WC-3	0.5727	2472	PPB	SMPL	-7471B		25 Sep 2025 13:20:46
U	Q3187-01 - 1	228	3.0579	12919	PPB	SMPL	-7471B		25 Sep 2025 13:23:06
U	Q3188-02 - 1	RB22120	0.7606	3262	PPB	SMPL	-7471B		25 Sep 2025 13:25:24
U	Q3189-01 - 1	72-11989	0.8293	3551	PPB	SMPL	-7471B		25 Sep 2025 13:27:57
U	Q3178-01LX5 - 1		-0.0409	-107	PPB	SMPL	-7471B		25 Sep 2025 13:30:14
U	CCV41 - 1	CCV41	5.1910	21886	PPB	SMPL	-7471B		25 Sep 2025 13:32:32
U	CCB41 - 1	CCB41	-0.0799	-271	PPB	SMPL	-7471B		25 Sep 2025 13:34:48
U	Q3178-01A - 1		4.7352	19970	PPB	SMPL	-7471B		25 Sep 2025 13:37:26
U	CCV42 - 1	CCV42	5.1213	21593	PPB	SMPL	-7471B		25 Sep 2025 13:39:41
U	CCB42 - 1	CCB42	-0.0770	-259	PPB	SMPL	-7471B		25 Sep 2025 13:42:19